

Data Path : Z:\HPCHEM1\BNA M\Data\BM123015\
 Data File : BM003832.D
 Acq On : 28 Dec 2015 17:22
 Operator : SJ/UM
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.464

Quant Time: Dec 29 07:09:01 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 00:41:46 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.166	2.9	113	0.02
3	1,4-Dioxane	0.218	0.203	6.9	110	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
5	Naphthalene	0.901	0.981	-8.9	141	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.515	-7.5	139	0.00
7	2-Methylnaphthalene	0.605	0.678	-12.1	143	0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
9	Acenaphthylene	1.528	1.558	-2.0	131	0.00
10 C	Acenaphthene	1.232	1.357	-10.1	138	0.00
11	Fluorene	1.365	1.475	-8.1	135	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Pentachlorophenol	0.057	0.030	47.4#	61	0.00
14	Phenanthrene	1.032	1.093	-5.9	116	0.00
15	Anthracene	0.893	0.983	-10.1	121	0.00
16 SURR	Fluoranthene-d10	0.849	0.873	-2.8	112	0.00
17 C	Fluoranthene	1.120	1.064	5.0	102	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
19	Pyrene	1.080	1.464	-35.6#	102	0.00
20	Benzo(a)anthracene	0.977	1.225	-25.4#	97	0.00
21	Chrysene	1.119	1.209	-8.0	89	0.00
22 I	Perylene-d12	1.000	1.000	0.0	109	0.00
23	Benzo(b)fluoranthene	1.365	1.347	1.3	103	0.00
24	Benzo(k)fluoranthene	1.237	1.203	2.7	107	0.00
25 C	Benzo(a)pyrene	1.152	1.260	-9.4	118	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.500	-12.4	129	0.01
27	Dibenzo(a,h)anthracene	1.075	1.206	-12.2	129	0.00
28	Benzo(q,h,i)perylene	1.162	1.284	-10.5	128	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0