

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123015\
 Data File : BM003845.D
 Acq On : 29 Dec 2015 01:08
 Operator : SJ/UM
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.465

Quant Time: Dec 29 07:16:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 29 07:09:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.170	0.6	95	0.00
3	1,4-Dioxane	0.218	0.208	4.6	92	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
5	Naphthalene	0.901	0.993	-10.2	121	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.534	-11.5	122	0.00
7	2-Methylnaphthalene	0.605	0.700	-15.7	125	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
9	Acenaphthylene	1.528	1.571	-2.8	123	0.00
10 C	Acenaphthene	1.232	1.375	-11.6	130	0.00
11	Fluorene	1.365	1.547	-13.3	131	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
13	Pentachlorophenol	0.057	0.027	52.6#	59	0.00
14	Phenanthrene	1.032	1.127	-9.2	131	0.00
15	Anthracene	0.893	1.018	-14.0	137	0.00
16 SURR	Fluoranthene-d10	0.849	0.965	-13.7	134	0.00
17 C	Fluoranthene	1.120	1.344	-20.0	141	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
19	Pyrene	1.080	1.524	-41.1#	136	0.00
20	Benzo(a)anthracene	0.977	1.209	-23.7#	123	0.00
21	Chrysene	1.119	1.218	-8.8	115	0.00
22 I	Perylene-d12	1.000	1.000	0.0	110	0.00
23	Benzo(b)fluoranthene	1.365	1.466	-7.4	114	0.00
24	Benzo(k)fluoranthene	1.237	1.268	-2.5	114	0.00
25 C	Benzo(a)pyrene	1.152	1.263	-9.6	120	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.416	-6.1	123	0.00
27	Dibenzo(a,h)anthracene	1.075	1.129	-5.0	122	0.00
28	Benzo(q,h,i)perylene	1.162	1.224	-5.3	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0