

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123015\
 Data File : BM003852.D
 Acq On : 29 Dec 2015 05:54
 Operator : SJ/UM
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.466

Quant Time: Dec 29 15:34:53 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 29 08:19:02 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.170	0.6	95	0.00
3	1,4-Dioxane	0.218	0.205	6.0	91	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	110	-0.01
5	Naphthalene	0.901	0.980	-8.8	118	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.530	-10.6	120	0.00
7	2-Methylnaphthalene	0.605	0.699	-15.5	124	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
9	Acenaphthylene	1.528	1.570	-2.7	122	0.00
10 C	Acenaphthene	1.232	1.369	-11.1	129	0.00
11	Fluorene	1.365	1.558	-14.1	131	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
13	Pentachlorophenol	0.057	0.030	47.4#	67	0.00
14	Phenanthrene	1.032	1.122	-8.7	133	0.00
15	Anthracene	0.893	0.997	-11.6	138	0.00
16 SURR	Fluoranthene-d10	0.849	0.978	-15.2	140	0.00
17 C	Fluoranthene	1.120	1.358	-21.2#	146	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
19	Pyrene	1.080	1.485	-37.5#	141	0.00
20	Benzo(a)anthracene	0.977	1.206	-23.4#	131	0.00
21	Chrysene	1.119	1.235	-10.4	124	0.00
22 I	Perylene-d12	1.000	1.000	0.0	116	0.00
23	Benzo(b)fluoranthene	1.365	1.443	-5.7	118	0.00
24	Benzo(k)fluoranthene	1.237	1.331	-7.6	126	0.00
25 C	Benzo(a)pyrene	1.152	1.279	-11.0	128	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.384	-3.7	128	0.01
27	Dibenzo(a,h)anthracene	1.075	1.114	-3.6	127	0.00
28	Benzo(q,h,i)perylene	1.162	1.196	-2.9	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1