

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123020\
 Data File : BM028282.D
 Acq On : 30 Dec 2020 09:10
 Operator : JU/CG
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4001

Quant Time: Dec 30 10:02:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 30 10:02:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	1564	0.40	ng/ul	0.00
4) Naphthalene-d8	11.01	136	6048	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.81	164	3131	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.53	188	6261	0.40	ng/ul	0.00
17) Chrysene-d12	21.65	240	5603	0.40	ng/ul	0.00
23) Perylene-d12	24.00	264	5628	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.41	96	710	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.58	152	3732	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.53	212	6786	0.41	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.45	88	726	0.367	ng/ul#	69
5) Naphthalene	11.06	128	7221	0.423	ng/ul	99
7) 2-Methylnaphthalene	12.65	142	4675	0.415	ng/ul	99
8) 1-Methylnaphthalene	12.88	142	4459	0.415	ng/ul	99
10) Acenaphthylene	14.53	152	5637	0.446	ng/ul	99
11) Acenaphthene	14.87	153	4738	0.426	ng/ul	99
12) Fluorene	15.85	166	5243	0.415	ng/ul	98
14) Pentachlorophenol	17.18	266	680	0.428	ng/ul	97
15) Phenanthrene	17.57	178	8340	0.434	ng/ul	99
16) Anthracene	17.66	178	7353	0.436	ng/ul	99
19) Fluoranthene	19.56	202	9094	0.413	ng/ul	96
20) Pyrene	19.92	202	9230	0.420	ng/ul	96
21) Benzo(a)anthracene	21.63	228	8443	0.457	ng/ul	98
22) Chrysene	21.68	228	8753	0.440	ng/ul	99
24) Benzo(b)fluoranthene	23.28	252	9026	0.399	ng/ul	94
25) Benzo(k)fluoranthene	23.33	252	8830	0.406	ng/ul	96
26) Benzo(a)pyrene	23.90	252	7912	0.423	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	26.41	276	10228	0.451	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.42	278	8360	0.440	ng/ul	97
29) Benzo(g,h,i)perylene	27.15	276	8930	0.451	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123020\
Data File : BM028282.D
Acq On : 30 Dec 2020 09:10
Operator : JU/CG
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4001

Quant Time: Dec 30 10:02:54 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 30 10:02:26 2020
Response via : Initial Calibration

