

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123020\
 Data File : BM028289.D
 Acq On : 30 Dec 2020 17:18
 Operator : JU/CG
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4002

Quant Time: Dec 30 17:48:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 30 10:02:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	1536	0.40	ng/ul	0.00
4) Naphthalene-d8	11.01	136	5846	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.81	164	3020	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.53	188	6129	0.40	ng/ul	0.00
17) Chrysene-d12	21.65	240	5728	0.40	ng/ul	0.00
23) Perylene-d12	24.00	264	5673	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.41	96	704	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.58	152	3582	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.53	212	6686	0.39	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.45	88	739	0.380	ng/ul	94
5) Naphthalene	11.06	128	6927	0.420	ng/ul	99
7) 2-Methylnaphthalene	12.65	142	4438	0.408	ng/ul	100
8) 1-Methylnaphthalene	12.88	142	4229	0.407	ng/ul	99
10) Acenaphthylene	14.53	152	5346	0.438	ng/ul	98
11) Acenaphthene	14.87	153	4565	0.426	ng/ul	99
12) Fluorene	15.85	166	5066	0.416	ng/ul	99
14) Pentachlorophenol	17.18	266	661	0.425	ng/ul	98
15) Phenanthrene	17.57	178	8044	0.428	ng/ul	99
16) Anthracene	17.66	178	7093	0.430	ng/ul	98
19) Fluoranthene	19.56	202	8856	0.393	ng/ul	98
20) Pyrene	19.92	202	8973	0.399	ng/ul	96
21) Benzo(a)anthracene	21.63	228	8257	0.437	ng/ul	98
22) Chrysene	21.68	228	8767	0.431	ng/ul	99
24) Benzo(b)fluoranthene	23.29	252	9161	0.401	ng/ul	96
25) Benzo(k)fluoranthene	23.33	252	8772	0.400	ng/ul	97
26) Benzo(a)pyrene	23.90	252	8023	0.426	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	26.41	276	10394	0.454	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.43	278	8686	0.453	ng/ul	98
29) Benzo(a,h,i)perylene	27.15	276	9146	0.459	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123020\
Data File : BM028289.D
Acq On : 30 Dec 2020 17:18
Operator : JU/CG
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4002

Quant Time: Dec 30 17:48:32 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM121120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 30 10:02:26 2020
Response via : Initial Calibration

