


Data Path : Z:\HPCHEM1\BNA_M\Data\BM123114\
Data File : BM000094.D
Acq On : 31 Dec 2014 13:38
Operator : TP
Sample : DFTPP02
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM122914.M
Title : SVOA CALIBRATION
Last Update : Wed Dec 31 02:57:11 2014



AutoFind: Scans 1309, 1310, 1311; Background Corrected with Scan 1306

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.2	71507	PASS
68	69	0.00	2	1.6	1139	PASS
69	198	0.00	100	41.1	73056	PASS
70	69	0.00	2	0.6	423	PASS
127	198	10	80	46.6	82938	PASS
197	198	0.00	2	0.4	725	PASS
198	198	100	100	100.0	177818	PASS
199	198	5	9	6.9	12303	PASS
275	198	10	60	26.3	46717	PASS
365	198	1	100	3.9	6971	PASS
441	443	0.01	100	78.0	25737	PASS
442	198	50	100	97.3	173082	PASS
443	442	15	24	19.1	32991	PASS

m/z	Abundance
37.00	323.0
38.10	952.0
39.00	6756.0
40.00	1434.0
43.10	302.0
44.10	4198.0
49.00	663.0
50.10	23536.0
51.10	95944.0
52.00	4657.0
55.00	684.0
56.00	2720.0
57.10	6564.0
58.00	326.0
61.00	1290.0
62.10	1239.0
63.00	3997.0
63.90	600.0
65.10	1824.0
68.10	1678.0
69.00	94720.0
70.00	491.0
73.10	1032.0
74.10	10523.0
75.00	15320.0
76.00	5386.0
77.10	103304.0
78.00	6822.0
79.00	6836.0
80.00	6439.0
81.00	8122.0
82.00	1760.0
83.00	1928.0
85.00	1569.0
86.00	2556.0
87.00	1048.0
88.10	362.0
91.00	1896.0
92.00	2377.0
93.00	12393.0
94.00	723.0
95.90	853.0
98.00	10913.0
99.00	8772.0
100.00	799.0
101.00	3996.0
103.10	1710.0
104.00	3253.0
105.00	2899.0
105.90	854.0
107.00	35440.0
108.00	5131.0
109.10	820.0
110.00	52160.0
111.00	9028.0
112.10	1195.0
113.00	399.0
116.10	2279.0
117.00	26808.0
118.00	2079.0
119.00	405.0
120.20	350.0
122.00	1657.0
123.00	2825.0
124.10	1357.0
125.00	1408.0
126.20	366.0
127.00	93800.0
128.10	7320.0
129.00	39024.0
130.00	3240.0
131.00	653.0
133.10	401.0
134.10	813.0
135.00	3408.0
136.10	1497.0
137.00	2251.0
137.80	375.0
139.10	318.0
140.10	489.0
141.00	4395.0
142.00	1735.0
143.00	959.0
144.10	355.0
145.00	397.0
146.00	992.0
147.10	2641.0
148.00	6338.0
149.10	1640.0
150.00	506.0
151.00	681.0

Instrument :
BNA_M
ClientSampleId :

152.00	489.0
153.00	1857.0
154.00	1081.0
155.10	2476.0
156.10	3950.0
157.10	609.0
157.90	757.0
159.00	689.0
160.00	1825.0
161.00	2175.0
162.00	869.0
164.10	379.0
165.00	1809.0
166.00	1537.0
167.10	8399.0
168.10	5319.0
169.10	893.0
170.90	624.0
172.00	722.0
173.00	993.0
174.00	2327.0
175.10	3642.0
176.10	969.0
177.00	1978.0
178.10	639.0
179.00	6922.0
180.10	5765.0
181.10	2242.0
182.10	549.0
184.00	614.0
185.10	3996.0
186.10	26232.0
187.00	7223.0
188.10	1060.0
189.10	1829.0
190.90	1179.0
192.00	1993.0
193.10	2887.0
194.10	584.0
196.00	4015.0
197.10	720.0
198.00	154880.0
199.00	12724.0
200.00	1160.0
201.20	624.0
203.00	1549.0
204.00	6177.0
205.00	10465.0
206.10	43768.0
207.10	8078.0
208.00	2192.0
209.00	831.0
210.00	589.0
211.10	1638.0
215.10	617.0
216.00	1395.0
217.00	11489.0
218.00	1384.0
221.00	3419.0
221.90	1146.0
223.10	2426.0
224.10	22392.0
225.00	5502.0
225.90	552.0
227.00	9528.0
228.10	1165.0
229.10	2058.0
229.90	302.0
231.20	747.0
233.90	430.0
234.90	644.0
236.10	408.0
237.00	1288.0
239.00	435.0
240.90	562.0
242.00	831.0
243.10	986.0
244.10	14254.0
245.00	1833.0
246.00	2975.0
247.00	819.0
249.10	595.0
253.00	523.0
254.00	696.0
255.00	80600.0
256.00	12835.0
257.00	909.0
258.10	5351.0
259.10	897.0
265.10	2638.0
266.00	356.0
273.00	1965.0
274.00	5605.0
275.00	31904.0
276.00	3983.0

Instrument :
BNA_M
ClientSampleId :

277.00	3376.0
278.00	649.0
283.10	556.0
285.00	605.0
293.10	630.0
294.10	324.0
296.10	8911.0
297.10	1149.0
302.90	814.0
314.10	301.0
315.00	843.0
315.90	371.0
320.90	375.0
323.10	2631.0
324.10	582.0
327.00	793.0
333.00	300.0
334.10	1688.0
335.20	375.0
341.00	412.0
345.90	414.0
352.10	747.0
353.90	822.0
365.00	3569.0
366.00	504.0
372.00	1115.0
373.10	435.0
402.10	333.0
403.00	576.0
404.00	365.0
421.10	457.0
422.20	538.0
423.10	3112.0
424.00	1417.0
441.10	8781.0
442.00	54544.0
443.00	12351.0
444.10	721.0

Instrument :

BNA_M

ClientSampleId :

m/z	Abundance
37.10	359.0
38.10	902.0
39.10	6620.0
40.00	1435.0
41.10	380.0
44.00	4457.0
45.00	322.0
49.10	825.0
50.10	24528.0
51.10	102944.0
52.10	5208.0
55.00	349.0
56.10	3523.0
57.00	7238.0
61.00	1311.0
62.00	1498.0
63.10	4485.0
64.10	806.0
65.10	2143.0
68.00	1739.0
69.00	107288.0
70.10	778.0
73.10	1049.0
74.00	11825.0
75.00	17768.0
76.10	5958.0
77.10	114952.0
78.00	8041.0
79.00	8385.0
80.00	7164.0
81.00	9911.0
82.10	2266.0
83.00	2327.0
85.10	2078.0
86.10	3107.0
87.00	959.0
88.00	478.0
91.00	2232.0
92.00	2327.0
93.00	14927.0
94.00	1156.0
95.00	375.0
96.10	829.0
98.00	13138.0
99.00	9336.0
100.10	758.0
101.00	6043.0
103.00	1821.0
104.00	3870.0
105.00	3541.0
106.00	987.0
107.00	46008.0
108.00	6727.0
109.00	1276.0
110.00	69792.0
111.00	11751.0
112.00	1336.0
112.80	327.0
116.00	2524.0
117.00	37552.0
118.10	2697.0
119.00	341.0
120.10	491.0
122.00	2798.0
123.10	4106.0
124.10	2121.0
125.10	1697.0
126.10	588.0
127.00	130016.0
128.00	9774.0
129.00	58152.0
130.00	4884.0
131.00	1026.0
132.20	397.0
133.10	332.0
134.10	1771.0
135.10	4573.0
136.10	1968.0
137.10	2819.0
138.10	553.0
140.00	923.0
141.00	8141.0
142.00	2645.0
143.00	1791.0
144.00	424.0
145.00	481.0
146.10	1251.0
147.00	3949.0
148.00	9528.0
149.00	1871.0
150.10	437.0

Instrument :
BNA_M
ClientSampleId :

151.10	958.0
151.70	477.0
152.20	301.0
153.00	2957.0
154.00	2064.0
155.00	4368.0
156.10	7213.0
157.00	1171.0
158.10	1507.0
159.00	1141.0
160.00	2980.0
161.00	3603.0
162.00	1046.0
163.00	691.0
164.10	530.0
165.00	3231.0
166.00	2707.0
167.00	16162.0
168.00	7552.0
169.00	1610.0
170.00	491.0
171.00	880.0
172.10	1831.0
173.10	2357.0
174.10	3955.0
175.10	6807.0
176.10	1932.0
177.00	2938.0
178.00	1072.0
179.00	12626.0
180.00	9088.0
181.00	4090.0
182.10	751.0
183.10	413.0
184.00	1355.0
185.10	6517.0
186.10	45904.0
187.00	14481.0
188.10	1900.0
189.00	3304.0
190.10	346.0
191.20	1553.0
192.00	3971.0
193.00	4907.0
194.10	1154.0
196.00	7622.0
196.80	2175.0
198.00	301184.0
199.00	19712.0
200.00	1866.0
201.60	849.0
202.20	355.0
203.00	2819.0
204.00	13101.0
205.10	21688.0
206.10	90048.0
207.10	12133.0
208.10	3635.0
209.00	1312.0
210.00	1260.0
211.00	2882.0
212.00	699.0
215.10	1250.0
216.10	1855.0
217.00	24336.0
218.00	3189.0
219.00	301.0
220.10	641.0
221.10	7090.0
222.00	2723.0
223.10	4902.0
224.10	47208.0
225.10	13030.0
226.10	1464.0
227.10	24112.0
228.00	2600.0
229.00	4835.0
229.90	492.0
231.10	1752.0
234.00	1318.0
235.00	1625.0
236.10	974.0
237.00	1909.0
239.00	1184.0
240.10	604.0
241.00	1185.0
242.10	2432.0
243.10	3173.0
244.10	36032.0
245.00	4367.0
246.00	8700.0
247.10	1456.0
249.10	1061.0
251.00	574.0
252.20	415.0

Instrument :
BNA_M
ClientSampleId :

253.10	948.0
254.10	1692.0
255.00	194752.0
256.00	27448.0
257.10	2066.0
258.10	12608.0
259.00	1958.0
259.90	384.0
261.10	371.0
265.10	4711.0
266.10	425.0
271.20	565.0
272.10	505.0
273.10	6134.0
274.10	15529.0
275.10	82360.0
276.10	10306.0
277.00	8483.0
278.00	1353.0
279.00	362.0
281.10	374.0
282.10	370.0
283.10	1048.0
284.20	669.0
285.10	1316.0
288.90	389.0
292.10	533.0
293.00	2091.0
294.10	584.0
295.00	560.0
296.00	27824.0
297.00	3751.0
297.90	303.0
301.00	419.0
302.10	744.0
303.10	2656.0
304.00	835.0
308.00	347.0
314.00	1643.0
315.10	2439.0
316.10	1953.0
321.00	550.0
323.10	8701.0
324.10	1712.0
326.90	1361.0
328.00	862.0
332.10	538.0
332.90	657.0
334.10	5553.0
335.20	1376.0
341.20	1169.0
345.90	1654.0
352.10	2862.0
353.10	2320.0
354.10	2602.0
355.00	557.0
365.00	10895.0
366.10	1955.0
371.00	657.0
372.00	4955.0
373.20	1270.0
383.10	1516.0
384.00	386.0
390.10	345.0
402.00	1852.0
403.00	2585.0
404.10	1329.0
421.10	2728.0
422.10	2264.0
423.10	14394.0
424.10	3620.0
441.10	41800.0
442.10	284224.0
443.10	53176.0
444.10	5120.0

Instrument :
BNA_M
ClientSampleId :

m/z	Abundance
39.10	1265.0
40.00	1504.0
44.00	4926.0
50.10	3931.0
51.10	15634.0
52.10	1090.0
56.00	465.0
57.00	1388.0
62.10	448.0
62.90	630.0
64.90	464.0
69.00	17160.0
73.10	630.0
74.00	2311.0
75.00	2783.0
76.00	1300.0
77.00	19208.0
78.10	1487.0
79.10	1731.0
80.00	1076.0
81.00	1673.0
81.90	358.0
83.00	541.0
86.00	576.0
90.90	595.0
92.00	646.0
93.00	2844.0
98.10	2242.0
99.00	2019.0
101.00	1198.0
102.90	410.0
104.00	637.0
105.00	435.0
107.00	8910.0
108.00	843.0
109.00	358.0
110.00	13737.0
111.10	2023.0
112.20	450.0
116.20	586.0
117.00	6791.0
118.10	747.0
122.10	771.0
123.00	830.0
124.10	326.0
125.00	523.0
127.00	25000.0
128.00	1819.0
129.00	11538.0
130.00	741.0
134.10	436.0
135.00	868.0
135.80	300.0
136.90	540.0
141.00	1631.0
141.90	507.0
143.00	354.0
146.00	302.0
147.00	1033.0
148.00	1883.0
149.00	704.0
153.00	837.0
154.10	464.0
155.10	1317.0
156.00	1410.0
157.10	539.0
157.90	325.0
159.00	402.0
160.10	381.0
161.10	776.0
165.00	926.0
166.10	422.0
167.10	2998.0
168.00	1945.0
172.00	486.0
173.00	324.0
174.00	994.0
175.10	1573.0
175.90	468.0
176.90	801.0
178.10	356.0
179.00	2564.0
180.00	2811.0
181.10	1037.0
185.00	1656.0
186.00	10777.0
187.10	3352.0
187.90	546.0
189.10	845.0
191.00	594.0
192.10	969.0

Instrument :
BNA_M
ClientSampleId :

193.00	1473.0
193.90	482.0
196.10	2020.0
197.00	428.0
198.00	77392.0
199.00	4473.0
200.00	632.0
202.90	408.0
204.00	2944.0
205.10	5434.0
206.10	22920.0
207.10	4599.0
208.10	1314.0
209.20	382.0
209.90	522.0
210.90	760.0
215.00	330.0
216.10	672.0
217.00	5776.0
218.00	819.0
221.00	2312.0
221.80	881.0
223.00	1194.0
224.10	13089.0
225.10	2982.0
225.90	548.0
227.00	6762.0
228.00	775.0
228.90	1371.0
230.90	513.0
233.90	356.0
235.00	696.0
236.00	340.0
237.10	538.0
239.20	345.0
241.00	451.0
242.00	687.0
243.10	930.0
244.10	10828.0
245.00	1428.0
246.10	2488.0
247.00	486.0
249.10	496.0
254.00	633.0
255.10	62456.0
256.10	8941.0
257.00	812.0
258.00	3824.0
259.00	494.0
265.00	2054.0
266.00	422.0
273.10	2243.0
274.10	4317.0
275.10	25888.0
276.10	3652.0
277.10	2431.0
278.00	341.0
283.00	316.0
285.00	557.0
292.90	740.0
296.00	9747.0
297.00	1248.0
303.10	1393.0
304.10	301.0
314.00	463.0
315.00	1315.0
316.00	565.0
321.00	342.0
323.00	3282.0
324.10	466.0
327.10	578.0
328.00	640.0
333.10	420.0
334.00	2500.0
335.00	587.0
340.90	353.0
346.10	624.0
352.10	1192.0
353.00	713.0
354.00	1075.0
365.00	6450.0
366.00	1006.0
371.10	477.0
372.10	2163.0
373.00	693.0
383.00	551.0
389.90	445.0
402.10	1088.0
403.00	1510.0
404.10	561.0
405.20	310.0
421.10	1695.0
422.10	1247.0
423.10	7497.0
424.00	1863.0

Instrument :
BNA_M
ClientSampleId :

425.00	316.0
441.10	26632.0
442.10	180480.0
443.10	33448.0
444.10	3386.0

Instrument :
BNA_M
ClientSampleId :

Scan 1306 (17.614 min): BM000094.D
DFTPP02

m/z	Abundance
39.90	1256.0
44.00	4619.0
73.10	488.0
133.00	315.0
191.00	454.0
207.10	2145.0
208.00	333.0

Instrument :
BNA_M
ClientSampleId :