

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123115\
 Data File : BM003862.D
 Acq On : 29 Dec 2015 22:56
 Operator : SJ/UM
 Sample : G4849-02DL 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9N68DL

Manual Integrations
 APPROVED

Sohil
 12/30/2015 3:50:58 PM

Quant Time: Dec 30 10:32:29 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 30 03:30:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	4268	0.40	ng/ul	0.00
4) Naphthalene-d8	10.46	136	18586	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.33	164	10925	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.07	188	25277	0.40	ng/ul	0.00
18) Chrysene-d12	21.28	240	21528	0.40	ng/ul	0.00
22) Perylene-d12	23.51	264	18033	0.40	ng/ul	-0.01
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.14	96	5513	3.02	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.07	152	2085	0.09	ng/ul	0.00
16) Fluoranthene-d10	19.12	212	5317	0.10	ng/ul	0.00
Target Compounds						
5) Naphthalene	10.51	128	103755	2.48	ng/ul	98
7) 2-Methylnaphthalene	12.13	142	12186	0.43	ng/ul	99
10) Acenaphthene	14.38	153	6873	0.20	ng/ul#	82
11) Fluorene	15.38	166	9329m	0.25	ng/ul	
14) Phenanthrene	17.11	178	29725m	0.46	ng/ul	
15) Anthracene	17.21	178	3503	0.06	ng/ul#	92
17) Fluoranthene	19.14	202	13358	0.19	ng/ul	98
19) Pyrene	19.51	202	10156	0.17	ng/ul	97
20) Benzo(a)anthracene	21.26	228	1719	0.03	ng/ul#	89
21) Chrysene	21.31	228	1753m	0.03	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123115\
Data File : BM003862.D
Acq On : 29 Dec 2015 22:56
Operator : SJ/UM
Sample : G4849-02DL 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9N68DL

Manual Integrations
APPROVED

Sohil
12/30/2015 3:50:58 PM

Quant Time: Dec 30 10:32:29 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 30 03:30:26 2015
Response via : Initial Calibration

