

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123115\
 Data File : BM003866.D
 Acq On : 30 Dec 2015 01:19
 Operator : SJ/UM
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.435

Quant Time: Dec 30 05:57:07 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 30 03:30:26 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.172	-0.6	106	0.00
3	1,4-Dioxane	0.218	0.209	4.1	102	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
5	Naphthalene	0.901	0.992	-10.1	132	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.537	-12.1	134	0.00
7	2-Methylnaphthalene	0.605	0.707	-16.9	138	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
9	Acenaphthylene	1.528	1.566	-2.5	137	0.00
10 C	Acenaphthene	1.232	1.373	-11.4	145	0.00
11	Fluorene	1.365	1.552	-13.7	147	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
13	Pentachlorophenol	0.057	0.030	47.4#	77	0.00
14	Phenanthrene	1.032	1.128	-9.3	150	0.00
15	Anthracene	0.893	0.980	-9.7	151#	0.00
16 SURR	Fluoranthene-d10	0.849	1.036	-22.0#	165#	0.00
17 C	Fluoranthene	1.120	1.372	-22.5#	165#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
19	Pyrene	1.080	1.468	-35.9#	157#	0.00
20	Benzo(a)anthracene	0.977	1.152	-17.9	140	0.00
21	Chrysene	1.119	1.279	-14.3	145	0.00
22 I	Perylene-d12	1.000	1.000	0.0	132	0.00
23	Benzo(b)fluoranthene	1.365	1.403	-2.8	131	0.00
24	Benzo(k)fluoranthene	1.237	1.328	-7.4	143	0.00
25 C	Benzo(a)pyrene	1.152	1.271	-10.3	144	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.353	-1.3	142	0.00
27	Dibenzo(a,h)anthracene	1.075	1.088	-1.2	142	0.00
28	Benzo(q,h,i)perylene	1.162	1.169	-0.6	142	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1