

Data Path : Z:\HPCHEM1\BNA M\DATA\BM123115\
 Data File : BM003869.D
 Acq On : 30 Dec 2015 04:09
 Operator : SJ/UM
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.436

Quant Time: Dec 30 06:07:48 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 30 05:57:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.176	-2.9	125	0.00
3	1,4-Dioxane	0.218	0.209	4.1	119	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
5	Naphthalene	0.901	0.993	-10.2	149	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.528	-10.2	148	0.00
7	2-Methylnaphthalene	0.605	0.696	-15.0	153#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	145	0.00
9	Acenaphthylene	1.528	1.535	-0.5	145	0.00
10 C	Acenaphthene	1.232	1.362	-10.6	156#	0.00
11	Fluorene	1.365	1.536	-12.5	157#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	145	0.00
13	Pentachlorophenol	0.057	0.029	49.1#	75	0.00
14	Phenanthrene	1.032	1.119	-8.4	153#	0.00
15	Anthracene	0.893	0.954	-6.8	152#	0.00
16 SURR	Fluoranthene-d10	0.849	0.914	-7.7	150#	0.00
17 C	Fluoranthene	1.120	1.216	-8.6	150#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
19	Pyrene	1.080	1.530	-41.7#	139	0.00
20	Benzo(a)anthracene	0.977	1.127	-15.4	117	0.01
21	Chrysene	1.119	1.228	-9.7	118	0.01
22 I	Perylene-d12	1.000	1.000	0.0	123	0.00
23	Benzo(b)fluoranthene	1.365	1.375	-0.7	120	0.00
24	Benzo(k)fluoranthene	1.237	1.315	-6.3	132	0.00
25 C	Benzo(a)pyrene	1.152	1.269	-10.2	135	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.455	-9.0	143	0.00
27	Dibenzo(a,h)anthracene	1.075	1.158	-7.7	141	0.00
28	Benzo(q,h,i)perylene	1.162	1.257	-8.2	142	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0