

Data File : BM024401.D
Acq On : 31 Dec 2019 18:16
Operator : JU
Sample : K6457-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Jan 01 04:46:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 01 04:44:04 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	181008	20.00	ng/ul	0.00
18) Naphthalene-d8	10.17	136	732460	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.07	164	428356	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.83	188	837955	20.00	ng/ul	0.00
78) Chrysene-d12	21.05	240	796440	20.00	ng/ul	0.00
86) Perylene-d12	23.17	264	970487	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	10.17	131	686	0.05	ng/ul	-0.15
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
58) Fluorene-d10	0.00	176	0	0.00	ng/ul	
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
71) Anthracene-d10	16.83	188	837955	21.98	ng/ul	-0.09
79) Pyrene-d10	19.27	212	615	0.02	ng/ul	0.03
90) Benzo(a)pyrene-d12	23.17	264	936884	19.44	ng/ul	0.13

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM024401.D
Acq On : 31 Dec 2019 18:16
Operator : JU
Sample : K6457-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Jan 01 04:46:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 01 04:44:04 2020
Response via : Initial Calibration

