


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123119\
Data File : BM024387.D
Acq On : 31 Dec 2019 09:14
Operator : JU
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
Title : SVOA CALIBRATION
Last Update : Thu Jan 02 02:37:24 2020



AutoFind: Scans 2449, 2450, 2451; Background Corrected with Scan 2444

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.7	70216	PASS
68	69	0.00	2	1.8	1599	PASS
69	198	0.00	100	54.9	88210	PASS
70	69	0.00	2	0.5	439	PASS
127	198	10	80	56.6	90874	PASS
197	198	0.00	2	0.7	1092	PASS
198	198	100	100	100.0	160565	PASS
199	198	5	9	6.8	10971	PASS
275	198	10	60	30.3	48640	PASS
365	198	1	100	4.3	6900	PASS
441	198	0.01	100	8.3	13287	PASS
442	198	50	100	55.8	89626	PASS
443	442	15	24	19.7	17694	PASS

Scan 2444 (17.263 min): BM024387.D
DFTPP08

m/z	Abundance
40.00	1306.0
41.00	452.0
44.00	5218.0
49.10	339.0
73.00	1170.0
77.00	339.0
95.90	316.0
167.00	623.0
207.10	1864.0
280.90	560.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	374.0
38.00	911.0
39.10	4899.0
39.90	1403.0
41.00	381.0
43.00	372.0
44.00	4803.0
48.90	517.0
50.00	15373.0
51.00	57384.0
52.00	3312.0
55.00	676.0
55.90	1990.0
57.00	4453.0
61.00	944.0
61.90	932.0
63.00	2260.0
64.00	386.0
65.00	1254.0
68.10	1247.0
69.00	69944.0
70.00	435.0
73.00	1966.0
74.00	7189.0
75.00	11763.0
76.00	3881.0
77.00	75416.0
78.10	5811.0
79.00	6003.0
80.00	3855.0
81.00	5963.0
82.00	1442.0
83.00	1250.0
84.10	465.0
85.10	851.0
85.90	1661.0
87.00	775.0
87.90	379.0
91.00	1457.0
92.00	1557.0
93.00	8599.0
94.10	636.0
97.10	328.0
98.00	6333.0
99.00	6878.0
100.00	675.0
101.00	2329.0
103.00	776.0
104.00	1916.0
104.90	1674.0
106.00	842.0
107.00	21928.0
108.10	3172.0
109.00	588.0
110.00	37056.0
111.00	5834.0
112.00	855.0
116.00	1128.0
117.00	20152.0
118.00	1647.0
119.00	550.0
120.10	317.0
122.00	1456.0
123.00	1942.0
124.00	1274.0
125.00	882.0
125.90	427.0
127.00	66408.0
128.00	5358.0
129.00	27792.0
129.90	2487.0
131.00	567.0
132.20	394.0
133.00	588.0
134.00	801.0
135.00	2647.0
136.00	763.0
137.00	1772.0
139.90	437.0
141.00	3393.0
142.00	996.0
143.00	1052.0
146.00	645.0
147.00	2065.0
148.00	4606.0
149.00	1303.0
150.00	376.0
150.80	432.0
151.30	311.0
153.00	1161.0
153.90	831.0
155.00	2011.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

156.00	3154.0
157.20	517.0
157.90	589.0
158.90	476.0
160.00	1098.0
161.00	1635.0
162.00	408.0
163.00	338.0
163.90	325.0
165.00	1468.0
166.00	1359.0
167.00	7859.0
168.00	6645.0
168.90	924.0
170.80	380.0
172.10	537.0
173.00	993.0
174.00	1329.0
175.00	2382.0
176.00	620.0
176.90	1615.0
177.90	530.0
178.90	5780.0
180.00	4205.0
181.00	1451.0
183.90	369.0
185.00	2160.0
186.00	18712.0
187.00	4971.0
188.00	1074.0
188.90	1355.0
190.90	1028.0
192.00	1804.0
193.00	1694.0
193.90	509.0
194.90	425.0
196.00	3473.0
196.80	624.0
197.10	619.0
197.90	111904.0
198.90	7921.0
200.00	587.0
201.30	509.0
202.90	1059.0
204.00	4200.0
205.00	7549.0
206.00	32616.0
207.00	5888.0
208.10	1545.0
208.90	439.0
210.00	788.0
210.90	927.0
215.00	391.0
216.00	800.0
217.00	8411.0
217.90	1076.0
219.00	303.0
220.90	3873.0
221.80	1120.0
223.10	2055.0
224.00	17896.0
225.00	4620.0
226.00	851.0
227.00	8091.0
228.00	1288.0
229.00	1810.0
231.00	577.0
233.90	644.0
235.00	606.0
236.10	407.0
237.00	1362.0
239.00	478.0
240.90	613.0
242.00	945.0
243.00	1149.0
244.00	13765.0
245.10	1750.0
246.00	2982.0
246.90	666.0
249.00	601.0
254.10	819.0
255.00	72784.0
256.00	10938.0
256.90	1008.0
258.00	4604.0
259.00	923.0
265.00	1816.0
265.90	477.0
273.00	2269.0
274.00	5879.0
275.00	32104.0
276.00	4309.0
277.00	3058.0
278.00	672.0
281.00	811.0
282.00	363.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

283.00	472.0
283.90	404.0
285.10	413.0
293.00	686.0
296.00	8800.0
297.00	1566.0
303.00	1009.0
303.90	335.0
314.00	426.0
315.00	839.0
316.00	682.0
323.00	2523.0
324.00	628.0
327.00	587.0
328.00	409.0
334.00	1535.0
335.10	414.0
345.90	466.0
352.10	530.0
353.00	581.0
354.00	690.0
354.90	439.0
365.00	3802.0
366.00	648.0
372.00	1089.0
373.00	370.0
382.90	450.0
402.00	594.0
403.10	703.0
421.10	518.0
422.00	532.0
423.00	2961.0
424.00	1333.0
424.80	344.0
441.10	6087.0
442.10	43000.0
443.10	8668.0
444.00	1089.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.10	532.0
38.00	1330.0
39.00	6722.0
40.00	1151.0
41.10	448.0
43.00	368.0
44.00	5040.0
44.90	327.0
49.00	964.0
50.00	21232.0
51.00	82648.0
52.00	4146.0
53.00	333.0
54.90	668.0
56.00	2640.0
57.00	5664.0
58.00	492.0
61.00	1395.0
62.00	1751.0
63.00	3659.0
64.10	495.0
65.10	1623.0
68.10	2017.0
69.00	104608.0
69.90	515.0
73.00	2029.0
74.00	11121.0
75.00	15880.0
76.10	5250.0
77.00	115504.0
78.00	8597.0
79.00	7867.0
80.00	5698.0
81.00	8307.0
82.00	1792.0
83.10	1622.0
84.00	584.0
85.00	1416.0
86.00	2283.0
87.00	776.0
88.00	512.0
89.00	311.0
91.00	2246.0
92.00	2031.0
93.00	14116.0
94.00	1051.0
94.80	431.0
95.90	1025.0
97.10	346.0
98.00	10473.0
99.00	9710.0
100.00	733.0
101.00	4253.0
103.00	1466.0
104.00	3021.0
105.00	2769.0
106.00	720.0
107.00	33576.0
108.00	4778.0
109.10	885.0
110.00	54888.0
111.00	8790.0
112.00	1358.0
112.90	540.0
116.00	1611.0
117.00	29568.0
118.00	2493.0
118.90	464.0
120.00	529.0
122.00	2150.0
123.00	3504.0
124.00	1757.0
125.00	1126.0
126.20	709.0
127.00	108304.0
128.00	8540.0
129.00	42880.0
130.00	3583.0
131.00	845.0
132.30	467.0
132.90	428.0
133.90	1608.0
135.00	3956.0
135.90	1224.0
137.00	2379.0
137.90	414.0
139.90	545.0
141.00	5062.0
142.00	1397.0
143.10	1338.0
144.00	350.0
144.90	319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

146.00	1134.0
147.00	3097.0
148.00	7259.0
149.00	2143.0
150.00	424.0
151.00	807.0
151.90	396.0
153.00	1703.0
153.90	1431.0
155.00	3134.0
156.00	4434.0
157.00	923.0
158.00	1145.0
158.90	1107.0
159.90	1948.0
161.00	2635.0
161.90	846.0
164.90	2197.0
166.00	1556.0
167.00	13253.0
168.00	9042.0
168.90	1269.0
169.90	338.0
171.00	462.0
171.90	1246.0
173.00	1260.0
174.00	2619.0
175.00	4634.0
175.90	1092.0
176.90	2361.0
177.80	855.0
179.00	7287.0
180.00	6719.0
181.00	2663.0
182.00	510.0
183.90	596.0
185.00	3800.0
186.00	29336.0
187.00	8639.0
187.90	1491.0
189.00	2434.0
190.00	307.0
191.00	1237.0
192.00	2640.0
193.00	3561.0
194.00	532.0
195.00	399.0
196.00	5692.0
197.10	1139.0
197.90	186112.0
199.00	12437.0
200.00	910.0
201.60	871.0
203.00	1784.0
204.00	7509.0
205.00	13745.0
206.00	53440.0
207.00	9743.0
208.00	2635.0
209.00	881.0
209.90	866.0
211.10	1911.0
211.90	371.0
215.00	524.0
216.00	1206.0
217.00	14592.0
218.00	1794.0
221.00	7194.0
222.00	2070.0
223.00	4122.0
224.00	29976.0
225.00	7588.0
226.00	1138.0
227.00	13337.0
227.90	2005.0
229.00	2584.0
230.00	542.0
231.00	1045.0
233.00	312.0
234.00	913.0
235.00	927.0
236.00	803.0
237.00	1887.0
239.00	699.0
240.00	812.0
240.90	1200.0
242.10	1501.0
243.00	1745.0
244.00	23240.0
245.00	3358.0
246.00	4864.0
246.90	940.0
249.00	648.0
252.00	339.0
252.90	436.0
253.90	610.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

254.20	791.0
255.00	129664.0
256.00	19008.0
257.00	1658.0
258.00	8511.0
258.90	1336.0
260.00	321.0
265.00	3710.0
266.00	845.0
266.90	324.0
271.00	308.0
271.90	304.0
273.00	4171.0
274.00	9716.0
275.00	53872.0
276.00	7892.0
277.00	6136.0
278.00	1067.0
280.90	816.0
283.00	683.0
283.90	584.0
285.00	929.0
293.00	1703.0
294.00	453.0
294.90	502.0
296.00	18760.0
297.00	2884.0
303.00	2002.0
304.00	467.0
307.90	303.0
314.00	702.0
314.90	1586.0
316.00	1082.0
320.90	379.0
323.00	4454.0
324.10	997.0
327.00	1438.0
327.80	533.0
332.00	473.0
332.90	715.0
334.00	4071.0
335.10	742.0
341.00	788.0
346.00	1198.0
352.00	1532.0
353.00	1157.0
354.00	1603.0
355.10	368.0
365.00	7763.0
366.00	830.0
370.90	456.0
372.10	2374.0
373.10	457.0
383.00	683.0
402.00	1087.0
403.10	1326.0
403.90	480.0
421.00	919.0
422.10	891.0
423.10	5780.0
424.00	2779.0
425.10	541.0
441.10	14535.0
442.10	99552.0
443.10	19520.0
444.10	1858.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	411.0
38.00	1215.0
39.00	5374.0
40.00	1736.0
41.00	427.0
43.00	300.0
44.00	4858.0
45.10	385.0
49.00	615.0
50.00	18760.0
51.00	70616.0
52.00	3836.0
55.00	672.0
56.00	2394.0
57.00	5741.0
58.10	315.0
61.00	1091.0
61.90	1053.0
63.00	3012.0
63.90	353.0
65.00	1915.0
67.10	330.0
68.00	1535.0
69.00	90080.0
70.00	367.0
73.00	1818.0
74.00	10110.0
75.00	14392.0
76.00	5524.0
77.10	102112.0
78.10	7352.0
79.00	6893.0
80.00	5429.0
81.00	6936.0
82.00	1924.0
83.00	1595.0
84.00	561.0
85.00	1180.0
86.00	1897.0
86.90	649.0
88.00	390.0
89.00	355.0
91.00	1985.0
92.00	1886.0
93.00	12586.0
94.00	863.0
95.00	305.0
95.90	1216.0
98.00	9380.0
99.00	7992.0
99.90	562.0
101.00	3655.0
101.90	422.0
102.90	1130.0
104.00	2728.0
105.00	2400.0
105.90	676.0
107.00	29568.0
108.00	5227.0
109.10	1068.0
110.00	51200.0
111.00	7701.0
112.00	1071.0
113.00	350.0
116.10	1537.0
117.00	26592.0
118.00	2133.0
119.00	435.0
120.10	491.0
121.20	311.0
122.00	2474.0
123.00	3107.0
123.90	1395.0
125.00	1151.0
126.00	541.0
127.00	97912.0
128.00	7155.0
129.00	38376.0
130.00	3150.0
131.00	971.0
132.10	393.0
132.90	444.0
134.00	1571.0
134.90	3241.0
136.00	1311.0
137.00	1832.0
137.90	460.0
140.10	559.0
141.00	5485.0
142.00	1730.0
142.90	1363.0
144.10	399.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

146.00	810.0
147.00	3175.0
148.00	7777.0
149.00	1558.0
149.90	435.0
151.00	814.0
152.00	580.0
153.00	1876.0
154.00	1534.0
155.00	3055.0
156.00	4508.0
157.00	923.0
158.00	778.0
159.00	851.0
159.90	2036.0
161.00	2548.0
162.00	746.0
165.00	1854.0
166.00	1854.0
167.00	11312.0
168.00	7914.0
169.00	1064.0
169.90	455.0
170.90	452.0
171.90	1068.0
173.00	1484.0
174.00	2435.0
175.00	4139.0
176.00	1162.0
177.00	2058.0
178.00	856.0
179.00	8355.0
180.00	5619.0
180.90	2271.0
182.00	459.0
183.90	640.0
185.00	3751.0
186.00	29512.0
187.00	8130.0
187.90	1123.0
189.00	2211.0
191.00	1549.0
192.00	2763.0
193.00	2966.0
194.00	816.0
195.10	347.0
196.00	4220.0
197.20	1520.0
197.90	183680.0
198.90	12556.0
200.10	1092.0
201.30	572.0
201.70	612.0
203.00	1534.0
204.00	7763.0
205.10	13122.0
206.00	54848.0
207.00	9318.0
208.00	2699.0
209.00	1035.0
210.20	876.0
210.90	2147.0
211.90	432.0
214.90	628.0
216.00	1669.0
216.90	14754.0
218.00	2052.0
221.00	7116.0
221.90	1766.0
223.00	3281.0
224.00	28200.0
225.10	7334.0
226.00	1045.0
227.00	13898.0
228.00	2146.0
229.00	3028.0
230.00	325.0
231.00	1003.0
233.90	860.0
234.90	892.0
236.00	593.0
237.00	1997.0
238.00	310.0
238.90	448.0
240.10	646.0
240.90	655.0
242.00	1817.0
243.00	1896.0
244.00	24816.0
245.10	2854.0
246.00	5508.0
247.00	1236.0
249.00	911.0
253.00	849.0
254.00	990.0
255.00	138176.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

256.00	19168.0
257.00	1849.0
258.00	9298.0
258.90	1048.0
265.00	3921.0
265.90	808.0
271.00	328.0
272.00	365.0
273.00	4258.0
274.00	10221.0
275.00	59944.0
276.00	7278.0
277.00	5923.0
277.90	1497.0
281.00	707.0
282.00	325.0
283.00	973.0
284.10	544.0
285.00	1030.0
291.90	375.0
293.00	1410.0
293.90	481.0
296.00	21296.0
297.00	2985.0
302.00	322.0
303.00	2016.0
303.90	575.0
308.10	304.0
310.00	370.0
314.00	1266.0
315.00	2541.0
316.00	1192.0
321.00	771.0
323.10	5658.0
324.00	1032.0
326.90	1628.0
327.90	720.0
332.20	308.0
333.00	786.0
334.00	3593.0
335.10	1114.0
341.00	676.0
345.90	1363.0
352.00	1424.0
353.00	1355.0
354.00	1997.0
355.10	305.0
365.00	9135.0
365.90	783.0
371.00	665.0
372.10	2946.0
373.00	811.0
382.90	864.0
390.00	303.0
402.10	1177.0
403.10	1304.0
404.20	463.0
421.00	995.0
422.00	769.0
423.00	7898.0
424.10	2618.0
425.00	488.0
441.10	19240.0
442.10	126328.0
443.10	24896.0
444.00	2090.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08