

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010323\
 Data File : BP013375.D
 Acq On : 03 Jan 2023 11:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733

Quant Time: Jan 04 01:06:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	499409	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	2131837	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	1443209	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.334	188	3399647	20.000	ng/u1	0.00
79) Chrysene-d12	21.422	240	3559996	20.000	ng/u1	0.00
88) Perylene-d12	23.892	264	3660424	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.323	96	95612	8.186	ng/uL	0.00
4) Pyridine-d5	3.746	84	686623	20.694	ng/u1	0.00
7) Phenol-d5	7.093	99	819609	20.149	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	526815	21.469	ng/u1	0.00
11) 2-Chlorophenol-d4	7.458	132	646168	20.179	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	659128	19.776	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	314735	20.094	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	349907	19.843	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	694781	20.285	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	984871	20.368	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	2149411	19.379	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	2444035	20.053	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	389375	19.516	ng/u1	0.00
60) Fluorene-d10	15.569	176	1885906	20.098	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.693	200	401285	18.342	ng/u1	0.00
73) Anthracene-d10	17.434	188	3063703	19.975	ng/u1	0.00
81) Pyrene-d10	19.669	212	3826648	18.727	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	3611710	19.796	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	99574	8.166	ng/uL#	92
5) Pyridine	3.770	79	692174	20.990	ng/u1	92
6) Benzaldehyde	7.069	77	326671	20.509	ng/u1	96
8) Phenol	7.117	94	841386	20.457	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.352	93	688391	20.392	ng/u1	94
12) 2-Chlorophenol	7.493	128	657857	20.033	ng/u1	95
13) 2-Methylphenol	8.375	108	629311	19.626	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.458	45	1071720	23.136	ng/u1#	96
16) Acetophenone	8.764	105	1068283	20.903	ng/u1#	94
17) N-Nitroso-di-n-propyla...	8.746	70	554741	21.586	ng/u1	92
18) 4-Methylphenol	8.711	108	705345	19.860	ng/u1	100
19) Hexachloroethane	9.016	117	259721	19.792	ng/u1	98
22) Nitrobenzene	9.146	77	801988	21.820	ng/u1	96
23) Isophorone	9.669	82	1492503	20.082	ng/u1	97
25) 2-Nitrophenol	9.858	139	383325	20.063	ng/u1	94
26) 2,4-Dimethylphenol	9.916	107	773186	20.740	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.152	93	951252	20.018	ng/u1	99
29) 2,4-Dichlorophenol	10.393	162	682474	20.341	ng/u1	99
30) Naphthalene	10.793	128	2220467	19.916	ng/u1	99
32) 4-Chloroaniline	10.910	127	989325	20.676	ng/u1	98
33) Hexachlorobutadiene	11.075	225	470941	20.443	ng/u1	99
34) Caprolactam	11.693	113	114946	10.048	ng/u1	87
35) 4-Chloro-3-methylphenol	12.028	107	717951	20.968	ng/u1	100
36) 2-Methylnaphthalene	12.404	142	1525256	19.941	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010323\
 Data File : BP013375.D
 Acq On : 03 Jan 2023 11:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733


Quant Time: Jan 04 01:06:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	1568271	19.937	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.769	216	944825	20.331	ng/ul	99
40) Hexachlorocyclopentadiene	12.746	237	512704	16.954	ng/ul	99
41) 2,4,6-Trichlorophenol	13.010	196	601863	20.152	ng/ul	99
42) 2,4,5-Trichlorophenol	13.081	196	660175	20.579	ng/ul	98
43) 1,1'-Biphenyl	13.410	154	2155234	19.884	ng/ul	100
44) 2-Chloronaphthalene	13.451	162	1679561	19.688	ng/ul	100
45) 2-Nitroaniline	13.663	65	467839	22.619	ng/ul	92
47) Dimethylphthalate	14.034	163	2161780	19.664	ng/ul	100
48) 2,6-Dinitrotoluene	14.157	165	414003	20.264	ng/ul	96
50) Acenaphthylene	14.299	152	2617844	19.979	ng/ul	99
51) 3-Nitroaniline	14.487	138	346706	17.972	ng/ul	97
52) Acenaphthene	14.640	153	1803542	19.863	ng/ul	100
53) 2,4-Dinitrophenol	14.693	184	230903	16.534	ng/ul	92
55) 4-Nitrophenol	14.793	109	303109	22.265	ng/ul	91
56) Dibenzofuran	14.975	168	2612245	20.234	ng/ul	100
57) 2,4-Dinitrotoluene	14.946	165	617780	20.817	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.204	232	593838	20.284	ng/ul	95
59) Diethylphthalate	15.393	149	2056633	19.245	ng/ul	100
61) Fluorene	15.628	166	2144505	20.713	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.616	204	1136071	20.749	ng/ul	98
63) 4-Nitroaniline	15.651	138	318184	18.744	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.710	198	402063	18.806	ng/ul	98
67) N-Nitrosodiphenylamine	15.834	169	1804580	19.146	ng/ul	100
68) 4-Bromophenyl-phenylether	16.516	248	705530	19.110	ng/ul	97
69) Hexachlorobenzene	16.634	284	838572	19.129	ng/ul	98
70) Atrazine	16.793	200	699566	19.047	ng/ul	99
71) Pentachlorophenol	16.981	266	475515	17.911	ng/ul	99
72) Phenanthrene	17.381	178	3514109	19.845	ng/ul	100
74) Anthracene	17.469	178	3566350	20.171	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.375	216	942053	19.422	ng/ul	99
76) Pentachlorobenzene	14.893	250	956373	19.339	ng/ul	100
77) Carbazole	17.740	167	3154827	21.187	ng/ul	99
78) Di-n-butylphthalate	18.281	149	3440915	18.979	ng/ul	99
80) Fluoranthene	19.339	202	4401605	18.978	ng/ul	100
82) Pyrene	19.698	202	4613135	19.282	ng/ul	99
83) Butylbenzylphthalate	20.557	149	1626170	17.594	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.333	252	1561526	19.416	ng/ul	100
85) Benzo(a)anthracene	21.404	228	4745640	20.081	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.316	149	2376787	18.230	ng/ul	100
87) Chrysene	21.463	228	4437560	19.856	ng/ul	100
89) Di-n-octyl phthalate	22.263	149	3931183	19.170	ng/ul	100
90) Benzo(b)fluoranthene	23.133	252	4648868	19.917	ng/ul	100
91) Benzo(k)fluoranthene	23.186	252	4674590	20.196	ng/ul	100
93) Benzo(a)pyrene	23.780	252	4043925	19.897	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.468	276	4672509	18.457	ng/ul	98
95) Dibenzo(a,h)anthracene	26.486	278	3896546	18.590	ng/ul	99
96) Benzo(g,h,i)perylene	27.263	276	3654605	17.982	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 04 01:06:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 28 00:16:35 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.166 ng/uL
RT: 3.358 min Scan# 29
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 88 Resp: 99574
Ion Ratio Lower Upper
88 100
43 33.7 22.4 33.6#
58 78.3 58.2 87.4



#5
Pyridine
Concen: 20.990 ng/uL
RT: 3.770 min Scan# 99
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 79 Resp: 692174
Ion Ratio Lower Upper
79 100
52 82.9 60.7 91.1
51 39.9 28.9 43.3









#6
Benzaldehyde
Concen: 20.509 ng/ul
RT: 7.069 min Scan# 660
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 77 Resp: 326671
Ion Ratio Lower Upper
77 100
105 96.2 81.5 122.3
106 94.0 76.9 115.3



#8
Phenol
Concen: 20.457 ng/ul
RT: 7.117 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 94 Resp: 841386
Ion Ratio Lower Upper
94 100
65 29.5 21.1 31.7
66 38.1 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.392 ng/ul
RT: 7.352 min Scan# 708
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 93 Resp: 688391
Ion Ratio Lower Upper
93 100
63 82.1 59.8 89.8
95 33.2 26.4 39.6



#12
2-Chlorophenol
Concen: 20.033 ng/ul
RT: 7.493 min Scan# 732
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:128 Resp: 657857
Ion Ratio Lower Upper
128 100
64 48.7 34.9 52.3
130 33.3 25.8 38.6









#13
2-Methylphenol
Concen: 19.626 ng/ul
RT: 8.375 min Scan# 882
Delta R.T. -0.006 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:108 Resp: 629311
Ion Ratio Lower Upper
108 100
107 89.6 69.1 103.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.136 ng/ul
RT: 8.458 min Scan# 896
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 45 Resp: 1071720
Ion Ratio Lower Upper
45 100
77 12.6 11.0 16.6
79 9.2 9.3 13.9#









#16
Acetophenone
Concen: 20.903 ng/ul
RT: 8.764 min Scan# 948
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:105 Resp: 1068283
Ion Ratio Lower Upper
105 100
77 78.4 59.3 88.9
51 29.8 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 21.586 ng/ul
RT: 8.746 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 70 Resp: 554741
Ion Ratio Lower Upper
70 100
42 65.2 45.5 68.3
101 9.1 7.9 11.9
130 20.1 17.4 26.2









#18
4-Methylphenol
Concen: 19.860 ng/ul
RT: 8.711 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:108 Resp: 705345
Ion Ratio Lower Upper
108 100
107 109.3 87.4 131.2



#19
Hexachloroethane
Concen: 19.792 ng/ul
RT: 9.016 min Scan# 991
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:117 Resp: 259721
Ion Ratio Lower Upper
117 100
201 119.3 93.7 140.5
199 72.5 58.6 88.0









#22
Nitrobenzene
Concen: 21.820 ng/ul
RT: 9.146 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion: 77 Resp: 801988
Ion Ratio Lower Upper
77 100
123 44.6 38.3 57.5
65 13.8 10.3 15.5



#23
Isophorone
Concen: 20.082 ng/ul
RT: 9.669 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 82 Resp: 1492503
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.2
138 17.5 15.3 22.9









#25
2-Nitrophenol
Concen: 20.063 ng/ul
RT: 9.858 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:139 Resp: 383325
Ion Ratio Lower Upper
139 100
65 51.0 36.6 54.8
109 28.2 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.740 ng/ul
RT: 9.916 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:107 Resp: 773186
Ion Ratio Lower Upper
107 100
121 50.5 41.8 62.6
122 86.7 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.018 ng/ul
 RT: 10.152 min Scan# 1184
 Delta R.T. 0.000 min
 Lab File: BP013375.D
 Acq: 03 Jan 2023 11:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733


Tgt Ion: 93 Resp: 951252
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.0 39.0
 123 10.9 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.341 ng/ul
 RT: 10.393 min Scan# 1225
 Delta R.T. 0.000 min
 Lab File: BP013375.D
 Acq: 03 Jan 2023 11:35

Tgt Ion:162 Resp: 682474
 Ion Ratio Lower Upper
 162 100
 164 64.9 52.5 78.7
 98 34.0 26.6 39.8









#30
Naphthalene
Concen: 19.916 ng/ul
RT: 10.793 min Scan# 1293
Delta R.T. -0.006 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:128 Resp: 2220467
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 12.7 10.4 15.6



#32
4-Chloroaniline
Concen: 20.676 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:127 Resp: 989325
Ion Ratio Lower Upper
127 100
129 32.9 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.443 ng/ul
RT: 11.075 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:225 Resp: 470941
Ion Ratio Lower Upper
225 100
223 63.1 49.2 73.8
227 62.4 50.2 75.2



#34
Caprolactam
Concen: 10.048 ng/ul
RT: 11.693 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:113 Resp: 114946
Ion Ratio Lower Upper
113 100
55 196.0 141.9 212.9
56 139.6 99.6 149.4









#35
 4-Chloro-3-methylphenol
 Concen: 20.968 ng/ul
 RT: 12.028 min Scan# 1503
 Delta R.T. 0.000 min
 Lab File: BP013375.D
 Acq: 03 Jan 2023 11:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020733


Tgt Ion:107 Resp: 717951
 Ion Ratio Lower Upper
 107 100
 144 26.5 21.4 32.2
 142 82.0 65.8 98.6



#36
 2-Methylnaphthalene
 Concen: 19.941 ng/ul
 RT: 12.404 min Scan# 1567
 Delta R.T. 0.000 min
 Lab File: BP013375.D
 Acq: 03 Jan 2023 11:35

Tgt Ion:142 Resp: 1525256
 Ion Ratio Lower Upper
 142 100
 141 88.6 70.8 106.2









#37
1-Methylnaphthalene
Concen: 19.937 ng/ul
RT: 12.622 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:142 Resp: 1568271
Ion Ratio Lower Upper
142 100
141 94.2 72.7 109.1
116 3.8 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.331 ng/ul
RT: 12.769 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:216 Resp: 944825
Ion Ratio Lower Upper
216 100
214 78.1 61.9 92.9
179 19.6 15.0 22.6
108 14.4 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 16.954 ng/ul

RT: 12.746 min Scan# 1625

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:237 Resp: 512704

Ion Ratio Lower Upper

237 100

235 62.5 49.7 74.5

272 12.6 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 20.152 ng/ul

RT: 13.010 min Scan# 1670

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:196 Resp: 601863


Ion Ratio Lower Upper

196 100

198 97.4 77.0 115.6

200 31.5 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 20.579 ng/ul

RT: 13.081 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:196 Resp: 660175

Ion Ratio Lower Upper

196 100

198 94.4 77.1 115.7

200 31.5 25.4 38.2



#43

1,1'-Biphenyl

Concen: 19.884 ng/ul

RT: 13.410 min Scan# 1738

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:154 Resp: 2155234


Ion Ratio Lower Upper

154 100

153 39.4 31.5 47.3

76 12.6 9.6 14.4









#44
2-Chloronaphthalene
Concen: 19.688 ng/ul
RT: 13.451 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:162 Resp: 1679561
Ion Ratio Lower Upper
162 100
164 33.1 26.3 39.5
127 36.5 29.3 43.9



#45
2-Nitroaniline
Concen: 22.619 ng/ul
RT: 13.663 min Scan# 1781
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion: 65 Resp: 467839
Ion Ratio Lower Upper
65 100
92 66.6 56.4 84.6
138 102.9 90.3 135.5









#47
Dimethylphthalate
Concen: 19.664 ng/ul
RT: 14.034 min Scan# 1844
Delta R.T. 0.006 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:163 Resp: 2161780
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.264 ng/ul
RT: 14.157 min Scan# 1865
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:165 Resp: 414003
Ion Ratio Lower Upper
165 100
89 57.3 43.0 64.4
121 20.1 16.1 24.1









#50
Acenaphthylene
Concen: 19.979 ng/ul
RT: 14.299 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:152 Resp: 2617844
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.6
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 17.972 ng/ul
RT: 14.487 min Scan# 1921
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:138 Resp: 346706
Ion Ratio Lower Upper
138 100
108 10.6 8.6 13.0
92 113.5 88.4 132.6





#52

Acenaphthene

Concen: 19.863 ng/ul

RT: 14.640 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:153 Resp: 1803542

Ion Ratio Lower Upper

153 100

152 48.2 38.2 57.4

154 88.1 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 16.534 ng/ul

RT: 14.693 min Scan# 1956

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:184 Resp: 230903


Ion Ratio Lower Upper

184 100

63 68.9 48.4 72.6

154 61.0 45.5 68.3









#55
4-Nitrophenol
Concen: 22.265 ng/ul
RT: 14.793 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:109 Resp: 303109
Ion Ratio Lower Upper
109 100
139 135.7 120.8 181.2
65 141.0 107.9 161.9



#56
Dibenzofuran
Concen: 20.234 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:168 Resp: 2612245
Ion Ratio Lower Upper
168 100
139 34.7 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.817 ng/ul
RT: 14.946 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:165 Resp: 617780
Ion Ratio Lower Upper
165 100
63 44.5 31.8 47.6
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.284 ng/ul
RT: 15.204 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:232 Resp: 593838
Ion Ratio Lower Upper
232 100
131 35.9 31.8 47.6
130 2.1 1.8 2.6
166 26.5 22.9 34.3









#59
Diethylphthalate
Concen: 19.245 ng/ul
RT: 15.393 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:149 Resp: 2056633
Ion Ratio Lower Upper
149 100
177 23.1 18.4 27.6
150 12.6 9.9 14.9



#61
Fluorene
Concen: 20.713 ng/ul
RT: 15.628 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:166 Resp: 2144505
Ion Ratio Lower Upper
166 100
165 97.1 77.4 116.2
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.749 ng/ul
RT: 15.616 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:204 Resp: 1136071
Ion Ratio Lower Upper
204 100
206 33.2 26.3 39.5
141 53.0 40.4 60.6



#63
4-Nitroaniline
Concen: 18.744 ng/ul
RT: 15.651 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:138 Resp: 318184
Ion Ratio Lower Upper
138 100
92 52.3 39.0 58.4
108 80.9 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.806 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:198 Resp: 402063
Ion Ratio Lower Upper
198 100
182 4.5 3.3 4.9
77 21.8 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 19.146 ng/ul
RT: 15.834 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:169 Resp: 1804580
Ion Ratio Lower Upper
169 100
168 66.6 53.0 79.6
167 36.7 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.110 ng/ul
RT: 16.516 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:248 Resp: 705530
Ion Ratio Lower Upper
248 100
250 98.6 77.4 116.0
141 64.2 48.3 72.5



#69
Hexachlorobenzene
Concen: 19.129 ng/ul
RT: 16.634 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:284 Resp: 838572
Ion Ratio Lower Upper
284 100
142 30.0 22.6 34.0
249 30.6 24.6 37.0





#70

Atrazine

Concen: 19.047 ng/ul

RT: 16.793 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:200 Resp: 699566

Ion Ratio Lower Upper

200 100

173 25.3 20.5 30.7

215 51.1 39.9 59.9



#71

Pentachlorophenol

Concen: 17.911 ng/ul

RT: 16.981 min Scan# 2345

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:266 Resp: 475515


Ion Ratio Lower Upper

266 100

264 64.5 50.9 76.3

268 64.5 51.4 77.2





#72

Phenanthrene

Concen: 19.845 ng/ul

RT: 17.381 min Scan# 2413

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:178 Resp: 3514109

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 19.1 15.4 23.0



#74

Anthracene

Concen: 20.171 ng/ul

RT: 17.469 min Scan# 2428

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:178 Resp: 3566350


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.3 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.422 ng/uL
RT: 13.375 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD020733


Tgt Ion:216 Resp: 942053
Ion Ratio Lower Upper
216 100
214 76.6 62.2 93.2
218 48.3 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.339 ng/uL
RT: 14.893 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP013375.D
Acq: 03 Jan 2023 11:35

Tgt Ion:250 Resp: 956373
Ion Ratio Lower Upper
250 100
248 63.8 50.6 76.0
252 63.8 51.1 76.7





#77

Carbazole

Concen: 21.187 ng/ul

RT: 17.740 min Scan# 2474

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:167 Resp: 3154827

Ion Ratio Lower Upper

167 100

166 21.5 17.2 25.8

139 12.5 9.4 14.2



#78

Di-n-butylphthalate

Concen: 18.979 ng/ul

RT: 18.281 min Scan# 2566

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:149 Resp: 3440915


Ion Ratio Lower Upper

149 100

150 9.2 7.5 11.3

104 5.1 3.8 5.6





#80

Fluoranthene

Concen: 18.978 ng/ul

RT: 19.339 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:202 Resp: 4401605

Ion Ratio Lower Upper

202 100

101 10.1 8.2 12.2

100 7.9 6.1 9.1



#82

Pyrene

Concen: 19.282 ng/ul

RT: 19.698 min Scan# 2807

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:202 Resp: 4613135


Ion Ratio Lower Upper

202 100

101 11.4 9.4 14.2

100 9.2 7.8 11.8





#83

Butylbenzylphthalate

Concen: 17.594 ng/ul

RT: 20.557 min Scan# 2953

Delta R.T. -0.006 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:149 Resp: 1626170

Ion Ratio Lower Upper

149 100

91 72.6 53.4 80.0

206 22.2 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 19.416 ng/ul

RT: 21.333 min Scan# 3085

Delta R.T. -0.006 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:252 Resp: 1561526


Ion Ratio Lower Upper

252 100

254 64.6 51.8 77.8

126 9.8 7.4 11.2





#85

Benzo(a)anthracene

Concen: 20.081 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:228 Resp: 4745640

Ion Ratio Lower Upper

228 100

229 20.0 16.1 24.1

226 27.3 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.230 ng/ul

RT: 21.316 min Scan# 3082

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:149 Resp: 2376787


Ion Ratio Lower Upper

149 100

167 28.3 22.5 33.7

279 5.0 3.8 5.8





#87

Chrysene

Concen: 19.856 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:228 Resp: 4437560

Ion Ratio Lower Upper

228 100

226 30.8 24.5 36.7

229 20.0 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 19.170 ng/ul




RT: 22.263 min Scan# 3243


Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Tgt Ion:149 Resp: 3931183





#90

Benzo(b)fluoranthene

Concen: 19.917 ng/ul

RT: 23.133 min Scan# 3391

Delta R.T. -0.006 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:252 Resp: 4648868

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 8.5 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.196 ng/ul

RT: 23.186 min Scan# 3400

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:252 Resp: 4674590


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.0 6.8 10.2





#93

Benzo(a)pyrene

Concen: 19.897 ng/ul

RT: 23.780 min Scan# 3501

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:252 Resp: 4043925

Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 9.3 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.457 ng/ul

RT: 26.468 min Scan# 3958

Delta R.T. -0.006 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35




Tgt Ion:276 Resp: 4672509


Ion Ratio Lower Upper

276 100

138 20.0 17.1 25.7

277 25.3 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 18.590 ng/ul

RT: 26.486 min Scan# 3961

Delta R.T. -0.006 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD020733





Tgt Ion:278 Resp: 3896546

Ion Ratio Lower Upper

278 100

139 13.0 11.1 16.7

279 24.0 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 17.982 ng/ul

RT: 27.263 min Scan# 4093

Delta R.T. 0.000 min

Lab File: BP013375.D

Acq: 03 Jan 2023 11:35

Tgt Ion:276 Resp: 3654605

Ion Ratio Lower Upper

276 100

138 17.8 15.4 23.2

277 23.0 19.1 28.7

