

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010323\
 Data File : BP013382.D
 Acq On : 03 Jan 2023 17:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020734

Quant Time: Jan 04 01:10:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	550845	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	2446063	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	1672235	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	3851614	20.000	ng/u1	0.00
79) Chrysene-d12	21.416	240	3395070	20.000	ng/u1	0.00
88) Perylene-d12	23.874	264	2727208	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	100883	7.831	ng/uL	0.00
4) Pyridine-d5	3.746	84	739713	20.212	ng/u1	0.00
7) Phenol-d5	7.087	99	909034	20.260	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.252	67	592890	21.905	ng/u1	0.00
11) 2-Chlorophenol-d4	7.457	132	714989	20.243	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	743813	20.232	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	363527	20.228	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	409912	20.260	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.357	165	800064	20.358	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	1124520	20.269	ng/u1	0.00
46) Dimethylphthalate-d6	13.981	166	2533625	19.714	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	2854981	20.216	ng/u1	0.00
54) 4-Nitrophenol-d4	14.775	143	437144	18.909	ng/u1	0.00
60) Fluorene-d10	15.563	176	2222081	20.437	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	469301	18.934	ng/u1	0.00
73) Anthracene-d10	17.427	188	3448710	19.847	ng/u1	0.00
81) Pyrene-d10	19.657	212	4115155	21.117	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.715	264	2712516	19.954	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	107872	8.021	ng/uL	93
5) Pyridine	3.764	79	757693	20.831	ng/u1	91
6) Benzaldehyde	7.063	77	365770	20.819	ng/u1	95
8) Phenol	7.116	94	930668	20.515	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.346	93	776442	20.853	ng/u1	94
12) 2-Chlorophenol	7.487	128	728838	20.122	ng/u1	95
13) 2-Methylphenol	8.369	108	709650	20.065	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.463	45	1228080	24.036	ng/u1	97
16) Acetophenone	8.757	105	1225300	21.737	ng/u1#	92
17) N-Nitroso-di-n-propyla...	8.740	70	636421	22.452	ng/u1	90
18) 4-Methylphenol	8.704	108	801197	20.452	ng/u1	99
19) Hexachloroethane	9.010	117	292340	20.197	ng/u1	100
22) Nitrobenzene	9.140	77	921504	21.851	ng/u1	96
23) Isophorone	9.663	82	1766426	20.715	ng/u1	97
25) 2-Nitrophenol	9.851	139	444891	20.294	ng/u1	92
26) 2,4-Dimethylphenol	9.910	107	887348	20.745	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.146	93	1106861	20.300	ng/u1	100
29) 2,4-Dichlorophenol	10.387	162	781018	20.288	ng/u1	99
30) Naphthalene	10.787	128	2535548	19.820	ng/u1	100
32) 4-Chloroaniline	10.904	127	1136084	20.693	ng/u1	99
33) Hexachlorobutadiene	11.069	225	534012	20.203	ng/u1	96
34) Caprolactam	11.687	113	142467	10.854	ng/u1	86
35) 4-Chloro-3-methylphenol	12.022	107	829940	21.125	ng/u1	99
36) 2-Methylnaphthalene	12.398	142	1778729	20.268	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010323\
 Data File : BP013382.D
 Acq On : 03 Jan 2023 17:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020734

Quant Time: Jan 04 01:10:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	1822006	20.188	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.763	216	1091358	20.268	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	605308	17.275	ng/ul	99
41) 2,4,6-Trichlorophenol	13.004	196	698467	20.184	ng/ul	99
42) 2,4,5-Trichlorophenol	13.075	196	751225	20.210	ng/ul	98
43) 1,1'-Biphenyl	13.404	154	2492781	19.848	ng/ul	99
44) 2-Chloronaphthalene	13.445	162	1971474	19.945	ng/ul	100
45) 2-Nitroaniline	13.657	65	557940	23.281	ng/ul	92
47) Dimethylphthalate	14.028	163	2514147	19.737	ng/ul	99
48) 2,6-Dinitrotoluene	14.151	165	489492	20.678	ng/ul	99
50) Acenaphthylene	14.292	152	3041726	20.034	ng/ul	100
51) 3-Nitroaniline	14.481	138	403200	18.038	ng/ul	94
52) Acenaphthene	14.634	153	2103468	19.993	ng/ul	100
53) 2,4-Dinitrophenol	14.686	184	275017	16.996	ng/ul	95
55) 4-Nitrophenol	14.786	109	336668	21.343	ng/ul	90
56) Dibenzofuran	14.969	168	3018025	20.176	ng/ul	100
57) 2,4-Dinitrotoluene	14.939	165	726652	21.132	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.198	232	682971	20.133	ng/ul	96
59) Diethylphthalate	15.386	149	2427486	19.604	ng/ul	100
61) Fluorene	15.622	166	2502474	20.860	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.610	204	1326409	20.907	ng/ul	98
63) 4-Nitroaniline	15.645	138	371108	18.867	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.704	198	459929	18.988	ng/ul	97
67) N-Nitrosodiphenylamine	15.828	169	2104785	19.711	ng/ul	100
68) 4-Bromophenyl-phenylether	16.510	248	833121	19.918	ng/ul	100
69) Hexachlorobenzene	16.628	284	971265	19.556	ng/ul	98
70) Atrazine	16.780	200	799407	19.211	ng/ul	98
71) Pentachlorophenol	16.975	266	547247	18.195	ng/ul	99
72) Phenanthrene	17.369	178	4009735	19.987	ng/ul	100
74) Anthracene	17.463	178	4042673	20.182	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.369	216	1076496	19.590	ng/ul	100
76) Pentachlorobenzene	14.886	250	1106467	19.749	ng/ul	99
77) Carbazole	17.733	167	3516811	20.847	ng/ul	99
78) Di-n-butylphthalate	18.275	149	3921146	19.090	ng/ul	99
80) Fluoranthene	19.333	202	4743269	21.445	ng/ul	99
82) Pyrene	19.686	202	4889536	21.430	ng/ul	99
83) Butylbenzylphthalate	20.551	149	1675120	19.005	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.327	252	1380202	17.995	ng/ul	99
85) Benzo(a)anthracene	21.398	228	4517021	20.042	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.304	149	2306540	18.551	ng/ul	98
87) Chrysene	21.451	228	4206559	19.736	ng/ul	100
89) Di-n-octyl phthalate	22.251	149	3425049	22.417	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	3689924	21.218	ng/ul	100
91) Benzo(k)fluoranthene	23.174	252	3707057	21.497	ng/ul	99
93) Benzo(a)pyrene	23.768	252	3046241	20.117	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.451	276	3326335	17.635	ng/ul	99
95) Dibenzo(a,h)anthracene	26.462	278	2782181	17.815	ng/ul	100
96) Benzo(g,h,i)perylene	27.239	276	2681581	17.709	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010323\
 Data File : BP013382.D
 Acq On : 03 Jan 2023 17:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020734

Quant Time: Jan 04 01:10:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.021 ng/uL
RT: 3.352 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion: 88 Resp: 107872
Ion Ratio Lower Upper
88 100
43 32.9 22.4 33.6
58 77.7 58.2 87.4



#5
Pyridine
Concen: 20.831 ng/uL
RT: 3.764 min Scan# 98
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion: 79 Resp: 757693
Ion Ratio Lower Upper
79 100
52 85.3 60.7 91.1
51 38.8 28.9 43.3









#6
Benzaldehyde
Concen: 20.819 ng/ul
RT: 7.063 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion: 77 Resp: 365770
Ion Ratio Lower Upper
77 100
105 95.7 81.5 122.3
106 93.4 76.9 115.3



#8
Phenol
Concen: 20.515 ng/ul
RT: 7.116 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion: 94 Resp: 930668
Ion Ratio Lower Upper
94 100
65 29.8 21.1 31.7
66 37.4 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.853 ng/ul
RT: 7.346 min Scan# 708
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion: 93 Resp: 776442
Ion Ratio Lower Upper
93 100
63 81.6 59.8 89.8
95 32.8 26.4 39.6



#12
2-Chlorophenol
Concen: 20.122 ng/ul
RT: 7.487 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:128 Resp: 728838
Ion Ratio Lower Upper
128 100
64 49.5 34.9 52.3
130 32.3 25.8 38.6





#13

2-Methylphenol

Concen: 20.065 ng/ul

RT: 8.369 min Scan# 881

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :





SSTD020734

Tgt Ion:108 Resp: 709650

Ion Ratio Lower Upper

108 100

107 87.3 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.036 ng/ul

RT: 8.463 min Scan# 897

Delta R.T. 0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion: 45 Resp: 1228080


Ion Ratio Lower Upper

45 100

77 12.9 11.0 16.6

79 10.2 9.3 13.9









#16
Acetophenone
Concen: 21.737 ng/ul
RT: 8.757 min Scan# 948
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:105 Resp: 1225300
Ion Ratio Lower Upper
105 100
77 79.5 59.3 88.9
51 30.5 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 22.452 ng/ul
RT: 8.740 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion: 70 Resp: 636421
Ion Ratio Lower Upper
70 100
42 67.5 45.5 68.3
101 9.5 7.9 11.9
130 21.2 17.4 26.2





#18

4-Methylphenol

Concen: 20.452 ng/ul

RT: 8.704 min Scan# 91

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :





SSTD020734

Tgt Ion:108 Resp: 801197

Ion Ratio Lower Upper

108 100

107 108.0 87.4 131.2



#19

Hexachloroethane

Concen: 20.197 ng/ul

RT: 9.010 min Scan# 990

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:117 Resp: 292340


Ion Ratio Lower Upper

117 100

201 117.2 93.7 140.5

199 73.5 58.6 88.0









#22
Nitrobenzene
Concen: 21.851 ng/ul
RT: 9.140 min Scan# 1101
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion: 77 Resp: 921504
Ion Ratio Lower Upper
77 100
123 44.7 38.3 57.5
65 13.8 10.3 15.5



#23
Isophorone
Concen: 20.715 ng/ul
RT: 9.663 min Scan# 1101
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion: 82 Resp: 1766426
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.1 15.3 22.9









#25
2-Nitrophenol
Concen: 20.294 ng/ul
RT: 9.851 min Scan# 1133
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:139 Resp: 444891
Ion Ratio Lower Upper
139 100
65 52.2 36.6 54.8
109 30.2 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.745 ng/ul
RT: 9.910 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:107 Resp: 887348
Ion Ratio Lower Upper
107 100
121 51.4 41.8 62.6
122 86.2 72.6 108.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.300 ng/ul

RT: 10.146 min Scan# 1183

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion: 93 Resp: 1106861

Ion Ratio Lower Upper

93 100

95 32.6 26.0 39.0

123 11.4 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 20.288 ng/ul

RT: 10.387 min Scan# 1224

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:162 Resp: 781018


Ion Ratio Lower Upper

162 100

164 66.2 52.5 78.7

98 33.9 26.6 39.8









#30
Naphthalene
Concen: 19.820 ng/ul
RT: 10.787 min Scan# 1292
Delta R.T. -0.012 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:128 Resp: 2535548
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 20.693 ng/ul
RT: 10.904 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:127 Resp: 1136084
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4





#33

Hexachlorobutadiene

Concen: 20.203 ng/ul

RT: 11.069 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:225 Resp: 534012

Ion Ratio Lower Upper

225 100

223 63.6 49.2 73.8

227 66.5 50.2 75.2



#34

Caprolactam

Concen: 10.854 ng/ul

RT: 11.687 min Scan# 1445

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:113 Resp: 142467


Ion Ratio Lower Upper

113 100

55 195.1 141.9 212.9

56 141.4 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 21.125 ng/ul
RT: 12.022 min Scan# 1502
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:107 Resp: 829940
Ion Ratio Lower Upper
107 100
144 26.3 21.4 32.2
142 81.9 65.8 98.6



#36
2-Methylnaphthalene
Concen: 20.268 ng/ul
RT: 12.398 min Scan# 1566
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:142 Resp: 1778729
Ion Ratio Lower Upper
142 100
141 88.2 70.8 106.2





#37

1-Methylnaphthalene

Concen: 20.188 ng/ul

RT: 12.616 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:142 Resp: 1822006

Ion Ratio Lower Upper

142 100

141 91.3 72.7 109.1

116 3.8 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.268 ng/ul

RT: 12.763 min Scan# 1628

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Tgt Ion:216 Resp: 1091358




Ion Ratio Lower Upper


216 100

214 76.3 61.9 92.9

179 19.0 15.0 22.6

108 15.0 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 17.275 ng/ul

RT: 12.740 min Scan# 1625

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:237 Resp: 605308

Ion Ratio Lower Upper

237 100

235 63.2 49.7 74.5

272 13.7 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 20.184 ng/ul

RT: 13.004 min Scan# 1669

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:196 Resp: 698467


Ion Ratio Lower Upper

196 100

198 97.2 77.0 115.6

200 30.8 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 20.210 ng/ul
RT: 13.075 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:196 Resp: 751225
Ion Ratio Lower Upper
196 100
198 97.7 77.1 115.7
200 30.4 25.4 38.2



#43
1,1'-Biphenyl
Concen: 19.848 ng/ul
RT: 13.404 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:154 Resp: 2492781
Ion Ratio Lower Upper
154 100
153 39.9 31.5 47.3
76 12.6 9.6 14.4









#44
2-Chloronaphthalene
Concen: 19.945 ng/ul
RT: 13.445 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:162 Resp: 1971474
Ion Ratio Lower Upper
162 100
164 32.8 26.3 39.5
127 36.3 29.3 43.9



#45
2-Nitroaniline
Concen: 23.281 ng/ul
RT: 13.657 min Scan# 1780
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion: 65 Resp: 557940
Ion Ratio Lower Upper
65 100
92 66.6 56.4 84.6
138 102.4 90.3 135.5









#47
Dimethylphthalate
Concen: 19.737 ng/ul
RT: 14.028 min Scan# 1843
Delta R.T. -0.000 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:163 Resp: 2514147
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.0
164 10.0 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.678 ng/ul
RT: 14.151 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:165 Resp: 489492
Ion Ratio Lower Upper
165 100
89 54.9 43.0 64.4
121 20.2 16.1 24.1









#50
Acenaphthylene
Concen: 20.034 ng/ul
RT: 14.292 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:152 Resp: 3041726
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 18.038 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:138 Resp: 403200
Ion Ratio Lower Upper
138 100
108 11.0 8.6 13.0
92 117.6 88.4 132.6









#52
Acenaphthene
Concen: 19.993 ng/ul
RT: 14.634 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:153 Resp: 2103468
Ion Ratio Lower Upper
153 100
152 47.4 38.2 57.4
154 88.4 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 16.996 ng/ul
RT: 14.686 min Scan# 1955
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:184 Resp: 275017
Ion Ratio Lower Upper
184 100
63 67.3 48.4 72.6
154 57.9 45.5 68.3









#55
4-Nitrophenol
Concen: 21.343 ng/ul
RT: 14.786 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:109 Resp: 336668
Ion Ratio Lower Upper
109 100
139 136.0 120.8 181.2
65 143.5 107.9 161.9



#56
Dibenzofuran
Concen: 20.176 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:168 Resp: 3018025
Ion Ratio Lower Upper
168 100
139 34.8 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 21.132 ng/ul
RT: 14.939 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:165 Resp: 726652
Ion Ratio Lower Upper
165 100
63 42.8 31.8 47.6
182 3.0 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.133 ng/ul
RT: 15.198 min Scan# 2042
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:232 Resp: 682971
Ion Ratio Lower Upper
232 100
131 37.1 31.8 47.6
130 1.9 1.8 2.6
166 27.1 22.9 34.3









#59
Diethylphthalate
Concen: 19.604 ng/ul
RT: 15.386 min Scan# 2075
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:149 Resp: 2427486
Ion Ratio Lower Upper
149 100
177 23.1 18.4 27.6
150 12.4 9.9 14.9



#61
Fluorene
Concen: 20.860 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:166 Resp: 2502474
Ion Ratio Lower Upper
166 100
165 97.4 77.4 116.2
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.907 ng/ul
RT: 15.610 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:204 Resp: 1326409
Ion Ratio Lower Upper
204 100
206 33.0 26.3 39.5
141 52.1 40.4 60.6



#63
4-Nitroaniline
Concen: 18.867 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:138 Resp: 371108
Ion Ratio Lower Upper
138 100
92 52.3 39.0 58.4
108 79.7 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.988 ng/ul
RT: 15.704 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:198 Resp: 459929
Ion Ratio Lower Upper
198 100
182 4.2 3.3 4.9
77 22.6 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 19.711 ng/ul
RT: 15.828 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:169 Resp: 2104785
Ion Ratio Lower Upper
169 100
168 66.3 53.0 79.6
167 36.6 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.918 ng/ul
RT: 16.510 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:248 Resp: 833121
Ion Ratio Lower Upper
248 100
250 96.7 77.4 116.0
141 60.6 48.3 72.5



#69
Hexachlorobenzene
Concen: 19.556 ng/ul
RT: 16.628 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:284 Resp: 971265
Ion Ratio Lower Upper
284 100
142 30.0 22.6 34.0
249 31.6 24.6 37.0





#70

Atrazine

Concen: 19.211 ng/ul

RT: 16.780 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:200 Resp: 799407

Ion Ratio Lower Upper

200 100

173 25.1 20.5 30.7

215 48.5 39.9 59.9



#71

Pentachlorophenol

Concen: 18.195 ng/ul

RT: 16.975 min Scan# 2344

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:266 Resp: 547247


Ion Ratio Lower Upper

266 100

264 64.0 50.9 76.3

268 65.2 51.4 77.2





#72

Phenanthrene

Concen: 19.987 ng/ul

RT: 17.369 min Scan# 2411

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:178 Resp: 4009735

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.0 15.4 23.0



#74

Anthracene

Concen: 20.182 ng/ul

RT: 17.463 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:178 Resp: 4042673


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.4 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.590 ng/uL
RT: 13.369 min Scan# 1732
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SSTD020734


Tgt Ion:216 Resp: 1076496
Ion Ratio Lower Upper
216 100
214 78.2 62.2 93.2
218 48.6 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.749 ng/uL
RT: 14.886 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP013382.D
Acq: 03 Jan 2023 17:31

Tgt Ion:250 Resp: 1106467
Ion Ratio Lower Upper
250 100
248 63.9 50.6 76.0
252 63.5 51.1 76.7





#77

Carbazole

Concen: 20.847 ng/ul

RT: 17.733 min Scan# 2473

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:167 Resp: 3516811

Ion Ratio Lower Upper

167 100

166 21.6 17.2 25.8

139 12.3 9.4 14.2



#78

Di-n-butylphthalate

Concen: 19.090 ng/ul

RT: 18.275 min Scan# 2565

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:149 Resp: 3921146


Ion Ratio Lower Upper

149 100

150 9.2 7.5 11.3

104 4.9 3.8 5.6





#80

Fluoranthene

Concen: 21.445 ng/ul

RT: 19.333 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:202 Resp: 4743269

Ion Ratio Lower Upper

202 100

101 9.7 8.2 12.2

100 7.6 6.1 9.1



#82

Pyrene

Concen: 21.430 ng/ul

RT: 19.686 min Scan# 2805

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:202 Resp: 4889536


Ion Ratio Lower Upper

202 100

101 12.4 9.4 14.2

100 10.2 7.8 11.8





#83

Butylbenzylphthalate

Concen: 19.005 ng/ul

RT: 20.551 min Scan# 2953

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:149 Resp: 1675120

Ion Ratio Lower Upper

149 100

91 69.3 53.4 80.0

206 22.9 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.995 ng/ul

RT: 21.327 min Scan# 3084

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:252 Resp: 1380202


Ion Ratio Lower Upper

252 100

254 63.9 51.8 77.8

126 9.4 7.4 11.2





#85

Benzo(a)anthracene

Concen: 20.042 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:228 Resp: 4517021

Ion Ratio Lower Upper

228 100

229 20.0 16.1 24.1

226 27.6 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.551 ng/ul

RT: 21.304 min Scan# 3080

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:149 Resp: 2306540


Ion Ratio Lower Upper

149 100

167 27.0 22.5 33.7

279 4.2 3.8 5.8





#87

Chrysene

Concen: 19.736 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:228 Resp: 4206559

Ion Ratio Lower Upper

228 100

226 30.5 24.5 36.7

229 19.6 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 22.417 ng/ul




RT: 22.251 min Scan# 3241


Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Tgt Ion:149 Resp: 3425049





#90

Benzo(b)fluoranthene

Concen: 21.218 ng/ul

RT: 23.121 min Scan# 3389

Delta R.T. -0.018 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:252 Resp: 3689924

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.3 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 21.497 ng/ul

RT: 23.174 min Scan# 3398

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:252 Resp: 3707057


Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 8.0 6.8 10.2





#93

Benzo(a)pyrene

Concen: 20.117 ng/ul

RT: 23.768 min Scan# 34

Delta R.T. -0.012 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:252 Resp: 3046241

Ion Ratio Lower Upper

252 100

253 21.1 17.5 26.3

125 9.1 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.635 ng/ul

RT: 26.451 min Scan# 3955

Delta R.T. -0.024 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31




Tgt Ion:276 Resp: 3326335


Ion Ratio Lower Upper

276 100

138 20.8 17.1 25.7

277 25.3 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 17.815 ng/ul

RT: 26.462 min Scan# 3957

Delta R.T. -0.029 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SSTD020734





Tgt Ion:278 Resp: 2782181

Ion Ratio Lower Upper

278 100

139 13.8 11.1 16.7

279 23.6 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 17.709 ng/ul

RT: 27.239 min Scan# 4089

Delta R.T. -0.024 min

Lab File: BP013382.D

Acq: 03 Jan 2023 17:31

Tgt Ion:276 Resp: 2681581

Ion Ratio Lower Upper

276 100

138 17.6 15.4 23.2

277 24.1 19.1 28.7

