

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013384.D
 Acq On : 04 Jan 2023 09:54
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735

Quant Time: Jan 05 04:04:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	492532	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	2212582	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	1540529	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	3551465	20.000	ng/u1	0.00
79) Chrysene-d12	21.410	240	3235235	20.000	ng/u1	0.00
88) Perylene-d12	23.874	264	2770297	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	92299	8.012	ng/uL	0.00
4) Pyridine-d5	3.746	84	678702	20.741	ng/u1	0.00
7) Phenol-d5	7.087	99	816003	20.340	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.252	67	539053	22.274	ng/u1	0.00
11) 2-Chlorophenol-d4	7.452	132	641490	20.312	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	673964	20.503	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	328495	20.207	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	365605	19.977	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.358	165	724320	20.376	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	1021174	20.348	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	2306098	19.478	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	2588866	19.899	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	400565	18.808	ng/u1	0.00
60) Fluorene-d10	15.563	176	2010212	20.069	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	423264	18.520	ng/u1	0.00
73) Anthracene-d10	17.428	188	3192563	19.926	ng/u1	0.00
81) Pyrene-d10	19.657	212	3819372	20.567	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.716	264	2734543	19.804	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	97388	8.099	ng/uL#	88
5) Pyridine	3.764	79	684561	21.049	ng/u1	86
6) Benzaldehyde	7.064	77	332279	21.152	ng/u1	93
8) Phenol	7.117	94	840907	20.731	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.346	93	704401	21.158	ng/u1	94
12) 2-Chlorophenol	7.487	128	655477	20.240	ng/u1	94
13) 2-Methylphenol	8.369	108	644465	20.379	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.458	45	1130959	24.756	ng/u1	96
16) Acetophenone	8.758	105	1116902	22.160	ng/u1#	94
17) N-Nitroso-di-n-propyla...	8.740	70	584587	23.065	ng/u1	92
18) 4-Methylphenol	8.699	108	731650	20.888	ng/u1	97
19) Hexachloroethane	9.011	117	263536	20.363	ng/u1	98
22) Nitrobenzene	9.134	77	839985	22.020	ng/u1	93
23) Isophorone	9.663	82	1611058	20.886	ng/u1	97
25) 2-Nitrophenol	9.852	139	398167	20.079	ng/u1	92
26) 2,4-Dimethylphenol	9.911	107	809996	20.935	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.146	93	1018432	20.649	ng/u1	99
29) 2,4-Dichlorophenol	10.381	162	722349	20.744	ng/u1	98
30) Naphthalene	10.787	128	2306427	19.932	ng/u1	100
32) 4-Chloroaniline	10.905	127	1037092	20.883	ng/u1	100
33) Hexachlorobutadiene	11.069	225	491272	20.547	ng/u1	98
34) Caprolactam	11.687	113	132306	11.143	ng/u1	84
35) 4-Chloro-3-methylphenol	12.022	107	757986	21.330	ng/u1	100
36) 2-Methylnaphthalene	12.393	142	1613471	20.325	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013384.D
 Acq On : 04 Jan 2023 09:54
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735

Quant Time: Jan 05 04:04:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	1656872	20.295	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.763	216	995233	20.063	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	541440	16.773	ng/ul	99
41) 2,4,6-Trichlorophenol	13.004	196	640252	20.083	ng/ul	98
42) 2,4,5-Trichlorophenol	13.075	196	696838	20.349	ng/ul	99
43) 1,1'-Biphenyl	13.399	154	2291843	19.809	ng/ul	98
44) 2-Chloronaphthalene	13.446	162	1807466	19.849	ng/ul	100
45) 2-Nitroaniline	13.652	65	506022	22.920	ng/ul	88
47) Dimethylphthalate	14.022	163	2293681	19.546	ng/ul	100
48) 2,6-Dinitrotoluene	14.151	165	436187	20.001	ng/ul	99
50) Acenaphthylene	14.293	152	2792623	19.966	ng/ul	100
51) 3-Nitroaniline	14.481	138	361285	17.545	ng/ul	95
52) Acenaphthene	14.634	153	1947930	20.097	ng/ul	99
53) 2,4-Dinitrophenol	14.687	184	240818	16.155	ng/ul	93
55) 4-Nitrophenol	14.787	109	314331	21.631	ng/ul	91
56) Dibenzofuran	14.969	168	2788332	20.234	ng/ul	100
57) 2,4-Dinitrotoluene	14.934	165	645469	20.376	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.193	232	628661	20.117	ng/ul	99
59) Diethylphthalate	15.387	149	2203512	19.316	ng/ul	100
61) Fluorene	15.622	166	2304816	20.855	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.610	204	1223825	20.940	ng/ul	100
63) 4-Nitroaniline	15.646	138	331434	18.291	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.698	198	421410	18.869	ng/ul	94
67) N-Nitrosodiphenylamine	15.828	169	1922489	19.525	ng/ul	99
68) 4-Bromophenyl-phenylether	16.510	248	751548	19.486	ng/ul	99
69) Hexachlorobenzene	16.622	284	888314	19.397	ng/ul	97
70) Atrazine	16.781	200	737143	19.212	ng/ul	99
71) Pentachlorophenol	16.975	266	500559	18.049	ng/ul	99
72) Phenanthrene	17.369	178	3688287	19.939	ng/ul	99
74) Anthracene	17.463	178	3733716	20.215	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.369	216	1000142	19.738	ng/ul	100
76) Pentachlorobenzene	14.887	250	1016476	19.676	ng/ul	99
77) Carbazole	17.734	167	3243777	20.853	ng/ul	100
78) Di-n-butylphthalate	18.275	149	3619919	19.113	ng/ul	99
80) Fluoranthene	19.334	202	4393363	20.844	ng/ul	99
82) Pyrene	19.686	202	4599337	21.154	ng/ul	100
83) Butylbenzylphthalate	20.551	149	1551375	18.470	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.328	252	1347535	18.437	ng/ul	99
85) Benzo(a)anthracene	21.398	228	4297024	20.007	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.304	149	2191349	18.495	ng/ul	99
87) Chrysene	21.451	228	4076397	20.071	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	3322456	21.408	ng/ul	100
90) Benzo(b)fluoranthene	23.122	252	3747212	21.212	ng/ul	100
91) Benzo(k)fluoranthene	23.169	252	3583444	20.457	ng/ul	100
93) Benzo(a)pyrene	23.763	252	3078094	20.011	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.445	276	3468241	18.102	ng/ul	99
95) Dibenzo(a,h)anthracene	26.457	278	2913342	18.365	ng/ul	100
96) Benzo(g,h,i)perylene	27.233	276	2789595	18.136	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
Data File : BP013384.D
Acq On : 04 Jan 2023 09:54
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020735

Quant Time: Jan 05 04:04:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 04:02:47 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.099 ng/uL
RT: 3.352 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 88 Resp: 97388
Ion Ratio Lower Upper
88 100
43 34.3 22.4 33.6#
58 83.3 58.2 87.4



#5
Pyridine
Concen: 21.049 ng/uL
RT: 3.764 min Scan# 98
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion: 79 Resp: 684561
Ion Ratio Lower Upper
79 100
52 89.9 60.7 91.1
51 42.0 28.9 43.3









#6
Benzaldehyde
Concen: 21.152 ng/ul
RT: 7.064 min Scan# 659
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 77 Resp: 332279
Ion Ratio Lower Upper
77 100
105 91.5 81.5 122.3
106 91.8 76.9 115.3



#8
Phenol
Concen: 20.731 ng/ul
RT: 7.117 min Scan# 668
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion: 94 Resp: 840907
Ion Ratio Lower Upper
94 100
65 29.7 21.1 31.7
66 37.0 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.158 ng/ul
RT: 7.346 min Scan# 707
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 93 Resp: 704401
Ion Ratio Lower Upper
93 100
63 81.9 59.8 89.8
95 32.5 26.4 39.6



#12
2-Chlorophenol
Concen: 20.240 ng/ul
RT: 7.487 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion: 128 Resp: 655477
Ion Ratio Lower Upper
128 100
64 50.3 34.9 52.3
130 32.8 25.8 38.6





#13

2-Methylphenol

Concen: 20.379 ng/ul

RT: 8.369 min Scan# 881

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :





SSTD020735

Tgt Ion:108 Resp: 644465

Ion Ratio Lower Upper

108 100

107 89.7 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.756 ng/ul

RT: 8.458 min Scan# 896

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion: 45 Resp: 1130959


Ion Ratio Lower Upper

45 100

77 12.7 11.0 16.6

79 9.7 9.3 13.9









#16
Acetophenone
Concen: 22.160 ng/ul
RT: 8.758 min Scan# 947
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:105 Resp: 1116902
Ion Ratio Lower Upper
105 100
77 77.0 59.3 88.9
51 30.9 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 23.065 ng/ul
RT: 8.740 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion: 70 Resp: 584587
Ion Ratio Lower Upper
70 100
42 65.2 45.5 68.3
101 9.4 7.9 11.9
130 20.0 17.4 26.2









#18
4-Methylphenol
Concen: 20.888 ng/ul
RT: 8.699 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:108 Resp: 731650
Ion Ratio Lower Upper
108 100
107 112.7 87.4 131.2



#19
Hexachloroethane
Concen: 20.363 ng/ul
RT: 9.011 min Scan# 990
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:117 Resp: 263536
Ion Ratio Lower Upper
117 100
201 119.7 93.7 140.5
199 73.1 58.6 88.0









#22
Nitrobenzene
Concen: 22.020 ng/ul
RT: 9.134 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion: 77 Resp: 839985
Ion Ratio Lower Upper
77 100
123 42.4 38.3 57.5
65 13.2 10.3 15.5



#23
Isophorone
Concen: 20.886 ng/ul
RT: 9.663 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion: 82 Resp: 1611058
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.3 15.3 22.9









#25
2-Nitrophenol
Concen: 20.079 ng/ul
RT: 9.852 min Scan# 1133
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:139 Resp: 398167
Ion Ratio Lower Upper
139 100
65 52.4 36.6 54.8
109 28.7 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.935 ng/ul
RT: 9.911 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:107 Resp: 809996
Ion Ratio Lower Upper
107 100
121 50.1 41.8 62.6
122 86.1 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.649 ng/ul
 RT: 10.146 min Scan# 1183
 Delta R.T. 0.000 min
 Lab File: BP013384.D
 Acq: 04 Jan 2023 09:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735


Tgt Ion: 93 Resp: 1018432
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.0 39.0
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.744 ng/ul
 RT: 10.381 min Scan# 1223
 Delta R.T. 0.000 min
 Lab File: BP013384.D
 Acq: 04 Jan 2023 09:54

Tgt Ion:162 Resp: 722349
 Ion Ratio Lower Upper
 162 100
 164 64.0 52.5 78.7
 98 34.0 26.6 39.8









#30
Naphthalene
Concen: 19.932 ng/ul
RT: 10.787 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:128 Resp: 2306427
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 20.883 ng/ul
RT: 10.905 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:127 Resp: 1037092
Ion Ratio Lower Upper
127 100
129 32.2 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.547 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:225 Resp: 491272
Ion Ratio Lower Upper
225 100
223 60.6 49.2 73.8
227 64.9 50.2 75.2



#34
Caprolactam
Concen: 11.143 ng/ul
RT: 11.687 min Scan# 1445
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:113 Resp: 132306
Ion Ratio Lower Upper
113 100
55 197.6 141.9 212.9
56 145.0 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 21.330 ng/ul
RT: 12.022 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:107 Resp: 757986
Ion Ratio Lower Upper
107 100
144 27.5 21.4 32.2
142 82.2 65.8 98.6



#36
2-Methylnaphthalene
Concen: 20.325 ng/ul
RT: 12.393 min Scan# 1565
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:142 Resp: 1613471
Ion Ratio Lower Upper
142 100
141 89.2 70.8 106.2





#37

1-Methylnaphthalene

Concen: 20.295 ng/ul

RT: 12.616 min Scan# 1603

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:142 Resp: 1656872

Ion Ratio Lower Upper

142 100

141 92.7 72.7 109.1

116 3.7 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.063 ng/ul

RT: 12.763 min Scan# 1628

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Tgt Ion:216 Resp: 995233




Ion Ratio Lower Upper


216 100

214 78.6 61.9 92.9

179 19.2 15.0 22.6

108 14.7 11.0 16.4





#40

Hexachlorocyclopentadiene

Concen: 16.773 ng/ul

RT: 12.740 min Scan# 1624

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:237 Resp: 541440

Ion Ratio Lower Upper

237 100

235 63.0 49.7 74.5

272 13.0 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 20.083 ng/ul

RT: 13.004 min Scan# 1669

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:196 Resp: 640252


Ion Ratio Lower Upper

196 100

198 98.4 77.0 115.6

200 31.9 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 20.349 ng/ul

RT: 13.075 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:196 Resp: 696838

Ion Ratio Lower Upper

196 100

198 96.9 77.1 115.7

200 31.5 25.4 38.2



#43

1,1'-Biphenyl

Concen: 19.809 ng/ul

RT: 13.399 min Scan# 1736

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:154 Resp: 2291843


Ion Ratio Lower Upper

154 100

153 40.2 31.5 47.3

76 13.4 9.6 14.4





#44

2-Chloronaphthalene

Concen: 19.849 ng/ul

RT: 13.446 min Scan# 1744

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:162 Resp: 1807466

Ion Ratio Lower Upper

162 100

164 33.0 26.3 39.5

127 36.8 29.3 43.9



#45

2-Nitroaniline

Concen: 22.920 ng/ul

RT: 13.652 min Scan# 1779

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion: 65 Resp: 506022


Ion Ratio Lower Upper

65 100

92 64.9 56.4 84.6

138 96.8 90.3 135.5









#47
Dimethylphthalate
Concen: 19.546 ng/ul
RT: 14.022 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:163 Resp: 2293681
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.3 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.001 ng/ul
RT: 14.151 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:165 Resp: 436187
Ion Ratio Lower Upper
165 100
89 52.8 43.0 64.4
121 19.4 16.1 24.1









#50
Acenaphthylene
Concen: 19.966 ng/ul
RT: 14.293 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:152 Resp: 2792623
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 17.545 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:138 Resp: 361285
Ion Ratio Lower Upper
138 100
108 10.8 8.6 13.0
92 116.5 88.4 132.6









#52
Acenaphthene
Concen: 20.097 ng/ul
RT: 14.634 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:153 Resp: 1947930
Ion Ratio Lower Upper
153 100
152 47.1 38.2 57.4
154 87.5 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 16.155 ng/ul
RT: 14.687 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:184 Resp: 240818
Ion Ratio Lower Upper
184 100
63 69.3 48.4 72.6
154 58.7 45.5 68.3









#55
4-Nitrophenol
Concen: 21.631 ng/ul
RT: 14.787 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:109 Resp: 314331
Ion Ratio Lower Upper
109 100
139 131.8 120.8 181.2
65 137.3 107.9 161.9



#56
Dibenzofuran
Concen: 20.234 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:168 Resp: 2788332
Ion Ratio Lower Upper
168 100
139 34.7 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 20.376 ng/ul
RT: 14.934 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:165 Resp: 645469
Ion Ratio Lower Upper
165 100
63 46.5 31.8 47.6
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.117 ng/ul
RT: 15.193 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:232 Resp: 628661
Ion Ratio Lower Upper
232 100
131 38.8 31.8 47.6
130 2.1 1.8 2.6
166 28.5 22.9 34.3









#59
Diethylphthalate
Concen: 19.316 ng/ul
RT: 15.387 min Scan# 2074
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:149 Resp: 2203512
Ion Ratio Lower Upper
149 100
177 23.2 18.4 27.6
150 12.5 9.9 14.9



#61
Fluorene
Concen: 20.855 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:166 Resp: 2304816
Ion Ratio Lower Upper
166 100
165 97.4 77.4 116.2
167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.940 ng/ul
RT: 15.610 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:204 Resp: 1223825
Ion Ratio Lower Upper
204 100
206 33.1 26.3 39.5
141 50.8 40.4 60.6



#63
4-Nitroaniline
Concen: 18.291 ng/ul
RT: 15.646 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:138 Resp: 331434
Ion Ratio Lower Upper
138 100
92 52.5 39.0 58.4
108 81.8 55.1 82.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.869 ng/ul
 RT: 15.698 min Scan# 2127
 Delta R.T. 0.000 min
 Lab File: BP013384.D
 Acq: 04 Jan 2023 09:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020735


Tgt Ion:198 Resp: 421410
 Ion Ratio Lower Upper
 198 100
 182 4.6 3.3 4.9
 77 24.3 16.6 25.0



#67
 N-Nitrosodiphenylamine
 Concen: 19.525 ng/ul
 RT: 15.828 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BP013384.D
 Acq: 04 Jan 2023 09:54

Tgt Ion:169 Resp: 1922489
 Ion Ratio Lower Upper
 169 100
 168 66.9 53.0 79.6
 167 36.3 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 19.486 ng/ul
RT: 16.510 min Scan# 2265
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:248 Resp: 751548
Ion Ratio Lower Upper
248 100
250 98.1 77.4 116.0
141 60.1 48.3 72.5



#69
Hexachlorobenzene
Concen: 19.397 ng/ul
RT: 16.622 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:284 Resp: 888314
Ion Ratio Lower Upper
284 100
142 30.7 22.6 34.0
249 30.2 24.6 37.0





#70

Atrazine

Concen: 19.212 ng/ul

RT: 16.781 min Scan# 2311

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:200 Resp: 737143

Ion Ratio Lower Upper

200 100

173 24.8 20.5 30.7

215 49.7 39.9 59.9



#71

Pentachlorophenol

Concen: 18.049 ng/ul

RT: 16.975 min Scan# 2344

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:266 Resp: 500559


Ion Ratio Lower Upper

266 100

264 62.9 50.9 76.3

268 63.9 51.4 77.2





#72

Phenanthrene

Concen: 19.939 ng/ul

RT: 17.369 min Scan# 2411

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:178 Resp: 3688287

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 18.9 15.4 23.0



#74

Anthracene

Concen: 20.215 ng/ul

RT: 17.463 min Scan# 2427

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:178 Resp: 3733716


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.3 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.738 ng/uL
RT: 13.369 min Scan# 1731
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:216 Resp: 1000142
Ion Ratio Lower Upper
216 100
214 77.6 62.2 93.2
218 48.7 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.676 ng/uL
RT: 14.887 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:250 Resp: 1016476
Ion Ratio Lower Upper
250 100
248 63.6 50.6 76.0
252 64.9 51.1 76.7





#77

Carbazole

Concen: 20.853 ng/ul

RT: 17.734 min Scan# 2473

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:167 Resp: 3243777

Ion Ratio Lower Upper

167 100

166 21.6 17.2 25.8

139 12.0 9.4 14.2



#78

Di-n-butylphthalate

Concen: 19.113 ng/ul

RT: 18.275 min Scan# 2565

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:149 Resp: 3619919


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 4.9 3.8 5.6





#80

Fluoranthene

Concen: 20.844 ng/ul

RT: 19.334 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:202 Resp: 4393363

Ion Ratio Lower Upper

202 100

101 9.6 8.2 12.2

100 7.6 6.1 9.1



#82

Pyrene

Concen: 21.154 ng/ul

RT: 19.686 min Scan# 2805

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:202 Resp: 4599337


Ion Ratio Lower Upper

202 100

101 11.8 9.4 14.2

100 10.0 7.8 11.8





#83

Butylbenzylphthalate

Concen: 18.470 ng/ul

RT: 20.551 min Scan# 2952

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:149 Resp: 1551375

Ion Ratio Lower Upper

149 100

91 70.4 53.4 80.0

206 23.3 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.437 ng/ul

RT: 21.328 min Scan# 3084

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:252 Resp: 1347535


Ion Ratio Lower Upper

252 100

254 63.8 51.8 77.8

126 8.8 7.4 11.2









#85
Benzo(a)anthracene
Concen: 20.007 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54




Instrument :
BNA_P
ClientSampleId :
SSTD020735


Tgt Ion:228 Resp: 4297024
Ion Ratio Lower Upper
228 100
229 20.0 16.1 24.1
226 27.3 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.495 ng/ul
RT: 21.304 min Scan# 3080
Delta R.T. 0.000 min
Lab File: BP013384.D
Acq: 04 Jan 2023 09:54

Tgt Ion:149 Resp: 2191349
Ion Ratio Lower Upper
149 100
167 27.8 22.5 33.7
279 4.4 3.8 5.8





#87

Chrysene

Concen: 20.071 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:228 Resp: 4076397

Ion Ratio Lower Upper

228 100

226 30.1 24.5 36.7

229 19.4 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 21.408 ng/ul




RT: 22.251 min Scan# 3241


Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Tgt Ion:149 Resp: 3322456





#90

Benzo(b)fluoranthene

Concen: 21.212 ng/ul

RT: 23.122 min Scan# 3389

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:252 Resp: 3747212

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 8.3 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.457 ng/ul

RT: 23.169 min Scan# 3397

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:252 Resp: 3583444


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.4 6.8 10.2





#93

Benzo(a)pyrene

Concen: 20.011 ng/ul

RT: 23.763 min Scan# 3498

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:252 Resp: 3078094

Ion Ratio Lower Upper

252 100

253 22.1 17.5 26.3

125 9.5 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.102 ng/ul

RT: 26.445 min Scan# 3954

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54




Tgt Ion:276 Resp: 3468241


Ion Ratio Lower Upper

276 100

138 20.5 17.1 25.7

277 24.9 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 18.365 ng/ul

RT: 26.457 min Scan# 3956

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Instrument :

BNA_P

ClientSampleId :

SSTD020735





Tgt Ion:278 Resp: 2913342

Ion Ratio Lower Upper

278 100

139 13.9 11.1 16.7

279 23.9 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 18.136 ng/ul

RT: 27.233 min Scan# 4088

Delta R.T. 0.000 min

Lab File: BP013384.D

Acq: 04 Jan 2023 09:54

Tgt Ion:276 Resp: 2789595

Ion Ratio Lower Upper

276 100

138 18.0 15.4 23.2

277 23.5 19.1 28.7

