

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013390.D
 Acq On : 04 Jan 2023 14:49
 Operator : CG/JU
 Sample : PB 149982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Jan 05 04:07:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	589021	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	2456627	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.569	164	1619075	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.327	188	3671486	20.000	ng/ul	0.00
79) Chrysene-d12	21.415	240	3071121	20.000	ng/ul	0.00
88) Perylene-d12	23.880	264	2693520	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	80387	5.835	ng/uL	0.00
4) Pyridine-d5	3.746	84	1108545	28.327	ng/ul	0.00
7) Phenol-d5	7.087	99	1308178	27.267	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.252	67	872989	30.164	ng/ul	0.00
11) 2-Chlorophenol-d4	7.457	132	1051885	27.851	ng/ul	0.00
15) 4-Methylphenol-d8	8.634	113	1031713	26.245	ng/ul	0.00
21) Nitrobenzene-d5	9.093	128	502023	27.814	ng/ul	0.00
24) 2-Nitrophenol-d4	9.816	143	566342	27.871	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	1016563	25.756	ng/ul	0.00
31) 4-Chloroaniline-d4	10.875	131	1443767	25.911	ng/ul	0.00
46) Dimethylphthalate-d6	13.975	166	3349947	26.922	ng/ul	0.00
49) Acenaphthylene-d8	14.263	160	3692805	27.008	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	463233	20.696	ng/ul	0.00
60) Fluorene-d10	15.563	176	2967088	28.185	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	492930	20.863	ng/ul	0.00
73) Anthracene-d10	17.427	188	4666383	28.172	ng/ul	0.00
81) Pyrene-d10	19.657	212	5451583	30.925	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.721	264	3828009	28.513	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	120	0.008	ng/ul#	16
5) Pyridine	3.752	79	522	0.013	ng/ul#	1
6) Benzaldehyde	7.057	77	691	0.037	ng/ul#	22
8) Phenol	7.122	94	540	0.011	ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.357	93	397	0.010	ng/ul#	71
12) 2-Chlorophenol	7.487	128	576	0.015	ng/ul#	6
13) 2-Methylphenol	8.381	108	127	0.003	ng/ul	87
14) 2,2'-oxybis(1-Chloropr...	8.469	45	68	0.001	ng/ul#	1
16) Acetophenone	8.763	105	659	0.011	ng/ul#	1
17) N-Nitroso-di-n-propyla...	8.751	70	121	0.004	ng/ul#	1
18) 4-Methylphenol	8.634	108	32	0.001	ng/ul	97
19) Hexachloroethane	9.028	117	145	0.009	ng/ul#	1
22) Nitrobenzene	9.140	77	1410	0.033	ng/ul#	54
23) Isophorone	9.663	82	564	0.007	ng/ul#	63
25) 2-Nitrophenol	9.851	139	248	0.011	ng/ul#	1
26) 2,4-Dimethylphenol	9.904	107	411	0.010	ng/ul#	77
27) Bis(2-Chloroethoxy)met...	10.157	93	638	0.012	ng/ul#	85
29) 2,4-Dichlorophenol	10.363	162	345	0.009	ng/ul#	1
30) Naphthalene	10.787	128	1158	0.009	ng/ul#	89
32) 4-Chloroaniline	10.804	127	71	0.001	ng/ul	92
33) Hexachlorobutadiene	11.063	225	195	0.007	ng/ul#	20
34) Caprolactam	11.692	113	166	0.013	ng/ul#	66
35) 4-Chloro-3-methylphenol	11.981	107	102	0.003	ng/ul	89
36) 2-Methylnaphthalene	12.428	142	170	0.002	ng/ul	94


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013390.D
 Acq On : 04 Jan 2023 14:49
 Operator : CG/JU
 Sample : PB 149982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Jan 05 04:07:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	669	0.007	ng/ul#	76
39) 1,2,4,5-Tetrachloroben...	12.763	216	634	0.012	ng/ul#	49
40) Hexachlorocyclopentadiene	12.722	237	18	0.001	ng/ul#	57
41) 2,4,6-Trichlorophenol	12.969	196	24	0.001	ng/ul	95
42) 2,4,5-Trichlorophenol	13.086	196	324	0.009	ng/ul#	75
43) 1,1'-Biphenyl	13.410	154	911	0.007	ng/ul#	66
44) 2-Chloronaphthalene	13.451	162	744	0.008	ng/ul#	60
45) 2-Nitroaniline	13.651	65	151	0.007	ng/ul#	36
47) Dimethylphthalate	14.022	163	1298	0.011	ng/ul#	47
48) 2,6-Dinitrotoluene	14.157	165	112	0.005	ng/ul#	61
50) Acenaphthylene	14.286	152	1678	0.011	ng/ul#	60
51) 3-Nitroaniline	14.486	138	151	0.007	ng/ul#	58
52) Acenaphthene	14.639	153	945	0.009	ng/ul	97
53) 2,4-Dinitrophenol	14.704	184	201	0.013	ng/ul	94
55) 4-Nitrophenol	14.792	109	297	0.019	ng/ul#	1
56) Dibenzofuran	15.086	168	166	0.001	ng/ul	88
57) 2,4-Dinitrotoluene	14.933	165	651	0.020	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.192	232	426	0.013	ng/ul#	1
59) Diethylphthalate	15.386	149	1629	0.014	ng/ul#	84
61) Fluorene	15.616	166	1635	0.014	ng/ul#	82
62) 4-Chlorophenyl-phenyle...	15.610	204	603	0.010	ng/ul#	62
63) 4-Nitroaniline	15.645	138	68	0.004	ng/ul#	4
66) 4,6-Dinitro-2-methylph...	15.686	198	2173	0.094	ng/ul#	1
67) N-Nitrosodiphenylamine	15.828	169	1029	0.010	ng/ul	94
68) 4-Bromophenyl-phenylether	16.516	248	199	0.005	ng/ul#	36
69) Hexachlorobenzene	16.627	284	678	0.014	ng/ul#	74
70) Atrazine	16.786	200	199	0.005	ng/ul#	89
71) Pentachlorophenol	16.992	266	88	0.003	ng/ul	92
72) Phenanthrene	17.369	178	2372	0.012	ng/ul#	86
74) Anthracene	17.469	178	1906	0.010	ng/ul#	87
75) 1,2,3,4-Tetrachloroben...	13.434	216	13	0.000	ng/uL	87
76) Pentachlorobenzene	14.881	250	474	0.009	ng/ul#	91
77) Carbazole	17.745	167	1241	0.008	ng/ul#	95
78) Di-n-butylphthalate	18.274	149	3788	0.019	ng/ul#	88
80) Fluoranthene	19.327	202	2317	0.012	ng/ul#	64
82) Pyrene	19.686	202	2393	0.012	ng/ul#	48
83) Butylbenzylphthalate	20.551	149	1070	0.013	ng/ul#	72
84) 3,3'-Dichlorobenzidine	21.327	252	288	0.004	ng/ul#	53
85) Benzo(a)anthracene	21.410	228	10835	0.053	ng/ul#	77
86) Bis(2-ethylhexyl)phtha...	21.304	149	3599	0.032	ng/ul#	83
87) Chrysene	21.451	228	3575	0.019	ng/ul#	87
89) Di-n-octyl phthalate	22.251	149	4057	0.027	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	4329	0.025	ng/ul#	33
91) Benzo(k)fluoranthene	23.174	252	3022	0.018	ng/ul#	1
93) Benzo(a)pyrene	23.768	252	1734	0.012	ng/ul#	1
94) Indeno(1,2,3-cd)pyrene	26.450	276	776	0.004	ng/ul#	55
95) Dibenzo(a,h)anthracene	26.468	278	903	0.006	ng/ul#	1
96) Benzo(g,h,i)perylene	27.227	276	475	0.003	ng/ul#	59





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 0.008 ng/uL
RT: 3.358 min Scan# 29
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 88 Resp: 120
Ion Ratio Lower Upper
88 100
43 98.4 22.4 33.6#
58 18.6 58.2 87.4#



#5
Pyridine
Concen: 0.013 ng/uL
RT: 3.752 min Scan# 96
Delta R.T. -0.012 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion: 79 Resp: 522
Ion Ratio Lower Upper
79 100
52 19728.3 60.7 91.1#
51 712.3 28.9 43.3#









#6
Benzaldehyde
Concen: 0.037 ng/ul
RT: 7.057 min Scan# 658
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 77 Resp: 691
Ion Ratio Lower Upper
77 100
105 18.9 81.5 122.3#
106 24.7 76.9 115.3#



#8
Phenol
Concen: 0.011 ng/ul
RT: 7.122 min Scan# 669
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion: 94 Resp: 540
Ion Ratio Lower Upper
94 100
65 76.2 21.1 31.7#
66 205.8 27.0 40.6#









#10
Bis(2-Chloroethyl)ether
Concen: 0.010 ng/ul
RT: 7.357 min Scan# 709
Delta R.T. 0.011 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 93 Resp: 397
Ion Ratio Lower Upper
93 100
63 49.8 59.8 89.8#
95 49.0 26.4 39.6#



#12
2-Chlorophenol
Concen: 0.015 ng/ul
RT: 7.487 min Scan# 731
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion: 128 Resp: 576
Ion Ratio Lower Upper
128 100
64 121.0 34.9 52.3#
130 0.0 25.8 38.6#





#13

2-Methylphenol

Concen: 0.003 ng/ul

RT: 8.381 min Scan# 881

Delta R.T. 0.011 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :





SBLK982

Tgt Ion:108 Resp: 127

Ion Ratio Lower Upper

108 100

107 74.3 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.001 ng/ul

RT: 8.469 min Scan# 898

Delta R.T. 0.011 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion: 45 Resp: 68


Ion Ratio Lower Upper

45 100

77 133.8 11.0 16.6#

79 90.3 9.3 13.9#









#16
Acetophenone
Concen: 0.011 ng/ul
RT: 8.763 min Scan# 947
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 105 Resp: 659
Ion Ratio Lower Upper
105 100
77 172.9 59.3 88.9#
51 253.8 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 0.004 ng/ul
RT: 8.751 min Scan# 946
Delta R.T. 0.011 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion: 70 Resp: 121
Ion Ratio Lower Upper
70 100
42 211.1 45.5 68.3#
101 35.2 7.9 11.9#
130 22.7 17.4 26.2









#18
4-Methylphenol
Concen: 0.001 ng/ul
RT: 8.634 min Scan# 91
Delta R.T. -0.065 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:108 Resp: 32
Ion Ratio Lower Upper
108 100
107 106.5 87.4 131.2



#19
Hexachloroethane
Concen: 0.009 ng/ul
RT: 9.028 min Scan# 993
Delta R.T. 0.017 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:117 Resp: 145
Ion Ratio Lower Upper
117 100
201 0.0 93.7 140.5#
199 0.0 58.6 88.0#





#22

Nitrobenzene

Concen: 0.033 ng/ul

RT: 9.140 min Scan# 1101

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion: 77 Resp: 1410

Ion Ratio Lower Upper

77 100

123 18.7 38.3 57.5#

65 34.8 10.3 15.5#



#23

Isophorone

Concen: 0.007 ng/ul

RT: 9.663 min Scan# 1101

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion: 82 Resp: 564


Ion Ratio Lower Upper

82 100

95 19.1 5.4 8.2#

138 2.4 15.3 22.9#









#25
2-Nitrophenol
Concen: 0.011 ng/ul
RT: 9.851 min Scan# 1133
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:139 Resp: 248
Ion Ratio Lower Upper
139 100
65 312.5 36.6 54.8#
109 59.8 21.6 32.4#



#26
2,4-Dimethylphenol
Concen: 0.010 ng/ul
RT: 9.904 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:107 Resp: 411
Ion Ratio Lower Upper
107 100
121 39.6 41.8 62.6#
122 65.8 72.6 108.8#









#27
Bis(2-Chloroethoxy)methane
Concen: 0.012 ng/ul
RT: 10.157 min Scan# 1185
Delta R.T. 0.011 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 93 Resp: 638
Ion Ratio Lower Upper
93 100
95 38.8 26.0 39.0
123 20.9 9.2 13.8#



#29
2,4-Dichlorophenol
Concen: 0.009 ng/ul
RT: 10.363 min Scan# 1220
Delta R.T. -0.018 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:162 Resp: 345
Ion Ratio Lower Upper
162 100
164 10257.0 52.5 78.7#
98 757.8 26.6 39.8#









#30
Naphthalene
Concen: 0.009 ng/ul
RT: 10.787 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:128 Resp: 1158
Ion Ratio Lower Upper
128 100
129 8.2 8.7 13.1#
127 18.6 10.4 15.6#



#32
4-Chloroaniline
Concen: 0.001 ng/ul
RT: 10.804 min Scan# 1295
Delta R.T. -0.100 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:127 Resp: 71
Ion Ratio Lower Upper
127 100
129 27.5 25.6 38.4









#33
Hexachlorobutadiene
Concen: 0.007 ng/ul
RT: 11.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:225 Resp: 195
Ion Ratio Lower Upper
225 100
223 127.8 49.2 73.8#
227 5.3 50.2 75.2#



#34
Caprolactam
Concen: 0.013 ng/ul
RT: 11.692 min Scan# 1446
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:113 Resp: 166
Ion Ratio Lower Upper
113 100
55 110.4 141.9 212.9#
56 107.1 99.6 149.4









#35
 4-Chloro-3-methylphenol
 Concen: 0.003 ng/ul
 RT: 11.981 min Scan# 1495
 Delta R.T. -0.042 min
 Lab File: BP013390.D
 Acq: 04 Jan 2023 14:49




Instrument :
 BNA_P
 ClientSampleId :
 SBLK982


Tgt Ion:	107	Resp:	102
Ion	Ratio	Lower	Upper
107	100		
144	26.7	21.4	32.2
142	95.3	65.8	98.6



#36
 2-Methylnaphthalene
 Concen: 0.002 ng/ul
 RT: 12.428 min Scan# 1571
 Delta R.T. 0.035 min
 Lab File: BP013390.D
 Acq: 04 Jan 2023 14:49

Tgt Ion:	142	Resp:	170
Ion	Ratio	Lower	Upper
142	100		
141	82.6	70.8	106.2





#37

1-Methylnaphthalene

Concen: 0.007 ng/ul

RT: 12.616 min Scan# 1603

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:142 Resp: 669

Ion Ratio Lower Upper

142 100

141 113.5 72.7 109.1#

116 13.7 2.7 4.1#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 0.012 ng/ul

RT: 12.763 min Scan# 1628

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Tgt Ion:216 Resp: 634




Ion Ratio Lower Upper


216 100

214 24.0 61.9 92.9#

179 26.1 15.0 22.6#

108 30.8 11.0 16.4#









#40
Hexachlorocyclopentadiene
Concen: 0.001 ng/ul
RT: 12.722 min Scan# 1621
Delta R.T. -0.018 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:237 Resp: 18
Ion Ratio Lower Upper
237 100
235 90.5 49.7 74.5#
272 42.9 10.8 16.2#



#41
2,4,6-Trichlorophenol
Concen: 0.001 ng/ul
RT: 12.969 min Scan# 1663
Delta R.T. -0.036 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:196 Resp: 24
Ion Ratio Lower Upper
196 100
198 91.3 77.0 115.6
200 34.8 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 0.009 ng/ul

RT: 13.086 min Scan# 1

Delta R.T. 0.011 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:196 Resp: 324

Ion Ratio Lower Upper

196 100

198 68.3 77.1 115.7#

200 24.8 25.4 38.2#



#43

1,1'-Biphenyl

Concen: 0.007 ng/ul

RT: 13.410 min Scan# 1738

Delta R.T. 0.011 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:154 Resp: 911


Ion Ratio Lower Upper

154 100

153 16.2 31.5 47.3#

76 3.8 9.6 14.4#





#44

2-Chloronaphthalene

Concen: 0.008 ng/ul

RT: 13.451 min Scan# 1745

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:162 Resp: 744

Ion Ratio Lower Upper

162 100

164 71.8 26.3 39.5#

127 44.8 29.3 43.9#



#45

2-Nitroaniline

Concen: 0.007 ng/ul

RT: 13.651 min Scan# 1779

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion: 65 Resp: 151


Ion Ratio Lower Upper

65 100

92 82.4 56.4 84.6

138 11.0 90.3 135.5#









#47
Dimethylphthalate
Concen: 0.011 ng/ul
RT: 14.022 min Scan# 1842
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:163 Resp: 1298
Ion Ratio Lower Upper
163 100
194 11.4 3.4 5.0#
164 34.9 8.3 12.5#



#48
2,6-Dinitrotoluene
Concen: 0.005 ng/ul
RT: 14.157 min Scan# 1865
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:165 Resp: 112
Ion Ratio Lower Upper
165 100
89 84.6 43.0 64.4#
121 33.2 16.1 24.1#









#50
Acenaphthylene
Concen: 0.011 ng/ul
RT: 14.286 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:152 Resp: 1678
Ion Ratio Lower Upper
152 100
151 43.7 15.8 23.6#
153 21.6 10.6 16.0#



#51
3-Nitroaniline
Concen: 0.007 ng/ul
RT: 14.486 min Scan# 1921
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:138 Resp: 151
Ion Ratio Lower Upper
138 100
108 52.1 8.6 13.0#
92 148.5 88.4 132.6#









#52
Acenaphthene
Concen: 0.009 ng/ul
RT: 14.639 min Scan# 1946
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:153	Resp:	945
Ion Ratio	Lower	Upper
153	100	
152	47.2	38.2 57.4
154	84.4	70.8 106.2



#53
2,4-Dinitrophenol
Concen: 0.013 ng/ul
RT: 14.704 min Scan# 1958
Delta R.T. 0.017 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:184	Resp:	201
Ion Ratio	Lower	Upper
184	100	
63	63.1	48.4 72.6
154	50.0	45.5 68.3









#55
4-Nitrophenol
Concen: 0.019 ng/ul
RT: 14.792 min Scan# 1972
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:109 Resp: 297
Ion Ratio Lower Upper
109 100
139 71.8 120.8 181.2#
65 485.2 107.9 161.9#



#56
Dibenzofuran
Concen: 0.001 ng/ul
RT: 15.086 min Scan# 2023
Delta R.T. 0.117 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:168 Resp: 166
Ion Ratio Lower Upper
168 100
139 41.6 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 0.020 ng/ul
RT: 14.933 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:165 Resp: 651
Ion Ratio Lower Upper
165 100
63 51.7 31.8 47.6#
182 18.1 2.1 3.1#



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.013 ng/ul
RT: 15.192 min Scan# 2041
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:232 Resp: 426
Ion Ratio Lower Upper
232 100
131 89.5 31.8 47.6#
130 24.9 1.8 2.6#
166 112.5 22.9 34.3#









#59
Diethylphthalate
Concen: 0.014 ng/ul
RT: 15.386 min Scan# 2074
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:149 Resp: 1629
Ion Ratio Lower Upper
149 100
177 32.1 18.4 27.6#
150 17.1 9.9 14.9#



#61
Fluorene
Concen: 0.014 ng/ul
RT: 15.616 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:166 Resp: 1635
Ion Ratio Lower Upper
166 100
165 116.3 77.4 116.2#
167 16.8 10.6 16.0#









#62
4-Chlorophenyl-phenylether
Concen: 0.010 ng/ul
RT: 15.610 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:204 Resp: 603
Ion Ratio Lower Upper
204 100
206 63.6 26.3 39.5#
141 69.9 40.4 60.6#



#63
4-Nitroaniline
Concen: 0.004 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:138 Resp: 68
Ion Ratio Lower Upper
138 100
92 150.0 39.0 58.4#
108 21.4 55.1 82.7#









#66
4,6-Dinitro-2-methylphenol
Concen: 0.094 ng/ul
RT: 15.686 min Scan# 2125
Delta R.T. -0.012 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:198 Resp: 2173
Ion Ratio Lower Upper
198 100
182 488.3 3.3 4.9#
77 1615.3 16.6 25.0#



#67
N-Nitrosodiphenylamine
Concen: 0.010 ng/ul
RT: 15.828 min Scan# 2149
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:169 Resp: 1029
Ion Ratio Lower Upper
169 100
168 62.1 53.0 79.6
167 40.3 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 0.005 ng/ul
RT: 16.516 min Scan# 219
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:248 Resp: 199
Ion Ratio Lower Upper
248 100
250 52.2 77.4 116.0#
141 130.7 48.3 72.5#



#69
Hexachlorobenzene
Concen: 0.014 ng/ul
RT: 16.627 min Scan# 2285
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:284 Resp: 678
Ion Ratio Lower Upper
284 100
142 26.4 22.6 34.0
249 56.9 24.6 37.0#









#70
Atrazine
Concen: 0.005 ng/ul
RT: 16.786 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:200 Resp: 199
Ion Ratio Lower Upper
200 100
173 40.5 20.5 30.7#
215 51.1 39.9 59.9



#71
Pentachlorophenol
Concen: 0.003 ng/ul
RT: 16.992 min Scan# 2347
Delta R.T. 0.017 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:266 Resp: 88
Ion Ratio Lower Upper
266 100
264 71.3 50.9 76.3
268 69.1 51.4 77.2





#72

Phenanthrene

Concen: 0.012 ng/ul

RT: 17.369 min Scan# 2411

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:178 Resp: 2372

Ion Ratio Lower Upper

178 100

179 26.6 12.5 18.7#

176 20.9 15.4 23.0



#74

Anthracene

Concen: 0.010 ng/ul

RT: 17.469 min Scan# 2428

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:178 Resp: 1906


Ion Ratio Lower Upper

178 100

179 24.7 12.5 18.7#

176 20.8 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.000 ng/uL

RT: 13.434 min Scan# 11

Delta R.T. 0.064 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:216 Resp: 13

Ion Ratio Lower Upper

216 100

214 64.3 62.2 93.2

218 42.9 38.9 58.3



#76

Pentachlorobenzene

Concen: 0.009 ng/uL

RT: 14.881 min Scan# 1988

Delta R.T. -0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:250 Resp: 474


Ion Ratio Lower Upper

250 100

248 63.2 50.6 76.0

252 50.2 51.1 76.7#









#77
 Carbazole
 Concen: 0.008 ng/ul
 RT: 17.745 min Scan# 2473
 Delta R.T. 0.011 min
 Lab File: BP013390.D
 Acq: 04 Jan 2023 14:49




Instrument :
 BNA_P
 ClientSampleId :
 SBLK982


Tgt Ion:167 Resp: 1241
 Ion Ratio Lower Upper
 167 100
 166 19.8 17.2 25.8
 139 14.3 9.4 14.2#



#78
 Di-n-butylphthalate
 Concen: 0.019 ng/ul
 RT: 18.274 min Scan# 2565
 Delta R.T. -0.000 min
 Lab File: BP013390.D
 Acq: 04 Jan 2023 14:49

Tgt Ion:149 Resp: 3788
 Ion Ratio Lower Upper
 149 100
 150 7.2 7.5 11.3#
 104 13.1 3.8 5.6#









#80
Fluoranthene
Concen: 0.012 ng/ul
RT: 19.327 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:202 Resp: 2317
Ion Ratio Lower Upper
202 100
101 24.3 8.2 12.2#
100 19.1 6.1 9.1#



#82
Pyrene
Concen: 0.012 ng/ul
RT: 19.686 min Scan# 2805
Delta R.T. -0.000 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:202 Resp: 2393
Ion Ratio Lower Upper
202 100
101 22.4 9.4 14.2#
100 40.2 7.8 11.8#





#83

Butylbenzylphthalate

Concen: 0.013 ng/ul

RT: 20.551 min Scan# 2952

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:149 Resp: 1070

Ion Ratio Lower Upper

149 100

91 46.7 53.4 80.0#

206 39.6 17.9 26.9#



#84

3,3'-Dichlorobenzidine

Concen: 0.004 ng/ul

RT: 21.327 min Scan# 3084

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:252 Resp: 288


Ion Ratio Lower Upper

252 100

254 41.9 51.8 77.8#

126 72.9 7.4 11.2#





#85

Benzo(a)anthracene

Concen: 0.053 ng/ul

RT: 21.410 min Scan# 3096

Delta R.T. 0.011 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:228 Resp: 10835

Ion Ratio Lower Upper

228 100

229 26.0 16.1 24.1#

226 11.6 22.0 33.0#



#86

Bis(2-ethylhexyl)phthalate

Concen: 0.032 ng/ul

RT: 21.304 min Scan# 3080

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:149 Resp: 3599


Ion Ratio Lower Upper

149 100

167 18.4 22.5 33.7#

279 7.4 3.8 5.8#





#87

Chrysene

Concen: 0.019 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:228 Resp: 3575

Ion Ratio Lower Upper

228 100

226 23.1 24.5 36.7#

229 14.9 16.0 24.0#



#89

Di-n-octyl phthalate

Concen: 0.027 ng/ul




RT: 22.251 min Scan# 3241


Delta R.T. -0.000 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Tgt Ion:149 Resp: 4057





#90

Benzo(b)fluoranthene

Concen: 0.025 ng/ul

RT: 23.127 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:252 Resp: 4329

Ion Ratio Lower Upper

252 100

253 46.2 17.7 26.5#

125 47.8 6.8 10.2#



#91

Benzo(k)fluoranthene

Concen: 0.018 ng/ul

RT: 23.174 min Scan# 3398

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:252 Resp: 3022


Ion Ratio Lower Upper

252 100

253 60.4 17.7 26.5#

125 81.9 6.8 10.2#





#93

Benzo(a)pyrene

Concen: 0.012 ng/ul

RT: 23.768 min Scan# 3498

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:252 Resp: 1734

Ion Ratio Lower Upper

252 100

253 65.6 17.5 26.3#

125 65.8 7.5 11.3#



#94

Indeno(1,2,3-cd)pyrene

Concen: 0.004 ng/ul

RT: 26.450 min Scan# 3955

Delta R.T. 0.006 min

Lab File: BP013390.D

Acq: 04 Jan 2023 14:49




Tgt Ion:276 Resp: 776


Ion Ratio Lower Upper

276 100

138 39.1 17.1 25.7#

277 50.5 20.0 30.0#









#95
Dibenzo(a,h)anthracene
Concen: 0.006 ng/ul
RT: 26.468 min Scan# 3956
Delta R.T. 0.011 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Instrument :
BNA_P
ClientSampleId :
SBLK982

Tgt Ion:278 Resp: 903
Ion Ratio Lower Upper
278 100
139 28.3 11.1 16.7#
279 95.7 19.0 28.6#



#96
Benzo(g,h,i)perylene
Concen: 0.003 ng/ul
RT: 27.227 min Scan# 4087
Delta R.T. -0.006 min
Lab File: BP013390.D
Acq: 04 Jan 2023 14:49

Tgt Ion:276 Resp: 475
Ion Ratio Lower Upper
276 100
138 51.0 15.4 23.2#
277 15.6 19.1 28.7#

