

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013392.D
 Acq On : 04 Jan 2023 16:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Quant Time: Jan 05 04:08:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	475084	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	1995272	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	1253017	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.328	188	2393913	20.000	ng/ul	0.00
79) Chrysene-d12	21.416	240	1477654	20.000	ng/ul	0.00
88) Perylene-d12	23.880	264	1675132	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	91769	8.259	ng/uL	0.00
4) Pyridine-d5	3.746	84	676703	21.439	ng/ul	0.00
7) Phenol-d5	7.087	99	781851	20.205	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	517150	22.154	ng/ul	0.00
11) 2-Chlorophenol-d4	7.458	132	609935	20.023	ng/ul	0.00
15) 4-Methylphenol-d8	8.640	113	613099	19.336	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	291398	19.877	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	321184	19.461	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	645376	20.132	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	827759	18.291	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	1769819	18.378	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	2086320	19.716	ng/ul	0.00
54) 4-Nitrophenol-d4	14.775	143	268747	15.514	ng/ul	0.00
60) Fluorene-d10	15.569	176	1542518	18.934	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	269433	17.490	ng/ul	0.00
73) Anthracene-d10	17.428	188	2113314	19.568	ng/ul	0.00
81) Pyrene-d10	19.663	212	1965624	23.175	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.721	264	1573711	18.848	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	99146	8.548	ng/uL	91
5) Pyridine	3.770	79	685238	21.844	ng/ul	92
6) Benzaldehyde	7.063	77	319483	21.085	ng/ul	92
8) Phenol	7.116	94	807477	20.638	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.352	93	661287	20.592	ng/ul	93
12) 2-Chlorophenol	7.493	128	633052	20.265	ng/ul	95
13) 2-Methylphenol	8.375	108	593920	19.471	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.458	45	1065188	24.173	ng/ul#	95
16) Acetophenone	8.758	105	1013967	20.856	ng/ul#	91
17) N-Nitroso-di-n-propyla...	8.740	70	524527	21.455	ng/ul	90
18) 4-Methylphenol	8.705	108	661673	19.584	ng/ul	100
19) Hexachloroethane	9.010	117	254150	20.359	ng/ul	96
22) Nitrobenzene	9.140	77	773779	22.493	ng/ul	93
23) Isophorone	9.669	82	1374317	19.758	ng/ul	97
25) 2-Nitrophenol	9.852	139	358205	20.031	ng/ul#	87
26) 2,4-Dimethylphenol	9.910	107	725624	20.797	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.152	93	895067	20.124	ng/ul	100
29) 2,4-Dichlorophenol	10.387	162	633246	20.166	ng/ul	98
30) Naphthalene	10.793	128	2092585	20.053	ng/ul	100
32) 4-Chloroaniline	10.904	127	839117	18.737	ng/ul	99
33) Hexachlorobutadiene	11.069	225	446515	20.710	ng/ul	98
34) Caprolactam	11.693	113	156442	14.611	ng/ul	86
35) 4-Chloro-3-methylphenol	12.022	107	626880	19.562	ng/ul	99
36) 2-Methylnaphthalene	12.399	142	1389422	19.409	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013392.D
 Acq On : 04 Jan 2023 16:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Quant Time: Jan 05 04:08:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	1419345	19.279	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.763	216	856290	21.223	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	467357	17.800	ng/ul	100
41) 2,4,6-Trichlorophenol	13.004	196	529041	20.403	ng/ul	99
42) 2,4,5-Trichlorophenol	13.075	196	556799	19.991	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	1919120	20.393	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	1522998	20.563	ng/ul	99
45) 2-Nitroaniline	13.657	65	391943	21.826	ng/ul	90
47) Dimethylphthalate	14.028	163	1757999	18.419	ng/ul	100
48) 2,6-Dinitrotoluene	14.151	165	327609	18.469	ng/ul	96
50) Acenaphthylene	14.293	152	2283781	20.075	ng/ul	99
51) 3-Nitroaniline	14.481	138	270127	16.128	ng/ul	92
52) Acenaphthene	14.634	153	1574397	19.971	ng/ul	100
53) 2,4-Dinitrophenol	14.692	184	163164	13.457	ng/ul	93
55) 4-Nitrophenol	14.787	109	211641	17.906	ng/ul	88
56) Dibenzofuran	14.969	168	2233999	19.931	ng/ul	100
57) 2,4-Dinitrotoluene	14.940	165	444492	17.251	ng/ul#	87
58) 2,3,4,6-Tetrachlorophenol	15.198	232	466026	18.334	ng/ul	98
59) Diethylphthalate	15.387	149	1585699	17.090	ng/ul	100
61) Fluorene	15.622	166	1782467	19.829	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.616	204	937028	19.711	ng/ul	100
63) 4-Nitroaniline	15.645	138	225490	15.300	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.704	198	271750	18.051	ng/ul#	96
67) N-Nitrosodiphenylamine	15.828	169	1424075	21.457	ng/ul	99
68) 4-Bromophenyl-phenylether	16.516	248	537544	20.677	ng/ul	98
69) Hexachlorobenzene	16.628	284	626163	20.285	ng/ul	97
70) Atrazine	16.787	200	445614	17.230	ng/ul	100
71) Pentachlorophenol	16.975	266	310473	16.608	ng/ul	98
72) Phenanthrene	17.375	178	2502217	20.068	ng/ul	99
74) Anthracene	17.463	178	2495107	20.041	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.369	216	837622	24.524	ng/ul	99
76) Pentachlorobenzene	14.892	250	804584	23.105	ng/ul	99
77) Carbazole	17.733	167	1979012	18.874	ng/ul	99
78) Di-n-butylphthalate	18.275	149	1887694	14.786	ng/ul	99
80) Fluoranthene	19.333	202	2324691	24.149	ng/ul	99
82) Pyrene	19.692	202	2427007	24.440	ng/ul	98
83) Butylbenzylphthalate	20.557	149	552104	14.392	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.333	252	541393	16.218	ng/ul	100
85) Benzo(a)anthracene	21.398	228	1917485	19.547	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.310	149	677211	12.514	ng/ul	98
87) Chrysene	21.451	228	1874162	20.203	ng/ul	99
89) Di-n-octyl phthalate	22.257	149	1073015	11.434	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	2004030	18.761	ng/ul	99
91) Benzo(k)fluoranthene	23.174	252	1939843	18.314	ng/ul	99
93) Benzo(a)pyrene	23.768	252	1805084	19.407	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.451	276	2745661	23.699	ng/ul	99
95) Dibenzo(a,h)anthracene	26.468	278	2282544	23.796	ng/ul	99
96) Benzo(g,h,i)perylene	27.245	276	2282453	24.540	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013392.D
 Acq On : 04 Jan 2023 16:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Quant Time: Jan 05 04:08:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.548 ng/uL
RT: 3.358 min Scan# 29
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 88 Resp: 99146
Ion Ratio Lower Upper
88 100
43 32.0 22.4 33.6
58 80.5 58.2 87.4



#5
Pyridine
Concen: 21.844 ng/uL
RT: 3.770 min Scan# 99
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion: 79 Resp: 685238
Ion Ratio Lower Upper
79 100
52 82.8 60.7 91.1
51 40.8 28.9 43.3









#6
Benzaldehyde
Concen: 21.085 ng/ul
RT: 7.063 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 77 Resp: 319483
Ion Ratio Lower Upper
77 100
105 92.1 81.5 122.3
106 89.9 76.9 115.3



#8
Phenol
Concen: 20.638 ng/ul
RT: 7.116 min Scan# 668
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion: 94 Resp: 807477
Ion Ratio Lower Upper
94 100
65 30.2 21.1 31.7
66 37.9 27.0 40.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.592 ng/ul
 RT: 7.352 min Scan# 708
 Delta R.T. 0.006 min
 Lab File: BP013392.D
 Acq: 04 Jan 2023 16:49

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736

Tgt Ion: 93 Resp: 661287




Ion	Ratio	Lower	Upper
93	100		
63	82.6	59.8	89.8
95	32.2	26.4	39.6




#12
 2-Chlorophenol
 Concen: 20.265 ng/ul
 RT: 7.493 min Scan# 732
 Delta R.T. 0.006 min
 Lab File: BP013392.D
 Acq: 04 Jan 2023 16:49

Tgt Ion: 128 Resp: 633052

Ion	Ratio	Lower	Upper
128	100		
64	49.2	34.9	52.3
130	32.5	25.8	38.6





#13

2-Methylphenol

Concen: 19.471 ng/ul

RT: 8.375 min Scan# 881

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :





SSTD020736

Tgt Ion:108 Resp: 593920

Ion Ratio Lower Upper

108 100

107 88.7 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.173 ng/ul

RT: 8.458 min Scan# 896

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion: 45 Resp: 1065188


Ion Ratio Lower Upper

45 100

77 12.6 11.0 16.6

79 9.1 9.3 13.9#









#16
Acetophenone
Concen: 20.856 ng/ul
RT: 8.758 min Scan# 947
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:105 Resp: 1013967
Ion Ratio Lower Upper
105 100
77 79.5 59.3 88.9
51 32.3 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 21.455 ng/ul
RT: 8.740 min Scan# 944
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion: 70 Resp: 524527
Ion Ratio Lower Upper
70 100
42 67.4 45.5 68.3
101 9.5 7.9 11.9
130 20.0 17.4 26.2









#18
4-Methylphenol
Concen: 19.584 ng/ul
RT: 8.705 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:108 Resp: 661673
Ion Ratio Lower Upper
108 100
107 109.2 87.4 131.2



#19
Hexachloroethane
Concen: 20.359 ng/ul
RT: 9.010 min Scan# 990
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:117 Resp: 254150
Ion Ratio Lower Upper
117 100
201 112.9 93.7 140.5
199 68.7 58.6 88.0









#22
Nitrobenzene
Concen: 22.493 ng/ul
RT: 9.140 min Scan# 1100
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion: 77 Resp: 773779
Ion Ratio Lower Upper
77 100
123 42.2 38.3 57.5
65 13.7 10.3 15.5



#23
Isophorone
Concen: 19.758 ng/ul
RT: 9.669 min Scan# 1102
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion: 82 Resp: 1374317
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.2
138 17.0 15.3 22.9









#25
2-Nitrophenol
Concen: 20.031 ng/ul
RT: 9.852 min Scan# 1133
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:139 Resp: 358205
Ion Ratio Lower Upper
139 100
65 56.4 36.6 54.8#
109 30.5 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.797 ng/ul
RT: 9.910 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:107 Resp: 725624
Ion Ratio Lower Upper
107 100
121 50.3 41.8 62.6
122 85.3 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.124 ng/ul
 RT: 10.152 min Scan# 1184
 Delta R.T. 0.006 min
 Lab File: BP013392.D
 Acq: 04 Jan 2023 16:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020736


Tgt Ion: 93 Resp: 895067
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.0 39.0
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.166 ng/ul
 RT: 10.387 min Scan# 1224
 Delta R.T. 0.006 min
 Lab File: BP013392.D
 Acq: 04 Jan 2023 16:49

Tgt Ion:162 Resp: 633246
 Ion Ratio Lower Upper
 162 100
 164 66.3 52.5 78.7
 98 35.3 26.6 39.8









#30
Naphthalene
Concen: 20.053 ng/ul
RT: 10.793 min Scan# 1292
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:128 Resp: 2092585
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 18.737 ng/ul
RT: 10.904 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:127 Resp: 839117
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.710 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:225 Resp: 446515
Ion Ratio Lower Upper
225 100
223 63.8 49.2 73.8
227 62.8 50.2 75.2



#34
Caprolactam
Concen: 14.611 ng/ul
RT: 11.693 min Scan# 1446
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:113 Resp: 156442
Ion Ratio Lower Upper
113 100
55 193.6 141.9 212.9
56 145.5 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 19.562 ng/ul
RT: 12.022 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:107 Resp: 626880
Ion Ratio Lower Upper
107 100
144 26.6 21.4 32.2
142 80.9 65.8 98.6



#36
2-Methylnaphthalene
Concen: 19.409 ng/ul
RT: 12.399 min Scan# 1566
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:142 Resp: 1389422
Ion Ratio Lower Upper
142 100
141 89.4 70.8 106.2





#37

1-Methylnaphthalene

Concen: 19.279 ng/ul

RT: 12.616 min Scan# 1603

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:142 Resp: 1419345

Ion Ratio Lower Upper

142 100

141 91.0 72.7 109.1

116 3.8 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.223 ng/ul

RT: 12.763 min Scan# 1628

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Tgt Ion:216 Resp: 856290




Ion Ratio Lower Upper


216 100

214 77.6 61.9 92.9

179 19.0 15.0 22.6

108 15.2 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 17.800 ng/ul
RT: 12.740 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:237 Resp: 467357
Ion Ratio Lower Upper
237 100
235 61.9 49.7 74.5
272 13.1 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.403 ng/ul
RT: 13.004 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:196 Resp: 529041
Ion Ratio Lower Upper
196 100
198 97.6 77.0 115.6
200 31.2 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 19.991 ng/ul

RT: 13.075 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:196 Resp: 556799

Ion Ratio Lower Upper

196 100

198 95.4 77.1 115.7

200 31.0 25.4 38.2



#43

1,1'-Biphenyl

Concen: 20.393 ng/ul

RT: 13.404 min Scan# 1737

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:154 Resp: 1919120


Ion Ratio Lower Upper

154 100

153 39.9 31.5 47.3

76 13.1 9.6 14.4





#44

2-Chloronaphthalene

Concen: 20.563 ng/ul

RT: 13.451 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:162 Resp: 1522998

Ion Ratio Lower Upper

162 100

164 32.9 26.3 39.5

127 35.4 29.3 43.9



#45

2-Nitroaniline

Concen: 21.826 ng/ul

RT: 13.657 min Scan# 1780

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion: 65 Resp: 391943


Ion Ratio Lower Upper

65 100

92 65.9 56.4 84.6

138 98.9 90.3 135.5









#47
Dimethylphthalate
Concen: 18.419 ng/ul
RT: 14.028 min Scan# 1842
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:163 Resp: 1757999
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.2 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 18.469 ng/ul
RT: 14.151 min Scan# 1864
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:165 Resp: 327609
Ion Ratio Lower Upper
165 100
89 57.0 43.0 64.4
121 21.8 16.1 24.1









#50
Acenaphthylene
Concen: 20.075 ng/ul
RT: 14.293 min Scan# 1888
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:152 Resp: 2283781
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.6
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 16.128 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:138 Resp: 270127
Ion Ratio Lower Upper
138 100
108 10.7 8.6 13.0
92 119.8 88.4 132.6









#52
Acenaphthene
Concen: 19.971 ng/ul
RT: 14.634 min Scan# 1946
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:153 Resp: 1574397
Ion Ratio Lower Upper
153 100
152 47.6 38.2 57.4
154 88.7 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 13.457 ng/ul
RT: 14.692 min Scan# 1956
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:184 Resp: 163164
Ion Ratio Lower Upper
184 100
63 70.2 48.4 72.6
154 57.3 45.5 68.3









#55
4-Nitrophenol
Concen: 17.906 ng/ul
RT: 14.787 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:109 Resp: 211641
Ion Ratio Lower Upper
109 100
139 131.0 120.8 181.2
65 144.6 107.9 161.9



#56
Dibenzofuran
Concen: 19.931 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:168 Resp: 2233999
Ion Ratio Lower Upper
168 100
139 34.8 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 17.251 ng/ul
RT: 14.940 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:165 Resp: 444492
Ion Ratio Lower Upper
165 100
63 47.9 31.8 47.6#
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.334 ng/ul
RT: 15.198 min Scan# 2042
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:232 Resp: 466026
Ion Ratio Lower Upper
232 100
131 37.7 31.8 47.6
130 2.0 1.8 2.6
166 28.3 22.9 34.3









#59
Diethylphthalate
Concen: 17.090 ng/ul
RT: 15.387 min Scan# 2074
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:149 Resp: 1585699
Ion Ratio Lower Upper
149 100
177 23.0 18.4 27.6
150 12.5 9.9 14.9



#61
Fluorene
Concen: 19.829 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:166 Resp: 1782467
Ion Ratio Lower Upper
166 100
165 98.4 77.4 116.2
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.711 ng/ul
RT: 15.616 min Scan# 2112
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:204 Resp: 937028
Ion Ratio Lower Upper
204 100
206 32.3 26.3 39.5
141 50.4 40.4 60.6



#63
4-Nitroaniline
Concen: 15.300 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:138 Resp: 225490
Ion Ratio Lower Upper
138 100
92 54.4 39.0 58.4
108 81.0 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.051 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:198 Resp: 271750
Ion Ratio Lower Upper
198 100
182 5.1 3.3 4.9#
77 22.6 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 21.457 ng/ul
RT: 15.828 min Scan# 2149
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:169 Resp: 1424075
Ion Ratio Lower Upper
169 100
168 67.3 53.0 79.6
167 36.6 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 20.677 ng/ul
RT: 16.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:248 Resp: 537544
Ion Ratio Lower Upper
248 100
250 95.4 77.4 116.0
141 58.4 48.3 72.5



#69
Hexachlorobenzene
Concen: 20.285 ng/ul
RT: 16.628 min Scan# 2285
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:284 Resp: 626163
Ion Ratio Lower Upper
284 100
142 30.6 22.6 34.0
249 31.4 24.6 37.0





#70

Atrazine

Concen: 17.230 ng/ul

RT: 16.787 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:200 Resp: 445614

Ion Ratio Lower Upper

200 100

173 25.4 20.5 30.7

215 49.7 39.9 59.9



#71

Pentachlorophenol

Concen: 16.608 ng/ul

RT: 16.975 min Scan# 2344

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:266 Resp: 310473


Ion Ratio Lower Upper

266 100

264 65.5 50.9 76.3

268 64.9 51.4 77.2





#72

Phenanthrene

Concen: 20.068 ng/ul

RT: 17.375 min Scan# 2411

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:178 Resp: 2502217

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.8 15.4 23.0



#74

Anthracene

Concen: 20.041 ng/ul

RT: 17.463 min Scan# 2427

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:178 Resp: 2495107


Ion Ratio Lower Upper

178 100

179 15.7 12.5 18.7

176 18.4 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.524 ng/uL

RT: 13.369 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:216 Resp: 837622

Ion Ratio Lower Upper

216 100

214 79.4 62.2 93.2

218 48.9 38.9 58.3



#76

Pentachlorobenzene

Concen: 23.105 ng/uL

RT: 14.892 min Scan# 1990

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:250 Resp: 804584


Ion Ratio Lower Upper

250 100

248 62.8 50.6 76.0

252 63.5 51.1 76.7





#77

Carbazole

Concen: 18.874 ng/ul

RT: 17.733 min Scan# 2473

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:167 Resp: 1979012

Ion Ratio Lower Upper

167 100

166 21.7 17.2 25.8

139 12.3 9.4 14.2



#78

Di-n-butylphthalate

Concen: 14.786 ng/ul

RT: 18.275 min Scan# 2565

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:149 Resp: 1887694


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 5.1 3.8 5.6





#80

Fluoranthene

Concen: 24.149 ng/ul

RT: 19.333 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:202 Resp: 2324691

Ion Ratio Lower Upper

202 100

101 10.0 8.2 12.2

100 7.8 6.1 9.1



#82

Pyrene

Concen: 24.440 ng/ul

RT: 19.692 min Scan# 2806

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:202 Resp: 2427007


Ion Ratio Lower Upper

202 100

101 10.9 9.4 14.2

100 8.9 7.8 11.8





#83

Butylbenzylphthalate

Concen: 14.392 ng/ul

RT: 20.557 min Scan# 2952

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:149 Resp: 552104

Ion Ratio Lower Upper

149 100

91 68.2 53.4 80.0

206 23.9 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 16.218 ng/ul

RT: 21.333 min Scan# 3085

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:252 Resp: 541393


Ion Ratio Lower Upper

252 100

254 64.9 51.8 77.8

126 8.7 7.4 11.2









#85
Benzo(a)anthracene
Concen: 19.547 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. -0.000 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:228 Resp: 1917485
Ion Ratio Lower Upper
228 100
229 20.2 16.1 24.1
226 27.3 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 12.514 ng/ul
RT: 21.310 min Scan# 3081
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:149 Resp: 677211
Ion Ratio Lower Upper
149 100
167 27.2 22.5 33.7
279 4.5 3.8 5.8





#87

Chrysene

Concen: 20.203 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:228 Resp: 1874162

Ion Ratio Lower Upper

228 100

226 29.8 24.5 36.7

229 19.5 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 11.434 ng/ul




RT: 22.257 min Scan# 3242


Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Tgt Ion:149 Resp: 1073015









#90
Benzo(b)fluoranthene
Concen: 18.761 ng/ul
RT: 23.127 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49




Instrument :
BNA_P
ClientSampleId :
SSTD020736


Tgt Ion:252 Resp: 2004030
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.6 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 18.314 ng/ul
RT: 23.174 min Scan# 3398
Delta R.T. 0.006 min
Lab File: BP013392.D
Acq: 04 Jan 2023 16:49

Tgt Ion:252 Resp: 1939843
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.3 6.8 10.2





#93

Benzo(a)pyrene

Concen: 19.407 ng/ul

RT: 23.768 min Scan# 3498

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:252 Resp: 1805084

Ion Ratio Lower Upper

252 100

253 22.3 17.5 26.3

125 9.5 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.699 ng/ul

RT: 26.451 min Scan# 3955

Delta R.T. 0.006 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49




Tgt Ion:276 Resp: 2745661


Ion Ratio Lower Upper

276 100

138 20.6 17.1 25.7

277 25.1 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 23.796 ng/ul

RT: 26.468 min Scan# 3956

Delta R.T. 0.012 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Instrument :

BNA_P

ClientSampleId :

SSTD020736





Tgt Ion:278 Resp: 2282544

Ion Ratio Lower Upper

278 100

139 13.2 11.1 16.7

279 23.7 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 24.540 ng/ul

RT: 27.245 min Scan# 4090

Delta R.T. 0.012 min

Lab File: BP013392.D

Acq: 04 Jan 2023 16:49

Tgt Ion:276 Resp: 2282453

Ion Ratio Lower Upper

276 100

138 17.8 15.4 23.2

277 23.9 19.1 28.7

