

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
 Data File : BP013390.D
 Acq On : 04 Jan 2023 14:49
 Operator : CG/JU
 Sample : PB149982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Jan 05 04:07:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 04:02:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	589021	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	2456627	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	1619075	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.327	188	3671486	20.000	ng/u1	0.00
79) Chrysene-d12	21.415	240	3071121	20.000	ng/u1	0.00
88) Perylene-d12	23.880	264	2693520	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	80387	5.835	ng/uL	0.00
4) Pyridine-d5	3.746	84	1108545	28.327	ng/u1	0.00
7) Phenol-d5	7.087	99	1308178	27.267	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.252	67	872989	30.164	ng/u1	0.00
11) 2-Chlorophenol-d4	7.457	132	1051885	27.851	ng/u1	0.00
15) 4-Methylphenol-d8	8.634	113	1031713	26.245	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	502023	27.814	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	566342	27.871	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	1016563	25.756	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	1443767	25.911	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	3349947	26.922	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	3692805	27.008	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	463233	20.696	ng/u1	0.00
60) Fluorene-d10	15.563	176	2967088	28.185	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	492930	20.863	ng/u1	0.00
73) Anthracene-d10	17.427	188	4666383	28.172	ng/u1	0.00
81) Pyrene-d10	19.657	212	5451583	30.925	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.721	264	3828009	28.513	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010423\
Data File : BP013390.D
Acq On : 04 Jan 2023 14:49
Operator : CG/JU
Sample : PB149982BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK982

Quant Time: Jan 05 04:07:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 04:02:47 2023
Response via : Initial Calibration

