

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010524\
 Data File : BP019287.D
 Acq On : 05 Jan 2024 18:46
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020740

Quant Time: Jan 05 23:21:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	99447	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.593	136	371115	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.440	164	243169	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.198	188	581446	20.000	ng/ul	0.00
79) Chrysene-d12	21.298	240	659303	20.000	ng/ul	-0.01
88) Perylene-d12	23.663	264	811475	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	21968	7.544	ng/uL	0.00
4) Pyridine-d5	3.652	84	151204	20.689	ng/ul	0.00
7) Phenol-d5	6.987	99	184492	20.687	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	106853	20.711	ng/ul	0.00
11) 2-Chlorophenol-d4	7.328	132	146203	20.335	ng/ul	0.00
15) 4-Methylphenol-d8	8.528	113	135951	19.581	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	65164	20.312	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	75673	21.826	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	144765	20.677	ng/ul	0.00
31) 4-Chloroaniline-d4	10.746	131	168917	19.615	ng/ul	0.00
46) Dimethylphthalate-d6	13.857	166	407725	18.983	ng/ul	-0.01
49) Acenaphthylene-d8	14.134	160	467397	20.055	ng/ul	0.00
54) 4-Nitrophenol-d4	14.687	143	71662	19.758	ng/ul	0.02
60) Fluorene-d10	15.434	176	361031	19.457	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.575	200	74029	19.421	ng/ul	0.00
73) Anthracene-d10	17.292	188	600804	19.786	ng/ul	-0.01
81) Pyrene-d10	19.539	212	774043	19.264	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.510	264	896278	19.215	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	24273	7.582	ng/uL	98
5) Pyridine	3.670	79	156781	21.134	ng/ul	93
6) Benzaldehyde	6.940	77	88651	21.409	ng/ul	97
8) Phenol	7.011	94	183574	20.840	ng/ul	89
10) Bis(2-Chloroethyl)ether	7.228	93	143400	20.120	ng/ul	100
12) 2-Chlorophenol	7.364	128	140712	19.122	ng/ul	96
13) 2-Methylphenol	8.258	108	128416	19.707	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.322	45	169193	20.124	ng/ul	99
16) Acetophenone	8.628	105	207208	19.926	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.611	70	100888	19.456	ng/ul#	94
18) 4-Methylphenol	8.587	108	135712	19.349	ng/ul	97
19) Hexachloroethane	8.863	117	57601	18.662	ng/ul	90
22) Nitrobenzene	9.005	77	163704	20.887	ng/ul	97
23) Isophorone	9.528	82	280607	19.854	ng/ul	98
25) 2-Nitrophenol	9.716	139	78742	21.271	ng/ul	93
26) 2,4-Dimethylphenol	9.787	107	142106	21.805	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.016	93	171649	18.375	ng/ul	97
29) 2,4-Dichlorophenol	10.252	162	138561	20.253	ng/ul	97
30) Naphthalene	10.640	128	431568	19.288	ng/ul	99
32) 4-Chloroaniline	10.769	127	164622	19.922	ng/ul	100
33) Hexachlorobutadiene	10.916	225	111087	21.098	ng/ul	98
34) Caprolactam	11.563	113	36650	20.281	ng/ul	88
35) 4-Chloro-3-methylphenol	11.910	107	144160	21.465	ng/ul	97
36) 2-Methylnaphthalene	12.257	142	296528	19.656	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010524\
 Data File : BP019287.D
 Acq On : 05 Jan 2024 18:46
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020740


Quant Time: Jan 05 23:21:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 03:48:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	299562	19.691	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.628	216	193885	19.359	ng/ul	99
40) Hexachlorocyclopentadiene	12.593	237	106829	19.218	ng/ul	97
41) 2,4,6-Trichlorophenol	12.881	196	121698	20.362	ng/ul	99
42) 2,4,5-Trichlorophenol	12.957	196	133319	21.013	ng/ul	98
43) 1,1'-Biphenyl	13.269	154	396010	18.669	ng/ul	99
44) 2-Chloronaphthalene	13.316	162	325308	19.112	ng/ul	98
45) 2-Nitroaniline	13.534	65	89057	21.795	ng/ul	92
47) Dimethylphthalate	13.904	163	401203	18.923	ng/ul	100
48) 2,6-Dinitrotoluene	14.034	165	83714	21.487	ng/ul	99
50) Acenaphthylene	14.163	152	517131	19.722	ng/ul	99
51) 3-Nitroaniline	14.363	138	78902	20.716	ng/ul	90
52) Acenaphthene	14.504	153	343122	19.307	ng/ul	100
53) 2,4-Dinitrophenol	14.581	184	49146	19.580	ng/ul	93
55) 4-Nitrophenol	14.704	109	68312	23.567	ng/ul	85
56) Dibenzofuran	14.840	168	490470	19.326	ng/ul	94
57) 2,4-Dinitrotoluene	14.822	165	122704	21.415	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.075	232	118562	21.389	ng/ul#	97
59) Diethylphthalate	15.269	149	397839	19.353	ng/ul	100
61) Fluorene	15.493	166	401067	19.744	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.487	204	219965	19.829	ng/ul	96
63) 4-Nitroaniline	15.534	138	77973	20.348	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.587	198	78819	19.530	ng/ul#	92
67) N-Nitrosodiphenylamine	15.710	169	339516	19.132	ng/ul	99
68) 4-Bromophenyl-phenylether	16.387	248	145387	20.118	ng/ul	98
69) Hexachlorobenzene	16.498	284	181033	20.498	ng/ul	97
70) Atrazine	16.669	200	96948	19.358	ng/ul	99
71) Pentachlorophenol	16.851	266	105389	20.324	ng/ul	98
72) Phenanthrene	17.239	178	667960	18.956	ng/ul	99
74) Anthracene	17.334	178	683269	19.478	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.234	216	193655	19.155	ng/uL	98
76) Pentachlorobenzene	14.757	250	200172	19.925	ng/uL	98
77) Carbazole	17.616	167	611422	20.954	ng/ul	99
78) Di-n-butylphthalate	18.157	149	715233	19.764	ng/ul	99
80) Fluoranthene	19.216	202	872805	18.930	ng/ul	97
82) Pyrene	19.569	202	910269	18.843	ng/ul	97
83) Butylbenzylphthalate	20.445	149	354303	20.533	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.222	252	355936	21.581	ng/ul	99
85) Benzo(a)anthracene	21.286	228	1004036	19.275	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.210	149	518579	20.409	ng/ul	99
87) Chrysene	21.333	228	928801	19.121	ng/ul	99
89) Di-n-octyl phthalate	22.116	149	915185	19.722	ng/ul	100
90) Benzo(b)fluoranthene	22.945	252	1068096	18.739	ng/ul	98
91) Benzo(k)fluoranthene	22.992	252	1028293	18.360	ng/ul	99
93) Benzo(a)pyrene	23.563	252	1007422	18.778	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.121	276	1302493	18.939	ng/ul	95
95) Dibenzo(a,h)anthracene	26.139	278	1085390	18.972	ng/ul#	97
96) Benzo(g,h,i)perylene	26.874	276	1040636	18.882	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 05 23:21:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP122723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 03:48:11 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.582 ng/uL
RT: 3.264 min Scan# 30
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion: 88	Resp:	24273
Ion	Ratio	Lower Upper
88	100	
43	27.7	21.0 31.4
58	69.8	56.6 84.8



#5
Pyridine
Concen: 21.134 ng/uL
RT: 3.670 min Scan# 99
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion: 79	Resp:	156781
Ion	Ratio	Lower Upper
79	100	
52	72.5	52.4 78.6
51	34.9	25.9 38.9









#6
Benzaldehyde
Concen: 21.409 ng/ul
RT: 6.940 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion: 77 Resp: 88651
Ion Ratio Lower Upper
77 100
105 97.0 78.6 117.8
106 97.2 74.8 112.2



#8
Phenol
Concen: 20.840 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. 0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion: 94 Resp: 183574
Ion Ratio Lower Upper
94 100
65 31.0 21.1 31.7
66 43.7 29.3 43.9









#10
Bis(2-Chloroethyl)ether
Concen: 20.120 ng/ul
RT: 7.228 min Scan# 704
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion: 93 Resp: 143400
Ion Ratio Lower Upper
93 100
63 74.6 59.4 89.0
95 33.1 26.6 40.0



#12
2-Chlorophenol
Concen: 19.122 ng/ul
RT: 7.364 min Scan# 727
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:128 Resp: 140712
Ion Ratio Lower Upper
128 100
64 48.3 35.9 53.9
130 32.5 25.3 37.9





#13

2-Methylphenol

Concen: 19.707 ng/ul

RT: 8.258 min Scan# 81

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :





SSTD020740

Tgt Ion:108 Resp: 128416

Ion Ratio Lower Upper

108 100

107 87.8 72.6 109.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.124 ng/ul

RT: 8.322 min Scan# 890

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion: 45 Resp: 169193


Ion Ratio Lower Upper

45 100

77 16.5 13.6 20.4

79 12.1 9.7 14.5









#16
Acetophenone
Concen: 19.926 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:105 Resp: 207208
Ion Ratio Lower Upper
105 100
77 77.9 59.8 89.6
51 27.2 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 19.456 ng/ul
RT: 8.611 min Scan# 939
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion: 70 Resp: 100888
Ion Ratio Lower Upper
70 100
42 55.8 40.6 61.0
101 11.4 7.3 10.9#
130 21.5 16.1 24.1









#18
4-Methylphenol
Concen: 19.349 ng/ul
RT: 8.587 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:108 Resp: 135712
Ion Ratio Lower Upper
108 100
107 113.4 88.1 132.1



#19
Hexachloroethane
Concen: 18.662 ng/ul
RT: 8.863 min Scan# 982
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:117 Resp: 57601
Ion Ratio Lower Upper
117 100
201 126.0 92.5 138.7
199 80.3 56.9 85.3









#22
Nitrobenzene
Concen: 20.887 ng/ul
RT: 9.005 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion: 77 Resp: 163704
Ion Ratio Lower Upper
77 100
123 42.2 35.5 53.3
65 13.9 10.3 15.5



#23
Isophorone
Concen: 19.854 ng/ul
RT: 9.528 min Scan# 1095
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion: 82 Resp: 280607
Ion Ratio Lower Upper
82 100
95 8.0 5.9 8.9
138 17.6 14.7 22.1









#25
2-Nitrophenol
Concen: 21.271 ng/ul
RT: 9.716 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:139 Resp: 78742
Ion Ratio Lower Upper
139 100
65 53.1 38.8 58.2
109 36.8 26.2 39.2



#26
2,4-Dimethylphenol
Concen: 21.805 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:107 Resp: 142106
Ion Ratio Lower Upper
107 100
121 50.2 39.8 59.6
122 80.9 66.4 99.6





#27

Bis(2-Chloroethoxy)methane

Concen: 18.375 ng/ul

RT: 10.016 min Scan# 1178

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion: 93 Resp: 171649

Ion Ratio Lower Upper

93 100

95 31.6 26.7 40.1

123 11.0 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 20.253 ng/ul

RT: 10.252 min Scan# 1218

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:162 Resp: 138561


Ion Ratio Lower Upper

162 100

164 63.5 52.6 78.8

98 35.4 26.7 40.1





#30

Naphthalene

Concen: 19.288 ng/ul

RT: 10.640 min Scan# 1285

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:128 Resp: 431568

Ion Ratio Lower Upper

128 100

129 10.9 8.6 13.0

127 12.8 10.6 15.8



#32

4-Chloroaniline

Concen: 19.922 ng/ul

RT: 10.769 min Scan# 1306

Delta R.T. 0.000 min

Lab File: BP019287.D




Acq: 05 Jan 2024 18:46


Tgt Ion:127 Resp: 164622

Ion Ratio Lower Upper

127 100

129 32.3 25.6 38.4





#33

Hexachlorobutadiene

Concen: 21.098 ng/ul

RT: 10.916 min Scan# 11

Delta R.T. -0.018 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:225 Resp: 111087

Ion Ratio Lower Upper

225 100

223 62.1 50.5 75.7

227 64.3 53.0 79.4



#34

Caprolactam

Concen: 20.281 ng/ul

RT: 11.563 min Scan# 1441

Delta R.T. 0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:113 Resp: 36650


Ion Ratio Lower Upper

113 100

55 164.3 122.6 184.0

56 133.6 93.0 139.4









#35
4-Chloro-3-methylphenol
Concen: 21.465 ng/ul
RT: 11.910 min Scan# 1500
Delta R.T. 0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:107 Resp: 144160
Ion Ratio Lower Upper
107 100
144 25.9 21.0 31.6
142 78.3 64.8 97.2



#36
2-Methylnaphthalene
Concen: 19.656 ng/ul
RT: 12.257 min Scan# 1559
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:142 Resp: 296528
Ion Ratio Lower Upper
142 100
141 86.8 71.7 107.5









#37
1-Methylnaphthalene
Concen: 19.691 ng/ul
RT: 12.475 min Scan# 1596
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:142 Resp: 299562
Ion Ratio Lower Upper
142 100
141 92.1 74.4 111.6
116 4.2 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.359 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:216 Resp: 193885
Ion Ratio Lower Upper
216 100
214 77.9 62.6 94.0
179 19.0 15.8 23.8
108 14.3 11.9 17.9





#40

Hexachlorocyclopentadiene

Concen: 19.218 ng/ul

RT: 12.593 min Scan# 1616

Delta R.T. -0.018 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:237 Resp: 106829

Ion Ratio Lower Upper

237 100

235 61.9 51.2 76.8

272 12.4 10.9 16.3



#41

2,4,6-Trichlorophenol

Concen: 20.362 ng/ul

RT: 12.881 min Scan# 1665

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:196 Resp: 121698


Ion Ratio Lower Upper

196 100

198 96.4 77.4 116.0

200 31.8 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 21.013 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:196 Resp: 133319
Ion Ratio Lower Upper
196 100
198 98.8 77.3 115.9
200 32.5 25.7 38.5



#43
1,1'-Biphenyl
Concen: 18.669 ng/ul
RT: 13.269 min Scan# 1731
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:154 Resp: 396010
Ion Ratio Lower Upper
154 100
153 40.8 31.9 47.9
76 12.5 9.5 14.3





#44

2-Chloronaphthalene

Concen: 19.112 ng/ul

RT: 13.316 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:162 Resp: 325308

Ion Ratio Lower Upper

162 100

164 33.5 26.2 39.4

127 37.8 28.8 43.2



#45

2-Nitroaniline

Concen: 21.795 ng/ul

RT: 13.534 min Scan# 1776

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion: 65 Resp: 89057


Ion Ratio Lower Upper

65 100

92 69.3 60.2 90.4

138 106.0 92.9 139.3









#47
Dimethylphthalate
Concen: 18.923 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:163 Resp: 401203
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.2
164 10.2 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 21.487 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:165 Resp: 83714
Ion Ratio Lower Upper
165 100
89 56.8 45.1 67.7
121 20.0 16.7 25.1









#50
Acenaphthylene
Concen: 19.722 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:152 Resp: 517131
Ion Ratio Lower Upper
152 100
151 20.3 15.9 23.9
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 20.716 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:138 Resp: 78902
Ion Ratio Lower Upper
138 100
108 13.0 9.8 14.6
92 127.1 91.9 137.9





#52

Acenaphthene

Concen: 19.307 ng/ul

RT: 14.504 min Scan# 1941

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:153 Resp: 343122

Ion Ratio Lower Upper

153 100

152 47.2 38.1 57.1

154 87.8 70.1 105.1



#53

2,4-Dinitrophenol

Concen: 19.580 ng/ul

RT: 14.581 min Scan# 1954

Delta R.T. 0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:184 Resp: 49146


Ion Ratio Lower Upper

184 100

63 56.3 48.5 72.7

154 67.5 49.3 73.9









#55
4-Nitrophenol
Concen: 23.567 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.024 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:109 Resp: 68312
Ion Ratio Lower Upper
109 100
139 103.4 102.6 154.0
65 106.5 92.9 139.3



#56
Dibenzofuran
Concen: 19.326 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:168 Resp: 490470
Ion Ratio Lower Upper
168 100
139 37.4 27.4 41.0









#57
2,4-Dinitrotoluene
Concen: 21.415 ng/ul
RT: 14.822 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:165 Resp: 122704
Ion Ratio Lower Upper
165 100
63 45.1 31.5 47.3
182 2.4 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.389 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:232 Resp: 118562
Ion Ratio Lower Upper
232 100
131 38.3 29.3 43.9
130 2.3 1.4 2.2#
166 26.8 22.6 33.8









#59
Diethylphthalate
Concen: 19.353 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:149 Resp: 397839
Ion Ratio Lower Upper
149 100
177 23.6 18.7 28.1
150 12.4 9.8 14.8



#61
Fluorene
Concen: 19.744 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:166 Resp: 401067
Ion Ratio Lower Upper
166 100
165 95.7 79.3 118.9
167 12.9 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.829 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.012 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:204 Resp: 219965
Ion Ratio Lower Upper
204 100
206 33.0 26.3 39.5
141 56.9 42.2 63.2



#63
4-Nitroaniline
Concen: 20.348 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:138 Resp: 77973
Ion Ratio Lower Upper
138 100
92 51.7 42.2 63.4
108 93.1 66.2 99.4









#66
4,6-Dinitro-2-methylphenol
Concen: 19.530 ng/ul
RT: 15.587 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:198 Resp: 78819
Ion Ratio Lower Upper
198 100
182 3.7 4.0 6.0#
77 26.1 17.4 26.2



#67
N-Nitrosodiphenylamine
Concen: 19.132 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:169 Resp: 339516
Ion Ratio Lower Upper
169 100
168 66.3 53.8 80.8
167 36.5 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.118 ng/ul
RT: 16.387 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:248 Resp: 145387
Ion Ratio Lower Upper
248 100
250 96.9 78.3 117.5
141 63.1 48.6 72.8



#69
Hexachlorobenzene
Concen: 20.498 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:284 Resp: 181033
Ion Ratio Lower Upper
284 100
142 25.4 22.9 34.3
249 30.1 23.7 35.5









#70
Atrazine
Concen: 19.358 ng/ul
RT: 16.669 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46




Instrument :
BNA_P
ClientSampleId :
SSTD020740


Tgt Ion:200 Resp: 96948
Ion Ratio Lower Upper
200 100
173 24.8 20.2 30.2
215 48.7 39.9 59.9



#71
Pentachlorophenol
Concen: 20.324 ng/ul
RT: 16.851 min Scan# 2340
Delta R.T. 0.000 min
Lab File: BP019287.D
Acq: 05 Jan 2024 18:46

Tgt Ion:266 Resp: 105389
Ion Ratio Lower Upper
266 100
264 65.2 50.0 75.0
268 65.5 52.8 79.2





#72

Phenanthrene

Concen: 18.956 ng/ul

RT: 17.239 min Scan# 2406

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:178 Resp: 667960

Ion Ratio Lower Upper

178 100

179 15.3 11.9 17.9

176 19.1 15.0 22.6



#74

Anthracene

Concen: 19.478 ng/ul

RT: 17.334 min Scan# 2422

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:178 Resp: 683269


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.8 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.155 ng/uL

RT: 13.234 min Scan# 1725

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:216 Resp: 193655

Ion Ratio Lower Upper

216 100

214 78.7 61.6 92.4

218 49.3 38.9 58.3



#76

Pentachlorobenzene

Concen: 19.925 ng/uL

RT: 14.757 min Scan# 1984

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:250 Resp: 200172


Ion Ratio Lower Upper

250 100

248 64.5 50.4 75.6

252 63.2 52.1 78.1





#77

Carbazole

Concen: 20.954 ng/ul

RT: 17.616 min Scan# 2470

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:167 Resp: 611422

Ion Ratio Lower Upper

167 100

166 22.0 17.4 26.0

139 12.7 9.8 14.6



#78

Di-n-butylphthalate

Concen: 19.764 ng/ul

RT: 18.157 min Scan# 2562

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:149 Resp: 715233


Ion Ratio Lower Upper

149 100

150 9.1 7.3 10.9

104 5.5 3.9 5.9





#80

Fluoranthene

Concen: 18.930 ng/ul

RT: 19.216 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:202 Resp: 872805

Ion Ratio Lower Upper

202 100

101 8.5 7.8 11.6

100 6.6 6.0 9.0



#82

Pyrene

Concen: 18.843 ng/ul

RT: 19.569 min Scan# 2802

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:202 Resp: 910269


Ion Ratio Lower Upper

202 100

101 10.4 9.2 13.8

100 8.4 7.8 11.6





#83

Butylbenzylphthalate

Concen: 20.533 ng/ul

RT: 20.445 min Scan# 2951

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:149 Resp: 354303

Ion Ratio Lower Upper

149 100

91 72.8 55.6 83.4

206 22.6 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 21.581 ng/ul

RT: 21.222 min Scan# 3083

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:252 Resp: 355936


Ion Ratio Lower Upper

252 100

254 64.6 51.5 77.3

126 8.8 7.6 11.4





#85

Benzo(a)anthracene

Concen: 19.275 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:228 Resp: 1004036

Ion Ratio Lower Upper

228 100

229 19.6 15.5 23.3

226 27.0 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.409 ng/ul

RT: 21.210 min Scan# 3081

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:149 Resp: 518579


Ion Ratio Lower Upper

149 100

167 28.0 22.5 33.7

279 5.8 4.0 6.0





#87

Chrysene

Concen: 19.121 ng/ul

RT: 21.333 min Scan# 3103

Delta R.T. -0.012 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:228 Resp: 928801

Ion Ratio Lower Upper

228 100

226 30.1 23.4 35.2

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 19.722 ng/ul




RT: 22.116 min Scan# 3235


Delta R.T. -0.018 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Tgt Ion:149 Resp: 915185





#90

Benzo(b)fluoranthene

Concen: 18.739 ng/ul

RT: 22.945 min Scan# 3377

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:252 Resp: 1068096

Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.8

125 7.2 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 18.360 ng/ul

RT: 22.992 min Scan# 3384

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:252 Resp: 1028293


Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 7.0 6.7 10.1





#93

Benzo(a)pyrene

Concen: 18.778 ng/ul

RT: 23.563 min Scan# 3482

Delta R.T. -0.006 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:252 Resp: 1007422

Ion Ratio Lower Upper

252 100

253 22.7 17.5 26.3

125 7.9 7.4 11.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.939 ng/ul

RT: 26.121 min Scan# 3916

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46




Tgt Ion:276 Resp: 1302493


Ion Ratio Lower Upper

276 100

138 17.1 16.6 25.0

277 25.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 18.972 ng/ul

RT: 26.139 min Scan# 3920

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Instrument :

BNA_P

ClientSampleId :

SSTD020740





Tgt Ion:278 Resp: 1085390

Ion Ratio Lower Upper

278 100

139 10.9 11.0 16.4#

279 23.7 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 18.882 ng/ul

RT: 26.874 min Scan# 4044

Delta R.T. 0.000 min

Lab File: BP019287.D

Acq: 05 Jan 2024 18:46

Tgt Ion:276 Resp: 1040636

Ion Ratio Lower Upper

276 100

138 15.4 15.2 22.8

277 24.2 19.4 29.0

