

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010722\
 Data File : BP008723.D
 Acq On : 07 Jan 2022 13:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 07 16:27:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 07 07:45:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	65	0.00
2	1,4-Dioxane	40.000	41.307	-3.3	76	0.00
3	Pyridine	40.000	44.050	-10.1	77	0.00
4	n-Nitrosodimethylamine	40.000	41.685	-4.2	75	0.00
5 S	2-Fluorophenol	80.000	88.908	-11.1	81	0.00
6	Aniline	40.000	36.708	8.2	66	0.00
7 S	Phenol-d6	80.000	74.098	7.4	66	0.00
8	2-Chlorophenol	40.000	37.754	5.6	68	0.00
9	Benzaldehyde	40.000	36.314	9.2	66	0.00
10 C	Phenol	40.000	36.776	8.1	66	0.00
11	bis(2-Chloroethyl)ether	40.000	37.171	7.1	67	0.00
12	1,3-Dichlorobenzene	40.000	37.613	6.0	68	0.00
13 C	1,4-Dichlorobenzene	40.000	38.004	5.0	69	0.00
14	1,2-Dichlorobenzene	40.000	38.296	4.3	70	0.00
15	Benzyl Alcohol	40.000	38.120	4.7	68	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.099	12.3	64	0.00
17	2-Methylphenol	40.000	38.472	3.8	68	0.00
18	Hexachloroethane	40.000	37.024	7.4	67	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.752	8.1	66	0.00
20	3+4-Methylphenols	40.000	38.288	4.3	69	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	75	0.00
22	Acetophenone	40.000	31.828	20.4	68	0.00
23 S	Nitrobenzene-d5	80.000	63.410	20.7	67	0.00
24	Nitrobenzene	40.000	31.778	20.6	67	0.00
25	Isophorone	40.000	32.343	19.1	68	0.00
26 C	2-Nitrophenol	40.000	33.997	15.0	70	0.00
27	2,4-Dimethylphenol	40.000	32.679	18.3	68	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.091	7.3	79	0.00
29 C	2,4-Dichlorophenol	40.000	38.820	2.9	81	0.00
30	1,2,4-Trichlorobenzene	40.000	39.086	2.3	83	0.00
31	Naphthalene	40.000	38.384	4.0	81	0.00
32	Benzoic acid	40.000	34.220	14.5	70	-0.01
33	4-Chloroaniline	40.000	38.240	4.4	81	0.00
34 C	Hexachlorobutadiene	40.000	39.368	1.6	84	0.00
35	Caprolactam	40.000	38.459	3.9	80	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.020	-2.6	87	0.00
37	2-Methylnaphthalene	40.000	40.930	-2.3	87	0.00
38	1-Methylnaphthalene	40.000	40.640	-1.6	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.049	4.9	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.718	10.7	79	0.00
42 S	2,4,6-Tribromophenol	80.000	71.477	10.7	83	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.932	5.2	85	0.00
44	2,4,5-Trichlorophenol	40.000	38.194	4.5	86	0.00
45 S	2-Fluorobiphenyl	80.000	75.775	5.3	87	0.00
46	1,1'-Biphenyl	40.000	38.253	4.4	88	0.00
47	2-Chloronaphthalene	40.000	38.257	4.4	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010722\
 Data File : BP008723.D
 Acq On : 07 Jan 2022 13:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 07 16:27:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 07 07:45:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.310	1.7	86	0.00
49	Acenaphthylene	40.000	38.636	3.4	88	0.00
50	Dimethylphthalate	40.000	38.050	4.9	88	0.00
51	2,6-Dinitrotoluene	40.000	38.858	2.9	88	0.00
52 C	Acenaphthene	40.000	38.583	3.5	87	0.00
53	3-Nitroaniline	40.000	39.138	2.2	86	0.00
54 P	2,4-Dinitrophenol	40.000	36.897	7.8	80	0.00
55	Dibenzofuran	40.000	38.571	3.6	88	0.00
56 P	4-Nitrophenol	40.000	38.622	3.4	84	0.00
57	2,4-Dinitrotoluene	40.000	39.783	0.5	89	0.00
58	Fluorene	40.000	36.688	8.3	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.230	4.4	86	0.00
60	Diethylphthalate	40.000	38.197	4.5	87	0.00
61	4-Chlorophenyl-phenylether	40.000	36.382	9.0	90	0.00
62	4-Nitroaniline	40.000	36.199	9.5	87	0.00
63	Azobenzene	40.000	32.443	18.9	83	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.607	1.0	77	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.151	4.6	82	0.00
67	4-Bromophenyl-phenylether	40.000	39.161	2.1	81	0.00
68	Hexachlorobenzene	40.000	40.492	-1.2	83	0.00
69	Atrazine	40.000	38.873	2.8	82	0.00
70 C	Pentachlorophenol	40.000	37.200	7.0	74	0.00
71	Phenanthrene	40.000	38.846	2.9	83	0.00
72	Anthracene	40.000	38.878	2.8	82	0.00
73	Carbazole	40.000	38.958	2.6	82	0.00
74	Di-n-butylphthalate	40.000	39.456	1.4	83	0.00
75 C	Fluoranthene	40.000	40.589	-1.5	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	79	0.00
77	Benzydine	40.000	36.049	9.9	79	0.00
78	Pyrene	40.000	37.469	6.3	83	0.00
79 S	Terphenyl-d14	80.000	80.594	-0.7	84	0.00
80	Butylbenzylphthalate	40.000	38.020	4.9	83	0.00
81	Benzo(a)anthracene	40.000	38.725	3.2	85	0.00
82	3,3'-Dichlorobenzidine	40.000	39.158	2.1	85	0.00
83	Chrysene	40.000	38.840	2.9	86	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.720	3.2	85	0.00
85 c	Di-n-octyl phthalate	40.000	39.114	2.2	86	0.00
86 I	Perylene-d12	20.000	20.000	0.0	79	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.968	0.1	88	0.00
88	Benzo(b)fluoranthene	40.000	37.932	5.2	86	0.00
89	Benzo(k)fluoranthene	40.000	39.631	0.9	87	0.00
90 C	Benzo(a)pyrene	40.000	38.775	3.1	86	0.00
91	Dibenzo(a,h)anthracene	40.000	40.139	-0.3	88	0.00
92	Benzo(g,h,i)perylene	40.000	40.325	-0.8	89	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010722\
Data File : BP008723.D
Acq On : 07 Jan 2022 13:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jan 07 16:27:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP010622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 07 07:45:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0