

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010824\
 Data File : BP019302.D
 Acq On : 08 Jan 2024 18:27
 Operator : MA/JU
 Sample : SST01077
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010645

Quant Time: Jan 09 01:55:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 01:51:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	139070	20.000	ng/ul	0.00
20) Naphthalene-d8	10.581	136	544813	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.428	164	378769	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.186	188	904018	20.000	ng/ul	0.00
79) Chrysene-d12	21.286	240	796285	20.000	ng/ul	0.00
88) Perylene-d12	23.639	264	753810	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	13534	3.324	ng/uL	0.00
4) Pyridine-d5	3.646	84	92496	9.050	ng/ul	0.00
7) Phenol-d5	6.975	99	115159	9.234	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	74618	10.342	ng/ul	0.00
11) 2-Chlorophenol-d4	7.322	132	84667	8.421	ng/ul	0.00
15) 4-Methylphenol-d8	8.510	113	93185	9.597	ng/ul	0.00
21) Nitrobenzene-d5	8.952	128	42068	8.932	ng/ul	0.00
24) 2-Nitrophenol-d4	9.669	143	49575	9.740	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	101405	9.866	ng/ul	0.00
31) 4-Chloroaniline-d4	10.734	131	126709	10.023	ng/ul	0.00
46) Dimethylphthalate-d6	13.845	166	318621	9.524	ng/ul	0.00
49) Acenaphthylene-d8	14.122	160	337166	9.288	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	42558	7.533	ng/ul	0.00
60) Fluorene-d10	15.422	176	267225	9.246	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	53228	8.981	ng/ul	0.00
73) Anthracene-d10	17.286	188	423450	8.969	ng/ul	0.00
81) Pyrene-d10	19.527	212	497807	10.258	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.486	264	384036	8.863	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	14559	3.252	ng/uL	96
5) Pyridine	3.664	79	97869	9.434	ng/ul	93
6) Benzaldehyde	6.934	77	65285	11.274	ng/ul	98
8) Phenol	6.999	94	118634	9.631	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.222	93	99810	10.014	ng/ul	98
12) 2-Chlorophenol	7.352	128	88596	8.610	ng/ul	98
13) 2-Methylphenol	8.246	108	89815	9.856	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.310	45	122221	10.395	ng/ul	99
16) Acetophenone	8.616	105	156903	10.790	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.599	70	85288	11.762	ng/ul	97
18) 4-Methylphenol	8.575	108	95710	9.758	ng/ul	100
19) Hexachloroethane	8.852	117	40579	9.401	ng/ul	92
22) Nitrobenzene	8.993	77	121818	10.587	ng/ul	97
23) Isophorone	9.516	82	238562	11.498	ng/ul	99
25) 2-Nitrophenol	9.704	139	52437	9.649	ng/ul	99
26) 2,4-Dimethylphenol	9.775	107	115602	12.083	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.004	93	138020	10.065	ng/ul	100
29) 2,4-Dichlorophenol	10.240	162	98312	9.788	ng/ul	99
30) Naphthalene	10.628	128	294178	8.956	ng/ul	100
32) 4-Chloroaniline	10.757	127	124718	10.281	ng/ul	99
33) Hexachlorobutadiene	10.904	225	85581	11.072	ng/ul	97
34) Caprolactam	11.551	113	28705	10.820	ng/ul	81
35) 4-Chloro-3-methylphenol	11.893	107	110351	11.192	ng/ul	99
36) 2-Methylnaphthalene	12.245	142	216236	9.764	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010824\
 Data File : BP019302.D
 Acq On : 08 Jan 2024 18:27
 Operator : MA/JU
 Sample : SSTD01077
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010645


Quant Time: Jan 09 01:55:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 01:51:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.463	142	208430	9.333	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.616	216	149486	9.583	ng/ul	99
40) Hexachlorocyclopentadiene	12.587	237	77259	8.923	ng/ul	96
41) 2,4,6-Trichlorophenol	12.869	196	91582	9.837	ng/ul	97
42) 2,4,5-Trichlorophenol	12.945	196	97892	9.906	ng/ul	98
43) 1,1'-Biphenyl	13.263	154	293403	8.880	ng/ul	99
44) 2-Chloronaphthalene	13.304	162	235759	8.892	ng/ul	100
45) 2-Nitroaniline	13.522	65	65265	10.254	ng/ul	98
47) Dimethylphthalate	13.892	163	313776	9.501	ng/ul	99
48) 2,6-Dinitrotoluene	14.022	165	60017	9.890	ng/ul	96
50) Acenaphthylene	14.151	152	367983	9.010	ng/ul	98
51) 3-Nitroaniline	14.357	138	51057	8.606	ng/ul	96
52) Acenaphthene	14.492	153	245679	8.875	ng/ul	99
53) 2,4-Dinitrophenol	14.569	184	29298	7.494	ng/ul	95
55) 4-Nitrophenol	14.681	109	47020	10.414	ng/ul	97
56) Dibenzofuran	14.828	168	361754	9.151	ng/ul	99
57) 2,4-Dinitrotoluene	14.816	165	88713	9.940	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.063	232	92391	10.701	ng/ul#	99
59) Diethylphthalate	15.257	149	306290	9.565	ng/ul	100
61) Fluorene	15.481	166	296867	9.383	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.475	204	173818	10.060	ng/ul	99
63) 4-Nitroaniline	15.522	138	42494	7.119	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.581	198	58617	9.341	ng/ul	96
67) N-Nitrosodiphenylamine	15.698	169	250277	9.071	ng/ul	99
68) 4-Bromophenyl-phenylether	16.375	248	114752	10.213	ng/ul	96
69) Hexachlorobenzene	16.486	284	134009	9.759	ng/ul	97
70) Atrazine	16.657	200	107713	13.833	ng/ul	98
71) Pentachlorophenol	16.839	266	68880	8.544	ng/ul	98
72) Phenanthrene	17.228	178	475093	8.672	ng/ul	100
74) Anthracene	17.322	178	483922	8.873	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.222	216	148761	9.464	ng/ul	99
76) Pentachlorobenzene	14.745	250	151662	9.710	ng/ul	98
77) Carbazole	17.604	167	397788	8.768	ng/ul	99
78) Di-n-butylphthalate	18.151	149	504243	8.962	ng/ul	99
80) Fluoranthene	19.204	202	588442	10.567	ng/ul	99
82) Pyrene	19.557	202	601557	10.311	ng/ul	100
83) Butylbenzylphthalate	20.433	149	206235	9.896	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.210	252	171809	8.625	ng/ul	97
85) Benzo(a)anthracene	21.269	228	582392	9.257	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.198	149	289349	9.429	ng/ul	99
87) Chrysene	21.321	228	533119	9.087	ng/ul	99
89) Di-n-octyl phthalate	22.104	149	456415	10.588	ng/ul	100
90) Benzo(b)fluoranthene	22.921	252	503598	9.511	ng/ul	99
91) Benzo(k)fluoranthene	22.968	252	501786	9.644	ng/ul	99
93) Benzo(a)pyrene	23.533	252	450199	9.034	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.080	276	529443	8.287	ng/ul	99
95) Dibenzo(a,h)anthracene	26.098	278	434066	8.168	ng/ul	98
96) Benzo(g,h,i)perylene	26.827	276	430396	8.407	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 09 01:55:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 09 01:51:15 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.252 ng/uL
RT: 3.258 min Scan# 29
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:	88	Resp:	14559
Ion	Ratio	Lower	Upper
88	100		
43	25.0	22.0	33.0
58	65.2	54.6	82.0



#5
Pyridine
Concen: 9.434 ng/uL
RT: 3.664 min Scan# 98
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:	79	Resp:	97869
Ion	Ratio	Lower	Upper
79	100		
52	70.8	50.7	76.1
51	34.1	26.8	40.2









#6
Benzaldehyde
Concen: 11.274 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion: 77 Resp: 65285
Ion Ratio Lower Upper
77 100
105 92.7 75.3 112.9
106 86.3 71.5 107.3



#8
Phenol
Concen: 9.631 ng/ul
RT: 6.999 min Scan# 665
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion: 94 Resp: 118634
Ion Ratio Lower Upper
94 100
65 31.0 25.3 37.9
66 45.9 36.8 55.2









#10
Bis(2-Chloroethyl)ether
Concen: 10.014 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. 0.006 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion: 93 Resp: 99810
Ion Ratio Lower Upper
93 100
63 73.4 60.7 91.1
95 32.6 26.2 39.2



#12
2-Chlorophenol
Concen: 8.610 ng/ul
RT: 7.352 min Scan# 725
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:128 Resp: 88596
Ion Ratio Lower Upper
128 100
64 50.4 41.3 61.9
130 31.0 25.9 38.9





#13

2-Methylphenol

Concen: 9.856 ng/ul

RT: 8.246 min Scan# 877

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :





SSTD010645

Tgt Ion:108 Resp: 89815

Ion Ratio Lower Upper

108 100

107 91.7 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.395 ng/ul

RT: 8.310 min Scan# 888

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion: 45 Resp: 122221


Ion Ratio Lower Upper

45 100

77 17.5 14.2 21.4

79 13.6 10.5 15.7









#16
Acetophenone
Concen: 10.790 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:105 Resp: 156903
Ion Ratio Lower Upper
105 100
77 81.9 66.1 99.1
51 26.5 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 11.762 ng/ul
RT: 8.599 min Scan# 937
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion: 70 Resp: 85288
Ion Ratio Lower Upper
70 100
42 50.3 38.5 57.7
101 8.8 7.9 11.9
130 16.2 14.3 21.5









#18
4-Methylphenol
Concen: 9.758 ng/ul
RT: 8.575 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:108 Resp: 95710
Ion Ratio Lower Upper
108 100
107 115.8 92.6 139.0



#19
Hexachloroethane
Concen: 9.401 ng/ul
RT: 8.852 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:117 Resp: 40579
Ion Ratio Lower Upper
117 100
201 116.2 101.8 152.6
199 73.3 62.6 93.8









#22
Nitrobenzene
Concen: 10.587 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion: 77 Resp: 121818
Ion Ratio Lower Upper
77 100
123 39.9 30.0 45.0
65 13.2 10.5 15.7



#23
Isophorone
Concen: 11.498 ng/ul
RT: 9.516 min Scan# 1093
Delta R.T. -0.006 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion: 82 Resp: 238562
Ion Ratio Lower Upper
82 100
95 7.8 6.6 10.0
138 16.1 13.0 19.4









#25
2-Nitrophenol
Concen: 9.649 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:139 Resp: 52437
Ion Ratio Lower Upper
139 100
65 56.7 45.6 68.4
109 39.5 32.6 49.0



#26
2,4-Dimethylphenol
Concen: 12.083 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:107 Resp: 115602
Ion Ratio Lower Upper
107 100
121 44.4 35.3 52.9
122 73.3 60.6 91.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.065 ng/ul
 RT: 10.004 min Scan# 1176
 Delta R.T. -0.000 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010645

Tgt Ion: 93 Resp: 138020




Ion	Ratio	Lower	Upper
93	100		
95	32.2	25.9	38.9
123	10.1	7.9	11.9




#29
 2,4-Dichlorophenol
 Concen: 9.788 ng/ul
 RT: 10.240 min Scan# 1216
 Delta R.T. -0.000 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Tgt Ion:162 Resp: 98312

Ion	Ratio	Lower	Upper
162	100		
164	64.9	52.3	78.5
98	36.7	28.3	42.5









#30
Naphthalene
Concen: 8.956 ng/ul
RT: 10.628 min Scan# 1282
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:128 Resp: 294178
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.1 10.6 15.8



#32
4-Chloroaniline
Concen: 10.281 ng/ul
RT: 10.757 min Scan# 1304
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:127 Resp: 124718
Ion Ratio Lower Upper
127 100
129 32.7 25.8 38.8





#33

Hexachlorobutadiene

Concen: 11.072 ng/ul

RT: 10.904 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:225 Resp: 85581

Ion Ratio Lower Upper

225 100

223 65.1 49.9 74.9

227 65.1 51.0 76.6



#34

Caprolactam

Concen: 10.820 ng/ul

RT: 11.551 min Scan# 1439

Delta R.T. 0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:113 Resp: 28705


Ion Ratio Lower Upper

113 100

55 141.4 133.3 199.9

56 113.0 108.0 162.0









#35
4-Chloro-3-methylphenol
Concen: 11.192 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:107 Resp: 110351
Ion Ratio Lower Upper
107 100
144 22.4 17.8 26.8
142 71.1 56.3 84.5



#36
2-Methylnaphthalene
Concen: 9.764 ng/ul
RT: 12.245 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:142 Resp: 216236
Ion Ratio Lower Upper
142 100
141 88.7 71.1 106.7









#37
1-Methylnaphthalene
Concen: 9.333 ng/ul
RT: 12.463 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:142 Resp: 208430
Ion Ratio Lower Upper
142 100
141 92.6 74.4 111.6
116 4.3 3.9 5.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.583 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:216 Resp: 149486
Ion Ratio Lower Upper
216 100
214 78.4 63.6 95.4
179 18.1 15.1 22.7
108 13.4 11.3 16.9









#40
Hexachlorocyclopentadiene
Concen: 8.923 ng/ul
RT: 12.587 min Scan# 1615
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:237 Resp: 77259
Ion Ratio Lower Upper
237 100
235 65.8 50.0 75.0
272 13.7 11.4 17.2



#41
2,4,6-Trichlorophenol
Concen: 9.837 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. 0.006 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:196 Resp: 91582
Ion Ratio Lower Upper
196 100
198 101.0 78.2 117.2
200 31.7 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 9.906 ng/ul

RT: 12.945 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:196 Resp: 97892

Ion Ratio Lower Upper

196 100

198 95.6 78.9 118.3

200 31.5 25.0 37.6



#43

1,1'-Biphenyl

Concen: 8.880 ng/ul

RT: 13.263 min Scan# 1730

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:154 Resp: 293403


Ion Ratio Lower Upper

154 100

153 41.0 32.6 48.8

76 12.1 9.3 13.9









#44
2-Chloronaphthalene
Concen: 8.892 ng/ul
RT: 13.304 min Scan# 1737
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:162 Resp: 235759
Ion Ratio Lower Upper
162 100
164 33.3 26.6 40.0
127 36.9 29.4 44.2



#45
2-Nitroaniline
Concen: 10.254 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion: 65 Resp: 65265
Ion Ratio Lower Upper
65 100
92 66.6 52.1 78.1
138 90.9 74.8 112.2









#47
Dimethylphthalate
Concen: 9.501 ng/ul
RT: 13.892 min Scan# 1837
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:163 Resp: 313776
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 10.7 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 9.890 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:165 Resp: 60017
Ion Ratio Lower Upper
165 100
89 65.6 49.9 74.9
121 21.9 16.5 24.7





#50

Acenaphthylene

Concen: 9.010 ng/ul

RT: 14.151 min Scan# 1881

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:152 Resp: 367983

Ion Ratio Lower Upper

152 100

151 19.1 15.8 23.6

153 13.9 10.3 15.5



#51

3-Nitroaniline

Concen: 8.606 ng/ul

RT: 14.357 min Scan# 1916

Delta R.T. 0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:138 Resp: 51057


Ion Ratio Lower Upper

138 100

108 12.0 10.6 15.8

92 127.5 105.8 158.6









#52
 Acenaphthene
 Concen: 8.875 ng/ul
 RT: 14.492 min Scan# 1939
 Delta R.T. -0.000 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010645

Tgt Ion:153 Resp: 245679




Ion	Ratio	Lower	Upper
153	100		
152	45.6	38.2	57.2
154	86.6	68.9	103.3




#53
 2,4-Dinitrophenol
 Concen: 7.494 ng/ul
 RT: 14.569 min Scan# 1952
 Delta R.T. -0.000 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Tgt Ion:184 Resp: 29298

Ion	Ratio	Lower	Upper
184	100		
63	53.0	43.6	65.4
154	59.3	52.1	78.1









#55
4-Nitrophenol
Concen: 10.414 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:109 Resp: 47020
Ion Ratio Lower Upper
109 100
139 94.3 73.2 109.8
65 116.0 90.9 136.3



#56
Dibenzofuran
Concen: 9.151 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:168 Resp: 361754
Ion Ratio Lower Upper
168 100
139 37.3 30.3 45.5









#57
2,4-Dinitrotoluene
Concen: 9.940 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:165 Resp: 88713
Ion Ratio Lower Upper
165 100
63 50.8 41.3 61.9
182 2.7 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.701 ng/ul
RT: 15.063 min Scan# 2036
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:232 Resp: 92391
Ion Ratio Lower Upper
232 100
131 34.5 27.9 41.9
130 1.8 1.8 2.8#
166 25.3 20.9 31.3









#59
Diethylphthalate
Concen: 9.565 ng/ul
RT: 15.257 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:149 Resp: 306290
Ion Ratio Lower Upper
149 100
177 23.5 18.8 28.2
150 13.3 10.2 15.4



#61
Fluorene
Concen: 9.383 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:166 Resp: 296867
Ion Ratio Lower Upper
166 100
165 98.6 78.6 118.0
167 12.6 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 10.060 ng/ul
RT: 15.475 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:204 Resp: 173818
Ion Ratio Lower Upper
204 100
206 33.2 27.0 40.4
141 53.0 41.5 62.3



#63
4-Nitroaniline
Concen: 7.119 ng/ul
RT: 15.522 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:138 Resp: 42494
Ion Ratio Lower Upper
138 100
92 59.5 48.9 73.3
108 115.0 90.4 135.6









#66
4,6-Dinitro-2-methylphenol
Concen: 9.341 ng/ul
RT: 15.581 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:198 Resp: 58617
Ion Ratio Lower Upper
198 100
182 3.3 3.2 4.8
77 23.3 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 9.071 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:169 Resp: 250277
Ion Ratio Lower Upper
169 100
168 67.9 54.8 82.2
167 37.3 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 10.213 ng/ul
RT: 16.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:248 Resp: 114752
Ion Ratio Lower Upper
248 100
250 99.2 75.6 113.4
141 63.6 48.7 73.1



#69
Hexachlorobenzene
Concen: 9.759 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. 0.006 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:284 Resp: 134009
Ion Ratio Lower Upper
284 100
142 27.6 23.6 35.4
249 31.7 26.1 39.1





#70

Atrazine

Concen: 13.833 ng/ul

RT: 16.657 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:200 Resp: 107713

Ion Ratio Lower Upper

200 100

173 21.8 19.4 29.0

215 49.3 39.7 59.5



#71

Pentachlorophenol

Concen: 8.544 ng/ul

RT: 16.839 min Scan# 2338

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:266 Resp: 68880


Ion Ratio Lower Upper

266 100

264 64.9 49.4 74.2

268 64.9 51.6 77.4





#72

Phenanthrene

Concen: 8.672 ng/ul

RT: 17.228 min Scan# 2404

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:178 Resp: 475093

Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 19.1 15.4 23.2



#74

Anthracene

Concen: 8.873 ng/ul

RT: 17.322 min Scan# 2420

Delta R.T. 0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:178 Resp: 483922


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.3 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.464 ng/uL

RT: 13.222 min Scan# 1723

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:216 Resp: 148761

Ion Ratio Lower Upper

216 100

214 75.7 61.8 92.8

218 48.5 39.0 58.6



#76

Pentachlorobenzene

Concen: 9.710 ng/uL

RT: 14.745 min Scan# 1982

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:250 Resp: 151662


Ion Ratio Lower Upper

250 100

248 63.5 50.8 76.2

252 65.4 50.2 75.4





#77

Carbazole

Concen: 8.768 ng/ul

RT: 17.604 min Scan# 2468

Delta R.T. 0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:167 Resp: 397788

Ion Ratio Lower Upper

167 100

166 22.1 17.3 25.9

139 11.9 10.1 15.1



#78

Di-n-butylphthalate

Concen: 8.962 ng/ul

RT: 18.151 min Scan# 2561

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:149 Resp: 504243


Ion Ratio Lower Upper

149 100

150 9.0 7.0 10.6

104 6.1 4.6 6.8





#80

Fluoranthene

Concen: 10.567 ng/ul

RT: 19.204 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:202 Resp: 588442

Ion Ratio Lower Upper

202 100

101 7.5 5.9 8.9

100 5.4 4.6 7.0



#82

Pyrene

Concen: 10.311 ng/ul

RT: 19.557 min Scan# 2800

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:202 Resp: 601557


Ion Ratio Lower Upper

202 100

101 9.3 7.2 10.8

100 7.2 5.8 8.6





#83

Butylbenzylphthalate

Concen: 9.896 ng/ul

RT: 20.433 min Scan# 2949

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:149 Resp: 206235

Ion Ratio Lower Upper

149 100

91 76.6 61.7 92.5

206 23.4 18.6 27.8



#84

3,3'-Dichlorobenzidine

Concen: 8.625 ng/ul

RT: 21.210 min Scan# 3081

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:252 Resp: 171809


Ion Ratio Lower Upper

252 100

254 65.8 50.6 76.0

126 8.0 6.0 9.0





#85

Benzo(a)anthracene

Concen: 9.257 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:228 Resp: 582392

Ion Ratio Lower Upper

228 100

229 20.1 15.8 23.6

226 26.5 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.429 ng/ul

RT: 21.198 min Scan# 3079

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:149 Resp: 289349


Ion Ratio Lower Upper

149 100

167 28.2 22.9 34.3

279 5.6 4.6 7.0





#87

Chrysene

Concen: 9.087 ng/ul

RT: 21.321 min Scan# 3100

Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:228 Resp: 533119

Ion Ratio Lower Upper

228 100

226 29.2 23.9 35.9

229 19.2 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 10.588 ng/ul




RT: 22.104 min Scan# 3233


Delta R.T. -0.000 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Tgt Ion:149 Resp: 456415









#90
Benzo(b)fluoranthene
Concen: 9.511 ng/ul
RT: 22.921 min Scan# 3372
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27




Instrument :
BNA_P
ClientSampleId :
SSTD010645


Tgt Ion:252 Resp: 503598
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.8
125 5.8 4.9 7.3



#91
Benzo(k)fluoranthene
Concen: 9.644 ng/ul
RT: 22.968 min Scan# 3380
Delta R.T. -0.000 min
Lab File: BP019302.D
Acq: 08 Jan 2024 18:27

Tgt Ion:252 Resp: 501786
Ion Ratio Lower Upper
252 100
253 22.4 17.3 25.9
125 5.8 4.8 7.2





#93

Benzo(a)pyrene

Concen: 9.034 ng/ul

RT: 23.533 min Scan# 3477

Delta R.T. -0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27

Instrument :

BNA_P

ClientSampleId :

SSTD010645





Tgt Ion:252 Resp: 450199

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 6.6 5.2 7.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.287 ng/ul

RT: 26.080 min Scan# 3909

Delta R.T. -0.006 min

Lab File: BP019302.D

Acq: 08 Jan 2024 18:27




Tgt Ion:276 Resp: 529443


Ion Ratio Lower Upper

276 100

138 14.5 12.2 18.4

277 25.5 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 8.168 ng/ul
 RT: 26.098 min Scan# 3912
 Delta R.T. -0.000 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010645

Tgt Ion:	278	Resp:	434066
Ion	Ratio	Lower	Upper
278	100		
139	9.8	7.4	11.2
279	24.3	18.7	28.1



#96
 Benzo(g,h,i)perylene
 Concen: 8.407 ng/ul
 RT: 26.827 min Scan# 4036
 Delta R.T. -0.006 min
 Lab File: BP019302.D
 Acq: 08 Jan 2024 18:27

Tgt Ion:	276	Resp:	430396
Ion	Ratio	Lower	Upper
276	100		
138	14.2	10.7	16.1
277	24.2	19.2	28.8

