

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010824\
 Data File : BP019307.D
 Acq On : 08 Jan 2024 22:24
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV601

Quant Time: Jan 09 03:06:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	126901	20.000	ng/ul	0.00
20) Naphthalene-d8	10.581	136	520607	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.428	164	363251	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.186	188	913349	20.000	ng/ul	0.00
79) Chrysene-d12	21.286	240	995793	20.000	ng/ul	0.00
88) Perylene-d12	23.639	264	1025877	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	27781	8.278	ng/uL	0.00
4) Pyridine-d5	3.640	84	197708	20.373	ng/ul	0.00
7) Phenol-d5	6.975	99	246522	20.325	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	150478	20.742	ng/ul	0.00
11) 2-Chlorophenol-d4	7.316	132	176182	21.146	ng/ul	0.00
15) 4-Methylphenol-d8	8.510	113	199551	20.738	ng/ul	0.00
21) Nitrobenzene-d5	8.952	128	93545	21.358	ng/ul	0.00
24) 2-Nitrophenol-d4	9.669	143	110986	21.809	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.210	165	218902	21.291	ng/ul	0.00
31) 4-Chloroaniline-d4	10.734	131	270991	20.866	ng/ul	0.00
46) Dimethylphthalate-d6	13.845	166	679814	21.183	ng/ul	0.00
49) Acenaphthylene-d8	14.122	160	721120	20.962	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	111397	22.605	ng/ul	0.00
60) Fluorene-d10	15.422	176	573985	21.846	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	131608	19.517	ng/ul	0.00
73) Anthracene-d10	17.286	188	960816	21.092	ng/ul	0.00
81) Pyrene-d10	19.527	212	1233591	18.998	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.486	264	1186367	21.268	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	28944	8.326	ng/uL	96
5) Pyridine	3.658	79	200543	20.343	ng/ul	91
6) Benzaldehyde	6.934	77	138905	26.269	ng/ul	99
8) Phenol	6.999	94	249079	20.384	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.216	93	201928	20.602	ng/ul	98
12) 2-Chlorophenol	7.352	128	183605	21.321	ng/ul	100
13) 2-Methylphenol	8.246	108	189123	20.635	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.310	45	247492	21.221	ng/ul	99
16) Acetophenone	8.616	105	330620	21.571	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.599	70	174256	21.820	ng/ul	97
18) 4-Methylphenol	8.575	108	205684	21.081	ng/ul	98
19) Hexachloroethane	8.852	117	81347	20.913	ng/ul	97
22) Nitrobenzene	8.993	77	261204	21.318	ng/ul	98
23) Isophorone	9.522	82	506911	21.341	ng/ul	99
25) 2-Nitrophenol	9.704	139	115002	21.513	ng/ul	99
26) 2,4-Dimethylphenol	9.769	107	246123	21.354	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.004	93	292385	21.402	ng/ul	98
29) 2,4-Dichlorophenol	10.240	162	212628	21.460	ng/ul	99
30) Naphthalene	10.628	128	620589	21.155	ng/ul	100
32) 4-Chloroaniline	10.757	127	269184	21.186	ng/ul	100
33) Hexachlorobutadiene	10.904	225	174083	20.364	ng/ul	98
34) Caprolactam	11.546	113	68734	22.120	ng/ul	100
35) 4-Chloro-3-methylphenol	11.893	107	239275	21.752	ng/ul	96
36) 2-Methylnaphthalene	12.245	142	455380	21.335	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010824\
 Data File : BP019307.D
 Acq On : 08 Jan 2024 22:24
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV601

Quant Time: Jan 09 03:06:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.463	142	449803	21.539	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.616	216	330602	21.002	ng/ul	98
40) Hexachlorocyclopentadiene	12.587	237	182220	18.036	ng/ul	97
41) 2,4,6-Trichlorophenol	12.863	196	208633	21.537	ng/ul	99
42) 2,4,5-Trichlorophenol	12.945	196	223451	21.525	ng/ul	99
43) 1,1'-Biphenyl	13.263	154	646591	21.547	ng/ul	100
44) 2-Chloronaphthalene	13.304	162	523050	21.553	ng/ul	99
45) 2-Nitroaniline	13.522	65	162026	23.277	ng/ul	95
47) Dimethylphthalate	13.898	163	674463	21.146	ng/ul	100
48) 2,6-Dinitrotoluene	14.022	165	140053	22.206	ng/ul	98
50) Acenaphthylene	14.151	152	794161	21.077	ng/ul	100
51) 3-Nitroaniline	14.351	138	124266	22.589	ng/ul	99
52) Acenaphthene	14.492	153	526751	21.336	ng/ul	98
53) 2,4-Dinitrophenol	14.569	184	79760	19.287	ng/ul	98
55) 4-Nitrophenol	14.681	109	118598	22.285	ng/ul	98
56) Dibenzofuran	14.828	168	767238	21.706	ng/ul	100
57) 2,4-Dinitrotoluene	14.816	165	208384	23.548	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.063	232	207176	22.480	ng/ul	98
59) Diethylphthalate	15.257	149	661807	22.173	ng/ul	99
61) Fluorene	15.481	166	641762	22.167	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.475	204	374422	21.672	ng/ul	99
63) 4-Nitroaniline	15.522	138	117005	27.434	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.581	198	140688	19.685	ng/ul	94
67) N-Nitrosodiphenylamine	15.692	169	550912	20.461	ng/ul	100
68) 4-Bromophenyl-phenylether	16.375	248	247772	19.860	ng/ul	96
69) Hexachlorobenzene	16.481	284	292213	20.324	ng/ul	99
70) Atrazine	16.657	200	242410	20.654	ng/ul	100
71) Pentachlorophenol	16.839	266	175287	19.629	ng/ul	99
72) Phenanthrene	17.228	178	1081871	21.087	ng/ul	100
74) Anthracene	17.316	178	1104871	21.244	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.222	216	327094	19.342	ng/ul	99
76) Pentachlorobenzene	14.745	250	318444	19.384	ng/ul	99
77) Carbazole	17.598	167	954057	21.562	ng/ul	99
78) Di-n-butylphthalate	18.145	149	1163001	21.963	ng/ul	98
80) Fluoranthene	19.204	202	1433282	18.674	ng/ul	99
82) Pyrene	19.557	202	1472061	18.896	ng/ul	99
83) Butylbenzylphthalate	20.433	149	547142	20.472	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.210	252	494820	21.885	ng/ul	99
85) Benzo(a)anthracene	21.269	228	1588503	20.492	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	794198	20.702	ng/ul	100
87) Chrysene	21.321	228	1459036	20.591	ng/ul	100
89) Di-n-octyl phthalate	22.104	149	1370725	19.576	ng/ul	100
90) Benzo(b)fluoranthene	22.921	252	1540054	21.007	ng/ul	99
91) Benzo(k)fluoranthene	22.968	252	1491130	20.966	ng/ul	99
93) Benzo(a)pyrene	23.539	252	1391218	21.118	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.080	276	1535252	19.706	ng/ul	99
95) Dibenzo(a,h)anthracene	26.098	278	1271981	19.722	ng/ul	100
96) Benzo(g,h,i)perylene	26.833	276	1195573	19.108	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010824\
 Data File : BP019307.D
 Acq On : 08 Jan 2024 22:24
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV601

Quant Time: Jan 09 03:06:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.326 ng/uL
RT: 3.258 min Scan# 29
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:	88	Resp:	28944
Ion	Ratio	Lower	Upper
88	100		
43	24.6	22.0	33.0
58	66.0	54.6	82.0



#5
Pyridine
Concen: 20.343 ng/uL
RT: 3.658 min Scan# 97
Delta R.T. -0.006 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:	79	Resp:	200543
Ion	Ratio	Lower	Upper
79	100		
52	73.8	50.7	76.1
51	34.3	26.8	40.2









#6
Benzaldehyde
Concen: 26.269 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion: 77 Resp: 138905
Ion Ratio Lower Upper
77 100
105 93.1 75.3 112.9
106 88.1 71.5 107.3



#8
Phenol
Concen: 20.384 ng/ul
RT: 6.999 min Scan# 665
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 94 Resp: 249079
Ion Ratio Lower Upper
94 100
65 30.0 25.3 37.9
66 41.3 36.8 55.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.602 ng/ul
RT: 7.216 min Scan# 702
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion: 93 Resp: 201928
Ion Ratio Lower Upper
93 100
63 74.2 60.7 91.1
95 32.2 26.2 39.2



#12
2-Chlorophenol
Concen: 21.321 ng/ul
RT: 7.352 min Scan# 725
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 128 Resp: 183605
Ion Ratio Lower Upper
128 100
64 51.2 41.3 61.9
130 32.3 25.9 38.9









#13
2-Methylphenol
Concen: 20.635 ng/ul
RT: 8.246 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:108 Resp: 189123
Ion Ratio Lower Upper
108 100
107 90.8 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.221 ng/ul
RT: 8.310 min Scan# 888
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 45 Resp: 247492
Ion Ratio Lower Upper
45 100
77 18.1 14.2 21.4
79 13.5 10.5 15.7









#16
Acetophenone
Concen: 21.571 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:105 Resp: 330620
Ion Ratio Lower Upper
105 100
77 82.9 66.1 99.1
51 26.3 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 21.820 ng/ul
RT: 8.599 min Scan# 937
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 70 Resp: 174256
Ion Ratio Lower Upper
70 100
42 50.8 38.5 57.7
101 9.0 7.9 11.9
130 17.1 14.3 21.5









#18
4-Methylphenol
Concen: 21.081 ng/ul
RT: 8.575 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:108 Resp: 205684
Ion Ratio Lower Upper
108 100
107 113.4 92.6 139.0



#19
Hexachloroethane
Concen: 20.913 ng/ul
RT: 8.852 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:117 Resp: 81347
Ion Ratio Lower Upper
117 100
201 130.0 101.8 152.6
199 80.9 62.6 93.8









#22
Nitrobenzene
Concen: 21.318 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion: 77 Resp: 261204
Ion Ratio Lower Upper
77 100
123 39.0 30.0 45.0
65 13.4 10.5 15.7



#23
Isophorone
Concen: 21.341 ng/ul
RT: 9.522 min Scan# 1094
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 82 Resp: 506911
Ion Ratio Lower Upper
82 100
95 7.8 6.6 10.0
138 16.1 13.0 19.4









#25
2-Nitrophenol
Concen: 21.513 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:139 Resp: 115002
Ion Ratio Lower Upper
139 100
65 56.4 45.6 68.4
109 40.6 32.6 49.0



#26
2,4-Dimethylphenol
Concen: 21.354 ng/ul
RT: 9.769 min Scan# 1136
Delta R.T. -0.006 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:107 Resp: 246123
Ion Ratio Lower Upper
107 100
121 44.6 35.3 52.9
122 76.1 60.6 91.0





#27

Bis(2-Chloroethoxy)methane

Concen: 21.402 ng/ul

RT: 10.004 min Scan# 1176

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion: 93 Resp: 292385

Ion Ratio Lower Upper

93 100

95 31.2 25.9 38.9

123 9.8 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 21.460 ng/ul

RT: 10.240 min Scan# 1216

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:162 Resp: 212628


Ion Ratio Lower Upper

162 100

164 65.8 52.3 78.5

98 35.9 28.3 42.5









#30
Naphthalene
Concen: 21.155 ng/ul
RT: 10.628 min Scan# 1282
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:128 Resp: 620589
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 21.186 ng/ul
RT: 10.757 min Scan# 1304
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:127 Resp: 269184
Ion Ratio Lower Upper
127 100
129 32.1 25.8 38.8









#33
Hexachlorobutadiene
Concen: 20.364 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:225 Resp: 174083
Ion Ratio Lower Upper
225 100
223 64.3 49.9 74.9
227 62.7 51.0 76.6



#34
Caprolactam
Concen: 22.120 ng/ul
RT: 11.546 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:113 Resp: 68734
Ion Ratio Lower Upper
113 100
55 166.4 133.3 199.9
56 135.5 108.0 162.0









#35
4-Chloro-3-methylphenol
Concen: 21.752 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:107 Resp: 239275
Ion Ratio Lower Upper
107 100
144 23.5 17.8 26.8
142 74.1 56.3 84.5



#36
2-Methylnaphthalene
Concen: 21.335 ng/ul
RT: 12.245 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:142 Resp: 455380
Ion Ratio Lower Upper
142 100
141 88.7 71.1 106.7





#37

1-Methylnaphthalene

Concen: 21.539 ng/ul

RT: 12.463 min Scan# 1594

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:142 Resp: 449803

Ion Ratio Lower Upper

142 100

141 92.0 74.4 111.6

116 4.6 3.9 5.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.002 ng/ul

RT: 12.616 min Scan# 1620

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Tgt Ion:216 Resp: 330602




Ion Ratio Lower Upper


216 100

214 77.6 63.6 95.4

179 19.3 15.1 22.7

108 13.4 11.3 16.9









#40
Hexachlorocyclopentadiene
Concen: 18.036 ng/ul
RT: 12.587 min Scan# 1615
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:237 Resp: 182220
Ion Ratio Lower Upper
237 100
235 60.7 50.0 75.0
272 12.0 11.4 17.2



#41
2,4,6-Trichlorophenol
Concen: 21.537 ng/ul
RT: 12.863 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:196 Resp: 208633
Ion Ratio Lower Upper
196 100
198 96.2 78.2 117.2
200 30.6 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 21.525 ng/ul

RT: 12.945 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:196 Resp: 223451

Ion Ratio Lower Upper

196 100

198 97.7 78.9 118.3

200 32.8 25.0 37.6



#43

1,1'-Biphenyl

Concen: 21.547 ng/ul

RT: 13.263 min Scan# 1730

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:154 Resp: 646591


Ion Ratio Lower Upper

154 100

153 40.4 32.6 48.8

76 11.7 9.3 13.9









#44
2-Chloronaphthalene
Concen: 21.553 ng/ul
RT: 13.304 min Scan# 1737
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:162 Resp: 523050
Ion Ratio Lower Upper
162 100
164 32.8 26.6 40.0
127 36.6 29.4 44.2



#45
2-Nitroaniline
Concen: 23.277 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion: 65 Resp: 162026
Ion Ratio Lower Upper
65 100
92 68.2 52.1 78.1
138 99.3 74.8 112.2









#47
Dimethylphthalate
Concen: 21.146 ng/ul
RT: 13.898 min Scan# 1838
Delta R.T. 0.006 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:163 Resp: 674463
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 10.1 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 22.206 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:165 Resp: 140053
Ion Ratio Lower Upper
165 100
89 60.9 49.9 74.9
121 21.1 16.5 24.7









#50
Acenaphthylene
Concen: 21.077 ng/ul
RT: 14.151 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:152 Resp: 794161
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 22.589 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:138 Resp: 124266
Ion Ratio Lower Upper
138 100
108 12.3 10.6 15.8
92 130.9 105.8 158.6









#52
Acenaphthene
Concen: 21.336 ng/ul
RT: 14.492 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:153 Resp: 526751
Ion Ratio Lower Upper
153 100
152 46.3 38.2 57.2
154 87.3 68.9 103.3



#53
2,4-Dinitrophenol
Concen: 19.287 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:184 Resp: 79760
Ion Ratio Lower Upper
184 100
63 56.3 43.6 65.4
154 66.2 52.1 78.1









#55
4-Nitrophenol
Concen: 22.285 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:109 Resp: 118598
Ion Ratio Lower Upper
109 100
139 94.9 73.2 109.8
65 112.6 90.9 136.3



#56
Dibenzofuran
Concen: 21.706 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:168 Resp: 767238
Ion Ratio Lower Upper
168 100
139 38.1 30.3 45.5





#57

2,4-Dinitrotoluene

Concen: 23.548 ng/ul

RT: 14.816 min Scan# 1994

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:165 Resp: 208384

Ion Ratio Lower Upper

165 100

63 50.7 41.3 61.9

182 2.7 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.480 ng/ul

RT: 15.063 min Scan# 2036

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Tgt Ion:232 Resp: 207176




Ion Ratio Lower Upper


232 100

131 33.7 27.9 41.9

130 2.2 1.8 2.8

166 25.1 20.9 31.3









#59
Diethylphthalate
Concen: 22.173 ng/ul
RT: 15.257 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:149 Resp: 661807
Ion Ratio Lower Upper
149 100
177 23.8 18.8 28.2
150 12.9 10.2 15.4



#61
Fluorene
Concen: 22.167 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:166 Resp: 641762
Ion Ratio Lower Upper
166 100
165 97.5 78.6 118.0
167 13.3 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.672 ng/ul
RT: 15.475 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:204 Resp: 374422
Ion Ratio Lower Upper
204 100
206 32.7 27.0 40.4
141 51.6 41.5 62.3



#63
4-Nitroaniline
Concen: 27.434 ng/ul
RT: 15.522 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:138 Resp: 117005
Ion Ratio Lower Upper
138 100
92 61.2 48.9 73.3
108 115.4 90.4 135.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.685 ng/ul
RT: 15.581 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:198 Resp: 140688
Ion Ratio Lower Upper
198 100
182 3.8 3.2 4.8
77 22.2 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 20.461 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:169 Resp: 550912
Ion Ratio Lower Upper
169 100
168 68.3 54.8 82.2
167 36.6 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 19.860 ng/ul
RT: 16.375 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:248 Resp: 247772
Ion Ratio Lower Upper
248 100
250 98.9 75.6 113.4
141 63.4 48.7 73.1



#69
Hexachlorobenzene
Concen: 20.324 ng/ul
RT: 16.481 min Scan# 2277
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:284 Resp: 292213
Ion Ratio Lower Upper
284 100
142 29.3 23.6 35.4
249 33.2 26.1 39.1





#70

Atrazine

Concen: 20.654 ng/ul

RT: 16.657 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:200 Resp: 242410

Ion Ratio Lower Upper

200 100

173 23.9 19.4 29.0

215 49.8 39.7 59.5



#71

Pentachlorophenol

Concen: 19.629 ng/ul

RT: 16.839 min Scan# 2338

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:266 Resp: 175287


Ion Ratio Lower Upper

266 100

264 63.4 49.4 74.2

268 64.3 51.6 77.4





#72

Phenanthrene

Concen: 21.087 ng/ul

RT: 17.228 min Scan# 2404

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:178 Resp: 1081871

Ion Ratio Lower Upper

178 100

179 15.1 12.1 18.1

176 19.1 15.4 23.2



#74

Anthracene

Concen: 21.244 ng/ul

RT: 17.316 min Scan# 2419

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:178 Resp: 1104871


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.7 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.342 ng/uL
RT: 13.222 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:216 Resp: 327094
Ion Ratio Lower Upper
216 100
214 77.5 61.8 92.8
218 47.5 39.0 58.6



#76
Pentachlorobenzene
Concen: 19.384 ng/uL
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:250 Resp: 318444
Ion Ratio Lower Upper
250 100
248 64.3 50.8 76.2
252 64.1 50.2 75.4





#77

Carbazole

Concen: 21.562 ng/ul

RT: 17.598 min Scan# 2467

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:167 Resp: 954057

Ion Ratio Lower Upper

167 100

166 21.4 17.3 25.9

139 12.1 10.1 15.1



#78

Di-n-butylphthalate

Concen: 21.963 ng/ul

RT: 18.145 min Scan# 2560

Delta R.T. -0.006 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:149 Resp: 1163001


Ion Ratio Lower Upper

149 100

150 9.6 7.0 10.6

104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 18.674 ng/ul
RT: 19.204 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24




Instrument :
BNA_P
ClientSampleId :
SICV601


Tgt Ion:202 Resp: 1433282
Ion Ratio Lower Upper
202 100
101 7.1 5.9 8.9
100 5.7 4.6 7.0



#82
Pyrene
Concen: 18.896 ng/ul
RT: 19.557 min Scan# 2800
Delta R.T. -0.000 min
Lab File: BP019307.D
Acq: 08 Jan 2024 22:24

Tgt Ion:202 Resp: 1472061
Ion Ratio Lower Upper
202 100
101 8.7 7.2 10.8
100 7.1 5.8 8.6





#83

Butylbenzylphthalate

Concen: 20.472 ng/ul

RT: 20.433 min Scan# 2949

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:149 Resp: 547142

Ion Ratio Lower Upper

149 100

91 74.2 61.7 92.5

206 24.2 18.6 27.8



#84

3,3'-Dichlorobenzidine

Concen: 21.885 ng/ul

RT: 21.210 min Scan# 3081

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:252 Resp: 494820


Ion Ratio Lower Upper

252 100

254 63.9 50.6 76.0

126 7.4 6.0 9.0





#85

Benzo(a)anthracene

Concen: 20.492 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:228 Resp: 1588503

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.6

226 27.4 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.702 ng/ul

RT: 21.192 min Scan# 3078

Delta R.T. -0.006 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:149 Resp: 794198


Ion Ratio Lower Upper

149 100

167 28.7 22.9 34.3

279 5.2 4.6 7.0





#87

Chrysene

Concen: 20.591 ng/ul

RT: 21.321 min Scan# 3100

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:228 Resp: 1459036

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.8 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 19.576 ng/ul




RT: 22.104 min Scan# 3233


Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Tgt Ion:149 Resp: 1370725





#90

Benzo(b)fluoranthene

Concen: 21.007 ng/ul

RT: 22.921 min Scan# 3372

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:252 Resp: 1540054

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 5.9 4.9 7.3



#91

Benzo(k)fluoranthene

Concen: 20.966 ng/ul

RT: 22.968 min Scan# 3380

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:252 Resp: 1491130


Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 6.1 4.8 7.2





#93

Benzo(a)pyrene

Concen: 21.118 ng/ul

RT: 23.539 min Scan# 3477

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:252 Resp: 1391218

Ion Ratio Lower Upper

252 100

253 21.4 17.3 25.9

125 6.7 5.2 7.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.706 ng/ul

RT: 26.080 min Scan# 3909

Delta R.T. -0.006 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24




Tgt Ion:276 Resp: 1535252


Ion Ratio Lower Upper

276 100

138 14.7 12.2 18.4

277 24.6 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 19.722 ng/ul

RT: 26.098 min Scan# 3912

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Instrument :

BNA_P

ClientSampleId :

SICV601





Tgt Ion:278 Resp: 1271981

Ion Ratio Lower Upper

278 100

139 9.7 7.4 11.2

279 23.4 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 19.108 ng/ul

RT: 26.833 min Scan# 4037

Delta R.T. -0.000 min

Lab File: BP019307.D

Acq: 08 Jan 2024 22:24

Tgt Ion:276 Resp: 1195573

Ion Ratio Lower Upper

276 100

138 12.9 10.7 16.1

277 23.5 19.2 28.8

