

Data File : BP001625.D
Acq On : 16 Jan 2020 00:09
Operator : JU
Sample : L1123-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SB-04-COMP

Quant Time: Jan 16 06:06:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 21:52:04 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.64	152	41189	20.00	ng	0.00
21) Naphthalene-d8	10.42	136	145032	20.00	ng	0.00
39) Acenaphthene-d10	14.29	164	95570	20.00	ng	0.00
64) Phenanthrene-d10	17.05	188	210853	20.00	ng	0.00
76) Chrysene-d12	21.16	240	171422	20.00	ng	0.00
87) Perylene-d12	23.39	264	179493	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	221547	105.37	ng	0.00
7) Phenol-d6	6.84	99	345981	102.96	ng	0.00
23) Nitrobenzene-d5	8.79	82	237816	71.65	ng	0.00
42) 2,4,6-Tribromophenol	15.79	330	103071	70.85	ng	0.00
45) 2-Fluorobiphenyl	12.90	172	442020	68.00	ng	0.00
79) Terphenyl-d14	19.64	244	613948	65.51	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Phenol	6.86	94	9797	2.818	ng	82
18) Hexachloroethane	8.75	117	57585	46.231	ng	# 1
24) Nitrobenzene	8.82	77	7753	2.309	ng	# 4
31) Naphthalene	10.46	128	29435	3.846	ng	# 88
32) Benzoic acid	9.78	122	172	4.794	ng	# 1
37) 2-Methylnaphthalene	12.09	142	88394	14.490	ng	97
38) 1-Methylnaphthalene	12.09	142	88394	15.058	ng	99
49) Acenaphthylene	14.00	152	37019	4.217	ng	# 94
50) Dimethylphthalate	13.75	163	48131	6.115	ng	98
52) Acenaphthene	14.35	154	18573	3.416	ng	94
55) Dibenzofuran	14.69	168	25337	2.756	ng	99
58) Fluorene	15.35	166	49681	6.678	ng	# 99
71) Phenanthrene	17.09	178	618121	53.905	ng	99
72) Anthracene	17.18	178	107755	9.653	ng	99
75) Fluoranthene	19.08	202	835792	55.393	ng	98
78) Pyrene	19.43	202	719101	57.106	ng	98
81) Benzo(a)anthracene	21.14	228	292938	24.537	ng	95
83) Chrysene	21.19	228	310965	26.728	ng	97
86) Indeno(1,2,3-cd)pyrene	25.65	276	164063	12.047	ng	99
88) Benzo(b)fluoranthene	22.72	252	472536	41.778	ng	# 96
89) Benzo(k)fluoranthene	22.72	252	472536	42.849	ng	# 97
90) Benzo(a)pyrene	23.30	252	264789	24.656	ng	97
91) Dibenzo(a,h)anthracene	25.65	278	42391	3.826	ng	# 88
92) Benzo(g,h,i)perylene	26.34	276	167020	15.163	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BP001625.D
Acq On : 16 Jan 2020 00:09
Operator : JU
Sample : L1123-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SB-04-COMP

Quant Time: Jan 16 06:06:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 21:52:04 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.64 min Scan# 825
 Delta R.T. 0.00 min
 Lab File: BP001625.D
 Acq: 16 Jan 2020 00:09




Instrument :
 BNA_P
 ClientSampleId :
 SB-04-COMP


Tgt Ion:152 Resp: 41189
 Ion Ratio Lower Upper
 152 100
 150 153.5 122.0 183.0
 115 64.4 50.7 76.1



#5
 2-Fluorophenol
 Concen: 105.365 ng
 RT: 5.26 min Scan# 421
 Delta R.T. 0.00 min
 Lab File: BP001625.D
 Acq: 16 Jan 2020 00:09

Tgt Ion:112 Resp: 221547
 Ion Ratio Lower Upper
 112 100
 64 74.3 52.8 79.2
 63 47.1 33.1 49.7





#7

Phenol-d6

Concen: 102.957 ng

RT: 6.84 min Scan# 689

Delta R.T. 0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP





Tgt Ion: 99 Resp: 345981

Ion Ratio Lower Upper

99 100

42 32.6 24.5 36.7

71 55.1 42.4 63.6



#10

Phenol

Concen: 2.818 ng

RT: 6.86 min Scan# 692

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09




Tgt Ion: 94 Resp: 9797


Ion Ratio Lower Upper

94 100

65 43.4 17.5 57.5

66 74.1 36.7 76.7









#18
Hexachloroethane
Concen: 46.231 ng
RT: 8.75 min Scan# 1013
Delta R.T. 0.04 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion:117 Resp: 57585
Ion Ratio Lower Upper
117 100
119 691.7 71.8 107.8#
201 0.0 97.6 146.4#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.42 min Scan# 1297
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion:136 Resp: 145032
Ion Ratio Lower Upper
136 100
137 11.6 8.6 12.8
54 12.7 9.2 13.8
68 7.2 4.1 6.1#









#23
Nitrobenzene-d5
Concen: 71.653 ng
RT: 8.79 min Scan# 1020
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion	Ratio	Lower	Upper
82	100		
128	34.2	28.4	42.6
54	68.9	53.4	80.2



#24
Nitrobenzene
Concen: 2.309 ng
RT: 8.82 min Scan# 1026
Delta R.T. -0.01 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion	Ratio	Lower	Upper
77	100		
123	10.3	31.1	46.7#
65	107.7	12.4	18.6#









#31
Naphthalene
Concen: 3.846 ng
RT: 10.46 min Scan# 1305
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion:128 Resp: 29435
Ion Ratio Lower Upper
128 100
129 18.5 8.9 13.3#
127 15.5 10.6 16.0



#32
Benzoic acid
Concen: 4.794 ng
RT: 9.78 min Scan# 1188
Delta R.T. 0.02 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion:122 Resp: 172
Ion Ratio Lower Upper
122 100
105 471.3 119.5 159.5#
77 402.8 102.8 142.8#









#37
 2-Methylnaphthalene
 Concen: 14.490 ng
 RT: 12.09 min Scan# 1581
 Delta R.T. 0.00 min
 Lab File: BP001625.D
 Acq: 16 Jan 2020 00:09




Instrument :
 BNA_P
 ClientSampleId :
 SB-04-COMP


Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.8	68.4	102.6
115	47.1	35.1	52.7



#38
 1-Methylnaphthalene
 Concen: 15.058 ng
 RT: 12.09 min Scan# 1581
 Delta R.T. -0.22 min
 Lab File: BP001625.D
 Acq: 16 Jan 2020 00:09

Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.8	70.9	106.3
116	5.1	4.1	6.1





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.29 min Scan# 1955

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:164 Resp: 95570

Ion Ratio Lower Upper

164 100

162 101.9 79.8 119.6




160 45.2 35.0 52.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 70.851 ng

RT: 15.79 min Scan# 2210

Delta R.T. -0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:330 Resp: 103071

Ion Ratio Lower Upper

330 100

332 96.3 77.8 116.6



141 38.1 28.7 43.1




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#45
2-Fluorobiphenyl
Concen: 68.002 ng
RT: 12.90 min Scan# 1720
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion:	172	Resp:	442020
Ion	Ratio	Lower	Upper
172	100		
171	34.6	27.6	41.4
170	23.4	18.3	27.5



#49
Acenaphthylene
Concen: 4.217 ng
RT: 14.00 min Scan# 1907
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion:	152	Resp:	37019
Ion	Ratio	Lower	Upper
152	100		
151	21.1	15.9	23.9
153	17.1	10.4	15.6#









#50
Dimethylphthalate
Concen: 6.115 ng
RT: 13.75 min Scan# 1864
Delta R.T. -0.01 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.6	5.4
164	10.6	7.5	11.3



#52
Acenaphthene
Concen: 3.416 ng
RT: 14.35 min Scan# 1966
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion	Ratio	Lower	Upper
154	100		
153	118.0	89.4	134.2
152	57.3	42.6	64.0





#55

Dibenzenofuran

Concen: 2.756 ng

RT: 14.69 min Scan# 2023

Delta R.T. -0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:168 Resp: 25337

Ion Ratio Lower Upper

168 100

139 37.9 30.2 45.4




169 14.2 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#58

Fluorene

Concen: 6.678 ng

RT: 15.35 min Scan# 2135

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:166 Resp: 49681

Ion Ratio Lower Upper

166 100

165 96.5 76.9 115.3



167 15.5 10.3 15.5#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.05 min Scan# 2425

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:188 Resp: 210853

Ion Ratio Lower Upper

188 100

94 7.0 5.9 8.9




80 8.4 6.8 10.2



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BP0

Ion 80.00 (79.70 to 80.70): BP0



#71

Phenanthrene

Concen: 53.905 ng

RT: 17.09 min Scan# 2432

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:178 Resp: 618121

Ion Ratio Lower Upper

178 100

176 19.3 14.9 22.3



179 15.9 12.5 18.7




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Anthracene

Concen: 9.653 ng

RT: 17.18 min Scan# 2447

Delta R.T. -0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:178 Resp: 107755

Ion Ratio Lower Upper

178 100

176 19.1 15.0 22.6




179 15.7 11.8 17.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#75

Fluoranthene

Concen: 55.393 ng

RT: 19.08 min Scan# 2770

Delta R.T. 0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:202 Resp: 835792

Ion Ratio Lower Upper

202 100

101 7.3 0.0 28.4



203 18.0 0.0 37.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.16 min Scan# 3123

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:240 Resp: 171422

Ion Ratio Lower Upper

240 100

120 6.6 5.7 8.5




236 27.5 20.4 30.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Pyrene

Concen: 57.106 ng

RT: 19.43 min Scan# 2829

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:202 Resp: 719101

Ion Ratio Lower Upper

202 100

200 20.9 17.0 25.6



203 18.2 13.7 20.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 65.507 ng

RT: 19.64 min Scan# 2865

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:244 Resp: 613948

Ion Ratio Lower Upper

244 100

212 8.0 6.3 9.5




122 7.0 5.8 8.6



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#81

Benzo(a)anthracene

Concen: 24.537 ng

RT: 21.14 min Scan# 3120

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:228 Resp: 292938

Ion Ratio Lower Upper

228 100

226 29.5 21.2 31.8



229 21.7 16.1 24.1




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#83

Chrysene

Concen: 26.728 ng

RT: 21.19 min Scan# 3129

Delta R.T. 0.00 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion: 228 Resp: 310965

Ion Ratio Lower Upper

228 100

226 29.5 23.2 34.8

229 21.3 15.2 22.8






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 12.047 ng

RT: 25.65 min Scan# 3887

Delta R.T. -0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion: 276 Resp: 164063

Ion Ratio Lower Upper

276 100

138 13.2 11.1 16.7

277 25.2 20.4 30.6






Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Perylene-d12
Concen: 20.000 ng
RT: 23.39 min Scan# 3503
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.2	19.0	28.6
265	22.6	18.2	27.4



#88
Benzo(b)fluoranthene
Concen: 41.778 ng
RT: 22.72 min Scan# 3389
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.6	26.4
125	7.5	4.6	7.0









#89
Benzo(k)fluoranthene
Concen: 42.849 ng
RT: 22.72 min Scan# 3389
Delta R.T. -0.05 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09




Instrument :
BNA_P
ClientSampleId :
SB-04-COMP


Tgt Ion:252 Resp: 472536
Ion Ratio Lower Upper
252 100
253 23.7 18.2 27.2
125 7.5 4.6 6.8#



#90
Benzo(a)pyrene
Concen: 24.656 ng
RT: 23.30 min Scan# 3487
Delta R.T. 0.00 min
Lab File: BP001625.D
Acq: 16 Jan 2020 00:09

Tgt Ion:252 Resp: 264789
Ion Ratio Lower Upper
252 100
253 23.4 17.6 26.4
125 7.3 5.0 7.4





#91

Dibenzo(a,h)anthracene

Concen: 3.826 ng

RT: 25.65 min Scan# 3887

Delta R.T. -0.02 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09

Instrument :

BNA_P

ClientSampleId :

SB-04-COMP


Tgt Ion:278 Resp: 42391

Ion Ratio Lower Upper

278 100

139 12.7 6.7 10.1#




279 28.8 18.6 28.0#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 15.163 ng

RT: 26.34 min Scan# 4005

Delta R.T. 0.01 min

Lab File: BP001625.D

Acq: 16 Jan 2020 00:09


Tgt Ion:276 Resp: 167020

Ion Ratio Lower Upper

276 100

277 23.3 18.6 28.0

138 13.2 9.4 14.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

