

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011924\
 Data File : BP019403.D
 Acq On : 19 Jan 2024 15:49
 Operator : MA/JU
 Sample : PB158501BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS501

Quant Time: Jan 19 17:31:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	128954	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.563	136	517971	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.416	164	351605	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.175	188	844933	20.000	ng/u1	-0.01
79) Chrysene-d12	21.280	240	758763	20.000	ng/u1	0.00
88) Perylene-d12	23.639	264	736762	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	16464	4.828	ng/uL	-0.01
4) Pyridine-d5	3.628	84	251798	25.534	ng/u1	-0.01
7) Phenol-d5	6.958	99	318537	25.844	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.110	67	192840	26.158	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.305	132	227743	26.899	ng/u1	-0.01
15) 4-Methylphenol-d8	8.499	113	253711	25.947	ng/u1	-0.02
21) Nitrobenzene-d5	8.940	128	118859	27.276	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.657	143	142647	28.173	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.199	165	285421	27.902	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.716	131	302992	23.449	ng/u1	-0.01
46) Dimethylphthalate-d6	13.840	166	831196	26.758	ng/u1	0.00
49) Acenaphthylene-d8	14.110	160	892701	26.809	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.657	143	116730	24.471	ng/u1	-0.01
60) Fluorene-d10	15.410	176	707682	27.826	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.557	200	169088	27.106	ng/u1	0.00
73) Anthracene-d10	17.275	188	1115449	26.470	ng/u1	-0.01
81) Pyrene-d10	19.522	212	1356794	27.423	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.486	264	1114085	27.810	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	37071	10.495	ng/uL	98
5) Pyridine	3.652	79	258612	25.816	ng/u1	96
6) Benzaldehyde	6.916	77	184348	34.308	ng/u1	97
8) Phenol	6.987	94	333755	26.879	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.205	93	264667	26.573	ng/u1	99
12) 2-Chlorophenol	7.334	128	241121	27.554	ng/u1	99
13) 2-Methylphenol	8.234	108	248412	26.673	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.299	45	323576	27.303	ng/u1	100
16) Acetophenone	8.604	105	426461	27.380	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.587	70	226413	27.900	ng/u1	97
18) 4-Methylphenol	8.563	108	270666	27.299	ng/u1	96
19) Hexachloroethane	8.834	117	109685	27.750	ng/u1	99
22) Nitrobenzene	8.981	77	341393	28.005	ng/u1	96
23) Isophorone	9.504	82	647132	27.383	ng/u1	99
25) 2-Nitrophenol	9.687	139	150536	28.304	ng/u1	95
26) 2,4-Dimethylphenol	9.757	107	308120	26.869	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.993	93	370761	27.277	ng/u1	99
29) 2,4-Dichlorophenol	10.228	162	279365	28.339	ng/u1	99
30) Naphthalene	10.616	128	799068	27.378	ng/u1	99
32) 4-Chloroaniline	10.740	127	269457	21.316	ng/u1	99
33) Hexachlorobutadiene	10.887	225	239100	28.112	ng/u1	99
34) Caprolactam	11.540	113	86159	27.869	ng/u1	98
35) 4-Chloro-3-methylphenol	11.881	107	313503	28.645	ng/u1	99
36) 2-Methylnaphthalene	12.228	142	587746	27.676	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011924\
 Data File : BP019403.D
 Acq On : 19 Jan 2024 15:49
 Operator : MA/JU
 Sample : PB158501BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS501


Quant Time: Jan 19 17:31:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 09 03:05:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.451	142	578828	27.858	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.598	216	443197	29.087	ng/u1	98
40) Hexachlorocyclopentadiene	12.569	237	256252	26.204	ng/u1	99
41) 2,4,6-Trichlorophenol	12.851	196	274636	29.290	ng/u1	99
42) 2,4,5-Trichlorophenol	12.934	196	293866	29.245	ng/u1	99
43) 1,1'-Biphenyl	13.251	154	784907	27.023	ng/u1	100
44) 2-Chloronaphthalene	13.287	162	638539	27.184	ng/u1	99
45) 2-Nitroaniline	13.510	65	200923	29.821	ng/u1	99
47) Dimethylphthalate	13.887	163	856008	27.727	ng/u1	99
48) 2,6-Dinitrotoluene	14.010	165	177228	29.031	ng/u1	96
50) Acenaphthylene	14.140	152	1003064	27.503	ng/u1	99
51) 3-Nitroaniline	14.345	138	144789	27.191	ng/u1	99
52) Acenaphthene	14.481	153	660061	27.621	ng/u1	99
53) 2,4-Dinitrophenol	14.557	184	98935	24.716	ng/u1	96
55) 4-Nitrophenol	14.675	109	134667	26.143	ng/u1	97
56) Dibenzofuran	14.816	168	969051	28.324	ng/u1	100
57) 2,4-Dinitrotoluene	14.804	165	259387	30.283	ng/u1	96
58) 2,3,4,6-Tetrachlorophenol	15.051	232	263315	29.518	ng/u1	98
59) Diethylphthalate	15.251	149	840374	29.089	ng/u1	99
61) Fluorene	15.469	166	806583	28.783	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.463	204	485163	29.012	ng/u1	98
63) 4-Nitroaniline	15.510	138	136117	32.973	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.569	198	177874	26.903	ng/u1	98
67) N-Nitrosodiphenylamine	15.686	169	676621	27.165	ng/u1	99
68) 4-Bromophenyl-phenylether	16.363	248	318612	27.607	ng/u1	95
69) Hexachlorobenzene	16.475	284	370233	27.836	ng/u1	98
70) Atrazine	16.651	200	250806	23.100	ng/u1	100
71) Pentachlorophenol	16.828	266	224031	27.119	ng/u1	98
72) Phenanthrene	17.216	178	1311903	27.641	ng/u1	100
74) Anthracene	17.310	178	1310445	27.237	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.210	216	431245	27.565	ng/uL	100
76) Pentachlorobenzene	14.734	250	421720	27.750	ng/uL	99
77) Carbazole	17.592	167	1100489	26.885	ng/u1	99
78) Di-n-butylphthalate	18.139	149	1366832	27.902	ng/u1	99
80) Fluoranthene	19.198	202	1629922	27.870	ng/u1	99
82) Pyrene	19.551	202	1635005	27.545	ng/u1	99
83) Butylbenzylphthalate	20.433	149	568792	27.930	ng/u1	93
84) 3,3'-Dichlorobenzidine	21.204	252	484694	28.134	ng/u1	100
85) Benzo(a)anthracene	21.269	228	1614593	27.335	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	796805	27.258	ng/u1	100
87) Chrysene	21.321	228	1472866	27.279	ng/u1	100
89) Di-n-octyl phthalate	22.098	149	1258719	25.031	ng/u1	100
90) Benzo(b)fluoranthene	22.921	252	1471317	27.945	ng/u1	99
91) Benzo(k)fluoranthene	22.968	252	1403591	27.479	ng/u1	99
93) Benzo(a)pyrene	23.533	252	1315005	27.795	ng/u1	100
94) Indeno(1,2,3-cd)pyrene	26.080	276	1536005	27.452	ng/u1	98
95) Dibenzo(a,h)anthracene	26.092	278	1268831	27.393	ng/u1	100
96) Benzo(g,h,i)perylene	26.833	276	1222640	27.209	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 19 17:31:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 09 03:05:24 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.495 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:	88	Resp:	37071
Ion	Ratio	Lower	Upper
88	100		
43	29.5	22.0	33.0
58	69.9	54.6	82.0



#5
Pyridine
Concen: 25.816 ng/uL
RT: 3.652 min Scan# 96
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:	79	Resp:	258612
Ion	Ratio	Lower	Upper
79	100		
52	67.6	50.7	76.1
51	35.0	26.8	40.2









#6
Benzaldehyde
Concen: 34.308 ng/ul
RT: 6.916 min Scan# 651
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion: 77 Resp: 184348
Ion Ratio Lower Upper
77 100
105 91.5 75.3 112.9
106 87.2 71.5 107.3



#8
Phenol
Concen: 26.879 ng/ul
RT: 6.987 min Scan# 663
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion: 94 Resp: 333755
Ion Ratio Lower Upper
94 100
65 31.0 25.3 37.9
66 47.1 36.8 55.2









#10
Bis(2-Chloroethyl)ether
Concen: 26.573 ng/ul
RT: 7.205 min Scan# 700
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
Client SampleId :
SLCS501


Tgt Ion: 93	Resp: 264667
Ion Ratio	Lower Upper
93 100	
63 75.0	60.7 91.1
95 32.2	26.2 39.2



#12
2-Chlorophenol
Concen: 27.554 ng/ul
RT: 7.334 min Scan# 722
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt	Ion:128	Resp:	241121
Ion	Ratio	Lower	Upper
128	100		
64	52.3	41.3	61.9
130	31.5	25.9	38.9





#13

2-Methylphenol

Concen: 26.673 ng/ul

RT: 8.234 min Scan# 875

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :





SLCS501

Tgt Ion:108 Resp: 248412

Ion Ratio Lower Upper

108 100

107 90.0 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.303 ng/ul

RT: 8.299 min Scan# 886

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion: 45 Resp: 323576


Ion Ratio Lower Upper

45 100

77 17.7 14.2 21.4

79 13.0 10.5 15.7









#16
Acetophenone
Concen: 27.380 ng/ul
RT: 8.604 min Scan# 938
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:105 Resp: 426461
Ion Ratio Lower Upper
105 100
77 85.2 66.1 99.1
51 26.5 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 27.900 ng/ul
RT: 8.587 min Scan# 935
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion: 70 Resp: 226413
Ion Ratio Lower Upper
70 100
42 50.9 38.5 57.7
101 9.2 7.9 11.9
130 18.4 14.3 21.5









#18
4-Methylphenol
Concen: 27.299 ng/ul
RT: 8.563 min Scan# 931
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:108 Resp: 270666
Ion Ratio Lower Upper
108 100
107 110.9 92.6 139.0



#19
Hexachloroethane
Concen: 27.750 ng/ul
RT: 8.834 min Scan# 977
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:117 Resp: 109685
Ion Ratio Lower Upper
117 100
201 127.9 101.8 152.6
199 77.4 62.6 93.8









#22
Nitrobenzene
Concen: 28.005 ng/ul
RT: 8.981 min Scan# 1002
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion	Ratio	Lower	Upper
77	100		
123	39.9	30.0	45.0
65	14.3	10.5	15.7



#23
Isophorone
Concen: 27.383 ng/ul
RT: 9.504 min Scan# 1091
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion	Ratio	Lower	Upper
82	100		
95	8.1	6.6	10.0
138	15.5	13.0	19.4









#25
2-Nitrophenol
Concen: 28.304 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:139	Resp:	150536
Ion	Ratio	Lower Upper
139	100	
65	61.6	45.6 68.4
109	42.9	32.6 49.0



#26
2,4-Dimethylphenol
Concen: 26.869 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:107	Resp:	308120
Ion	Ratio	Lower Upper
107	100	
121	44.6	35.3 52.9
122	74.7	60.6 91.0





#27

Bis(2-Chloroethoxy)methane

Concen: 27.277 ng/ul

RT: 9.993 min Scan# 1174

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion: 93 Resp: 370761

Ion Ratio Lower Upper

93 100

95 31.5 25.9 38.9

123 9.7 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 28.339 ng/ul

RT: 10.228 min Scan# 1214

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:162 Resp: 279365


Ion Ratio Lower Upper

162 100

164 65.7 52.3 78.5

98 37.1 28.3 42.5









#30
Naphthalene
Concen: 27.378 ng/ul
RT: 10.616 min Scan# 1280
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:128 Resp: 799068
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.8 10.6 15.8



#32
4-Chloroaniline
Concen: 21.316 ng/ul
RT: 10.740 min Scan# 1301
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49


Tgt Ion:127 Resp: 269457
Ion Ratio Lower Upper
127 100
129 32.0 25.8 38.8








#33
Hexachlorobutadiene
Concen: 28.112 ng/ul
RT: 10.887 min Scan# 1326
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Instrument :
BNA_P
ClientSampleId :
SLCS501







Tgt Ion:225 Resp: 239100
Ion Ratio Lower Upper
225 100
223 63.8 49.9 74.9
227 63.3 51.0 76.6



#34
Caprolactam
Concen: 27.869 ng/ul
RT: 11.540 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:113 Resp: 86159
Ion Ratio Lower Upper
113 100
55 163.0 133.3 199.9
56 134.5 108.0 162.0









#35
4-Chloro-3-methylphenol
Concen: 28.645 ng/ul
RT: 11.881 min Scan# 1495
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:107 Resp: 313503
Ion Ratio Lower Upper
107 100
144 23.6 17.8 26.8
142 71.1 56.3 84.5



#36
2-Methylnaphthalene
Concen: 27.676 ng/ul
RT: 12.228 min Scan# 1554
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:142 Resp: 587746
Ion Ratio Lower Upper
142 100
141 88.7 71.1 106.7









#37
1-Methylnaphthalene
Concen: 27.858 ng/ul
RT: 12.451 min Scan# 1592
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:142 Resp: 578828
Ion Ratio Lower Upper
142 100
141 93.8 74.4 111.6
116 4.8 3.9 5.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.087 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:216 Resp: 443197
Ion Ratio Lower Upper
216 100
214 77.2 63.6 95.4
179 19.9 15.1 22.7
108 13.9 11.3 16.9





#40

Hexachlorocyclopentadiene

Concen: 26.204 ng/ul

RT: 12.569 min Scan# 1612

Delta R.T. -0.018 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:237 Resp: 256252

Ion Ratio Lower Upper

237 100

235 63.1 50.0 75.0

272 12.6 11.4 17.2



#41

2,4,6-Trichlorophenol

Concen: 29.290 ng/ul

RT: 12.851 min Scan# 1660

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:196 Resp: 274636


Ion Ratio Lower Upper

196 100

198 98.2 78.2 117.2

200 32.5 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 29.245 ng/ul

RT: 12.934 min Scan# 1674

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:196 Resp: 293866

Ion Ratio Lower Upper

196 100

198 98.4 78.9 118.3

200 32.3 25.0 37.6



#43

1,1'-Biphenyl

Concen: 27.023 ng/ul

RT: 13.251 min Scan# 1728

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:154 Resp: 784907


Ion Ratio Lower Upper

154 100

153 40.6 32.6 48.8

76 11.5 9.3 13.9









#44
2-Chloronaphthalene
Concen: 27.184 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. -0.018 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:162 Resp: 638539
Ion Ratio Lower Upper
162 100
164 33.0 26.6 40.0
127 38.2 29.4 44.2



#45
2-Nitroaniline
Concen: 29.821 ng/ul
RT: 13.510 min Scan# 1772
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion: 65 Resp: 200923
Ion Ratio Lower Upper
65 100
92 66.1 52.1 78.1
138 93.6 74.8 112.2









#47
Dimethylphthalate
Concen: 27.727 ng/ul
RT: 13.887 min Scan# 1836
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:163 Resp: 856008
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 9.9 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 29.031 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:165 Resp: 177228
Ion Ratio Lower Upper
165 100
89 66.2 49.9 74.9
121 19.7 16.5 24.7





#50

Acenaphthylene

Concen: 27.503 ng/ul

RT: 14.140 min Scan# 1879

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:152 Resp: 1003064

Ion Ratio Lower Upper

152 100

151 19.5 15.8 23.6

153 13.3 10.3 15.5



#51

3-Nitroaniline

Concen: 27.191 ng/ul

RT: 14.345 min Scan# 1914

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:138 Resp: 144789


Ion Ratio Lower Upper

138 100

108 12.7 10.6 15.8

92 131.0 105.8 158.6









#52
Acenaphthene
Concen: 27.621 ng/ul
RT: 14.481 min Scan# 1937
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:153 Resp: 660061
Ion Ratio Lower Upper
153 100
152 47.0 38.2 57.2
154 87.1 68.9 103.3



#53
2,4-Dinitrophenol
Concen: 24.716 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:184 Resp: 98935
Ion Ratio Lower Upper
184 100
63 59.7 43.6 65.4
154 65.9 52.1 78.1









#55
4-Nitrophenol
Concen: 26.143 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:109 Resp: 134667
Ion Ratio Lower Upper
109 100
139 90.5 73.2 109.8
65 108.8 90.9 136.3



#56
Dibenzofuran
Concen: 28.324 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:168 Resp: 969051
Ion Ratio Lower Upper
168 100
139 38.2 30.3 45.5









#57
2,4-Dinitrotoluene
Concen: 30.283 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:165 Resp: 259387
Ion Ratio Lower Upper
165 100
63 54.8 41.3 61.9
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.518 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:232 Resp: 263315
Ion Ratio Lower Upper
232 100
131 33.3 27.9 41.9
130 1.9 1.8 2.8
166 25.1 20.9 31.3









#59
Diethylphthalate
Concen: 29.089 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:149 Resp: 840374
Ion Ratio Lower Upper
149 100
177 24.1 18.8 28.2
150 12.8 10.2 15.4



#61
Fluorene
Concen: 28.783 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:166 Resp: 806583
Ion Ratio Lower Upper
166 100
165 97.3 78.6 118.0
167 13.7 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 29.012 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:204 Resp: 485163
Ion Ratio Lower Upper
204 100
206 32.0 27.0 40.4
141 50.9 41.5 62.3



#63
4-Nitroaniline
Concen: 32.973 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:138 Resp: 136117
Ion Ratio Lower Upper
138 100
92 59.7 48.9 73.3
108 114.0 90.4 135.6









#66
4,6-Dinitro-2-methylphenol
Concen: 26.903 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:198 Resp: 177874
Ion Ratio Lower Upper
198 100
182 3.2 3.2 4.8
77 25.0 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 27.165 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:169 Resp: 676621
Ion Ratio Lower Upper
169 100
168 69.6 54.8 82.2
167 36.8 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 27.607 ng/ul
RT: 16.363 min Scan# 2257
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:248 Resp: 318612
Ion Ratio Lower Upper
248 100
250 100.2 75.6 113.4
141 62.7 48.7 73.1



#69
Hexachlorobenzene
Concen: 27.836 ng/ul
RT: 16.475 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:284 Resp: 370233
Ion Ratio Lower Upper
284 100
142 27.4 23.6 35.4
249 32.3 26.1 39.1





#70

Atrazine

Concen: 23.100 ng/ul

RT: 16.651 min Scan# 2306

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:200 Resp: 250806

Ion Ratio Lower Upper

200 100

173 23.7 19.4 29.0

215 49.6 39.7 59.5



#71

Pentachlorophenol

Concen: 27.119 ng/ul

RT: 16.828 min Scan# 2336

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:266 Resp: 224031


Ion Ratio Lower Upper

266 100

264 64.7 49.4 74.2

268 64.3 51.6 77.4





#72

Phenanthrene

Concen: 27.641 ng/ul

RT: 17.216 min Scan# 2402

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:178 Resp: 1311903

Ion Ratio Lower Upper

178 100

179 15.0 12.1 18.1

176 19.3 15.4 23.2



#74

Anthracene

Concen: 27.237 ng/ul

RT: 17.310 min Scan# 2418

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:178 Resp: 1310445


Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 18.7 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.565 ng/uL
RT: 13.210 min Scan# 1721
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:216 Resp: 431245
Ion Ratio Lower Upper
216 100
214 77.0 61.8 92.8
218 49.2 39.0 58.6



#76
Pentachlorobenzene
Concen: 27.750 ng/uL
RT: 14.734 min Scan# 1980
Delta R.T. -0.012 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:250 Resp: 421720
Ion Ratio Lower Upper
250 100
248 61.3 50.8 76.2
252 62.7 50.2 75.4





#77

Carbazole

Concen: 26.885 ng/ul

RT: 17.592 min Scan# 2466

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:167 Resp: 1100489

Ion Ratio Lower Upper

167 100

166 21.8 17.3 25.9

139 12.3 10.1 15.1



#78

Di-n-butylphthalate

Concen: 27.902 ng/ul

RT: 18.139 min Scan# 2559

Delta R.T. -0.012 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:149 Resp: 1366832


Ion Ratio Lower Upper

149 100

150 9.1 7.0 10.6

104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 27.870 ng/ul
RT: 19.198 min Scan# 2739
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:202 Resp: 1629922
Ion Ratio Lower Upper
202 100
101 6.9 5.9 8.9
100 5.3 4.6 7.0



#82
Pyrene
Concen: 27.545 ng/ul
RT: 19.551 min Scan# 2799
Delta R.T. -0.006 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:202 Resp: 1635005
Ion Ratio Lower Upper
202 100
101 8.3 7.2 10.8
100 6.9 5.8 8.6





#83

Butylbenzylphthalate

Concen: 27.930 ng/ul

RT: 20.433 min Scan# 2948

Delta R.T. -0.000 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:149 Resp: 568792

Ion Ratio Lower Upper

149 100

91 70.8 61.7 92.5

206 25.8 18.6 27.8



#84

3,3'-Dichlorobenzidine

Concen: 28.134 ng/ul

RT: 21.204 min Scan# 3080

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:252 Resp: 484694


Ion Ratio Lower Upper

252 100

254 63.5 50.6 76.0

126 7.6 6.0 9.0





#85

Benzo(a)anthracene

Concen: 27.335 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.000 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:228 Resp: 1614593

Ion Ratio Lower Upper

228 100

229 19.3 15.8 23.6

226 27.5 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 27.258 ng/ul

RT: 21.192 min Scan# 3078

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:149 Resp: 796805


Ion Ratio Lower Upper

149 100

167 28.7 22.9 34.3

279 5.5 4.6 7.0





#87

Chrysene

Concen: 27.279 ng/ul

RT: 21.321 min Scan# 3100

Delta R.T. -0.000 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:228 Resp: 1472866

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 25.031 ng/ul




RT: 22.098 min Scan# 3232


Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Tgt Ion:149 Resp: 1258719









#90
Benzo(b)fluoranthene
Concen: 27.945 ng/ul
RT: 22.921 min Scan# 3371
Delta R.T. -0.000 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49




Instrument :
BNA_P
ClientSampleId :
SLCS501


Tgt Ion:252 Resp: 1471317
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 5.3 4.9 7.3



#91
Benzo(k)fluoranthene
Concen: 27.479 ng/ul
RT: 22.968 min Scan# 3380
Delta R.T. -0.000 min
Lab File: BP019403.D
Acq: 19 Jan 2024 15:49

Tgt Ion:252 Resp: 1403591
Ion Ratio Lower Upper
252 100
253 22.1 17.3 25.9
125 5.4 4.8 7.2





#93

Benzo(a)pyrene

Concen: 27.795 ng/ul

RT: 23.533 min Scan# 3476

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:252 Resp: 1315005

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 6.3 5.2 7.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 27.452 ng/ul

RT: 26.080 min Scan# 3909

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49




Tgt Ion:276 Resp: 1536005


Ion Ratio Lower Upper

276 100

138 14.1 12.2 18.4

277 25.4 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 27.393 ng/ul

RT: 26.092 min Scan# 3911

Delta R.T. -0.006 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Instrument :

BNA_P

ClientSampleId :

SLCS501





Tgt Ion:278 Resp: 1268831

Ion Ratio Lower Upper

278 100

139 9.5 7.4 11.2

279 23.5 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 27.209 ng/ul

RT: 26.833 min Scan# 4037

Delta R.T. -0.000 min

Lab File: BP019403.D

Acq: 19 Jan 2024 15:49

Tgt Ion:276 Resp: 1222640

Ion Ratio Lower Upper

276 100

138 13.1 10.7 16.1

277 23.7 19.2 28.8

