

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013125\
 Data File : BP023820.D
 Acq On : 31 Jan 2025 11:16
 Operator : RC/JU
 Sample : PB166295BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK295

Quant Time: Jan 31 11:44:22 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 31 11:40:55 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	352524	20.000	ng/u1	0.00
20) Naphthalene-d8	10.557	136	1439295	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.404	164	987189	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	2107861	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	2175930	20.000	ng/u1	0.00
88) Perylene-d12	25.021	264	2349437	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	63382	7.470	ng/uL	0.00
4) Pyridine-d5	3.711	84	918770	37.527	ng/u1	0.00
7) Phenol-d5	6.969	99	1167997	37.439	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.116	67	569102	34.463	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	939714	38.358	ng/u1	0.00
15) 4-Methylphenol-d8	8.493	113	939372	36.860	ng/u1	0.00
21) Nitrobenzene-d5	8.928	128	412922	35.859	ng/u1	0.00
24) 2-Nitrophenol-d4	9.646	143	452900	36.450	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.193	165	876949	38.473	ng/u1	0.00
31) 4-Chloroaniline-d4	10.687	131	1217230	35.708	ng/u1	0.00
46) Dimethylphthalate-d6	13.810	166	2782962	37.794	ng/u1	0.00
49) Acenaphthylene-d8	14.093	160	3035778	36.554	ng/u1	0.00
54) 4-Nitrophenol-d4	14.634	143	463454	31.693	ng/u1	0.00
60) Fluorene-d10	15.410	176	2352907	37.944	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.522	200	340452	26.181	ng/u1	0.00
73) Anthracene-d10	17.292	188	3937047	38.041	ng/u1	-0.01
81) Pyrene-d10	19.663	212	4554413	39.063	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.792	264	4773689	38.943	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013125\
Data File : BP023820.D
Acq On : 31 Jan 2025 11:16
Operator : RC/JU
Sample : PB166295BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK295

Quant Time: Jan 31 11:44:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 31 11:40:55 2025
Response via : Initial Calibration

