

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
 Data File : BP009046.D
 Acq On : 07 Feb 2022 14:33
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020722

Quant Time: Feb 07 15:16:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 07 15:13:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	115	0.00
2	1,4-Dioxane	40.000	38.012	5.0	112	0.00
3	Pyridine	40.000	35.102	12.2	101	0.00
4	n-Nitrosodimethylamine	40.000	34.166	14.6	95	0.00
5 S	2-Fluorophenol	80.000	79.419	0.7	116	0.00
6	Aniline	40.000	38.867	2.8	116	0.00
7 S	Phenol-d6	80.000	78.367	2.0	116	0.00
8	2-Chlorophenol	40.000	38.746	3.1	116	0.00
9	Benzaldehyde	40.000	39.259	1.9	120	0.00
10 C	Phenol	40.000	39.082	2.3	116	0.00
11	bis(2-Chloroethyl)ether	40.000	39.221	1.9	118	0.00
12	1,3-Dichlorobenzene	40.000	38.552	3.6	116	0.00
13 C	1,4-Dichlorobenzene	40.000	38.463	3.8	115	0.00
14	1,2-Dichlorobenzene	40.000	38.291	4.3	115	0.00
15	Benzyl Alcohol	40.000	40.640	-1.6	119	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.746	3.1	119	0.00
17	2-Methylphenol	40.000	38.960	2.6	117	0.00
18	Hexachloroethane	40.000	39.065	2.3	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.481	1.3	119	0.00
20	3+4-Methylphenols	40.000	39.192	2.0	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	40.153	-0.4	118	0.00
23 S	Nitrobenzene-d5	80.000	80.213	-0.3	118	0.00
24	Nitrobenzene	40.000	39.883	0.3	118	0.00
25	Isophorone	40.000	39.390	1.5	121	0.00
26 C	2-Nitrophenol	40.000	41.494	-3.7	120	0.00
27	2,4-Dimethylphenol	40.000	39.990	0.0	119	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.560	1.1	120	0.00
29 C	2,4-Dichlorophenol	40.000	40.220	-0.5	119	0.00
30	1,2,4-Trichlorobenzene	40.000	39.158	2.1	118	0.00
31	Naphthalene	40.000	38.956	2.6	117	0.00
32	Benzoic acid	40.000	38.840	2.9	123	0.00
33	4-Chloroaniline	40.000	39.454	1.4	119	0.00
34 C	Hexachlorobutadiene	40.000	39.606	1.0	119	0.00
35	Caprolactam	40.000	36.866	7.8	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.890	2.8	121	0.00
37	2-Methylnaphthalene	40.000	38.508	3.7	119	0.00
38	1-Methylnaphthalene	40.000	38.328	4.2	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.864	-2.2	120	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.002	-0.0	131	0.00
42 S	2,4,6-Tribromophenol	80.000	75.714	5.4	119	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.513	-3.8	120	0.00
44	2,4,5-Trichlorophenol	40.000	40.932	-2.3	120	0.00
45 S	2-Fluorobiphenyl	80.000	81.122	-1.4	119	0.00
46	1,1'-Biphenyl	40.000	40.269	-0.7	119	0.00
47	2-Chloronaphthalene	40.000	40.249	-0.6	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
 Data File : BP009046.D
 Acq On : 07 Feb 2022 14:33
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020722

Quant Time: Feb 07 15:16:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 07 15:13:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.156	-2.9	122	0.00
49	Acenaphthylene	40.000	39.996	0.0	119	0.00
50	Dimethylphthalate	40.000	37.906	5.2	119	0.00
51	2,6-Dinitrotoluene	40.000	39.150	2.1	120	0.00
52 C	Acenaphthene	40.000	39.378	1.6	119	0.00
53	3-Nitroaniline	40.000	39.343	1.6	119	0.00
54 P	2,4-Dinitrophenol	40.000	38.375	4.1	117	0.00
55	Dibenzofuran	40.000	38.722	3.2	119	0.00
56 P	4-Nitrophenol	40.000	38.366	4.1	120	0.00
57	2,4-Dinitrotoluene	40.000	38.616	3.5	119	0.00
58	Fluorene	40.000	38.724	3.2	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.964	2.6	120	0.00
60	Diethylphthalate	40.000	36.689	8.3	118	0.00
61	4-Chlorophenyl-phenylether	40.000	38.620	3.5	120	0.00
62	4-Nitroaniline	40.000	38.259	4.4	115	0.00
63	Azobenzene	40.000	38.485	3.8	120	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	117	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.924	0.2	119	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.423	-3.6	119	0.00
67	4-Bromophenyl-phenylether	40.000	41.241	-3.1	121	0.00
68	Hexachlorobenzene	40.000	40.149	-0.4	118	0.00
69	Atrazine	40.000	39.397	1.5	116	0.00
70 C	Pentachlorophenol	40.000	40.576	-1.4	117	0.00
71	Phenanthrene	40.000	39.367	1.6	115	0.00
72	Anthracene	40.000	39.455	1.4	115	0.00
73	Carbazole	40.000	38.151	4.6	111	0.00
74	Di-n-butylphthalate	40.000	38.566	3.6	116	0.00
75 C	Fluoranthene	40.000	36.952	7.6	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	43.526	-8.8	88	0.00
78	Pyrene	40.000	39.125	2.2	105	0.00
79 S	Terphenyl-d14	80.000	77.636	3.0	104	0.00
80	Butylbenzylphthalate	40.000	38.631	3.4	102	0.00
81	Benzo(a)anthracene	40.000	39.838	0.4	96	0.00
82	3,3'-Dichlorobenzidine	40.000	42.044	-5.1	95	0.00
83	Chrysene	40.000	39.338	1.7	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.397	4.0	100	0.00
85 c	Di-n-octyl phthalate	40.000	41.611	-4.0	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.044	-7.6	97	0.00
88	Benzo(b)fluoranthene	40.000	37.655	5.9	95	0.00
89	Benzo(k)fluoranthene	40.000	38.036	4.9	95	0.00
90 C	Benzo(a)pyrene	40.000	39.572	1.1	95	0.00
91	Dibenzo(a,h)anthracene	40.000	42.586	-6.5	96	0.00
92	Benzo(g,h,i)perylene	40.000	42.506	-6.3	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
Data File : BP009046.D
Acq On : 07 Feb 2022 14:33
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP020722

Quant Time: Feb 07 15:16:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 07 15:13:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0