



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
 Data File : BP009038.D
 Acq On : 07 Feb 2022 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 07 15:13:17 2022






AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.2	99130	PASS
68	69	0.00	2	1.7	1809	PASS
69	198	0.00	100	26.9	105784	PASS
70	69	0.00	2	0.6	601	PASS
127	198	10	80	40.7	160021	PASS
197	198	0.00	2	0.5	2012	PASS
198	198	100	100	100.0	393219	PASS
199	198	5	9	6.8	26578	PASS
275	198	10	60	28.6	112419	PASS
365	198	1	100	3.9	15301	PASS
441	198	0.01	100	11.7	46201	PASS
442	442	100	100	100.0	299221	PASS
443	442	15	24	19.5	58219	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
Data File : BP009038.D
Acq On : 07 Feb 2022 10:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Feb 07 12:47:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 07 12:37:20 2022
Response via : Initial Calibration



TIC: BP009038.D\data.ms

(70) Pentachlorophenol (C)

16.869min (-0.000) 4722960.63 ng




response 460186

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	64.56
264.00	62.90	63.99
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020722\
 Data File : BP009038.D
 Acq On : 07 Feb 2022 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Feb 07 12:47:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 07 12:37:20 2022
 Response via : Initial Calibration



TIC: BP009038.D\data.ms

(77) Benzidine

19.404min (-0.000) 2393648.30 ng

response 3004316

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	13.74
183.00	11.90	11.33
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/7/2022	BNA_P	<u>BP009038.D</u>
Compound Name	Response	Retention Time
DDT	1464717	20.6
DDD	5304	20.18
DDE	2452	19.69
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7756	1472473	0.53