

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
 Data File : BP013787.D
 Acq On : 08 Feb 2023 09:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758

Quant Time: Feb 08 23:53:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	151231	20.000	ng/u1	0.00
20) Naphthalene-d8	10.951	136	682945	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	492387	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	1173845	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	1294218	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1396675	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.469	96	29628	8.120	ng/uL	0.00
4) Pyridine-d5	3.899	84	218741	20.221	ng/u1	0.00
7) Phenol-d5	7.269	99	264600	19.637	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	160569	20.336	ng/u1	0.00
11) 2-Chlorophenol-d4	7.646	132	204067	20.573	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	220582	20.028	ng/u1	0.00
21) Nitrobenzene-d5	9.299	128	95558	20.110	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	111541	20.330	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	233201	20.858	ng/u1	0.00
31) 4-Chloroaniline-d4	11.081	131	331158	20.896	ng/u1	0.00
46) Dimethylphthalate-d6	14.157	166	789233	20.351	ng/u1	0.00
49) Acenaphthylene-d8	14.451	160	859323	20.470	ng/u1	0.00
54) 4-Nitrophenol-d4	14.939	143	137756	19.938	ng/u1	0.00
60) Fluorene-d10	15.751	176	675641	20.454	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	139977	19.437	ng/u1	0.00
73) Anthracene-d10	17.610	188	1096360	20.426	ng/u1	0.00
81) Pyrene-d10	19.827	212	1408758	19.195	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	1406283	20.095	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	30713	7.841	ng/uL	93
5) Pyridine	3.922	79	213834	20.003	ng/u1	99
6) Benzaldehyde	7.252	77	122468	20.155	ng/u1	96
8) Phenol	7.293	94	275090	19.888	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.534	93	225117	20.311	ng/u1	98
12) 2-Chlorophenol	7.681	128	209688	20.603	ng/u1	100
13) 2-Methylphenol	8.563	108	210752	19.962	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.652	45	254404	20.179	ng/u1	98
16) Acetophenone	8.951	105	364011	20.610	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.934	70	181769	20.724	ng/u1	96
18) 4-Methylphenol	8.893	108	236220	20.266	ng/u1	97
19) Hexachloroethane	9.216	117	86254	20.696	ng/u1	98
22) Nitrobenzene	9.340	77	259952	20.794	ng/u1	100
23) Isophorone	9.869	82	529328	20.069	ng/u1	98
25) 2-Nitrophenol	10.057	139	123358	20.745	ng/u1	97
26) 2,4-Dimethylphenol	10.104	107	271332	20.494	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.346	93	324389	20.366	ng/u1	98
29) 2,4-Dichlorophenol	10.593	162	227990	20.713	ng/u1	98
30) Naphthalene	11.004	128	737799	20.294	ng/u1	99
32) 4-Chloroaniline	11.104	127	334381	20.929	ng/u1	97
33) Hexachlorobutadiene	11.287	225	157859	20.359	ng/u1	97
34) Caprolactam	11.881	113	50097	12.971	ng/u1	96
35) 4-Chloro-3-methylphenol	12.216	107	259957	20.457	ng/u1	99
36) 2-Methylnaphthalene	12.598	142	528547	20.678	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
 Data File : BP013787.D
 Acq On : 08 Feb 2023 09:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758


Quant Time: Feb 08 23:53:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	540931	20.482	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.957	216	316696	20.411	ng/ul	98
40) Hexachlorocyclopentadiene	12.940	237	181080	18.175	ng/ul	99
41) 2,4,6-Trichlorophenol	13.192	196	208102	20.438	ng/ul	97
42) 2,4,5-Trichlorophenol	13.263	196	222569	20.194	ng/ul	99
43) 1,1'-Biphenyl	13.592	154	757033	20.490	ng/ul	100
44) 2-Chloronaphthalene	13.640	162	589791	20.292	ng/ul	100
45) 2-Nitroaniline	13.839	65	142415	20.343	ng/ul	97
47) Dimethylphthalate	14.210	163	787435	20.472	ng/ul	100
48) 2,6-Dinitrotoluene	14.334	165	137007	20.387	ng/ul	98
50) Acenaphthylene	14.481	152	920161	20.318	ng/ul	99
51) 3-Nitroaniline	14.663	138	125112	18.830	ng/ul	93
52) Acenaphthene	14.822	153	637351	20.107	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	84201	17.570	ng/ul	91
55) 4-Nitrophenol	14.957	109	122630	20.228	ng/ul	97
56) Dibenzofuran	15.157	168	920144	20.247	ng/ul	99
57) 2,4-Dinitrotoluene	15.116	165	212583	21.218	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.381	232	214018	20.602	ng/ul	98
59) Diethylphthalate	15.569	149	793541	20.545	ng/ul	99
61) Fluorene	15.804	166	755415	20.376	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.792	204	395607	20.522	ng/ul	97
63) 4-Nitroaniline	15.822	138	124407	19.773	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.875	198	137290	19.446	ng/ul	99
67) N-Nitrosodiphenylamine	16.010	169	656355	20.571	ng/ul	99
68) 4-Bromophenyl-phenylether	16.692	248	250862	20.237	ng/ul	98
69) Hexachlorobenzene	16.810	284	292989	20.284	ng/ul	97
70) Atrazine	16.957	200	259085	20.306	ng/ul	99
71) Pentachlorophenol	17.157	266	179704	19.119	ng/ul	99
72) Phenanthrene	17.557	178	1269849	20.443	ng/ul	100
74) Anthracene	17.645	178	1278480	20.457	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.563	216	316623	20.206	ng/uL	98
76) Pentachlorobenzene	15.075	250	329453	20.374	ng/uL	98
77) Carbazole	17.910	167	1165041	21.163	ng/ul	100
78) Di-n-butylphthalate	18.439	149	1370532	20.527	ng/ul	99
80) Fluoranthene	19.504	202	1609887	19.161	ng/ul	99
82) Pyrene	19.857	202	1712930	19.122	ng/ul	98
83) Butylbenzylphthalate	20.710	149	587808	19.173	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.492	252	585715	20.108	ng/ul	98
85) Benzo(a)anthracene	21.568	228	1778391	20.110	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	926001	19.724	ng/ul	98
87) Chrysene	21.627	228	1712663	20.527	ng/ul	98
89) Di-n-octyl phthalate	22.445	149	1605129	16.702	ng/ul	100
90) Benzo(b)fluoranthene	23.363	252	1816808	19.508	ng/ul	99
91) Benzo(k)fluoranthene	23.415	252	1816428	19.659	ng/ul	100
93) Benzo(a)pyrene	24.039	252	1589595	20.138	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.851	276	1860851	21.017	ng/ul	98
95) Dibenzo(a,h)anthracene	26.868	278	1557642	21.061	ng/ul	99
96) Benzo(g,h,i)perylene	27.692	276	1468068	20.827	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 08 23:53:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 08 00:37:14 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.841 ng/uL
RT: 3.505 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 88 Resp: 30713
Ion Ratio Lower Upper
88 100
43 26.9 20.1 30.1
58 60.0 53.0 79.6



#5
Pyridine
Concen: 20.003 ng/uL
RT: 3.922 min Scan# 125
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion: 79 Resp: 213834
Ion Ratio Lower Upper
79 100
52 67.5 55.1 82.7
51 31.8 25.7 38.5









#6
Benzaldehyde
Concen: 20.155 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 77 Resp: 122468
Ion Ratio Lower Upper
77 100
105 96.3 79.5 119.3
106 90.5 76.5 114.7



#8
Phenol
Concen: 19.888 ng/ul
RT: 7.293 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion: 94 Resp: 275090
Ion Ratio Lower Upper
94 100
65 27.1 22.2 33.4
66 38.0 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 20.311 ng/ul
RT: 7.534 min Scan# 739
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 93 Resp: 225117
Ion Ratio Lower Upper
93 100
63 72.0 59.4 89.0
95 32.3 25.8 38.8



#12
2-Chlorophenol
Concen: 20.603 ng/ul
RT: 7.681 min Scan# 764
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:128 Resp: 209688
Ion Ratio Lower Upper
128 100
64 46.5 37.3 55.9
130 33.1 26.3 39.5





#13

2-Methylphenol

Concen: 19.962 ng/ul

RT: 8.563 min Scan# 914

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :





SSTD020758

Tgt Ion:108 Resp: 210752

Ion Ratio Lower Upper

108 100

107 86.7 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.179 ng/ul

RT: 8.652 min Scan# 929

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion: 45 Resp: 254404


Ion Ratio Lower Upper

45 100

77 17.5 13.4 20.2

79 12.6 11.0 16.4









#16
Acetophenone
Concen: 20.610 ng/ul
RT: 8.951 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:105 Resp: 364011
Ion Ratio Lower Upper
105 100
77 78.6 61.0 91.4
51 25.3 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.724 ng/ul
RT: 8.934 min Scan# 977
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion: 70 Resp: 181769
Ion Ratio Lower Upper
70 100
42 52.2 39.8 59.8
101 9.6 7.4 11.2
130 20.5 14.6 22.0









#18
4-Methylphenol
Concen: 20.266 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:108 Resp: 236220
Ion Ratio Lower Upper
108 100
107 110.0 90.9 136.3



#19
Hexachloroethane
Concen: 20.696 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:117 Resp: 86254
Ion Ratio Lower Upper
117 100
201 105.5 86.6 129.8
199 69.9 55.2 82.8









#22
Nitrobenzene
Concen: 20.794 ng/ul
RT: 9.340 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 77 Resp: 259952
Ion Ratio Lower Upper
77 100
123 43.8 35.1 52.7
65 13.0 10.6 16.0



#23
Isophorone
Concen: 20.069 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion: 82 Resp: 529328
Ion Ratio Lower Upper
82 100
95 8.0 6.0 9.0
138 17.8 13.7 20.5









#25
2-Nitrophenol
Concen: 20.745 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:139 Resp: 123358
Ion Ratio Lower Upper
139 100
65 53.4 39.9 59.9
109 34.1 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 20.494 ng/ul
RT: 10.104 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:107 Resp: 271332
Ion Ratio Lower Upper
107 100
121 48.5 37.8 56.6
122 80.6 63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 20.366 ng/ul

RT: 10.346 min Scan# 1217

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion: 93 Resp: 324389

Ion Ratio Lower Upper

93 100

95 32.4 25.0 37.6

123 11.2 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 20.713 ng/ul

RT: 10.593 min Scan# 1259

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:162 Resp: 227990


Ion Ratio Lower Upper

162 100

164 65.3 51.4 77.2

98 36.2 27.5 41.3









#30
Naphthalene
Concen: 20.294 ng/ul
RT: 11.004 min Scan# 1329
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:128 Resp: 737799
Ion Ratio Lower Upper
128 100
129 10.5 8.6 12.8
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 20.929 ng/ul
RT: 11.104 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:127 Resp: 334381
Ion Ratio Lower Upper
127 100
129 31.0 26.2 39.4









#33
Hexachlorobutadiene
Concen: 20.359 ng/ul
RT: 11.287 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:225 Resp: 157859
Ion Ratio Lower Upper
225 100
223 64.0 50.6 76.0
227 65.1 49.0 73.4



#34
Caprolactam
Concen: 12.971 ng/ul
RT: 11.881 min Scan# 1478
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:113 Resp: 50097
Ion Ratio Lower Upper
113 100
55 147.6 120.9 181.3
56 111.9 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 20.457 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:107 Resp: 259957
Ion Ratio Lower Upper
107 100
144 24.8 20.0 30.0
142 78.4 62.1 93.1



#36
2-Methylnaphthalene
Concen: 20.678 ng/ul
RT: 12.598 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:142 Resp: 528547
Ion Ratio Lower Upper
142 100
141 89.1 72.8 109.2









#37
1-Methylnaphthalene
Concen: 20.482 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:142 Resp: 540931
Ion Ratio Lower Upper
142 100
141 93.4 74.2 111.2
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.411 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:216 Resp: 316696
Ion Ratio Lower Upper
216 100
214 79.1 62.6 93.8
179 20.4 15.4 23.2
108 17.3 12.7 19.1





#40

Hexachlorocyclopentadiene

Concen: 18.175 ng/ul

RT: 12.940 min Scan# 1658

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:237 Resp: 181080

Ion Ratio Lower Upper

237 100

235 61.6 49.8 74.8

272 13.1 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 20.438 ng/ul

RT: 13.192 min Scan# 1701

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:196 Resp: 208102


Ion Ratio Lower Upper

196 100

198 97.4 75.1 112.7

200 30.6 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 20.194 ng/ul
RT: 13.263 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:196 Resp: 222569
Ion Ratio Lower Upper
196 100
198 95.1 77.0 115.4
200 30.8 25.8 38.6



#43
1,1'-Biphenyl
Concen: 20.490 ng/ul
RT: 13.592 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:154 Resp: 757033
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 12.4 9.9 14.9









#44
2-Chloronaphthalene
Concen: 20.292 ng/ul
RT: 13.640 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:162 Resp: 589791
Ion Ratio Lower Upper
162 100
164 32.6 26.1 39.1
127 38.5 31.1 46.7



#45
2-Nitroaniline
Concen: 20.343 ng/ul
RT: 13.839 min Scan# 1811
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion: 65 Resp: 142415
Ion Ratio Lower Upper
65 100
92 72.7 58.6 88.0
138 106.4 81.2 121.8





#47

Dimethylphthalate

Concen: 20.472 ng/ul

RT: 14.210 min Scan# 1874

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:163 Resp: 787435

Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.2

164 10.0 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 20.387 ng/ul

RT: 14.334 min Scan# 1895

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:165 Resp: 137007


Ion Ratio Lower Upper

165 100

89 57.8 47.5 71.3

121 21.5 16.6 24.8









#50
Acenaphthylene
Concen: 20.318 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:152 Resp: 920161
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.6
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 18.830 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:138 Resp: 125112
Ion Ratio Lower Upper
138 100
108 11.1 9.4 14.2
92 113.3 97.4 146.0









#52
Acenaphthene
Concen: 20.107 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:153 Resp: 637351
Ion Ratio Lower Upper
153 100
152 47.2 37.4 56.0
154 87.0 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 17.570 ng/ul
RT: 14.863 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:184 Resp: 84201
Ion Ratio Lower Upper
184 100
63 55.6 49.5 74.3
154 60.1 54.3 81.5









#55
4-Nitrophenol
Concen: 20.228 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:109 Resp: 122630
Ion Ratio Lower Upper
109 100
139 117.3 96.0 144.0
65 115.2 95.8 143.6



#56
Dibenzofuran
Concen: 20.247 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:168 Resp: 920144
Ion Ratio Lower Upper
168 100
139 37.9 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 21.218 ng/ul
RT: 15.116 min Scan# 2028
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:165 Resp: 212583
Ion Ratio Lower Upper
165 100
63 38.7 32.0 48.0
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.602 ng/ul
RT: 15.381 min Scan# 2073
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:232 Resp: 214018
Ion Ratio Lower Upper
232 100
131 39.8 33.0 49.6
130 2.1 1.5 2.3
166 29.6 22.6 33.8









#59
Diethylphthalate
Concen: 20.545 ng/ul
RT: 15.569 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:149 Resp: 793541
Ion Ratio Lower Upper
149 100
177 22.6 17.8 26.8
150 12.3 10.1 15.1



#61
Fluorene
Concen: 20.376 ng/ul
RT: 15.804 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:166 Resp: 755415
Ion Ratio Lower Upper
166 100
165 97.8 79.1 118.7
167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.522 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:204 Resp: 395607
Ion Ratio Lower Upper
204 100
206 32.2 26.7 40.1
141 58.8 44.6 66.8



#63
4-Nitroaniline
Concen: 19.773 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:138 Resp: 124407
Ion Ratio Lower Upper
138 100
92 55.5 43.8 65.8
108 86.5 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.446 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:198 Resp: 137290
Ion Ratio Lower Upper
198 100
182 4.2 4.1 6.1
77 24.9 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.571 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:169 Resp: 656355
Ion Ratio Lower Upper
169 100
168 66.3 53.5 80.3
167 37.4 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 20.237 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:248 Resp: 250862
Ion Ratio Lower Upper
248 100
250 97.0 76.0 114.0
141 71.6 56.0 84.0



#69
Hexachlorobenzene
Concen: 20.284 ng/ul
RT: 16.810 min Scan# 2316
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:284 Resp: 292989
Ion Ratio Lower Upper
284 100
142 33.9 25.0 37.4
249 31.6 24.9 37.3





#70

Atrazine

Concen: 20.306 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:200 Resp: 259085

Ion Ratio Lower Upper

200 100

173 26.1 20.4 30.6

215 48.9 39.3 58.9



#71

Pentachlorophenol

Concen: 19.119 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:266 Resp: 179704


Ion Ratio Lower Upper

266 100

264 64.0 50.5 75.7

268 64.4 51.0 76.6





#72

Phenanthrene

Concen: 20.443 ng/ul

RT: 17.557 min Scan# 2443

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:178 Resp: 1269849

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.2

176 19.1 15.2 22.8



#74

Anthracene

Concen: 20.457 ng/ul

RT: 17.645 min Scan# 2458

Delta R.T. -0.006 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:178 Resp: 1278480


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.5 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.206 ng/uL

RT: 13.563 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:216 Resp: 316623

Ion Ratio Lower Upper

216 100

214 77.4 62.6 94.0

218 46.8 39.0 58.4



#76

Pentachlorobenzene

Concen: 20.374 ng/uL

RT: 15.075 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:250 Resp: 329453


Ion Ratio Lower Upper

250 100

248 62.1 51.1 76.7

252 63.1 51.0 76.6





#77

Carbazole

Concen: 21.163 ng/ul

RT: 17.910 min Scan# 2503

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:167 Resp: 1165041

Ion Ratio Lower Upper

167 100

166 21.8 17.3 25.9

139 12.9 10.4 15.6



#78

Di-n-butylphthalate

Concen: 20.527 ng/ul

RT: 18.439 min Scan# 2593

Delta R.T. -0.006 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:149 Resp: 1370532


Ion Ratio Lower Upper

149 100

150 9.1 7.1 10.7

104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 19.161 ng/ul
RT: 19.504 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:202 Resp: 1609887
Ion Ratio Lower Upper
202 100
101 10.2 8.6 12.8
100 7.5 6.3 9.5



#82
Pyrene
Concen: 19.122 ng/ul
RT: 19.857 min Scan# 2834
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:202 Resp: 1712930
Ion Ratio Lower Upper
202 100
101 12.0 10.3 15.5
100 9.7 8.5 12.7





#83

Butylbenzylphthalate

Concen: 19.173 ng/ul

RT: 20.710 min Scan# 2979

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:149 Resp: 587808

Ion Ratio Lower Upper

149 100

91 70.4 57.7 86.5

206 21.9 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 20.108 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. -0.006 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:252 Resp: 585715


Ion Ratio Lower Upper

252 100

254 66.2 51.8 77.6

126 10.7 7.9 11.9









#85
Benzo(a)anthracene
Concen: 20.110 ng/ul
RT: 21.568 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:228 Resp: 1778391
Ion Ratio Lower Upper
228 100
229 19.9 15.4 23.2
226 26.7 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.724 ng/ul
RT: 21.468 min Scan# 3108
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:149 Resp: 926001
Ion Ratio Lower Upper
149 100
167 27.7 21.5 32.3
279 4.6 3.2 4.8





#87

Chrysene

Concen: 20.527 ng/ul

RT: 21.627 min Scan# 3135

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:228 Resp: 1712663

Ion Ratio Lower Upper

228 100

226 30.1 25.0 37.6

229 19.8 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 16.702 ng/ul




RT: 22.445 min Scan# 3274


Delta R.T. -0.006 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Tgt Ion:149 Resp: 1605129









#90
Benzo(b)fluoranthene
Concen: 19.508 ng/ul
RT: 23.363 min Scan# 3430
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:252 Resp: 1816808
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 8.9 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 19.659 ng/ul
RT: 23.415 min Scan# 3439
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:252 Resp: 1816428
Ion Ratio Lower Upper
252 100
253 21.5 17.1 25.7
125 8.9 7.2 10.8





#93

Benzo(a)pyrene

Concen: 20.138 ng/ul

RT: 24.039 min Scan# 3545

Delta R.T. -0.000 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:252 Resp: 1589595

Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.0

125 9.4 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.017 ng/ul

RT: 26.851 min Scan# 4023

Delta R.T. -0.006 min

Lab File: BP013787.D

Acq: 08 Feb 2023 09:23




Tgt Ion:276 Resp: 1860851


Ion Ratio Lower Upper

276 100

138 22.4 18.4 27.6

277 25.5 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 21.061 ng/ul
RT: 26.868 min Scan# 4026
Delta R.T. -0.006 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Instrument :
BNA_P
ClientSampleId :
SSTD020758

Tgt Ion:278 Resp: 1557642
Ion Ratio Lower Upper
278 100
139 14.6 12.0 18.0
279 24.1 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 20.827 ng/ul
RT: 27.692 min Scan# 4166
Delta R.T. -0.000 min
Lab File: BP013787.D
Acq: 08 Feb 2023 09:23

Tgt Ion:276 Resp: 1468068
Ion Ratio Lower Upper
276 100
138 20.5 15.9 23.9
277 24.0 19.0 28.6

