

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
 Data File : BP013789.D
 Acq On : 08 Feb 2023 10:31
 Operator : CG/JU
 Sample : 01455-05
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC-BLANK

Quant Time: Feb 08 23:54:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	159460	20.000	ng/ul	0.00
20) Naphthalene-d8	10.952	136	687843	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.757	164	512512	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.510	188	1276471	20.000	ng/ul	0.00
79) Chrysene-d12	21.586	240	1374486	20.000	ng/ul	0.00
88) Perylene-d12	24.151	264	1406607	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.376	96	40	0.010	ng/uL	-0.09
4) Pyridine-d5	3.905	84	255	0.022	ng/ul	0.00
7) Phenol-d5	7.264	99	7	0.000	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	41	0.005	ng/ul	0.00
11) 2-Chlorophenol-d4	7.640	132	15	0.001	ng/ul	0.00
15) 4-Methylphenol-d8	8.834	113	7	0.001	ng/ul	0.00
21) Nitrobenzene-d5	9.299	128	29	0.006	ng/ul	0.00
24) 2-Nitrophenol-d4	10.016	143	5	0.001	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.505	165	22	0.002	ng/ul	-0.06
31) 4-Chloroaniline-d4	11.093	131	16	0.001	ng/ul	0.00
46) Dimethylphthalate-d6	14.140	166	12	0.000	ng/ul	-0.02
49) Acenaphthylene-d8	14.451	160	78	0.002	ng/ul	0.00
54) 4-Nitrophenol-d4	14.940	143	19	0.003	ng/ul	0.00
60) Fluorene-d10	15.734	176	61	0.002	ng/ul	-0.02
65) 4,6-Dinitro-2-methylph...	15.863	200	6	0.001	ng/ul	0.00
73) Anthracene-d10	17.510	188	1276471	21.869	ng/ul	-0.11
81) Pyrene-d10	19.839	212	223	0.003	ng/ul	0.01
92) Benzo(a)pyrene-d12	24.151	264	1027152	14.574	ng/ul	0.16

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
Data File : BP013789.D
Acq On : 08 Feb 2023 10:31
Operator : CG/JU
Sample : 01455-05
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Feb 08 23:54:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 08 00:37:14 2023
Response via : Initial Calibration

