

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
 Data File : BP013801.D
 Acq On : 08 Feb 2023 17:31
 Operator : CG/JU
 Sample : PB150698BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS698

Quant Time: Feb 08 23:56:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	202177	20.000	ng/u1	0.00
20) Naphthalene-d8	10.951	136	819585	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	492704	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	985866	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	912426	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1060088	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.469	96	24360	4.994	ng/uL	0.00
4) Pyridine-d5	3.899	84	333195	23.040	ng/u1	0.00
7) Phenol-d5	7.269	99	441704	24.521	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	254882	24.146	ng/u1	0.00
11) 2-Chlorophenol-d4	7.646	132	339285	25.586	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	363532	24.690	ng/u1	0.00
21) Nitrobenzene-d5	9.298	128	165742	29.065	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	186525	28.330	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	359534	26.796	ng/u1	0.00
31) 4-Chloroaniline-d4	11.087	131	419326	22.048	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	964763	24.861	ng/u1	0.00
49) Acenaphthylene-d8	14.457	160	1119259	26.644	ng/u1	0.00
54) 4-Nitrophenol-d4	14.939	143	151072	21.852	ng/u1	0.00
60) Fluorene-d10	15.751	176	820659	24.828	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	140525	23.234	ng/u1	0.00
73) Anthracene-d10	17.610	188	1185947	26.308	ng/u1	0.00
81) Pyrene-d10	19.827	212	1346991	26.033	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	1439252	27.096	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	50002	9.549	ng/uL	97
5) Pyridine	3.922	79	345092	24.147	ng/u1	94
6) Benzaldehyde	7.252	77	259565	31.953	ng/u1	98
8) Phenol	7.293	94	463501	25.066	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.534	93	365939	24.697	ng/u1	98
12) 2-Chlorophenol	7.681	128	354868	26.082	ng/u1	98
13) 2-Methylphenol	8.563	108	360668	25.554	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.657	45	432951	25.688	ng/u1	99
16) Acetophenone	8.951	105	578842	24.515	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.940	70	283337	24.164	ng/u1	98
18) 4-Methylphenol	8.893	108	391538	25.126	ng/u1	96
19) Hexachloroethane	9.216	117	150150	26.949	ng/u1	99
22) Nitrobenzene	9.340	77	433114	28.869	ng/u1	99
23) Isophorone	9.869	82	820691	25.928	ng/u1	100
25) 2-Nitrophenol	10.057	139	205979	28.864	ng/u1	99
26) 2,4-Dimethylphenol	10.104	107	412237	25.945	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.345	93	509915	26.677	ng/u1	99
29) 2,4-Dichlorophenol	10.593	162	357888	27.094	ng/u1	98
30) Naphthalene	10.998	128	1174369	26.917	ng/u1	99
32) 4-Chloroaniline	11.110	127	429714	22.412	ng/u1	98
33) Hexachlorobutadiene	11.281	225	250692	26.942	ng/u1	97
34) Caprolactam	11.887	113	98347	21.218	ng/u1	96
35) 4-Chloro-3-methylphenol	12.216	107	376384	24.681	ng/u1	99
36) 2-Methylnaphthalene	12.598	142	762952	24.872	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020823\
 Data File : BP013801.D
 Acq On : 08 Feb 2023 17:31
 Operator : CG/JU
 Sample : PB150698BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS698


Quant Time: Feb 08 23:56:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	773135	24.394	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.957	216	456042	29.372	ng/u1	99
40) Hexachlorocyclopentadiene	12.939	237	258698	25.949	ng/u1	99
41) 2,4,6-Trichlorophenol	13.192	196	291670	28.627	ng/u1	99
42) 2,4,5-Trichlorophenol	13.263	196	312265	28.314	ng/u1	98
43) 1,1'-Biphenyl	13.592	154	1040415	28.142	ng/u1	99
44) 2-Chloronaphthalene	13.639	162	825981	28.400	ng/u1	100
45) 2-Nitroaniline	13.839	65	194987	27.835	ng/u1	95
47) Dimethylphthalate	14.210	163	983445	25.551	ng/u1	100
48) 2,6-Dinitrotoluene	14.334	165	180354	26.820	ng/u1	96
50) Acenaphthylene	14.481	152	1224814	27.028	ng/u1	99
51) 3-Nitroaniline	14.657	138	167735	25.229	ng/u1	97
52) Acenaphthene	14.822	153	851818	26.856	ng/u1	99
53) 2,4-Dinitrophenol	14.863	184	74455	15.526	ng/u1	94
55) 4-Nitrophenol	14.957	109	132843	21.899	ng/u1	96
56) Dibenzofuran	15.157	168	1186163	26.084	ng/u1	99
57) 2,4-Dinitrotoluene	15.116	165	255327	25.468	ng/u1	98
58) 2,3,4,6-Tetrachlorophenol	15.381	232	262385	25.242	ng/u1	98
59) Diethylphthalate	15.569	149	941696	24.365	ng/u1	99
61) Fluorene	15.804	166	933836	25.172	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.792	204	496822	25.756	ng/u1	99
63) 4-Nitroaniline	15.822	138	170758	27.122	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.875	198	140535	23.701	ng/u1	98
67) N-Nitrosodiphenylamine	16.010	169	790777	29.510	ng/u1	99
68) 4-Bromophenyl-phenylether	16.692	248	305665	29.359	ng/u1	99
69) Hexachlorobenzene	16.810	284	341206	28.126	ng/u1	97
70) Atrazine	16.957	200	259770	24.241	ng/u1	98
71) Pentachlorophenol	17.157	266	198851	25.189	ng/u1	99
72) Phenanthrene	17.557	178	1425392	27.323	ng/u1	100
74) Anthracene	17.645	178	1423040	27.112	ng/u1	100
75) 1,2,3,4-Tetrachloroben...	13.563	216	448396	34.071	ng/uL	99
76) Pentachlorobenzene	15.075	250	424382	31.248	ng/uL	99
77) Carbazole	17.910	167	1239293	26.804	ng/u1	99
78) Di-n-butylphthalate	18.439	149	1432633	25.548	ng/u1	100
80) Fluoranthene	19.504	202	1626053	27.451	ng/u1	100
82) Pyrene	19.857	202	1677080	26.555	ng/u1	99
83) Butylbenzylphthalate	20.710	149	566888	26.227	ng/u1	99
84) 3,3'-Dichlorobenzidine	21.492	252	537084	26.154	ng/u1	99
85) Benzo(a)anthracene	21.574	228	1692238	27.143	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	873421	26.389	ng/u1	99
87) Chrysene	21.627	228	1619362	27.530	ng/u1	97
89) Di-n-octyl phthalate	22.445	149	1518260	20.814	ng/u1	100
90) Benzo(b)fluoranthene	23.362	252	1810036	25.606	ng/u1	99
91) Benzo(k)fluoranthene	23.415	252	1844261	26.298	ng/u1	100
93) Benzo(a)pyrene	24.039	252	1654849	27.622	ng/u1	100
94) Indeno(1,2,3-cd)pyrene	26.856	276	2343709	34.876	ng/u1	100
95) Dibenzo(a,h)anthracene	26.874	278	1962894	34.967	ng/u1	99
96) Benzo(g,h,i)perylene	27.692	276	1947137	36.393	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 08 23:56:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 08 00:37:14 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.549 ng/uL
RT: 3.505 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion: 88	Resp: 50002
Ion Ratio	Lower Upper
88 100	
43 26.1	20.1 30.1
58 68.8	53.0 79.6



#5
Pyridine
Concen: 24.147 ng/uL
RT: 3.922 min Scan# 125
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 79	Resp: 345092
Ion Ratio	Lower Upper
79 100	
52 62.6	55.1 82.7
51 31.1	25.7 38.5









#6
Benzaldehyde
Concen: 31.953 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion: 77 Resp: 259565
Ion Ratio Lower Upper
77 100
105 97.1 79.5 119.3
106 93.1 76.5 114.7



#8
Phenol
Concen: 25.066 ng/ul
RT: 7.293 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 94 Resp: 463501
Ion Ratio Lower Upper
94 100
65 28.3 22.2 33.4
66 38.8 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 24.697 ng/ul
RT: 7.534 min Scan# 739
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion: 93 Resp: 365939
Ion Ratio Lower Upper
93 100
63 71.7 59.4 89.0
95 32.9 25.8 38.8



#12
2-Chlorophenol
Concen: 26.082 ng/ul
RT: 7.681 min Scan# 764
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:128 Resp: 354868
Ion Ratio Lower Upper
128 100
64 45.5 37.3 55.9
130 31.6 26.3 39.5









#13
2-Methylphenol
Concen: 25.554 ng/ul
RT: 8.563 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:108 Resp: 360668
Ion Ratio Lower Upper
108 100
107 88.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.688 ng/ul
RT: 8.657 min Scan# 930
Delta R.T. 0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 45 Resp: 432951
Ion Ratio Lower Upper
45 100
77 17.2 13.4 20.2
79 14.3 11.0 16.4









#16
Acetophenone
Concen: 24.515 ng/ul
RT: 8.951 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:105 Resp: 578842
Ion Ratio Lower Upper
105 100
77 76.5 61.0 91.4
51 24.7 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 24.164 ng/ul
RT: 8.940 min Scan# 978
Delta R.T. 0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 70 Resp: 283337
Ion Ratio Lower Upper
70 100
42 48.9 39.8 59.8
101 10.3 7.4 11.2
130 20.3 14.6 22.0









#18
4-Methylphenol
Concen: 25.126 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:108 Resp: 391538
Ion Ratio Lower Upper
108 100
107 109.3 90.9 136.3



#19
Hexachloroethane
Concen: 26.949 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:117 Resp: 150150
Ion Ratio Lower Upper
117 100
201 108.8 86.6 129.8
199 68.1 55.2 82.8









#22
Nitrobenzene
Concen: 28.869 ng/ul
RT: 9.340 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion: 77 Resp: 433114
Ion Ratio Lower Upper
77 100
123 44.9 35.1 52.7
65 13.1 10.6 16.0



#23
Isophorone
Concen: 25.928 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 82 Resp: 820691
Ion Ratio Lower Upper
82 100
95 7.7 6.0 9.0
138 17.2 13.7 20.5









#25
2-Nitrophenol
Concen: 28.864 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:139 Resp: 205979
Ion Ratio Lower Upper
139 100
65 51.1 39.9 59.9
109 33.8 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 25.945 ng/ul
RT: 10.104 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:107 Resp: 412237
Ion Ratio Lower Upper
107 100
121 48.6 37.8 56.6
122 82.6 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 26.677 ng/ul
 RT: 10.345 min Scan# 1217
 Delta R.T. -0.000 min
 Lab File: BP013801.D
 Acq: 08 Feb 2023 17:31




Instrument :
 BNA_P
 ClientSampleId :
 SLCS698


Tgt Ion: 93 Resp: 509915
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.0 37.6
 123 11.0 8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 27.094 ng/ul
 RT: 10.593 min Scan# 1259
 Delta R.T. -0.000 min
 Lab File: BP013801.D
 Acq: 08 Feb 2023 17:31

Tgt Ion:162 Resp: 357888
 Ion Ratio Lower Upper
 162 100
 164 65.9 51.4 77.2
 98 35.9 27.5 41.3









#30
Naphthalene
Concen: 26.917 ng/ul
RT: 10.998 min Scan# 1328
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:128 Resp: 1174369
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.1 10.6 15.8



#32
4-Chloroaniline
Concen: 22.412 ng/ul
RT: 11.110 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:127 Resp: 429714
Ion Ratio Lower Upper
127 100
129 31.8 26.2 39.4





#33

Hexachlorobutadiene

Concen: 26.942 ng/ul

RT: 11.281 min Scan# 1377

Delta R.T. -0.006 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:225 Resp: 250692

Ion Ratio Lower Upper

225 100

223 62.3 50.6 76.0

227 64.5 49.0 73.4



#34

Caprolactam

Concen: 21.218 ng/ul

RT: 11.887 min Scan# 1479

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:113 Resp: 98347


Ion Ratio Lower Upper

113 100

55 146.3 120.9 181.3

56 113.1 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 24.681 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:107 Resp: 376384
Ion Ratio Lower Upper
107 100
144 25.2 20.0 30.0
142 76.1 62.1 93.1



#36
2-Methylnaphthalene
Concen: 24.872 ng/ul
RT: 12.598 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:142 Resp: 762952
Ion Ratio Lower Upper
142 100
141 89.0 72.8 109.2









#37
1-Methylnaphthalene
Concen: 24.394 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:142 Resp: 773135
Ion Ratio Lower Upper
142 100
141 91.8 74.2 111.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.372 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:216 Resp: 456042
Ion Ratio Lower Upper
216 100
214 78.6 62.6 93.8
179 19.7 15.4 23.2
108 16.6 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 25.949 ng/ul
RT: 12.939 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:237 Resp: 258698
Ion Ratio Lower Upper
237 100
235 62.9 49.8 74.8
272 12.9 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 28.627 ng/ul
RT: 13.192 min Scan# 1701
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:196 Resp: 291670
Ion Ratio Lower Upper
196 100
198 94.3 75.1 112.7
200 30.9 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 28.314 ng/ul
RT: 13.263 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:196 Resp: 312265
Ion Ratio Lower Upper
196 100
198 98.1 77.0 115.4
200 31.1 25.8 38.6



#43
1,1'-Biphenyl
Concen: 28.142 ng/ul
RT: 13.592 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:154 Resp: 1040415
Ion Ratio Lower Upper
154 100
153 40.3 31.9 47.9
76 12.7 9.9 14.9









#44
2-Chloronaphthalene
Concen: 28.400 ng/ul
RT: 13.639 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:162 Resp: 825981
Ion Ratio Lower Upper
162 100
164 33.1 26.1 39.1
127 38.8 31.1 46.7



#45
2-Nitroaniline
Concen: 27.835 ng/ul
RT: 13.839 min Scan# 1811
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion: 65 Resp: 194987
Ion Ratio Lower Upper
65 100
92 72.8 58.6 88.0
138 109.2 81.2 121.8









#47
Dimethylphthalate
Concen: 25.551 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:163 Resp: 983445
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 26.820 ng/ul
RT: 14.334 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:165 Resp: 180354
Ion Ratio Lower Upper
165 100
89 55.6 47.5 71.3
121 19.8 16.6 24.8









#50
Acenaphthylene
Concen: 27.028 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:152 Resp: 1224814
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 13.4 10.4 15.6



#51
3-Nitroaniline
Concen: 25.229 ng/ul
RT: 14.657 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:138 Resp: 167735
Ion Ratio Lower Upper
138 100
108 11.2 9.4 14.2
92 124.8 97.4 146.0









#52
Acenaphthene
Concen: 26.856 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:153 Resp: 851818
Ion Ratio Lower Upper
153 100
152 48.1 37.4 56.0
154 88.6 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 15.526 ng/ul
RT: 14.863 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:184 Resp: 74455
Ion Ratio Lower Upper
184 100
63 55.8 49.5 74.3
154 63.8 54.3 81.5









#55
4-Nitrophenol
Concen: 21.899 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:109 Resp: 132843
Ion Ratio Lower Upper
109 100
139 124.9 96.0 144.0
65 116.7 95.8 143.6



#56
Dibenzofuran
Concen: 26.084 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:168 Resp: 1186163
Ion Ratio Lower Upper
168 100
139 36.7 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 25.468 ng/ul
RT: 15.116 min Scan# 2028
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:165 Resp: 255327
Ion Ratio Lower Upper
165 100
63 38.7 32.0 48.0
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 25.242 ng/ul
RT: 15.381 min Scan# 2073
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:232 Resp: 262385
Ion Ratio Lower Upper
232 100
131 40.0 33.0 49.6
130 1.9 1.5 2.3
166 28.7 22.6 33.8





#59

Diethylphthalate

Concen: 24.365 ng/ul

RT: 15.569 min Scan# 2105

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:149 Resp: 941696

Ion Ratio Lower Upper

149 100

177 22.6 17.8 26.8

150 12.5 10.1 15.1



#61

Fluorene

Concen: 25.172 ng/ul

RT: 15.804 min Scan# 2145

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:166 Resp: 933836


Ion Ratio Lower Upper

166 100

165 99.0 79.1 118.7

167 13.8 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 25.756 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:204 Resp: 496822
Ion Ratio Lower Upper
204 100
206 33.3 26.7 40.1
141 56.6 44.6 66.8



#63
4-Nitroaniline
Concen: 27.122 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:138 Resp: 170758
Ion Ratio Lower Upper
138 100
92 53.9 43.8 65.8
108 87.9 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 23.701 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:198 Resp: 140535
Ion Ratio Lower Upper
198 100
182 4.4 4.1 6.1
77 24.3 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 29.510 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:169 Resp: 790777
Ion Ratio Lower Upper
169 100
168 66.5 53.5 80.3
167 36.7 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 29.359 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:248 Resp: 305665
Ion Ratio Lower Upper
248 100
250 96.8 76.0 114.0
141 70.3 56.0 84.0



#69
Hexachlorobenzene
Concen: 28.126 ng/ul
RT: 16.810 min Scan# 2316
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:284 Resp: 341206
Ion Ratio Lower Upper
284 100
142 33.5 25.0 37.4
249 32.6 24.9 37.3





#70

Atrazine

Concen: 24.241 ng/ul

RT: 16.957 min Scan# 2341

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:200 Resp: 259770

Ion Ratio Lower Upper

200 100

173 24.7 20.4 30.6

215 50.4 39.3 58.9



#71

Pentachlorophenol

Concen: 25.189 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:266 Resp: 198851


Ion Ratio Lower Upper

266 100

264 63.9 50.5 75.7

268 64.5 51.0 76.6





#72

Phenanthrene

Concen: 27.323 ng/ul

RT: 17.557 min Scan# 2443

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:178 Resp: 1425392

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.3 15.2 22.8



#74

Anthracene

Concen: 27.112 ng/ul

RT: 17.645 min Scan# 2458

Delta R.T. -0.006 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:178 Resp: 1423040


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 18.6 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 34.071 ng/uL

RT: 13.563 min Scan# 1764

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:216 Resp: 448396

Ion Ratio Lower Upper

216 100

214 79.4 62.6 94.0

218 49.3 39.0 58.4



#76

Pentachlorobenzene

Concen: 31.248 ng/uL

RT: 15.075 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:250 Resp: 424382


Ion Ratio Lower Upper

250 100

248 62.6 51.1 76.7

252 63.5 51.0 76.6





#77

Carbazole

Concen: 26.804 ng/ul

RT: 17.910 min Scan# 2503

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:167 Resp: 1239293

Ion Ratio Lower Upper

167 100

166 21.1 17.3 25.9

139 12.4 10.4 15.6



#78

Di-n-butylphthalate

Concen: 25.548 ng/ul

RT: 18.439 min Scan# 2593

Delta R.T. -0.006 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:149 Resp: 1432633


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 5.1 4.1 6.1





#80

Fluoranthene

Concen: 27.451 ng/ul

RT: 19.504 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:202 Resp: 1626053

Ion Ratio Lower Upper

202 100

101 10.7 8.6 12.8

100 8.2 6.3 9.5



#82

Pyrene

Concen: 26.555 ng/ul

RT: 19.857 min Scan# 2834

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:202 Resp: 1677080


Ion Ratio Lower Upper

202 100

101 12.4 10.3 15.5

100 10.5 8.5 12.7









#83
Butylbenzylphthalate
Concen: 26.227 ng/ul
RT: 20.710 min Scan# 2979
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:149 Resp: 566888
Ion Ratio Lower Upper
149 100
91 72.2 57.7 86.5
206 22.1 16.7 25.1



#84
3,3'-Dichlorobenzidine
Concen: 26.154 ng/ul
RT: 21.492 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:252 Resp: 537084
Ion Ratio Lower Upper
252 100
254 63.7 51.8 77.6
126 10.8 7.9 11.9





#85

Benzo(a)anthracene

Concen: 27.143 ng/ul

RT: 21.574 min Scan# 3125

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:228 Resp: 1692238

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.2

226 27.0 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 26.389 ng/ul

RT: 21.468 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:149 Resp: 873421


Ion Ratio Lower Upper

149 100

167 27.4 21.5 32.3

279 4.6 3.2 4.8





#87

Chrysene

Concen: 27.530 ng/ul

RT: 21.627 min Scan# 3135

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:228 Resp: 1619362

Ion Ratio Lower Upper

228 100

226 29.8 25.0 37.6

229 19.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 20.814 ng/ul




RT: 22.445 min Scan# 3274


Delta R.T. -0.006 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Tgt Ion:149 Resp: 1518260









#90
Benzo(b)fluoranthene
Concen: 25.606 ng/uI
RT: 23.362 min Scan# 3430
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31




Instrument :
BNA_P
ClientSampleId :
SLCS698


Tgt Ion:252 Resp: 1810036
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 9.0 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 26.298 ng/uI
RT: 23.415 min Scan# 3439
Delta R.T. -0.000 min
Lab File: BP013801.D
Acq: 08 Feb 2023 17:31

Tgt Ion:252 Resp: 1844261
Ion Ratio Lower Upper
252 100
253 21.3 17.1 25.7
125 8.8 7.2 10.8





#93

Benzo(a)pyrene

Concen: 27.622 ng/ul

RT: 24.039 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:252 Resp: 1654849

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 10.1 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 34.876 ng/ul

RT: 26.856 min Scan# 4024

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31




Tgt Ion:276 Resp: 2343709


Ion Ratio Lower Upper

276 100

138 22.8 18.4 27.6

277 24.5 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 34.967 ng/ul

RT: 26.874 min Scan# 4026

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Instrument :

BNA_P

ClientSampleId :

SLCS698





Tgt Ion:278 Resp: 1962894

Ion Ratio Lower Upper

278 100

139 14.9 12.0 18.0

279 23.2 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 36.393 ng/ul

RT: 27.692 min Scan# 4166

Delta R.T. -0.000 min

Lab File: BP013801.D

Acq: 08 Feb 2023 17:31

Tgt Ion:276 Resp: 1947137

Ion Ratio Lower Upper

276 100

138 20.4 15.9 23.9

277 24.0 19.0 28.6

