

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP021723\  
 Data File : BP013848.D  
 Acq On : 17 Feb 2023 16:49  
 Operator : CG/JU  
 Sample : 01581-01  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 MH-A

Quant Time: Feb 17 22:04:59 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 17 18:05:40 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	8.111	152	133697	20.000 ng	0.00
21) Naphthalene-d8	10.934	136	483508	20.000 ng	0.00
39) Acenaphthene-d10	14.740	164	302601	20.000 ng	0.00
64) Phenanthrene-d10	17.498	188	694880	20.000 ng	0.00
76) Chrysene-d12	21.575	240	814115	20.000 ng	0.00
86) Perylene-d12	24.133	264	952493	20.000 ng	0.00
System Monitoring Compounds					
5) 2-Fluorophenol	5.646	112	824679	101.660 ng	0.00
7) Phenol-d6	7.258	99	1076364	103.602 ng	0.00
23) Nitrobenzene-d5	9.281	82	600386	67.053 ng	0.00
42) 2,4,6-Tribromophenol	16.234	330	439981	99.885 ng	0.00
45) 2-Fluorobiphenyl	13.369	172	1420340	61.651 ng	0.00
79) Terphenyl-d14	20.028	244	2597418	61.385 ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP021723\  
 Data File : BP013848.D  
 Acq On : 17 Feb 2023 16:49  
 Operator : CG/JU  
 Sample : 01581-01  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 MH-A

Quant Time: Feb 17 22:04:59 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 17 18:05:40 2023  
 Response via : Initial Calibration

