

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030223\
 Data File : BP013930.D
 Acq On : 02 Mar 2023 19:40
 Operator : CG/JU
 Sample : 01749-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CF-3

Quant Time: Mar 03 02:39:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.093	152	103488	20.000	ng	0.00
21) Naphthalene-d8	10.916	136	413857	20.000	ng	0.00
39) Acenaphthene-d10	14.728	164	254515	20.000	ng	0.00
64) Phenanthrene-d10	17.480	188	551060	20.000	ng	# 0.00
76) Chrysene-d12	21.563	240	591469	20.000	ng	# 0.00
86) Perylene-d12	24.109	264	715451	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.634	112	599798	95.521	ng	0.00
7) Phenol-d6	7.246	99	821532	102.156	ng	0.00
23) Nitrobenzene-d5	9.263	82	509195	66.440	ng	0.00
42) 2,4,6-Tribromophenol	16.222	330	364898	98.491	ng	0.00
45) 2-Fluorobiphenyl	13.357	172	1170323	60.397	ng	0.00
79) Terphenyl-d14	20.016	244	1994279	64.872	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030223\
Data File : BP013930.D
Acq On : 02 Mar 2023 19:40
Operator : CG/JU
Sample : 01749-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CF-3

Quant Time: Mar 03 02:39:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 02 04:26:35 2023
Response via : Initial Calibration

