

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030223\
 Data File : BP013936.D
 Acq On : 02 Mar 2023 23:04
 Operator : CG/JU
 Sample : 01749-14
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CF-16

Quant Time: Mar 03 02:42:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.093	152	99391	20.000	ng	0.00
21) Naphthalene-d8	10.916	136	390052	20.000	ng	0.00
39) Acenaphthene-d10	14.728	164	253954	20.000	ng	0.00
64) Phenanthrene-d10	17.481	188	590673	20.000	ng	0.00
76) Chrysene-d12	21.563	240	588719	20.000	ng	# 0.00
86) Perylene-d12	24.110	264	687263	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.634	112	523083	86.738	ng	0.00
7) Phenol-d6	7.246	99	685127	88.706	ng	0.00
23) Nitrobenzene-d5	9.263	82	441904	61.179	ng	0.00
42) 2,4,6-Tribromophenol	16.222	330	326247	88.253	ng	0.00
45) 2-Fluorobiphenyl	13.351	172	911888	47.163	ng	0.00
79) Terphenyl-d14	20.016	244	1628271	53.214	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030223\
Data File : BP013936.D
Acq On : 02 Mar 2023 23:04
Operator : CG/JU
Sample : 01749-14
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CF-16

Quant Time: Mar 03 02:42:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 02 04:26:35 2023
Response via : Initial Calibration

