

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
 Data File : BP013940.D
 Acq On : 03 Mar 2023 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	50	0.00
2	1,4-Dioxane	40.000	40.289	-0.7	54	0.00
3	Pyridine	40.000	42.578	-6.4	55	0.00
4	n-Nitrosodimethylamine	40.000	48.850	-22.1	63	0.00
5 S	2-Fluorophenol	80.000	84.639	-5.8	53	0.00
6	Aniline	40.000	41.340	-3.4	50	0.00
7 S	Phenol-d6	80.000	83.367	-4.2	50	-0.01
8	2-Chlorophenol	40.000	40.551	-1.4	51	0.00
9	Benzaldehyde	40.000	43.922	-9.8	57	0.00
10 C	Phenol	40.000	41.821	-4.6	50	0.00
11	bis(2-Chloroethyl)ether	40.000	40.760	-1.9	53	0.00
12	1,3-Dichlorobenzene	40.000	40.897	-2.2	50	0.00
13 C	1,4-Dichlorobenzene	40.000	41.272	-3.2	52	0.00
14	1,2-Dichlorobenzene	40.000	40.368	-0.9	49	0.00
15	Benzyl Alcohol	40.000	48.240	-20.6	56	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.372	-3.4	49	0.00
17	2-Methylphenol	40.000	43.242	-8.1	50	0.00
18	Hexachloroethane	40.000	41.909	-4.8	51	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	49.319	-23.3	55	-0.01
20	3+4-Methylphenols	40.000	44.034	-10.1	51	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	50	0.00
22	Acetophenone	40.000	44.906	-12.3	52	0.00
23 S	Nitrobenzene-d5	80.000	94.146	-17.7	55	0.00
24	Nitrobenzene	40.000	45.794	-14.5	53	0.00
25	Isophorone	40.000	47.110	-17.8	55	0.00
26 C	2-Nitrophenol	40.000	44.488	-11.2	56	0.00
27	2,4-Dimethylphenol	40.000	43.922	-9.8	51	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.173	-2.9	49	0.00
29 C	2,4-Dichlorophenol	40.000	44.972	-12.4	52	0.00
30	1,2,4-Trichlorobenzene	40.000	42.227	-5.6	52	0.00
31	Naphthalene	40.000	41.301	-3.3	51	0.00
32	Benzoic acid	40.000	39.797	0.5	50	-0.01
33	4-Chloroaniline	40.000	43.558	-8.9	52	0.00
34 C	Hexachlorobutadiene	40.000	43.998	-10.0	55	0.00
35	Caprolactam	40.000	44.995	-12.5	54	-0.01
36 C	4-Chloro-3-methylphenol	40.000	45.790	-14.5	55	0.00
37	2-Methylnaphthalene	40.000	40.745	-1.9	50	0.00
38	1-Methylnaphthalene	40.000	41.024	-2.6	51	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	52	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.281	-3.2	53	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.342	-10.9	55	0.00
42 S	2,4,6-Tribromophenol	80.000	95.376	-19.2	59	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.704	-11.8	55	0.00
44	2,4,5-Trichlorophenol	40.000	44.003	-10.0	55	0.00
45 S	2-Fluorobiphenyl	80.000	81.796	-2.2	53	0.00
46	1,1'-Biphenyl	40.000	40.683	-1.7	53	0.00
47	2-Chloronaphthalene	40.000	40.809	-2.0	53	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
 Data File : BP013940.D
 Acq On : 03 Mar 2023 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.468	-16.2	62	0.00
49	Acenaphthylene	40.000	42.003	-5.0	53	0.00
50	Dimethylphthalate	40.000	43.956	-9.9	56	0.00
51	2,6-Dinitrotoluene	40.000	42.551	-6.4	55	0.00
52 C	Acenaphthene	40.000	40.968	-2.4	52	0.00
53	3-Nitroaniline	40.000	41.244	-3.1	54	0.00
54 P	2,4-Dinitrophenol	40.000	45.414	-13.5	61	0.00
55	Dibenzofuran	40.000	41.555	-3.9	53	0.00
56 P	4-Nitrophenol	40.000	42.419	-6.0	53	0.00
57	2,4-Dinitrotoluene	40.000	43.943	-9.9	57	0.00
58	Fluorene	40.000	42.381	-6.0	54	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.703	-11.8	56	0.00
60	Diethylphthalate	40.000	45.286	-13.2	57	0.00
61	4-Chlorophenyl-phenylether	40.000	43.686	-9.2	56	0.00
62	4-Nitroaniline	40.000	42.367	-5.9	56	0.00
63	Azobenzene	40.000	45.312	-13.3	56	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	53	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.669	-9.2	60	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.846	-4.6	55	0.00
67	4-Bromophenyl-phenylether	40.000	43.924	-9.8	57	0.00
68	Hexachlorobenzene	40.000	43.918	-9.8	58	0.00
69	Atrazine	40.000	44.521	-11.3	57	0.00
70 C	Pentachlorophenol	40.000	45.818	-14.5	57	0.00
71	Phenanthrene	40.000	41.714	-4.3	55	0.00
72	Anthracene	40.000	42.587	-6.5	55	0.00
73	Carbazole	40.000	42.186	-5.5	54	0.00
74	Di-n-butylphthalate	40.000	46.874	-17.2	59	0.00
75 C	Fluoranthene	40.000	43.862	-9.7	58	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	56	0.00
77	Benzydine	40.000	48.665	-21.7	71	0.00
78	Pyrene	40.000	42.144	-5.4	56	0.00
79 S	Terphenyl-d14	80.000	88.585	-10.7	60	0.00
80	Butylbenzylphthalate	40.000	49.038	-22.6	63	0.00
81	Benzo(a)anthracene	40.000	42.386	-6.0	57	0.00
82	3,3'-Dichlorobenzidine	40.000	45.386	-13.5	59	0.00
83	Chrysene	40.000	41.864	-4.7	57	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	48.213	-20.5	61	0.00
85 c	Di-n-octyl phthalate	40.000	46.693	-16.7	62	0.00
86 I	Perylene-d12	20.000	20.000	0.0	57	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.569	-3.9	58	0.00
88	Benzo(b)fluoranthene	40.000	41.942	-4.9	59	0.00
89	Benzo(k)fluoranthene	40.000	41.437	-3.6	56	0.00
90 C	Benzo(a)pyrene	40.000	42.417	-6.0	58	0.00
91	Dibenzo(a,h)anthracene	40.000	41.098	-2.7	57	0.00
92	Benzo(g,h,i)perylene	40.000	41.251	-3.1	57	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
Data File : BP013940.D
Acq On : 03 Mar 2023 12:27
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 02 04:26:35 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0