

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
 Data File : BP013940.D
 Acq On : 03 Mar 2023 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	50	0.00
2	1,4-Dioxane	0.527	0.531	-0.8	54	0.00
3	Pyridine	1.422	1.514	-6.5	55	0.00
4	n-Nitrosodimethylamine	0.533	0.651	-22.1	63	0.00
5 S	2-Fluorophenol	1.214	1.284	-5.8	53	0.00
6	Aniline	2.019	2.087	-3.4	50#	0.00
7 S	Phenol-d6	1.554	1.620	-4.2	50#	-0.01
8	2-Chlorophenol	1.280	1.298	-1.4	51	0.00
9	Benzaldehyde	0.739	0.812	-9.9	57	0.00
10 C	Phenol	1.618	1.691	-4.5	50	0.00
11	bis(2-Chloroethyl)ether	1.295	1.319	-1.9	53	0.00
12	1,3-Dichlorobenzene	1.439	1.472	-2.3	50#	0.00
13 C	1,4-Dichlorobenzene	1.454	1.500	-3.2	52	0.00
14	1,2-Dichlorobenzene	1.357	1.370	-1.0	49#	0.00
15	Benzyl Alcohol	1.090	1.315	-20.6	56	0.00
16	2,2'-oxybis(1-Chloropropane	1.383	1.430	-3.4	49#	0.00
17	2-Methylphenol	1.101	1.190	-8.1	50#	0.00
18	Hexachloroethane	0.509	0.534	-4.9	51	0.00
19 P	n-Nitroso-di-n-propylamine	0.910	1.122	-23.3	55	-0.01
20	3+4-Methylphenols	1.482	1.631	-10.1	51	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	50	0.00
22	Acetophenone	0.504	0.566	-12.3	52	0.00
23 S	Nitrobenzene-d5	0.370	0.436	-17.8	55	0.00
24	Nitrobenzene	0.387	0.442	-14.2	53	0.00
25	Isophorone	0.714	0.841	-17.8	55	0.00
26 C	2-Nitrophenol	0.162	0.202	-24.7#	56	0.00
27	2,4-Dimethylphenol	0.298	0.328	-10.1	51	0.00
28	bis(2-Chloroethoxy)methane	0.456	0.469	-2.9	49#	0.00
29 C	2,4-Dichlorophenol	0.320	0.360	-12.5	52	0.00
30	1,2,4-Trichlorobenzene	0.376	0.397	-5.6	52	0.00
31	Naphthalene	1.097	1.132	-3.2	51	0.00
32	Benzoic acid	0.214	0.228	-6.5	50	-0.01
33	4-Chloroaniline	0.460	0.501	-8.9	52	0.00
34 C	Hexachlorobutadiene	0.250	0.275	-10.0	55	0.00
35	Caprolactam	0.102	0.115	-12.7	54	-0.01
36 C	4-Chloro-3-methylphenol	0.357	0.409	-14.6	55	0.00
37	2-Methylnaphthalene	0.794	0.809	-1.9	50	0.00
38	1-Methylnaphthalene	0.743	0.762	-2.6	51	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	52	0.00
40	1,2,4,5-Tetrachlorobenzene	0.721	0.744	-3.2	53	0.00
41 P	Hexachlorocyclopentadiene	0.383	0.425	-11.0	55	0.00
42 S	2,4,6-Tribromophenol	0.291	0.347	-19.2	59	0.00
43 C	2,4,6-Trichlorophenol	0.451	0.504	-11.8	55	0.00
44	2,4,5-Trichlorophenol	0.536	0.590	-10.1	55	0.00
45 S	2-Fluorobiphenyl	1.523	1.557	-2.2	53	0.00
46	1,1'-Biphenyl	1.617	1.644	-1.7	53	0.00
47	2-Chloronaphthalene	1.232	1.256	-1.9	53	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
 Data File : BP013940.D
 Acq On : 03 Mar 2023 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 02 04:26:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.307	0.405	-31.9#	62	0.00
49	Acenaphthylene	1.941	2.039	-5.0	53	0.00
50	Dimethylphthalate	1.627	1.788	-9.9	56	0.00
51	2,6-Dinitrotoluene	0.324	0.376	-16.0	55	0.00
52 C	Acenaphthene	1.181	1.210	-2.5	52	0.00
53	3-Nitroaniline	0.335	0.379	-13.1	54	0.00
54 P	2,4-Dinitrophenol	0.202	0.251	-24.3	61	0.00
55	Dibenzofuran	1.966	2.042	-3.9	53	0.00
56 P	4-Nitrophenol	0.291	0.309	-6.2	53	0.00
57	2,4-Dinitrotoluene	0.428	0.524	-22.4	57	0.00
58	Fluorene	1.541	1.633	-6.0	54	0.00
59	2,3,4,6-Tetrachlorophenol	0.462	0.516	-11.7	56	0.00
60	Diethylphthalate	1.587	1.797	-13.2	57	0.00
61	4-Chlorophenyl-phenylether	0.835	0.912	-9.2	56	0.00
62	4-Nitroaniline	0.339	0.400	-18.0	56	0.00
63	Azobenzene	1.436	1.626	-13.2	56	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	53	0.00
65	4,6-Dinitro-2-methylphenol	0.130	0.152	-16.9	60	0.00
66 c	n-Nitrosodiphenylamine	0.606	0.634	-4.6	55	0.00
67	4-Bromophenyl-phenylether	0.236	0.259	-9.7	57	0.00
68	Hexachlorobenzene	0.258	0.283	-9.7	58	0.00
69	Atrazine	0.207	0.230	-11.1	57	0.00
70 C	Pentachlorophenol	0.190	0.218	-14.7	57	0.00
71	Phenanthrene	1.120	1.168	-4.3	55	0.00
72	Anthracene	1.120	1.193	-6.5	55	0.00
73	Carbazole	1.009	1.064	-5.5	54	0.00
74	Di-n-butylphthalate	1.182	1.385	-17.2	59	0.00
75 C	Fluoranthene	1.397	1.532	-9.7	58	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
77	Benzydine	0.386	0.498	-29.0#	71	0.00
78	Pyrene	1.367	1.440	-5.3	56	0.00
79 S	Terphenyl-d14	1.040	1.151	-10.7	60	0.00
80	Butylbenzylphthalate	0.486	0.596	-22.6	63	0.00
81	Benzo(a)anthracene	1.432	1.518	-6.0	57	0.00
82	3,3'-Dichlorobenzidine	0.451	0.511	-13.3	59	0.00
83	Chrysene	1.387	1.452	-4.7	57	0.00
84	Bis(2-ethylhexyl)phthalate	0.757	0.913	-20.6	61	0.00
85 c	Di-n-octyl phthalate	1.350	1.575	-16.7	62	0.00
86 I	Perylene-d12	1.000	1.000	0.0	57	0.00
87	Indeno(1,2,3-cd)pyrene	1.556	1.617	-3.9	58	0.00
88	Benzo(b)fluoranthene	1.265	1.326	-4.8	59	0.00
89	Benzo(k)fluoranthene	1.284	1.330	-3.6	56	0.00
90 C	Benzo(a)pyrene	1.223	1.297	-6.1	58	0.00
91	Dibenzo(a,h)anthracene	1.304	1.339	-2.7	57	0.00
92	Benzo(g,h,i)perylene	1.289	1.329	-3.1	57	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030323\
Data File : BP013940.D
Acq On : 03 Mar 2023 12:27
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 04 00:54:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 02 04:26:35 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1