

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP030323\  
 Data File : BP013940.D  
 Acq On : 03 Mar 2023 12:27  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :

Quant Time: Mar 04 00:54:26 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 02 04:26:35 2023  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Christian  
 Giraldo

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.093  | 152  | 86840    | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.916 | 136  | 323576   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.728 | 164  | 210254   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.487 | 188  | 485671   | 20.000 | ng    | # 0.00   |
| 76) Chrysene-d12              | 21.563 | 240  | 530020   | 20.000 | ng    | # 0.00   |
| 86) Perylene-d12              | 24.104 | 264  | 615733   | 20.000 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 5) 2-Fluorophenol             | 5.628  | 112  | 445968   | 84.639 | ng    | 0.00     |
| 7) Phenol-d6                  | 7.240  | 99   | 562580   | 83.367 | ng    | -0.01    |
| 23) Nitrobenzene-d5           | 9.263  | 82   | 564140   | 94.146 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 16.222 | 330  | 291909   | 95.376 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 13.351 | 172  | 1309347  | 81.796 | ng    | 0.00     |
| 79) Terphenyl-d14             | 20.016 | 244  | 2440344  | 88.585 | ng    | 0.00     |
| Target Compounds              |        |      |          |        |       |          |
|                               |        |      |          |        |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.481  | 88   | 92270    | 40.289 | ng    | 97       |
| 3) Pyridine                   | 3.899  | 79   | 262872   | 42.578 | ng    | 98       |
| 4) n-Nitrosodimethylamine     | 3.799  | 42   | 113024   | 48.850 | ng    | 86       |
| 6) Aniline                    | 7.411  | 93   | 362473   | 41.340 | ng    | 97       |
| 8) 2-Chlorophenol             | 7.652  | 128  | 225375   | 40.551 | ng    | 97       |
| 9) Benzaldehyde               | 7.222  | 77   | 141012   | 43.922 | ng    | 94       |
| 10) Phenol                    | 7.269  | 94   | 293755   | 41.821 | ng    | 90       |
| 11) bis(2-Chloroethyl)ether   | 7.505  | 93   | 229126   | 40.760 | ng    | 96       |
| 12) 1,3-Dichlorobenzene       | 7.975  | 146  | 255582   | 40.897 | ng    | 97       |
| 13) 1,4-Dichlorobenzene       | 8.128  | 146  | 260531   | 41.272 | ng    | 99       |
| 14) 1,2-Dichlorobenzene       | 8.446  | 146  | 237878   | 40.368 | ng    | 99       |
| 15) Benzyl Alcohol            | 8.334  | 79   | 228397   | 48.240 | ng    | 96       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.616  | 45   | 248426   | 41.372 | ng    | 97       |
| 17) 2-Methylphenol            | 8.534  | 107  | 206692   | 43.242 | ng    | 93       |
| 18) Hexachloroethane          | 9.181  | 117  | 92671    | 41.909 | ng    | 97       |
| 19) n-Nitroso-di-n-propyla... | 8.899  | 70   | 194790   | 49.319 | ng    | 97       |
| 20) 3+4-Methylphenols         | 8.863  | 107  | 283286   | 44.034 | ng    | 96       |
| 22) Acetophenone              | 8.922  | 105  | 366499   | 44.906 | ng    | 96       |
| 24) Nitrobenzene              | 9.310  | 77   | 286355   | 45.794 | ng    | 97       |
| 25) Isophorone                | 9.834  | 82   | 543996   | 47.110 | ng    | 99       |
| 26) 2-Nitrophenol             | 10.022 | 139  | 130455   | 44.488 | ng    | 90       |
| 27) 2,4-Dimethylphenol        | 10.075 | 122  | 212105   | 43.922 | ng    | 98       |
| 28) bis(2-Chloroethoxy)met... | 10.316 | 93   | 303497   | 41.173 | ng    | 100      |
| 29) 2,4-Dichlorophenol        | 10.557 | 162  | 233009   | 44.972 | ng    | 99       |
| 30) 1,2,4-Trichlorobenzene    | 10.775 | 180  | 256735   | 42.227 | ng    | 98       |
| 31) Naphthalene               | 10.969 | 128  | 732693   | 41.301 | ng    | 99       |
| 32) Benzoic acid              | 10.199 | 122  | 147849   | 39.797 | ng    | 97       |
| 33) 4-Chloroaniline           | 11.075 | 127  | 324478   | 43.558 | ng    | 95       |
| 34) Hexachlorobutadiene       | 11.246 | 225  | 177766   | 43.998 | ng    | 99       |
| 35) Caprolactam               | 11.863 | 113  | 74431    | 44.995 | ng    | 95       |
| 36) 4-Chloro-3-methylphenol   | 12.187 | 107  | 264363   | 45.790 | ng    | 95       |
| 37) 2-Methylnaphthalene       | 12.563 | 142  | 523247   | 40.745 | ng    | 98       |
| 38) 1-Methylnaphthalene       | 12.781 | 142  | 493062m  | 41.024 | ng    |          |
| 40) 1,2,4,5-Tetrachloroben... | 12.928 | 216  | 313017   | 41.281 | ng    | 99       |
| 41) Hexachlorocyclopentadiene | 12.904 | 237  | 178594   | 44.342 | ng    | 97       |
| 43) 2,4,6-Trichlorophenol     | 13.169 | 196  | 212053   | 44.704 | ng    | 96       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP030323\  
 Data File : BP013940.D  
 Acq On : 03 Mar 2023 12:27  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :

Manual Integrations  
 APPROVED

Reviewed By :Christian  
 Giraldo

Quant Time: Mar 04 00:54:26 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 02 04:26:35 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |                                |
|-------------------------------|--------|------|----------|--------|-------|----------|--------------------------------|
| 44) 2,4,5-Trichlorophenol     | 13.240 | 196  | 248076   | 44.003 | ng    | 98       | 03/06/2023                     |
| 46) 1,1'-Biphenyl             | 13.563 | 154  | 691361   | 40.683 | ng    | 99       | Supervised By :Jagrut Upadhyay |
| 47) 2-Chloronaphthalene       | 13.610 | 162  | 528352   | 40.809 | ng    | 98       |                                |
| 48) 2-Nitroaniline            | 13.816 | 65   | 170468   | 46.468 | ng    | 96       |                                |
| 49) Acenaphthylene            | 14.451 | 152  | 857228   | 42.003 | ng    | 99       |                                |
| 50) Dimethylphthalate         | 14.181 | 163  | 752052   | 43.956 | ng    | 100      |                                |
| 51) 2,6-Dinitrotoluene        | 14.304 | 165  | 158281   | 42.551 | ng    | 95       | 03/06/2023                     |
| 52) Acenaphthene              | 14.793 | 154  | 508782   | 40.968 | ng    | 98       |                                |
| 53) 3-Nitroaniline            | 14.634 | 138  | 159387   | 41.244 | ng    | 95       |                                |
| 54) 2,4-Dinitrophenol         | 14.840 | 184  | 105696   | 45.414 | ng    | 89       |                                |
| 55) Dibenzofuran              | 15.128 | 168  | 858695   | 41.555 | ng    | 96       |                                |
| 56) 4-Nitrophenol             | 14.940 | 139  | 129727   | 42.419 | ng    | # 83     |                                |
| 57) 2,4-Dinitrotoluene        | 15.087 | 165  | 220178   | 43.943 | ng    | 90       |                                |
| 58) Fluorene                  | 15.781 | 166  | 686669   | 42.381 | ng    | 99       |                                |
| 59) 2,3,4,6-Tetrachlorophenol | 15.351 | 232  | 216958   | 44.703 | ng    | 98       |                                |
| 60) Diethylphthalate          | 15.540 | 149  | 755516   | 45.286 | ng    | 99       |                                |
| 61) 4-Chlorophenyl-phenyle... | 15.769 | 204  | 383515   | 43.686 | ng    | 98       |                                |
| 62) 4-Nitroaniline            | 15.798 | 138  | 168381   | 42.367 | ng    | # 82     |                                |
| 63) Azobenzene                | 16.057 | 77   | 683935   | 45.312 | ng    | 95       |                                |
| 65) 4,6-Dinitro-2-methylph... | 15.851 | 198  | 147992   | 43.669 | ng    | 95       |                                |
| 66) n-Nitrosodiphenylamine    | 15.981 | 169  | 615586   | 41.846 | ng    | 99       |                                |
| 67) 4-Bromophenyl-phenylether | 16.663 | 248  | 251791   | 43.924 | ng    | 97       |                                |
| 68) Hexachlorobenzene         | 16.787 | 284  | 274819   | 43.918 | ng    | 100      |                                |
| 69) Atrazine                  | 16.934 | 200  | 223301   | 44.521 | ng    | 99       |                                |
| 70) Pentachlorophenol         | 17.128 | 266  | 211380   | 45.818 | ng    | 98       |                                |
| 71) Phenanthrene              | 17.528 | 178  | 1134948  | 41.714 | ng    | 100      |                                |
| 72) Anthracene                | 17.616 | 178  | 1158592  | 42.587 | ng    | 99       |                                |
| 73) Carbazole                 | 17.887 | 167  | 1033223  | 42.186 | ng    | 99       |                                |
| 74) Di-n-butylphthalate       | 18.416 | 149  | 1345098  | 46.874 | ng    | 99       |                                |
| 75) Fluoranthene              | 19.481 | 202  | 1487724  | 43.862 | ng    | 98       |                                |
| 77) Benzidine                 | 19.651 | 184  | 527484   | 48.665 | ng    | 99       |                                |
| 78) Pyrene                    | 19.833 | 202  | 1526464  | 42.144 | ng    | 100      |                                |
| 80) Butylbenzylphthalate      | 20.686 | 149  | 631285   | 49.038 | ng    | 97       |                                |
| 81) Benzo(a)anthracene        | 21.545 | 228  | 1608717  | 42.386 | ng    | 100      |                                |
| 82) 3,3'-Dichlorobenzidine    | 21.469 | 252  | 541948   | 45.386 | ng    | 100      |                                |
| 83) Chrysene                  | 21.604 | 228  | 1539251  | 41.864 | ng    | 99       |                                |
| 84) Bis(2-ethylhexyl)phtha... | 21.439 | 149  | 967663   | 48.213 | ng    | 99       |                                |
| 85) Di-n-octyl phthalate      | 22.410 | 149  | 1669902  | 46.693 | ng    | 99       |                                |
| 87) Indeno(1,2,3-cd)pyrene    | 26.798 | 276  | 1991111  | 41.569 | ng    | 95       |                                |
| 88) Benzo(b)fluoranthene      | 23.327 | 252  | 1633322m | 41.942 | ng    |          |                                |
| 89) Benzo(k)fluoranthene      | 23.374 | 252  | 1637986  | 41.437 | ng    | 98       |                                |
| 90) Benzo(a)pyrene            | 23.992 | 252  | 1596916  | 42.417 | ng    | 98       |                                |
| 91) Dibenzo(a,h)anthracene    | 26.810 | 278  | 1649316  | 41.098 | ng    | 98       |                                |
| 92) Benzo(g,h,i)perylene      | 27.627 | 276  | 1636941  | 41.251 | ng    | # 95     |                                |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP030323\  
 Data File : BP013940.D  
 Acq On : 03 Mar 2023 12:27  
 Operator : CG/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1


Instrument :  
 BNA\_P  
 ClientSampleId :

Manual Integrations  
 APPROVED

Reviewed By :Christian  
 Giraldo


Quant Time: Mar 04 00:54:26 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 02 04:26:35 2023  
 Response via : Initial Calibration





#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 8.093 min Scan# 81  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:152 Resp: 86840  
Ion Ratio Lower Upper  
152 100  
150 156.2 122.8 184.2  
115 61.3 49.0 73.6


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#2  
1,4-Dioxane  
Concen: 40.289 ng  
RT: 3.481 min Scan# 50  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 88 Resp: 92270  
Ion Ratio Lower Upper  
88 100  
58 68.9 53.5 80.3  
43 27.9 20.6 30.8





#3  
Pyridine  
Concen: 42.578 ng  
RT: 3.899 min Scan# 104  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 79 Resp: 262872  
Ion Ratio Lower Upper  
79 100  
52 60.3 50.0 75.0  
51 34.2 26.7 40.1


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#4  
n-Nitrosodimethylamine  
Concen: 48.850 ng  
RT: 3.799 min Scan# 104  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion: 42 Resp: 113024  
Ion Ratio Lower Upper  
42 100  
74 131.1 119.3 178.9  
44 6.7 6.0 9.0





#5  
2-Fluorophenol  
Concen: 84.639 ng  
RT: 5.628 min Scan# 418  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:112 Resp: 445968  
Ion Ratio Lower Upper  
112 100  
64 59.3 44.6 66.8  
63 32.3 23.6 35.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#6  
Aniline  
Concen: 41.340 ng  
RT: 7.411 min Scan# 718  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion: 93 Resp: 362473  
Ion Ratio Lower Upper  
93 100  
66 39.3 29.6 44.4  
65 20.2 15.4 23.0






#7  
Phenol-d6  
Concen: 83.367 ng  
RT: 7.240 min Scan# 693  
Delta R.T. -0.012 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P



ClientSampleId :



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 99      | 100   |       |       |
| 42      | 18.5  | 12.5  | 18.7  |
| 71      | 41.5  | 27.8  | 41.8  |


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#8  
2-Chlorophenol  
Concen: 40.551 ng  
RT: 7.652 min Scan# 759  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 130     | 31.8  | 11.4  | 51.4  |
| 64      | 48.4  | 25.9  | 65.9  |





#9  
Benzaldehyde  
Concen: 43.922 ng  
RT: 7.222 min Scan# 61  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 77 Resp: 14101  
Ion Ratio Lower Upper  
77 100  
105 94.3 80.2 120.2  
106 88.8 74.8 114.8


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#10  
Phenol  
Concen: 41.821 ng  
RT: 7.269 min Scan# 694  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 94 Resp: 293755  
Ion Ratio Lower Upper  
94 100  
65 30.5 6.8 46.8  
66 44.1 17.4 57.4





#11  
bis(2-Chloroethyl)ether  
Concen: 40.760 ng  
RT: 7.505 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 93 Resp: 229120  
Ion Ratio Lower Upper  
93 100  
63 73.0 48.1 88.1  
95 32.2 12.6 52.6


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#12  
1,3-Dichlorobenzene  
Concen: 40.897 ng  
RT: 7.975 min Scan# 814  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:146 Resp: 255582  
Ion Ratio Lower Upper  
146 100  
148 64.4 51.4 77.2  
75 35.9 25.0 37.4





#13  
1,4-Dichlorobenzene  
Concen: 41.272 ng  
RT: 8.128 min Scan# 843  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:146 Resp: 26053  
Ion Ratio Lower Upper  
146 100  
148 64.0 50.4 75.6  
111 42.2 33.8 50.6


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#14  
1,2-Dichlorobenzene  
Concen: 40.368 ng  
RT: 8.446 min Scan# 894  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:146 Resp: 237878  
Ion Ratio Lower Upper  
146 100  
148 65.2 52.1 78.1  
111 45.2 34.9 52.3





#15  
Benzyl Alcohol  
Concen: 48.240 ng  
RT: 8.334 min Scan# 81  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 79 Resp: 22839  
Ion Ratio Lower Upper  
79 100  
108 72.2 60.7 91.1  
77 67.6 51.4 77.0


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 41.372 ng  
RT: 8.616 min Scan# 923  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion: 45 Resp: 248426  
Ion Ratio Lower Upper  
45 100  
77 19.1 0.0 37.5  
79 13.2 0.0 33.8






#17  
2-Methylphenol  
Concen: 43.242 ng  
RT: 8.534 min Scan# 909  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P



ClientSampleId :



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 107     | 100   |       |       |
| 108     | 107.6 | 88.9  | 133.3 |
| 77      | 53.5  | 37.1  | 55.7  |
| 79      | 54.7  | 36.7  | 55.1  |


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#18  
Hexachloroethane  
Concen: 41.909 ng  
RT: 9.181 min Scan# 1019  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 119     | 93.5  | 77.2  | 115.8 |
| 201     | 114.0 | 89.4  | 134.2 |





#19

n-Nitroso-di-n-propylamine

Concen: 49.319 ng

RT: 8.899 min Scan# 91

Delta R.T. -0.012 min


Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P

ClientSampleId :



Tgt Ion: 70 Resp: 194790

Ion Ratio Lower Upper

70 100

42 51.3 39.0 58.4



101 9.7 8.0 12.0

130 19.7 16.5 24.7

Manual Integrations

APPROVED


Reviewed By :Christian  
Giraldo



03/06/2023

Supervised By :Jagrut  
Upadhyay

03/06/2023



#20

3+4-Methylphenols

Concen: 44.034 ng

RT: 8.863 min Scan# 965

Delta R.T. -0.006 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Tgt Ion:107 Resp: 283286




Ion Ratio Lower Upper


107 100

108 87.4 67.7 107.7

77 39.3 13.4 53.4


79 34.0 9.5 49.5





#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.916 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:136 Resp: 323570  
Ion Ratio Lower Upper  
136 100  
137 11.1 8.5 12.7  
54 7.3 5.8 8.6  
68 5.3 4.6 6.8



Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#22  
Acetophenone  
Concen: 44.906 ng  
RT: 8.922 min Scan# 975  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:105 Resp: 366499  
Ion Ratio Lower Upper  
105 100  
71 1.0 0.7 1.1  
51 27.2 18.9 28.3  
120 21.4 17.4 26.0





#23  
Nitrobenzene-d5  
Concen: 94.146 ng  
RT: 9.263 min Scan# 1033  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 82 Resp: 564140  
Ion Ratio Lower Upper  
82 100  
128 38.8 32.9 49.3  
54 48.1 35.6 53.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#24  
Nitrobenzene  
Concen: 45.794 ng  
RT: 9.310 min Scan# 1041  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 77 Resp: 286355  
Ion Ratio Lower Upper  
77 100  
123 41.5 34.7 52.1  
65 13.8 10.5 15.7





#25  
Isophorone  
Concen: 47.110 ng  
RT: 9.834 min Scan# 1134  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 82 Resp: 543990  
Ion Ratio Lower Upper  
82 100  
95 8.1 5.8 8.8  
138 17.0 13.7 20.5



Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#26  
2-Nitrophenol  
Concen: 44.488 ng  
RT: 10.022 min Scan# 1162  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:139 Resp: 130455  
Ion Ratio Lower Upper  
139 100  
109 39.0 26.7 40.1  
65 55.7 39.4 59.2





#27  
2,4-Dimethylphenol  
Concen: 43.922 ng  
RT: 10.075 min Scan# 1171  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:122 Resp: 21210  
Ion Ratio Lower Upper  
122 100  
107 127.7 100.0 150.0  
121 61.0 47.2 70.8


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#28  
bis(2-Chloroethoxy)methane  
Concen: 41.173 ng  
RT: 10.316 min Scan# 1212  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 93 Resp: 303497  
Ion Ratio Lower Upper  
93 100  
95 32.6 26.1 39.1  
123 10.4 8.7 13.1





#29  
 2,4-Dichlorophenol  
 Concen: 44.972 ng  
 RT: 10.557 min Scan# 1253  
 Delta R.T. -0.006 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27


Instrument : BNA\_P  
 ClientSampleId :



Tgt Ion: 162 Resp: 233009



| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 162 | 100   |       |       |
| 164 | 64.6  | 44.4  | 84.4  |
| 98  | 36.8  | 14.7  | 54.7  |

Manual Integrations  
**APPROVED**  
 Reviewed By : Christian Giraldo



03/06/2023  
 Supervised By : Jagrut Upadhyay




03/06/2023




#30  
 1,2,4-Trichlorobenzene  
 Concen: 42.227 ng  
 RT: 10.775 min Scan# 1290  
 Delta R.T. -0.006 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27

Tgt Ion: 180 Resp: 256735


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 180 | 100   |       |       |
| 182 | 93.7  | 76.8  | 115.2 |
| 145 | 29.7  | 22.9  | 34.3  |





#31  
Naphthalene  
Concen: 41.301 ng  
RT: 10.969 min Scan# 1119  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:128 Resp: 73269  
Ion Ratio Lower Upper  
128 100  
129 10.8 8.6 12.8  
127 13.4 10.4 15.6


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#32  
Benzoic acid  
Concen: 39.797 ng  
RT: 10.199 min Scan# 1192  
Delta R.T. -0.012 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:122 Resp: 147849  
Ion Ratio Lower Upper  
122 100  
105 126.3 107.4 147.4  
77 102.4 76.0 116.0





#33  
4-Chloroaniline  
Concen: 43.558 ng  
RT: 11.075 min Scan# 1107  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 127   | Resp: | 324478 |
| Ion Ratio | Lower | Upper |        |
| 127       | 100   |       |        |
| 129       | 31.7  | 25.7  | 38.5   |
| 65        | 34.5  | 23.4  | 35.2   |
| 92        | 22.8  | 16.6  | 24.8   |


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#34  
Hexachlorobutadiene  
Concen: 43.998 ng  
RT: 11.246 min Scan# 1370  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 225   | Resp: | 177766 |
| Ion Ratio | Lower | Upper |        |
| 225       | 100   |       |        |
| 223       | 62.1  | 51.0  | 76.4   |
| 227       | 64.2  | 51.6  | 77.4   |





#35  
Caprolactam  
Concen: 44.995 ng  
RT: 11.863 min Scan# 1479  
Delta R.T. -0.012 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:113 Resp: 7443  
Ion Ratio Lower Upper  
113 100  
55 150.8 128.5 168.5  
56 122.3 93.4 133.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#36  
4-Chloro-3-methylphenol  
Concen: 45.790 ng  
RT: 12.187 min Scan# 1530  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:107 Resp: 264363  
Ion Ratio Lower Upper  
107 100  
144 24.2 20.4 30.6  
142 72.6 62.2 93.4





#37  
2-Methylnaphthalene  
Concen: 40.745 ng  
RT: 12.563 min Scan# 1594  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:142 Resp: 52324  
Ion Ratio Lower Upper  
142 100  
141 89.9 72.1 108.1  
115 41.6 30.2 45.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#38  
1-Methylnaphthalene  
Concen: 41.024 ng m  
RT: 12.781 min Scan# 1631  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Tgt Ion:142 Resp: 493062  
Ion Ratio Lower Upper  
142 100  
141 92.8 74.2 111.4  
116 4.9 3.3 4.9#






#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.728 min Scan# 1962  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P



ClientSampleId :



Tgt Ion:164 Resp: 210254  
Ion Ratio Lower Upper  
164 100  
162 101.9 81.7 122.5  
160 44.9 36.1 54.1


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 41.281 ng  
RT: 12.928 min Scan# 1656  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:216 Resp: 313017  
Ion Ratio Lower Upper  
216 100  
214 77.5 63.0 94.4  
179 20.6 16.0 24.0  
108 17.7 13.8 20.8





#41  
Hexachlorocyclopentadiene  
Concen: 44.342 ng  
RT: 12.904 min Scan# 1652  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 237 Resp: 178594  
Ion Ratio Lower Upper  
237 100  
235 62.0 44.3 84.3  
272 12.6 0.0 32.8


Manual Integrations  
APPROVED

Reviewed By : Christian  
Giraldo




03/06/2023  
Supervised By : Jagrut  
Upadhyay



03/06/2023




#42  
2,4,6-Tribromophenol  
Concen: 95.376 ng  
RT: 16.222 min Scan# 2216  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 330 Resp: 291909  
Ion Ratio Lower Upper  
330 100  
332 97.1 76.7 115.1  
141 33.0 25.1 37.7





#43  
2,4,6-Trichlorophenol  
Concen: 44.704 ng  
RT: 13.169 min Scan# 1709  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:196 Resp: 21205  
Ion Ratio Lower Upper  
196 100  
198 97.7 75.5 113.3  
200 32.8 24.1 36.1


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#44  
2,4,5-Trichlorophenol  
Concen: 44.003 ng  
RT: 13.240 min Scan# 1709  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:196 Resp: 248076  
Ion Ratio Lower Upper  
196 100  
198 97.9 76.1 114.1  
97 48.8 38.8 58.2  
132 23.8 20.5 30.7





#45  
2-Fluorobiphenyl  
Concen: 81.796 ng  
RT: 13.351 min Scan# 1728  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:172 Resp: 130934  
Ion Ratio Lower Upper  
172 100  
171 35.6 28.2 42.4  
170 23.9 19.0 28.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#46  
1,1'-Biphenyl  
Concen: 40.683 ng  
RT: 13.563 min Scan# 1764  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:154 Resp: 691361  
Ion Ratio Lower Upper  
154 100  
153 40.6 19.9 59.9  
76 11.8 0.0 32.8





#47  
2-Chloronaphthalene  
Concen: 40.809 ng  
RT: 13.610 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:162 Resp: 528352  
Ion Ratio Lower Upper  
162 100  
127 39.7 30.8 46.2  
164 32.3 26.3 39.5


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#48  
2-Nitroaniline  
Concen: 46.468 ng  
RT: 13.816 min Scan# 1807  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion: 65 Resp: 170468  
Ion Ratio Lower Upper  
65 100  
92 71.6 58.7 88.1  
138 103.0 87.3 130.9





#49  
Acenaphthylene  
Concen: 42.003 ng  
RT: 14.451 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:152 Resp: 857228  
Ion Ratio Lower Upper  
152 100  
151 20.1 15.8 23.8  
153 13.7 10.7 16.1


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#50  
Dimethylphthalate  
Concen: 43.956 ng  
RT: 14.181 min Scan# 1869  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:163 Resp: 752052  
Ion Ratio Lower Upper  
163 100  
194 4.3 3.4 5.2  
164 10.1 8.0 12.0





#51  
2,6-Dinitrotoluene  
Concen: 42.551 ng  
RT: 14.304 min Scan# 1892  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:165 Resp: 15828  
Ion Ratio Lower Upper  
165 100  
63 58.1 42.6 63.8  
89 60.8 47.1 70.7


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#52  
Acenaphthene  
Concen: 40.968 ng  
RT: 14.793 min Scan# 1973  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:154 Resp: 508782  
Ion Ratio Lower Upper  
154 100  
153 115.1 90.4 135.6  
152 53.7 42.6 64.0







#53  
3-Nitroaniline  
Concen: 41.244 ng  
RT: 14.634 min Scan# 1946  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Instrument :  
BNA\_P  
ClientSampleId :

Tgt Ion:138 Resp: 15938  
Ion Ratio Lower Upper  
138 100  
108 12.4 9.0 13.4  
92 127.2 97.1 145.7



Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#54  
2,4-Dinitrophenol  
Concen: 45.414 ng  
RT: 14.840 min Scan# 1981  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:184 Resp: 105696  
Ion Ratio Lower Upper  
184 100  
63 66.4 47.0 70.4  
154 73.8 51.5 77.3





#55  
Dibenzenofuran  
Concen: 41.555 ng  
RT: 15.128 min Scan# 2033  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:168 Resp: 85869  
Ion Ratio Lower Upper  
168 100  
139 40.3 29.8 44.8  
169 13.5 10.8 16.2


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#56  
4-Nitrophenol  
Concen: 42.419 ng  
RT: 14.940 min Scan# 1998  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:139 Resp: 129727  
Ion Ratio Lower Upper  
139 100  
109 98.6 58.2 98.2#  
65 106.5 74.9 114.9





#57

2,4-Dinitrotoluene

Concen: 43.943 ng

RT: 15.087 min Scan# 2026

Delta R.T. -0.006 min


Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P

ClientSampleId :



Tgt Ion:165 Resp: 220178

Ion Ratio Lower Upper

165 100

63 45.0 31.1 46.7



89 77.2 54.9 82.3

182 2.9 2.5 3.7

Manual Integrations

APPROVED


Reviewed By :Christian  
Giraldo



03/06/2023

Supervised By :Jagrut  
Upadhyay

03/06/2023



#58

Fluorene

Concen: 42.381 ng

RT: 15.781 min Scan# 2141

Delta R.T. 0.000 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27




Tgt Ion:166 Resp: 686669


Ion Ratio Lower Upper

166 100

165 98.5 78.1 117.1


167 12.9 10.7 16.1





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 44.703 ng  
RT: 15.351 min Scan# 2071  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:232 Resp: 216958  
Ion Ratio Lower Upper  
232 100  
131 41.5 31.7 47.5  
130 2.5 1.7 2.5  
166 29.1 23.3 34.9


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#60  
Diethylphthalate  
Concen: 45.286 ng  
RT: 15.540 min Scan# 2100  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:149 Resp: 755516  
Ion Ratio Lower Upper  
149 100  
177 22.7 18.4 27.6  
150 12.7 9.8 14.8





#61  
4-Chlorophenyl-phenylether  
Concen: 43.686 ng  
RT: 15.769 min Scan# 2139  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:204 Resp: 38351  
Ion Ratio Lower Upper  
204 100  
206 31.8 26.2 39.2  
141 58.2 45.3 67.9


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023




#62  
4-Nitroaniline  
Concen: 42.367 ng  
RT: 15.798 min Scan# 2144  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27




Tgt Ion:138 Resp: 168381  
Ion Ratio Lower Upper  
138 100  
92 56.2 32.0 72.0  
108 109.5 65.6 105.6#





#63  
Azobenzene  
Concen: 45.312 ng  
RT: 16.057 min Scan# 2191  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion: 77 Resp: 68393  
Ion Ratio Lower Upper  
77 100  
182 25.9 6.9 46.9  
105 20.0 1.1 41.1  
51 31.1 6.3 46.3


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.487 min Scan# 2431  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:188 Resp: 485671  
Ion Ratio Lower Upper  
188 100  
94 7.1 7.0 10.6  
80 7.6 7.8 11.8#






#65  
4,6-Dinitro-2-methylphenol  
Concen: 43.669 ng  
RT: 15.851 min Scan# 2153  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P



ClientSampleId :



Tgt Ion:198 Resp: 147992  
Ion Ratio Lower Upper  
198 100  
51 40.5 14.3 54.3  
105 40.9 20.5 60.5


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#66  
n-Nitrosodiphenylamine  
Concen: 41.846 ng  
RT: 15.981 min Scan# 2175  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:169 Resp: 615586  
Ion Ratio Lower Upper  
169 100  
168 67.6 53.0 79.6  
167 36.6 29.0 43.4





#67  
4-Bromophenyl-phenylether  
Concen: 43.924 ng  
RT: 16.663 min Scan# 2119  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:248 Resp: 25179  
Ion Ratio Lower Upper  
248 100  
250 97.0 78.1 117.1  
141 74.1 54.3 81.5


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#68  
Hexachlorobenzene  
Concen: 43.918 ng  
RT: 16.787 min Scan# 2312  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:284 Resp: 274819  
Ion Ratio Lower Upper  
284 100  
142 30.7 24.8 37.2  
249 32.2 25.8 38.6





#69

Atrazine

Concen: 44.521 ng

RT: 16.934 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :



BNA\_P

ClientSampleId :

Tgt Ion:200 Resp: 22330  
 Ion Ratio Lower Upper  
 200 100  
 173 24.2 5.2 45.2  
 215 48.4 28.4 68.4



Manual Integrations  
 APPROVED

Reviewed By :Christian  
 Giraldo



03/06/2023  
 Supervised By :Jagrut  
 Upadhyay

03/06/2023



#70

Pentachlorophenol

Concen: 45.818 ng




RT: 17.128 min Scan# 2370


Delta R.T. -0.006 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Tgt Ion:266 Resp: 211380  
 Ion Ratio Lower Upper  
 266 100  
 268 61.5 50.9 76.3  
 264 63.2 49.4 74.0





#71

Phenanthrene

Concen: 41.714 ng

RT: 17.528 min Scan# 2441

Delta R.T. -0.006 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P

ClientSampleId :

Tgt Ion:178 Resp: 113494

Ion Ratio Lower Upper

178 100



176 19.3 15.4 23.0

179 15.1 12.2 18.4


Manual Integrations

APPROVED


Reviewed By :Christian  
Giraldo



03/06/2023  
Supervised By :Jagrut  
Upadhyay



03/06/2023



#72

Anthracene

Concen: 42.587 ng

RT: 17.616 min Scan# 2453

Delta R.T. -0.006 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27




Tgt Ion:178 Resp: 1158592


Ion Ratio Lower Upper

178 100

176 18.9 14.7 22.1


179 15.4 12.4 18.6





#73  
 Carbazole  
 Concen: 42.186 ng  
 RT: 17.887 min Scan# 2499  
 Delta R.T. 0.000 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27


Instrument :  
 BNA\_P  
 ClientSampleId :



Tgt Ion:167 Resp: 103322  
 Ion Ratio Lower Upper  
 167 100  
 166 22.0 17.7 26.5  
 139 13.3 10.2 15.2



Manual Integrations  
 APPROVED

Reviewed By :Christian  
 Giraldo






03/06/2023  
 Supervised By :Jagrut  
 Upadhyay


03/06/2023



#74  
 Di-n-butylphthalate  
 Concen: 46.874 ng  
 RT: 18.416 min Scan# 2589  
 Delta R.T. 0.000 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27


Tgt Ion:149 Resp: 1345098  
 Ion Ratio Lower Upper  
 149 100  
 150 8.9 7.5 11.3  
 104 5.5 4.1 6.1





#75  
Fluoranthene  
Concen: 43.862 ng  
RT: 19.481 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:202 Resp: 1487724  
Ion Ratio Lower Upper  
202 100  
101 8.3 0.0 30.0  
203 17.5 0.0 37.5


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.563 min Scan# 3124  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:240 Resp: 530020  
Ion Ratio Lower Upper  
240 100  
120 7.9 7.9 11.9#  
236 26.2 20.9 31.3





#77  
Benzidine  
Concen: 48.665 ng  
RT: 19.651 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:184 Resp: 527484  
Ion Ratio Lower Upper  
184 100  
185 13.7 11.4 17.0  
183 11.5 9.6 14.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#78  
Pyrene  
Concen: 42.144 ng  
RT: 19.833 min Scan# 2830  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:202 Resp: 1526464  
Ion Ratio Lower Upper  
202 100  
200 21.3 16.9 25.3  
203 17.5 14.2 21.2






#79  
 Terphenyl-d14  
 Concen: 88.585 ng  
 RT: 20.016 min Scan# 21  
 Delta R.T. -0.006 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P



ClientSampleId :



Tgt Ion:244 Resp: 2440344  
 Ion Ratio Lower Upper  
 244 100  
 212 7.7 5.8 8.6  
 122 7.9 6.7 10.1


Manual Integrations  
 APPROVED

Reviewed By :Christian  
 Giraldo






03/06/2023  
 Supervised By :Jagrut  
 Upadhyay


03/06/2023



#80  
 Butylbenzylphthalate  
 Concen: 49.038 ng  
 RT: 20.686 min Scan# 2975  
 Delta R.T. 0.000 min  
 Lab File: BP013940.D  
 Acq: 03 Mar 2023 12:27


Tgt Ion:149 Resp: 631285  
 Ion Ratio Lower Upper  
 149 100  
 91 68.6 56.8 85.2  
 206 23.7 18.1 27.1





#81  
Benzo(a)anthracene  
Concen: 42.386 ng  
RT: 21.545 min Scan# 3108  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:228 Resp: 160871  
Ion Ratio Lower Upper  
228 100  
226 27.5 21.7 32.5  
229 19.7 15.8 23.6


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#82  
3,3'-Dichlorobenzidine  
Concen: 45.386 ng  
RT: 21.469 min Scan# 3108  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:252 Resp: 541948  
Ion Ratio Lower Upper  
252 100  
254 64.3 51.4 77.0  
126 8.5 7.4 11.0





#83

Chrysene

Concen: 41.864 ng

RT: 21.604 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P

ClientSampleId :

Tgt Ion:228 Resp: 153925

Ion Ratio Lower Upper

228 100



226 30.6 23.8 35.8

229 19.5 15.8 23.6

Manual Integrations



APPROVED

Reviewed By :Christian  
Giraldo



03/06/2023  
Supervised By :Jagrut  
Upadhyay

03/06/2023



#84

Bis(2-ethylhexyl)phthalate

Concen: 48.213 ng

RT: 21.439 min Scan# 3103

Delta R.T. -0.006 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27




Tgt Ion:149 Resp: 967663


Ion Ratio Lower Upper

149 100

167 27.2 22.1 33.1


279 4.9 3.9 5.9





#85  
Di-n-octyl phthalate  
Concen: 46.693 ng  
RT: 22.410 min Scan# 31  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:149 Resp: 166990  
Ion Ratio Lower Upper  
149 100  
167 1.6 1.3 1.9  
43 8.0 6.1 9.1


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 24.104 min Scan# 3556  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27

Tgt Ion:264 Resp: 615733  
Ion Ratio Lower Upper  
264 100  
260 23.2 19.4 29.0  
265 22.7 17.6 26.4





#87

Indeno(1,2,3-cd)pyrene

Concen: 41.569 ng

RT: 26.798 min Scan# 4019

Delta R.T. -0.006 min


Lab File: BP013940.D

Acq: 03 Mar 2023 12:27

Instrument :

BNA\_P

ClientSampleId :



Tgt Ion:276 Resp: 199111

Ion Ratio Lower Upper

276 100



138 17.9 17.8 26.8

277 25.6 20.2 30.2

Manual Integrations


APPROVED

Reviewed By :Christian  
Giraldo



03/06/2023  
Supervised By :Jagrut  
Upadhyay

03/06/2023



#88

Benzo(b)fluoranthene

Concen: 41.942 ng m

RT: 23.327 min Scan# 3424

Delta R.T. 0.000 min

Lab File: BP013940.D

Acq: 03 Mar 2023 12:27




Tgt Ion:252 Resp: 1633322


Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.6


125 7.0 7.1 10.7#





#89  
Benzo(k)fluoranthene  
Concen: 41.437 ng  
RT: 23.374 min Scan# 3436  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:252 Resp: 163798  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.5 26.3  
125 7.2 7.0 10.4


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo






03/06/2023  
Supervised By :Jagrut  
Upadhyay


03/06/2023



#90  
Benzo(a)pyrene  
Concen: 42.417 ng  
RT: 23.992 min Scan# 3537  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27


Tgt Ion:252 Resp: 1596916  
Ion Ratio Lower Upper  
252 100  
253 22.2 17.4 26.0  
125 8.2 7.7 11.5





#91  
Dibenzo(a,h)anthracene  
Concen: 41.098 ng  
RT: 26.810 min Scan# 4023  
Delta R.T. -0.006 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Instrument :  
BNA\_P  
ClientSampleId :



Tgt Ion:278 Resp: 1649310  
Ion Ratio Lower Upper  
278 100  
139 11.6 11.0 16.6  
279 23.4 18.6 27.8


Manual Integrations  
APPROVED

Reviewed By :Christian  
Giraldo




03/06/2023  
Supervised By :Jagrut  
Upadhyay

03/06/2023



#92  
Benzo(g,h,i)perylene  
Concen: 41.251 ng  
RT: 27.627 min Scan# 4155  
Delta R.T. 0.000 min  
Lab File: BP013940.D  
Acq: 03 Mar 2023 12:27



Tgt Ion:276 Resp: 1636941  
Ion Ratio Lower Upper  
276 100  
277 23.3 18.7 28.1  
138 14.9 15.5 23.3

