

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP030520\
 Data File : BP002071.D
 Acq On : 05 Mar 2020 15:58
 Operator : CG/JU
 Sample : L1780-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0DF5

Quant Time: Mar 05 16:52:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP022720MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 05 06:21:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	262614	20.00	ng/ul	0.00
18) Naphthalene-d8	10.41	136	1079714	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.27	164	680443	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.03	188	1491309	20.00	ng/ul	0.00
78) Chrysene-d12	21.12	240	1461700	20.00	ng/ul	0.00
86) Perylene-d12	23.33	264	1648026	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.16	96	28381	4.65	ng/uL	0.00
5) Phenol-d5	6.82	99	106488	5.32	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.98	67	277472	26.16	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	348472	20.11	ng/ul	0.00
13) 4-Methylphenol-d8	8.34	113	210861	12.81	ng/ul	0.00
19) Nitrobenzene-d5	8.78	128	195152	23.53	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	209849	22.49	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	406920	22.64	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	30399	1.54	ng/ul	0.00
44) Dimethylphthalate-d6	13.69	166	1428239	27.74	ng/ul	0.00
47) Acenaphthylene-d8	13.96	160	1678679	27.26	ng/ul	0.00
52) 4-Nitrophenol-d4	14.49	143	24145	2.45	ng/ul	0.00
58) Fluorene-d10	15.27	176	1299417	29.16	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.40	200	141447	17.20	ng/ul	0.00
71) Anthracene-d10	17.13	188	2155377	32.16	ng/ul	0.00
79) Pyrene-d10	19.37	212	2551695	33.35	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.19	264	2681290	31.80	ng/ul	0.00


Target Compounds				Ovalue
45) Dimethylphthalate	13.74	163	125109	2.354 ng/ul 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP030520\
Data File : BP002071.D
Acq On : 05 Mar 2020 15:58
Operator : CG/JU
Sample : L1780-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0DF5

Quant Time: Mar 05 16:52:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP022720MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 05 06:21:29 2020
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 2.354 ng/ul

RT: 13.74 min Scan# 1845

Delta R.T. 0.00 min

Lab File: BP002071.D

Acq: 05 Mar 2020 15:58

Instrument :

BNA_P

ClientSampleId :

C0DF5


Tot Ion:163 Resp: 125109

Ion Ratio Lower Upper

163 100

194 4.4 3.6 5.4

164 9.8 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

