

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP030620\
 Data File : BP002097.D
 Acq On : 06 Mar 2020 15:23
 Operator : CG/JU
 Sample : L1782-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AB4

Quant Time: Mar 06 17:27:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP022720MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 05 06:21:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	304932	20.00	ng/ul	0.00
18) Naphthalene-d8	10.41	136	1275318	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.27	164	827748	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.03	188	1820569	20.00	ng/ul	0.00
78) Chrysene-d12	21.12	240	1779901	20.00	ng/ul	0.00
86) Perylene-d12	23.34	264	1980253	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.16	96	34682	4.89	ng/uL	0.00
5) Phenol-d5	6.82	99	136682	5.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.98	67	360183	29.24	ng/ul	0.00
9) 2-Chlorophenol-d4	7.17	132	461824	22.95	ng/ul	0.00
13) 4-Methylphenol-d8	8.34	113	273862	14.33	ng/ul	0.00
19) Nitrobenzene-d5	8.78	128	267846	27.34	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	286807	26.02	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.03	165	550906	25.95	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	38634	1.66	ng/ul	0.00
44) Dimethylphthalate-d6	13.69	166	1987026	31.72	ng/ul	0.00
47) Acenaphthylene-d8	13.96	160	2248149	30.01	ng/ul	0.00
52) 4-Nitrophenol-d4	14.48	143	38664	3.23	ng/ul	0.00
58) Fluorene-d10	15.27	176	1778183	32.80	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.39	200	218064	21.73	ng/ul	0.00
71) Anthracene-d10	17.13	188	2869054	35.06	ng/ul	0.00
79) Pyrene-d10	19.37	212	3341201	35.86	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.20	264	3700184	36.52	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	13.73	163	68717	1.063	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP030620\
Data File : BP002097.D
Acq On : 06 Mar 2020 15:23
Operator : CG/JU
Sample : L1782-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AB4

Quant Time: Mar 06 17:27:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP022720MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 05 06:21:29 2020
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 1.063 ng/ul

RT: 13.73 min Scan# 1844

Delta R.T. -0.01 min

Lab File: BP002097.D

Acq: 06 Mar 2020 15:23

Instrument :

BNA_P

ClientSampleId :

C0AB4


Tot Ion:163 Resp: 68717

Ion Ratio Lower Upper

163 100

194 4.0 3.6 5.4

164 10.3 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

