

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/17/2023	BNA_P	<u>BP014114.D</u>
Compound Name	Response	Retention Time
DDT	1307047	20.839
DDD	18664	20.463
DDE	546	19.916
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
19210	1326257	1.45