

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031723\
 Data File : BP014115.D
 Acq On : 17 Mar 2023 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 00:22:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 18 00:19:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	137	0.00
2	1,4-Dioxane	40.000	37.229	6.9	119	0.00
3	Pyridine	40.000	39.459	1.4	123	0.00
4	n-Nitrosodimethylamine	40.000	39.359	1.6	122	0.00
5 S	2-Fluorophenol	80.000	78.274	2.2	124	0.00
6	Aniline	40.000	42.029	-5.1	135	0.00
7 S	Phenol-d6	80.000	84.111	-5.1	135	0.00
8	2-Chlorophenol	40.000	40.514	-1.3	131	0.00
9	Benzaldehyde	40.000	38.121	4.7	124	0.00
10 C	Phenol	40.000	42.456	-6.1	135	0.00
11	bis(2-Chloroethyl)ether	40.000	43.487	-8.7	142	0.00
12	1,3-Dichlorobenzene	40.000	40.268	-0.7	132	0.00
13 C	1,4-Dichlorobenzene	40.000	40.541	-1.4	132	0.00
14	1,2-Dichlorobenzene	40.000	40.863	-2.2	135	0.00
15	Benzyl Alcohol	40.000	43.468	-8.7	138	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	49.896	-24.7	169	0.00
17	2-Methylphenol	40.000	43.422	-8.6	141	0.00
18	Hexachloroethane	40.000	41.292	-3.2	132	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	47.236	-18.1	155	0.00
20	3+4-Methylphenols	40.000	44.132	-10.3	146	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	152	0.00
22	Acetophenone	40.000	40.910	-2.3	145	0.00
23 S	Nitrobenzene-d5	80.000	82.819	-3.5	141	0.00
24	Nitrobenzene	40.000	41.824	-4.6	148	0.00
25	Isophorone	40.000	42.306	-5.8	150	0.00
26 C	2-Nitrophenol	40.000	40.902	-2.3	141	0.00
27	2,4-Dimethylphenol	40.000	40.724	-1.8	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.379	-5.9	154	0.00
29 C	2,4-Dichlorophenol	40.000	40.038	-0.1	144	0.00
30	1,2,4-Trichlorobenzene	40.000	38.209	4.5	137	0.00
31	Naphthalene	40.000	40.394	-1.0	146	0.00
32	Benzoic acid	40.000	37.806	5.5	141	0.00
33	4-Chloroaniline	40.000	41.466	-3.7	147	0.00
34 C	Hexachlorobutadiene	40.000	36.886	7.8	130	0.00
35	Caprolactam	40.000	44.518	-11.3	165	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.961	-7.4	155	0.00
37	2-Methylnaphthalene	40.000	40.036	-0.1	147	0.00
38	1-Methylnaphthalene	40.000	40.290	-0.7	149	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	160	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.568	6.1	140	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.531	8.7	132	0.00
42 S	2,4,6-Tribromophenol	80.000	77.864	2.7	146	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.417	1.5	146	0.00
44	2,4,5-Trichlorophenol	40.000	39.887	0.3	150	0.00
45 S	2-Fluorobiphenyl	80.000	78.376	2.0	150	0.00
46	1,1'-Biphenyl	40.000	39.635	0.9	147	0.00
47	2-Chloronaphthalene	40.000	39.484	1.3	148	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031723\
 Data File : BP014115.D
 Acq On : 17 Mar 2023 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 00:22:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 18 00:19:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.540	-16.3	167	0.00
49	Acenaphthylene	40.000	40.509	-1.3	152	0.00
50	Dimethylphthalate	40.000	40.307	-0.8	159	0.00
51	2,6-Dinitrotoluene	40.000	41.736	-4.3	158	0.00
52 C	Acenaphthene	40.000	39.881	0.3	150	0.00
53	3-Nitroaniline	40.000	42.676	-6.7	159	0.00
54 P	2,4-Dinitrophenol	40.000	38.638	3.4	156	0.00
55	Dibenzofuran	40.000	40.300	-0.7	151	0.00
56 P	4-Nitrophenol	40.000	42.047	-5.1	162	0.00
57	2,4-Dinitrotoluene	40.000	42.670	-6.7	158	0.00
58	Fluorene	40.000	40.211	-0.5	153	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.573	3.6	146	0.00
60	Diethylphthalate	40.000	41.253	-3.1	161	0.00
61	4-Chlorophenyl-phenylether	40.000	38.957	2.6	148	0.00
62	4-Nitroaniline	40.000	44.366	-10.9	167	0.00
63	Azobenzene	40.000	44.522	-11.3	167	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	169	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.666	-1.7	160	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.289	-0.7	156	0.00
67	4-Bromophenyl-phenylether	40.000	37.755	5.6	148	0.00
68	Hexachlorobenzene	40.000	37.864	5.3	147	0.00
69	Atrazine	40.000	39.625	0.9	167	0.00
70 C	Pentachlorophenol	40.000	39.437	1.4	151	0.00
71	Phenanthrene	40.000	39.588	1.0	157	0.00
72	Anthracene	40.000	39.717	0.7	157	0.00
73	Carbazole	40.000	40.443	-1.1	163	0.00
74	Di-n-butylphthalate	40.000	41.703	-4.3	170	0.00
75 C	Fluoranthene	40.000	38.717	3.2	158	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	152	0.00
77	Benzydine	40.000	34.651	13.4	119	0.00
78	Pyrene	40.000	41.405	-3.5	157	0.00
79 S	Terphenyl-d14	80.000	81.122	-1.4	155	0.00
80	Butylbenzylphthalate	40.000	44.373	-10.9	160	0.00
81	Benzo(a)anthracene	40.000	39.728	0.7	142	0.00
82	3,3'-Dichlorobenzidine	40.000	39.536	1.2	134	0.00
83	Chrysene	40.000	39.447	1.4	140	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.128	-5.3	147	0.00
85 c	Di-n-octyl phthalate	40.000	39.298	1.8	129	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.635	8.4	92	0.00
88	Benzo(b)fluoranthene	40.000	41.791	-4.5	116	0.00
89	Benzo(k)fluoranthene	40.000	42.347	-5.9	118	0.00
90 C	Benzo(a)pyrene	40.000	40.524	-1.3	110	0.00
91	Dibenzo(a,h)anthracene	40.000	36.818	8.0	93	0.00
92	Benzo(g,h,i)perylene	40.000	36.128	9.7	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031723\
Data File : BP014115.D
Acq On : 17 Mar 2023 11:33
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 00:22:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 18 00:19:43 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0