

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040723\
 Data File : BP014616.D
 Acq On : 07 Apr 2023 21:47
 Operator : CG/JU
 Sample : 02215-21
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AE6

Quant Time: Apr 08 01:51:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 07 00:45:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	141054	20.000	ng/ul	0.00
20) Naphthalene-d8	10.799	136	621621	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.634	164	408285	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	863797	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	651223	20.000	ng/ul	0.00
88) Perylene-d12	23.986	264	722790	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	16026	4.406	ng/uL	0.00
4) Pyridine-d5	3.793	84	604	0.065	ng/ul	0.00
7) Phenol-d5	7.158	99	55726	4.284	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	249285	29.867	ng/ul	0.00
11) 2-Chlorophenol-d4	7.511	132	197046	20.957	ng/ul	0.00
15) 4-Methylphenol-d8	8.705	113	119182	11.358	ng/ul	0.00
21) Nitrobenzene-d5	9.158	128	132131	26.330	ng/ul	0.00
24) 2-Nitrophenol-d4	9.881	143	146893	25.563	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.428	165	233236	22.287	ng/ul	0.00
31) 4-Chloroaniline-d4	10.963	131	77469	5.620	ng/ul	0.01
46) Dimethylphthalate-d6	14.040	166	1012307	30.579	ng/ul	0.00
49) Acenaphthylene-d8	14.328	160	1094051	29.784	ng/ul	0.00
54) 4-Nitrophenol-d4	14.863	143	25	0.005	ng/ul	0.00
60) Fluorene-d10	15.628	176	835626	30.269	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	151295	23.804	ng/ul	0.00
73) Anthracene-d10	17.492	188	1320552	31.589	ng/ul	0.00
81) Pyrene-d10	19.722	212	1455123	33.266	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.822	264	1258798	32.780	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040723\
Data File : BP014616.D
Acq On : 07 Apr 2023 21:47
Operator : CG/JU
Sample : 02215-21
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AE6

Quant Time: Apr 08 01:51:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 07 00:45:34 2023
Response via : Initial Calibration

