

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040822\
 Data File : BP009737.D
 Acq On : 09 Apr 2022 02:39
 Operator : CG/JU
 Sample : N2266-03MSD
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD

Quant Time: Apr 09 05:52:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	84297	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	396854	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.604	164	280352	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.357	188	633699	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.439	240	713719	20.000	ng/u1	# 0.00
88) Perylene-d12	23.915	264	766363	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	12585m	6.663	ng/uL	0.00
4) Pyridine-d5	3.811	84	57378	11.018	ng/u1	0.00
7) Phenol-d5	7.122	99	90798	12.269	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	238313	54.089	ng/u1	0.00
11) 2-Chlorophenol-d4	7.504	132	234408	41.990	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	190757	30.877	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	166450	58.162	ng/u1	0.00
24) 2-Nitrophenol-d4	9.857	143	179171	57.105	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	315955	48.123	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	382816	40.938	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	1274701	57.883	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	1385844	53.118	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	58749	14.710	ng/u1	0.00
60) Fluorene-d10	15.598	176	1061773	57.313	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	240167	63.975	ng/u1	0.00
73) Anthracene-d10	17.457	188	1805179	59.496	ng/u1	0.00
81) Pyrene-d10	19.686	212	2295184	58.881	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.762	264	2458178	61.594	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	13034	6.769	ng/uL#	26
5) Pyridine	3.828	79	60188	11.077	ng/u1#	39
6) Benzaldehyde	7.104	77	247922m	54.623	ng/u1	
8) Phenol	7.152	94	90536	12.222	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.393	93	279968	48.518	ng/u1#	79
12) 2-Chlorophenol	7.534	128	214633	38.136	ng/u1	94
13) 2-Methylphenol	8.410	108	188017	32.548	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.504	45	415431	48.705	ng/u1#	84
16) Acetophenone	8.799	105	484328	49.757	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.787	70	254552	50.442	ng/u1#	71
18) 4-Methylphenol	8.740	108	178061	28.464	ng/u1	94
19) Hexachloroethane	9.063	117	109639	46.140	ng/u1	83
22) Nitrobenzene	9.175	77	373111	50.793	ng/u1#	84
23) Isophorone	9.710	82	736633	48.665	ng/u1	97
25) 2-Nitrophenol	9.887	139	168830	50.451	ng/u1#	88
26) 2,4-Dimethylphenol	9.951	107	308097	38.797	ng/u1#	80
27) Bis(2-Chloroethoxy)met...	10.193	93	414922	46.689	ng/u1#	98
29) 2,4-Dichlorophenol	10.428	162	282387	44.364	ng/u1#	84
30) Naphthalene	10.834	128	960958	47.083	ng/u1	97
32) 4-Chloroaniline	10.934	127	342689	37.032	ng/u1	98
33) Hexachlorobutadiene	11.134	225	207993	45.371	ng/u1	97
34) Caprolactam	11.710	113	15856m	7.532	ng/u1	
35) 4-Chloro-3-methylphenol	12.051	107	333513	43.242	ng/u1#	73
36) 2-Methylnaphthalene	12.439	142	703331	47.168	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040822\
 Data File : BP009737.D
 Acq On : 09 Apr 2022 02:39
 Operator : CG/JU
 Sample : N2266-03MSD
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD


Quant Time: Apr 09 05:52:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.657	142	720684	47.868	ng/ul#	93
39) 1,2,4,5-Tetrachloroben...	12.804	216	416023	48.502	ng/ul#	93
40) Hexachlorocyclopentadiene	12.792	237	227887	38.286	ng/ul	96
41) 2,4,6-Trichlorophenol	13.039	196	277408	50.643	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.104	196	309155	52.225	ng/ul#	87
43) 1,1'-Biphenyl	13.445	154	1000552	48.913	ng/ul	93
44) 2-Chloronaphthalene	13.486	162	776610	49.391	ng/ul	93
45) 2-Nitroaniline	13.681	65	291687	63.300	ng/ul#	80
47) Dimethylphthalate	14.063	163	1116705	52.686	ng/ul#	92
48) 2,6-Dinitrotoluene	14.175	165	240796	61.763	ng/ul	96
50) Acenaphthylene	14.328	152	1305635	50.710	ng/ul	95
51) 3-Nitroaniline	14.498	138	227648	56.167	ng/ul#	84
52) Acenaphthene	14.669	153	844141	50.074	ng/ul	96
53) 2,4-Dinitrophenol	14.704	184	146177	62.029	ng/ul#	83
55) 4-Nitrophenol	14.798	109	52227	14.206	ng/ul#	64
56) Dibenzofuran	15.004	168	1260604	50.809	ng/ul	96
57) 2,4-Dinitrotoluene	14.957	165	359142	62.734	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.228	232	296948	55.017	ng/ul#	86
59) Diethylphthalate	15.433	149	1178946	53.746	ng/ul	93
61) Fluorene	15.657	166	1065360	52.131	ng/ul	89
62) 4-Chlorophenyl-phenyle...	15.651	204	551187	50.725	ng/ul	97
63) 4-Nitroaniline	15.663	138	265487m	65.398	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.728	198	217609	58.316	ng/ul#	94
67) N-Nitrosodiphenylamine	15.857	169	948610	52.293	ng/ul	94
68) 4-Bromophenyl-phenylether	16.545	248	358959	52.330	ng/ul	96
69) Hexachlorobenzene	16.663	284	415345	52.658	ng/ul#	90
70) Atrazine	16.816	200	385148m	51.438	ng/ul	
71) Pentachlorophenol	17.004	266	290047	54.327	ng/ul#	82
72) Phenanthrene	17.398	178	1839718	54.388	ng/ul	98
74) Anthracene	17.492	178	1860585	54.253	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.410	216	451125	49.239	ng/uL#	85
76) Pentachlorobenzene	14.928	250	447624	50.124	ng/uL	90
77) Carbazole	17.757	167	1720226	55.505	ng/ul#	96
78) Di-n-butylphthalate	18.316	149	2149133	56.391	ng/ul#	92
80) Fluoranthene	19.363	202	2362053	52.645	ng/ul#	94
82) Pyrene	19.716	202	2403857	52.844	ng/ul#	91
83) Butylbenzylphthalate	20.586	149	1047773	56.147	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.351	252	856041	56.285	ng/ul#	96
85) Benzo(a)anthracene	21.421	228	2512009	55.519	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.351	149	1569463	55.913	ng/ul#	82
87) Chrysene	21.480	228	2421854	55.478	ng/ul	99
89) Di-n-octyl phthalate	22.310	149	2739496	51.021	ng/ul	100
90) Benzo(b)fluoranthene	23.162	252	2771196	54.426	ng/ul	99
91) Benzo(k)fluoranthene	23.210	252	2499541	53.797	ng/ul#	98
93) Benzo(a)pyrene	23.809	252	2635062	55.831	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.515	276	3123268	63.282	ng/ul	96
95) Dibenzo(a,h)anthracene	26.539	278	2689946	62.855	ng/ul#	96
96) Benzo(g,h,i)perylene	27.309	276	2664728	65.481	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.769 ng/uL
RT: 3.428 min Scan# 57
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 88 Resp: 13034
Ion Ratio Lower Upper
88 100
43 41.9 6.8 10.2
58 75.7 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#5
Pyridine
Concen: 11.077 ng/uL
RT: 3.828 min Scan# 126
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion: 79 Resp: 60188
Ion Ratio Lower Upper
79 100
52 92.9 35.5 53.3#
51 43.5 20.2 30.4#





#6
Benzaldehyde
Concen: 54.623 ng/ul m
RT: 7.104 min Scan# 611
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion: 77 Resp: 24792
Ion Ratio Lower Upper
77 100
105 92.2 77.0 115.4
106 88.7 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#8
Phenol
Concen: 12.222 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion: 94 Resp: 90536
Ion Ratio Lower Upper
94 100
65 31.9 19.0 28.6#
66 44.1 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 48.518 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 93 Resp: 279968
Ion Ratio Lower Upper
93 100
63 80.3 45.0 67.4
95 33.0 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#12
2-Chlorophenol
Concen: 38.136 ng/ul
RT: 7.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:128 Resp: 214633
Ion Ratio Lower Upper
128 100
64 52.2 45.3 67.9
130 32.9 29.0 43.6





#13
2-Methylphenol
Concen: 32.548 ng/ul
RT: 8.410 min Scan# 905
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion: 108 Resp: 18801
Ion Ratio Lower Upper
108 100
107 89.2 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 48.705 ng/ul
RT: 8.504 min Scan# 921
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39



Tgt Ion: 45 Resp: 415431
Ion Ratio Lower Upper
45 100
77 13.6 17.1 25.7#
79 10.4 13.2 19.8#









#16
Acetophenone
Concen: 49.757 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion: 105 Resp: 484328
Ion Ratio Lower Upper
105 100
77 79.5 53.4 80.2
51 30.4 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#17
N-Nitroso-di-n-propylamine
Concen: 50.442 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion: 70 Resp: 254552
Ion Ratio Lower Upper
70 100
42 65.8 31.8 47.8#
101 9.7 15.6 23.4#
130 21.1 13.7 20.5#





#18
4-Methylphenol
Concen: 28.464 ng/ul
RT: 8.740 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:108 Resp: 17806
Ion Ratio Lower Upper
108 100
107 110.9 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#19
Hexachloroethane
Concen: 46.140 ng/ul
RT: 9.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:117 Resp: 109639
Ion Ratio Lower Upper
117 100
201 106.9 106.5 159.7
199 67.3 59.0 88.6





#22
Nitrobenzene
Concen: 50.793 ng/ul
RT: 9.175 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion: 77 Resp: 37311
Ion Ratio Lower Upper
77 100
123 42.0 42.3 63.5
65 15.3 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#23
Isophorone
Concen: 48.665 ng/ul
RT: 9.710 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39



Tgt Ion: 82 Resp: 736633
Ion Ratio Lower Upper
82 100
95 7.9 7.8 11.8
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 50.451 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.012 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:139 Resp: 168830
Ion Ratio Lower Upper
139 100
65 61.8 49.4 74.2
109 37.3 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#26
2,4-Dimethylphenol
Concen: 38.797 ng/ul
RT: 9.951 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:107 Resp: 308097
Ion Ratio Lower Upper
107 100
121 47.2 34.1 51.1
122 79.3 46.3 69.5





#27
 Bis(2-Chloroethoxy)methane
 Concen: 46.689 ng/ul
 RT: 10.193 min Scan# 1208
 Delta R.T. 0.000 min
 Lab File: BP009737.D
 Acq: 09 Apr 2022 02:39




Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD



Tgt Ion: 93 Resp: 41492
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.4
 123 11.1 6.2 9.4




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022



#29
 2,4-Dichlorophenol
 Concen: 44.364 ng/ul
 RT: 10.428 min Scan# 1248
 Delta R.T. -0.006 min
 Lab File: BP009737.D
 Acq: 09 Apr 2022 02:39

Tgt Ion:162 Resp: 282387
 Ion Ratio Lower Upper
 162 100
 164 65.0 37.0 55.6#
 98 37.4 31.0 46.6









#30
Naphthalene
Concen: 47.083 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:128 Resp: 960958
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.1 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#32
4-Chloroaniline
Concen: 37.032 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:127 Resp: 342689
Ion Ratio Lower Upper
127 100
129 32.1 24.6 37.0





#33
Hexachlorobutadiene
Concen: 45.371 ng/ul
RT: 11.134 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:225 Resp: 20799
Ion Ratio Lower Upper
225 100
223 63.7 53.4 80.0
227 65.9 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#34
Caprolactam
Concen: 7.532 ng/ul m
RT: 11.710 min Scan# 1466
Delta R.T. 0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:113 Resp: 15856
Ion Ratio Lower Upper
113 100
55 178.5 58.2 87.2#
56 138.7 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 43.242 ng/ul
RT: 12.051 min Scan# 1524
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:107 Resp: 33351
Ion Ratio Lower Upper
107 100
144 23.9 15.8 23.8
142 72.2 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#36
2-Methylnaphthalene
Concen: 47.168 ng/ul
RT: 12.439 min Scan# 1590
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:142 Resp: 703331
Ion Ratio Lower Upper
142 100
141 88.3 64.3 96.5





#37
1-Methylnaphthalene
Concen: 47.868 ng/ul
RT: 12.657 min Scan# 1627
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:142 Resp: 720684
Ion Ratio Lower Upper
142 100
141 93.7 69.8 104.8
116 4.6 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 48.502 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:216 Resp: 416023
Ion Ratio Lower Upper
216 100
214 79.2 58.2 87.2
179 21.9 14.2 21.2#
108 17.3 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 38.286 ng/ul
RT: 12.792 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 237 Resp: 22788
Ion Ratio Lower Upper
237 100
235 62.8 53.2 79.8
272 12.7 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#41
2,4,6-Trichlorophenol
Concen: 50.643 ng/ul
RT: 13.039 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion: 196 Resp: 277408
Ion Ratio Lower Upper
196 100
198 96.2 64.1 96.1#
200 31.0 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 52.225 ng/ul
RT: 13.104 min Scan# 1703
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:196 Resp: 30915
Ion Ratio Lower Upper
196 100
198 97.0 68.6 103.0
200 31.2 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#43
1,1'-Biphenyl
Concen: 48.913 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Tgt Ion:154 Resp: 1000552
Ion Ratio Lower Upper
154 100
153 40.4 28.6 43.0
76 11.7 8.5 12.7









#44
2-Chloronaphthalene
Concen: 49.391 ng/ul
RT: 13.486 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:162 Resp: 776610
Ion Ratio Lower Upper
162 100
164 32.3 27.5 41.3
127 39.9 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#45
2-Nitroaniline
Concen: 63.300 ng/ul
RT: 13.681 min Scan# 1801
Delta R.T. -0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion: 65 Resp: 291687
Ion Ratio Lower Upper
65 100
92 64.3 77.9 116.9#
138 95.8 81.5 122.3





#47
Dimethylphthalate
Concen: 52.686 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:163 Resp: 111670
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 9.8 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#48
2,6-Dinitrotoluene
Concen: 61.763 ng/ul
RT: 14.175 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:165 Resp: 240796
Ion Ratio Lower Upper
165 100
89 61.1 51.7 77.5
121 21.7 18.3 27.5





#50
Acenaphthylene
Concen: 50.710 ng/ul
RT: 14.328 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:152 Resp: 130563
Ion Ratio Lower Upper
152 100
151 20.1 18.3 27.5
153 12.8 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#51
3-Nitroaniline
Concen: 56.167 ng/ul
RT: 14.498 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:138 Resp: 227648
Ion Ratio Lower Upper
138 100
108 12.6 15.8 23.8#
92 125.9 117.1 175.7





#52
Acenaphthene
Concen: 50.074 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:153 Resp: 84414
Ion Ratio Lower Upper
153 100
152 45.7 34.3 51.5
154 90.4 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#53
2,4-Dinitrophenol
Concen: 62.029 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:184 Resp: 146177
Ion Ratio Lower Upper
184 100
63 71.2 34.6 52.0#
154 72.3 57.7 86.5





#55
4-Nitrophenol
Concen: 14.206 ng/ul
RT: 14.798 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:109 Resp: 5222
Ion Ratio Lower Upper
109 100
139 105.4 61.4 92.2
65 116.6 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#56
Dibenzofuran
Concen: 50.809 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Tgt Ion:168 Resp: 1260604
Ion Ratio Lower Upper
168 100
139 39.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 62.734 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:165 Resp: 35914
Ion Ratio Lower Upper
165 100
63 50.3 29.5 44.3
182 3.0 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 55.017 ng/ul
RT: 15.228 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:232 Resp: 296948
Ion Ratio Lower Upper
232 100
131 43.2 25.4 38.2#
130 2.3 4.2 6.2#
166 29.7 28.1 42.1





#59
Diethylphthalate
Concen: 53.746 ng/ul
RT: 15.433 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:149 Resp: 1178940
Ion Ratio Lower Upper
149 100
177 22.3 22.0 33.0
150 12.7 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#61
Fluorene
Concen: 52.131 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:166 Resp: 1065360
Ion Ratio Lower Upper
166 100
165 98.6 70.3 105.5
167 13.4 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 50.725 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:204 Resp: 55118
Ion Ratio Lower Upper
204 100
206 33.6 27.6 41.4
141 61.5 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#63
4-Nitroaniline
Concen: 65.398 ng/ul m
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:138 Resp: 265487
Ion Ratio Lower Upper
138 100
92 53.2 41.6 62.4
108 93.3 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 58.316 ng/ul
RT: 15.728 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:198 Resp: 217609
Ion Ratio Lower Upper
198 100
182 3.3 6.0 9.0
77 24.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#67
N-Nitrosodiphenylamine
Concen: 52.293 ng/ul
RT: 15.857 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Tgt Ion:169 Resp: 948610
Ion Ratio Lower Upper
169 100
168 65.6 48.3 72.5
167 34.5 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 52.330 ng/ul
RT: 16.545 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:248 Resp: 358959
Ion Ratio Lower Upper
248 100
250 98.0 79.4 119.0
141 77.9 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#69
Hexachlorobenzene
Concen: 52.658 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:284 Resp: 415345
Ion Ratio Lower Upper
284 100
142 34.5 19.2 28.8#
249 33.0 27.4 41.2





#70
Atrazine
Concen: 51.438 ng/ul m
RT: 16.816 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:200 Resp: 38514
Ion Ratio Lower Upper
200 100
173 25.1 19.2 28.8
215 48.5 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#71
Pentachlorophenol
Concen: 54.327 ng/ul
RT: 17.004 min Scan# 2366
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:266 Resp: 290047
Ion Ratio Lower Upper
266 100
264 61.4 57.3 85.9
268 62.2 65.7 98.5





#72
Phenanthrene
Concen: 54.388 ng/u1
RT: 17.398 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:178 Resp: 1839718
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 19.1 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#74
Anthracene
Concen: 54.253 ng/u1
RT: 17.492 min Scan# 2449
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:178 Resp: 1860585
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.1 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 49.239 ng/uL
RT: 13.410 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:216 Resp: 451125
Ion Ratio Lower Upper
216 100
214 78.0 50.4 75.6
218 49.3 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#76
Pentachlorobenzene
Concen: 50.124 ng/uL
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:250 Resp: 447624
Ion Ratio Lower Upper
250 100
248 62.9 43.5 65.3
252 63.9 46.5 69.7





#77
 Carbazole
 Concen: 55.505 ng/ul
 RT: 17.757 min Scan# 2494
 Delta R.T. -0.000 min
 Lab File: BP009737.D
 Acq: 09 Apr 2022 02:39




Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD




Tgt Ion:167 Resp: 1720220
 Ion Ratio Lower Upper
 167 100
 166 21.0 15.8 23.8
 139 13.2 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022




#78
 Di-n-butylphthalate
 Concen: 56.391 ng/ul
 RT: 18.316 min Scan# 2589
 Delta R.T. -0.006 min
 Lab File: BP009737.D
 Acq: 09 Apr 2022 02:39




Tgt Ion:149 Resp: 2149133
 Ion Ratio Lower Upper
 149 100
 150 8.9 9.6 14.4#
 104 5.4 6.2 9.4#





#80
Fluoranthene
Concen: 52.645 ng/ul
RT: 19.363 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:202 Resp: 236205
Ion Ratio Lower Upper
202 100
101 9.2 9.6 14.4#
100 6.9 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#82
Pyrene
Concen: 52.844 ng/ul
RT: 19.716 min Scan# 2827
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:202 Resp: 2403857
Ion Ratio Lower Upper
202 100
101 10.2 10.6 15.8#
100 8.5 9.8 14.6#





#83
Butylbenzylphthalate
Concen: 56.147 ng/ul
RT: 20.586 min Scan# 2975
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:149 Resp: 104777
Ion Ratio Lower Upper
149 100
91 70.1 71.1 106.7
206 22.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#84
3,3'-Dichlorobenzidine
Concen: 56.285 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:252 Resp: 856041
Ion Ratio Lower Upper
252 100
254 65.1 49.8 74.6
126 8.8 4.3 6.5





#85
Benzo(a)anthracene
Concen: 55.519 ng/ul
RT: 21.421 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument : BNA_P
ClientSampleId : YAZ18MSD



Tgt Ion:228 Resp: 2512009
Ion Ratio Lower Upper
228 100
229 19.7 18.2 27.2
226 26.4 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 55.913 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. -0.012 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:149 Resp: 1569463
Ion Ratio Lower Upper
149 100
167 28.2 32.8 49.2#
279 4.9 5.6 8.4#





#87
Chrysene
Concen: 55.478 ng/ul
RT: 21.480 min Scan# 3127
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:228 Resp: 2421854
Ion Ratio Lower Upper
228 100
226 29.0 22.7 34.1
229 19.5 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#89
Di-n-octyl phthalate
Concen: 51.021 ng/ul
RT: 22.310 min Scan# 3268
Delta R.T. -0.012 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:149 Resp: 2739496





#90
Benzo(b)fluoranthene
Concen: 54.426 ng/ul
RT: 23.162 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:252 Resp: 2771190
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 7.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#91
Benzo(k)fluoranthene
Concen: 53.797 ng/ul
RT: 23.210 min Scan# 3421
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39


Tgt Ion:252 Resp: 2499541
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 7.9 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 55.831 ng/ul
RT: 23.809 min Scan# 3522
Delta R.T. -0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:252 Resp: 2635062
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 8.8 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 63.282 ng/ul
RT: 26.515 min Scan# 3983
Delta R.T. -0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Tgt Ion:276 Resp: 3123268
Ion Ratio Lower Upper
276 100
138 18.4 12.3 18.5
277 25.0 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 62.855 ng/ul
RT: 26.539 min Scan# 3986
Delta R.T. 0.006 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:278 Resp: 2689940
Ion Ratio Lower Upper
278 100
139 12.4 6.6 10.0
279 23.9 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#96
Benzo(g,h,i)perylene
Concen: 65.481 ng/ul
RT: 27.309 min Scan# 4118
Delta R.T. 0.000 min
Lab File: BP009737.D
Acq: 09 Apr 2022 02:39

Tgt Ion:276 Resp: 2664728
Ion Ratio Lower Upper
276 100
138 16.8 11.6 17.4
277 23.6 15.9 23.9

