

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040822\
 Data File : BP009741.D
 Acq On : 09 Apr 2022 04:55
 Operator : CG/JU
 Sample : PB143885BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS885

Quant Time: Apr 09 05:53:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	91448	20.000	ng/ul	0.00
20) Naphthalene-d8	10.781	136	401550	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.604	164	287121	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.357	188	639668	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.433	240	716920	20.000	ng/ul	#-0.01
88) Perylene-d12	23.910	264	737724	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	13986m	6.826	ng/uL	0.00
4) Pyridine-d5	3.805	84	195733	34.646	ng/ul	0.00
7) Phenol-d5	7.128	99	281353	35.044	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	173464	36.292	ng/ul	0.00
11) 2-Chlorophenol-d4	7.499	132	219066	36.173	ng/ul	0.00
15) 4-Methylphenol-d8	8.675	113	242569	36.193	ng/ul	0.00
21) Nitrobenzene-d5	9.134	128	115159	39.769	ng/ul	0.00
24) 2-Nitrophenol-d4	9.857	143	129441	40.773	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	244043	36.736	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	314741	33.264	ng/ul	0.00
46) Dimethylphthalate-d6	14.010	166	835753	37.056	ng/ul	-0.01
49) Acenaphthylene-d8	14.298	160	926722	34.683	ng/ul	0.00
54) 4-Nitrophenol-d4	14.787	143	158578	38.770	ng/ul	0.00
60) Fluorene-d10	15.598	176	700372	36.914	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	149637	39.488	ng/ul	0.00
73) Anthracene-d10	17.457	188	1124512	36.717	ng/ul	0.00
81) Pyrene-d10	19.680	212	1432166	36.577	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.751	264	1481102	38.552	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	26488	12.681	ng/uL#	18
5) Pyridine	3.822	79	193600	32.844	ng/ul#	37
6) Benzaldehyde	7.105	77	186429m	37.863	ng/ul	
8) Phenol	7.152	94	276316	34.383	ng/ul#	88
10) Bis(2-Chloroethyl)ether	7.393	93	212433	33.935	ng/ul#	79
12) 2-Chlorophenol	7.534	128	212456	34.797	ng/ul	95
13) 2-Methylphenol	8.410	108	214004	34.150	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.510	45	322819	34.887	ng/ul#	84
16) Acetophenone	8.793	105	354516	33.573	ng/ul#	80
17) N-Nitroso-di-n-propyla...	8.787	70	190256	34.753	ng/ul#	72
18) 4-Methylphenol	8.740	108	234399	34.540	ng/ul	93
19) Hexachloroethane	9.063	117	89932	34.887	ng/ul#	80
22) Nitrobenzene	9.175	77	272770	36.699	ng/ul#	85
23) Isophorone	9.704	82	531167	34.681	ng/ul	98
25) 2-Nitrophenol	9.887	139	128232	37.871	ng/ul#	84
26) 2,4-Dimethylphenol	9.951	107	271169	33.747	ng/ul#	81
27) Bis(2-Chloroethoxy)met...	10.187	93	304048	33.813	ng/ul#	98
29) 2,4-Dichlorophenol	10.422	162	228375	35.459	ng/ul#	85
30) Naphthalene	10.834	128	705424	34.158	ng/ul	98
32) 4-Chloroaniline	10.934	127	295336	31.542	ng/ul	97
33) Hexachlorobutadiene	11.128	225	161223	34.757	ng/ul	96
34) Caprolactam	11.704	113	83531m	39.218	ng/ul	
35) 4-Chloro-3-methylphenol	12.051	107	278807	35.727	ng/ul#	71
36) 2-Methylnaphthalene	12.440	142	510808	33.856	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040822\
 Data File : BP009741.D
 Acq On : 09 Apr 2022 04:55
 Operator : CG/JU
 Sample : PB143885BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS885


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022

Quant Time: Apr 09 05:53:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.657	142	518737	34.051	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.804	216	301569	34.330	ng/ul	93
40) Hexachlorocyclopentadiene	12.793	237	182648	29.962	ng/ul	96
41) 2,4,6-Trichlorophenol	13.040	196	202293	36.060	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.110	196	226322m	37.331	ng/ul	
43) 1,1'-Biphenyl	13.440	154	717736	34.260	ng/ul	93
44) 2-Chloronaphthalene	13.481	162	551314	34.236	ng/ul	94
45) 2-Nitroaniline	13.675	65	192088	40.703	ng/ul#	79
47) Dimethylphthalate	14.057	163	757192	34.882	ng/ul#	93
48) 2,6-Dinitrotoluene	14.175	165	158236	39.630	ng/ul	96
50) Acenaphthylene	14.328	152	910647	34.535	ng/ul	96
51) 3-Nitroaniline	14.498	138	156582	37.722	ng/ul#	82
52) Acenaphthene	14.669	153	592184	34.300	ng/ul	96
53) 2,4-Dinitrophenol	14.698	184	90510	37.502	ng/ul#	75
55) 4-Nitrophenol	14.798	109	146667m	38.953	ng/ul	
56) Dibenzofuran	15.004	168	873590	34.380	ng/ul	96
57) 2,4-Dinitrotoluene	14.957	165	235966	40.246	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.228	232	202369	36.610	ng/ul#	85
59) Diethylphthalate	15.428	149	797629	35.505	ng/ul#	92
61) Fluorene	15.651	166	720808	34.440	ng/ul	89
62) 4-Chlorophenyl-phenyle...	15.651	204	380767	34.215	ng/ul	97
63) 4-Nitroaniline	15.663	138	180748m	43.474	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.722	198	139338	36.992	ng/ul#	89
67) N-Nitrosodiphenylamine	15.857	169	639558	34.928	ng/ul	93
68) 4-Bromophenyl-phenylether	16.545	248	242571	35.033	ng/ul	95
69) Hexachlorobenzene	16.663	284	280171	35.189	ng/ul#	89
70) Atrazine	16.816	200	268764	35.560	ng/ul	92
71) Pentachlorophenol	17.004	266	189283	35.122	ng/ul#	84
72) Phenanthrene	17.398	178	1217450	35.656	ng/ul	98
74) Anthracene	17.492	178	1223655	35.348	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.410	216	322919	34.917	ng/uL#	86
76) Pentachlorobenzene	14.928	250	310165	34.408	ng/uL	90
77) Carbazole	17.751	167	1129480	36.104	ng/ul#	95
78) Di-n-butylphthalate	18.310	149	1428699	37.138	ng/ul#	93
80) Fluoranthene	19.357	202	1567165	34.773	ng/ul#	95
82) Pyrene	19.710	202	1594050	34.885	ng/ul#	92
83) Butylbenzylphthalate	20.580	149	683259	36.450	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.345	252	565454	37.013	ng/ul#	96
85) Benzo(a)anthracene	21.421	228	1643635	36.164	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.351	149	1023224	36.290	ng/ul#	82
87) Chrysene	21.474	228	1589739	36.254	ng/ul	99
89) Di-n-octyl phthalate	22.304	149	1744973	33.760	ng/ul	100
90) Benzo(b)fluoranthene	23.151	252	1767448	36.060	ng/ul	99
91) Benzo(k)fluoranthene	23.204	252	1610049	35.998	ng/ul	99
93) Benzo(a)pyrene	23.804	252	1676992	36.911	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.504	276	1902125	40.036	ng/ul#	96
95) Dibenzo(a,h)anthracene	26.527	278	1638182	39.765	ng/ul#	97
96) Benzo(g,h,i)perylene	27.298	276	1568445	40.038	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.681 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion: 88 Resp: 26488
Ion Ratio Lower Upper
88 100
43 37.8 6.8 10.2
58 83.6 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#5
Pyridine
Concen: 32.844 ng/uL
RT: 3.822 min Scan# 125
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Tgt Ion: 79 Resp: 193600
Ion Ratio Lower Upper
79 100
52 94.9 35.5 53.3#
51 44.8 20.2 30.4#





#6
Benzaldehyde
Concen: 37.863 ng/ul m
RT: 7.105 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion: 77 Resp: 186429
Ion Ratio Lower Upper
77 100
105 94.4 77.0 115.4
106 87.7 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#8
Phenol
Concen: 34.383 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Tgt Ion: 94 Resp: 276316
Ion Ratio Lower Upper
94 100
65 33.4 19.0 28.6#
66 47.8 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 33.935 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion: 93 Resp: 212438
Ion Ratio Lower Upper
93 100
63 80.0 45.0 67.4
95 32.6 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#12
2-Chlorophenol
Concen: 34.797 ng/ul
RT: 7.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:128 Resp: 212456
Ion Ratio Lower Upper
128 100
64 52.6 45.3 67.9
130 33.2 29.0 43.6





#13
2-Methylphenol
Concen: 34.150 ng/ul
RT: 8.410 min Scan# 905
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:108 Resp: 214004
Ion Ratio Lower Upper
108 100
107 89.4 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 34.887 ng/ul
RT: 8.510 min Scan# 922
Delta R.T. 0.000 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55



Tgt Ion: 45 Resp: 322819
Ion Ratio Lower Upper
45 100
77 13.4 17.1 25.7#
79 9.9 13.2 19.8#









#16
Acetophenone
Concen: 33.573 ng/ul
RT: 8.793 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion: 105 Resp: 354510
Ion Ratio Lower Upper
105 100
77 80.1 53.4 80.2
51 32.8 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#17
N-Nitroso-di-n-propylamine
Concen: 34.753 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion: 70 Resp: 190256
Ion Ratio Lower Upper
70 100
42 65.2 31.8 47.8#
101 10.7 15.6 23.4#
130 20.4 13.7 20.5









#18
4-Methylphenol
Concen: 34.540 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:108 Resp: 234399
Ion Ratio Lower Upper
108 100
107 111.7 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#19
Hexachloroethane
Concen: 34.887 ng/ul
RT: 9.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:117 Resp: 89932
Ion Ratio Lower Upper
117 100
201 103.0 106.5 159.7#
199 66.0 59.0 88.6





#22
Nitrobenzene
Concen: 36.699 ng/ul
RT: 9.175 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion: 77 Resp: 272770
Ion Ratio Lower Upper
77 100
123 43.0 42.3 63.5
65 15.2 5.8 8.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#23
Isophorone
Concen: 34.681 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion: 82 Resp: 531167
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 16.3 12.9 19.3









#25
2-Nitrophenol
Concen: 37.871 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:139 Resp: 12823
Ion Ratio Lower Upper
139 100
65 57.6 49.4 74.2
109 36.1 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#26
2,4-Dimethylphenol
Concen: 33.747 ng/ul
RT: 9.951 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:107 Resp: 271169
Ion Ratio Lower Upper
107 100
121 46.3 34.1 51.1
122 79.1 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 33.813 ng/ul
RT: 10.187 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion: 93 Resp: 304048
Ion Ratio Lower Upper
93 100
95 32.7 26.2 39.4
123 11.1 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#29
2,4-Dichlorophenol
Concen: 35.459 ng/ul
RT: 10.422 min Scan# 1247
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:162 Resp: 228375
Ion Ratio Lower Upper
162 100
164 63.7 37.0 55.6#
98 37.8 31.0 46.6









#30
Naphthalene
Concen: 34.158 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:128 Resp: 705424
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.7 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#32
4-Chloroaniline
Concen: 31.542 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:127 Resp: 295336
Ion Ratio Lower Upper
127 100
129 32.6 24.6 37.0





#33
Hexachlorobutadiene
Concen: 34.757 ng/ul
RT: 11.128 min Scan# 1367
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:225 Resp: 16122
Ion Ratio Lower Upper
225 100
223 63.3 53.4 80.0
227 63.7 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#34
Caprolactam
Concen: 39.218 ng/ul m
RT: 11.704 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:113 Resp: 83531
Ion Ratio Lower Upper
113 100
55 176.7 58.2 87.2#
56 146.3 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 35.727 ng/ul
RT: 12.051 min Scan# 1524
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:107 Resp: 27880
Ion Ratio Lower Upper
107 100
144 24.4 15.8 23.8
142 73.5 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#36
2-Methylnaphthalene
Concen: 33.856 ng/ul
RT: 12.440 min Scan# 1590
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:142 Resp: 510808
Ion Ratio Lower Upper
142 100
141 88.8 64.3 96.5





#37
1-Methylnaphthalene
Concen: 34.051 ng/ul
RT: 12.657 min Scan# 1627
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:142 Resp: 51873
Ion Ratio Lower Upper
142 100
141 93.1 69.8 104.8
116 4.4 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.330 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:216 Resp: 301569
Ion Ratio Lower Upper
216 100
214 79.1 58.2 87.2
179 20.6 14.2 21.2
108 17.0 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 29.962 ng/ul
RT: 12.793 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:237 Resp: 182648
Ion Ratio Lower Upper
237 100
235 63.0 53.2 79.8
272 12.5 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#41
2,4,6-Trichlorophenol
Concen: 36.060 ng/ul
RT: 13.040 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55



Tgt Ion:196 Resp: 202293
Ion Ratio Lower Upper
196 100
198 94.6 64.1 96.1
200 30.5 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 37.331 ng/ul m
RT: 13.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:196 Resp: 22632
Ion Ratio Lower Upper
196 100
198 95.8 68.6 103.0
200 30.8 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#43
1,1'-Biphenyl
Concen: 34.260 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:154 Resp: 717736
Ion Ratio Lower Upper
154 100
153 40.8 28.6 43.0
76 11.7 8.5 12.7









#44
2-Chloronaphthalene
Concen: 34.236 ng/ul
RT: 13.481 min Scan# 1768
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:162 Resp: 551314
Ion Ratio Lower Upper
162 100
164 32.0 27.5 41.3
127 41.5 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#45
2-Nitroaniline
Concen: 40.703 ng/ul
RT: 13.675 min Scan# 1800
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion: 65 Resp: 192088
Ion Ratio Lower Upper
65 100
92 63.9 77.9 116.9#
138 92.4 81.5 122.3









#47
Dimethylphthalate
Concen: 34.882 ng/ul
RT: 14.057 min Scan# 1865
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:163 Resp: 757192
Ion Ratio Lower Upper
163 100
194 4.0 4.1 6.1
164 10.6 11.3 16.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#48
2,6-Dinitrotoluene
Concen: 39.630 ng/ul
RT: 14.175 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:165 Resp: 158236
Ion Ratio Lower Upper
165 100
89 62.0 51.7 77.5
121 20.9 18.3 27.5









#50
Acenaphthylene
Concen: 34.535 ng/ul
RT: 14.328 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
Client Sample Id :
SLCS885

Tgt Ion:152 Resp: 91064
Ion Ratio Lower Upper
152 100
151 20.4 18.3 27.5
153 12.8 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#51
3-Nitroaniline
Concen: 37.722 ng/ul
RT: 14.498 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:138 Resp: 156582
Ion Ratio Lower Upper
138 100
108 12.3 15.8 23.8#
92 123.4 117.1 175.7









#52
Acenaphthene
Concen: 34.300 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:153 Resp: 592184
Ion Ratio Lower Upper
153 100
152 47.1 34.3 51.5
154 89.9 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#53
2,4-Dinitrophenol
Concen: 37.502 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:184 Resp: 90510
Ion Ratio Lower Upper
184 100
63 78.5 34.6 52.0#
154 78.6 57.7 86.5





#55
4-Nitrophenol
Concen: 38.953 ng/ul m
RT: 14.798 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:109 Resp: 14666
Ion Ratio Lower Upper
109 100
139 104.0 61.4 92.2
65 124.6 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#56
Dibenzofuran
Concen: 34.380 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:168 Resp: 873590
Ion Ratio Lower Upper
168 100
139 40.1 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 40.246 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:165 Resp: 235960
Ion Ratio Lower Upper
165 100
63 50.0 29.5 44.3#
182 3.1 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 36.610 ng/ul
RT: 15.228 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Tgt Ion:232 Resp: 202369
Ion Ratio Lower Upper
232 100
131 43.9 25.4 38.2#
130 2.3 4.2 6.2#
166 29.3 28.1 42.1





#59
Diethylphthalate
Concen: 35.505 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:149 Resp: 797629
Ion Ratio Lower Upper
149 100
177 21.6 22.0 33.0
150 12.7 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#61
Fluorene
Concen: 34.440 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:166 Resp: 720808
Ion Ratio Lower Upper
166 100
165 98.6 70.3 105.5
167 13.7 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 34.215 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:204 Resp: 38076
Ion Ratio Lower Upper
204 100
206 32.9 27.6 41.4
141 60.8 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#63
4-Nitroaniline
Concen: 43.474 ng/ul m
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55



Tgt Ion:138 Resp: 180748
Ion Ratio Lower Upper
138 100
92 52.9 41.6 62.4
108 96.6 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 36.992 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:198 Resp: 139338
Ion Ratio Lower Upper
198 100
182 4.1 6.0 9.0
77 27.4 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#67
N-Nitrosodiphenylamine
Concen: 34.928 ng/ul
RT: 15.857 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:169 Resp: 639558
Ion Ratio Lower Upper
169 100
168 67.1 48.3 72.5
167 35.3 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 35.033 ng/ul
RT: 16.545 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:248 Resp: 24257
Ion Ratio Lower Upper
248 100
250 97.4 79.4 119.0
141 76.5 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#69
Hexachlorobenzene
Concen: 35.189 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:284 Resp: 280171
Ion Ratio Lower Upper
284 100
142 33.9 19.2 28.8#
249 31.8 27.4 41.2





#70
Atrazine
Concen: 35.560 ng/ul
RT: 16.816 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55



Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:200 Resp: 268764
Ion Ratio Lower Upper
200 100
173 24.7 19.2 28.8
215 47.6 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#71
Pentachlorophenol
Concen: 35.122 ng/ul
RT: 17.004 min Scan# 2366
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Tgt Ion:266 Resp: 189283
Ion Ratio Lower Upper
266 100
264 61.3 57.3 85.9
268 65.6 65.7 98.5#





#72
Phenanthrene
Concen: 35.656 ng/u1
RT: 17.398 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885




Tgt Ion:178 Resp: 1217450
Ion Ratio Lower Upper
178 100
179 15.7 11.7 17.5
176 19.1 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022




#74
Anthracene
Concen: 35.348 ng/u1
RT: 17.492 min Scan# 2449
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55



Tgt Ion:178 Resp: 1223655
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.0 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 34.917 ng/uL

RT: 13.410 min Scan# 1755

Delta R.T. -0.006 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Instrument :

BNA_P

ClientSampleId :

SLCS885

Tgt Ion:216 Resp: 322919

Ion Ratio Lower Upper

216 100

214 76.9 50.4 75.6





218 48.9 35.2 52.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022

Supervised By :Yogesh Patel 04/12/2022



#76

Pentachlorobenzene

Concen: 34.408 ng/uL

RT: 14.928 min Scan# 2013

Delta R.T. -0.006 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55




Tgt Ion:250 Resp: 310165


Ion Ratio Lower Upper

250 100

248 63.0 43.5 65.3


252 63.8 46.5 69.7





#77
 Carbazole
 Concen: 36.104 ng/ul
 RT: 17.751 min Scan# 2493
 Delta R.T. -0.006 min
 Lab File: BP009741.D
 Acq: 09 Apr 2022 04:55




Instrument :
 BNA_P
 ClientSampleId :
 SLCS885



Tgt Ion:167 Resp: 1129480
 Ion Ratio Lower Upper
 167 100
 166 21.5 15.8 23.8
 139 13.1 8.3 12.5




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
 Supervised By :Yogesh Patel 04/12/2022



#78
 Di-n-butylphthalate
 Concen: 37.138 ng/ul
 RT: 18.310 min Scan# 2588
 Delta R.T. -0.012 min
 Lab File: BP009741.D
 Acq: 09 Apr 2022 04:55

Tgt Ion:149 Resp: 1428699
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.7 6.2 9.4#









#80
Fluoranthene
Concen: 34.773 ng/u1
RT: 19.357 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:202 Resp: 156716
Ion Ratio Lower Upper
202 100
101 9.1 9.6 14.4#
100 7.3 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#82
Pyrene
Concen: 34.885 ng/u1
RT: 19.710 min Scan# 2826
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:202 Resp: 1594050
Ion Ratio Lower Upper
202 100
101 10.3 10.6 15.8#
100 8.3 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 36.450 ng/ul

RT: 20.580 min Scan# 2975

Delta R.T. -0.012 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Instrument :

BNA_P

ClientSampleId :





SLCS885

Tgt Ion:	149	Resp:	68325
Ion Ratio	Lower	Upper	
149	100		
91	73.4	71.1	106.7
206	21.8	28.8	43.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#84

3,3'-Dichlorobenzidine

Concen: 37.013 ng/ul




RT: 21.345 min Scan# 3104


Delta R.T. -0.012 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Tgt Ion:	252	Resp:	565454
Ion Ratio	Lower	Upper	
252	100		
254	64.8	49.8	74.6
126	9.8	4.3	6.5





#85

Benzo(a)anthracene

Concen: 36.164 ng/ul

RT: 21.421 min Scan# 3117

Delta R.T. -0.006 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Instrument :

BNA_P

ClientSampleId :

SLCS885

Tgt Ion:228 Resp: 164363

Ion Ratio Lower Upper

228 100

229 19.5 18.2 27.2





226 25.9 22.8 34.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022

Supervised By :Yogesh Patel 04/12/2022



#86

Bis(2-ethylhexyl)phthalate

Concen: 36.290 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. -0.012 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55




Tgt Ion:149 Resp: 1023224


Ion Ratio Lower Upper

149 100

167 28.1 32.8 49.2#


279 5.3 5.6 8.4#





#87
Chrysene
Concen: 36.254 ng/ul
RT: 21.474 min Scan# 3127
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:228 Resp: 1589739
Ion Ratio Lower Upper
228 100
226 29.0 22.7 34.1
229 19.4 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#89
Di-n-octyl phthalate
Concen: 33.760 ng/ul
RT: 22.304 min Scan# 3267
Delta R.T. -0.018 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55


Tgt Ion:149 Resp: 1744973





#90
Benzo(b)fluoranthene
Concen: 36.060 ng/ul
RT: 23.151 min Scan# 3412
Delta R.T. -0.018 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55




Instrument :
BNA_P
ClientSampleId :
SLCS885



Tgt Ion:252 Resp: 1767448
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 8.0 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#91
Benzo(k)fluoranthene
Concen: 35.998 ng/ul
RT: 23.204 min Scan# 3420
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:252 Resp: 1610049
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 7.6 5.1 7.7





#93

Benzo(a)pyrene

Concen: 36.911 ng/ul

RT: 23.804 min Scan# 3522

Delta R.T. -0.012 min

Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Instrument :

BNA_P

ClientSampleId :

SLCS885





Tgt Ion:252 Resp: 1676992

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	8.8	6.4	9.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 40.036 ng/ul

RT: 26.504 min Scan# 3981




Delta R.T. -0.012 min


Lab File: BP009741.D

Acq: 09 Apr 2022 04:55

Tgt Ion:276 Resp: 1902125

Ion	Ratio	Lower	Upper
276	100		
138	18.8	12.3	18.5
277	25.1	19.3	28.9









#95
Dibenzo(a,h)anthracene
Concen: 39.765 ng/ul
RT: 26.527 min Scan# 3986
Delta R.T. -0.006 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Instrument :
BNA_P
ClientSampleId :
SLCS885

Tgt Ion:278 Resp: 163818
Ion Ratio Lower Upper
278 100
139 11.5 6.6 10.0
279 23.6 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/11/2022
Supervised By :Yogesh Patel 04/12/2022



#96
Benzo(g,h,i)perylene
Concen: 40.038 ng/ul
RT: 27.298 min Scan# 4116
Delta R.T. -0.012 min
Lab File: BP009741.D
Acq: 09 Apr 2022 04:55

Tgt Ion:276 Resp: 1568445
Ion Ratio Lower Upper
276 100
138 16.5 11.6 17.4
277 23.7 15.9 23.9

