

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP041025\  
 Data File : BP024248.D  
 Acq On : 10 Apr 2025 11:54  
 Operator : RC/JU  
 Sample : Q1745-04  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 IB-6A-WC

Quant Time: Apr 10 12:25:59 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP031225.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Apr 10 11:43:37 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min)  |
|-----------------------------|--------|------|----------|---------|-------|-----------|
| -----                       |        |      |          |         |       |           |
| Internal Standards          |        |      |          |         |       |           |
| 1) 1,4-Dichlorobenzene-d4   | 7.740  | 152  | 273317   | 20.000  | ng    | 0.00      |
| 21) Naphthalene-d8          | 10.516 | 136  | 1088733  | 20.000  | ng    | -0.01     |
| 39) Acenaphthene-d10        | 14.375 | 164  | 660298   | 20.000  | ng    | -0.02     |
| 64) Phenanthrene-d10        | 17.175 | 188  | 1305451  | 20.000  | ng    | -0.02     |
| 76) Chrysene-d12            | 21.627 | 240  | 1426132  | 20.000  | ng    | -0.02     |
| 86) Perylene-d12            | 24.986 | 264  | 1638352  | 20.000  | ng    | -0.02     |
| System Monitoring Compounds |        |      |          |         |       |           |
| 5) 2-Fluorophenol           | 5.358  | 112  | 2225506  | 129.979 | ng    | 0.00      |
| 7) Phenol-d6                | 6.922  | 99   | 2623928  | 114.599 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 8.887  | 82   | 1789182  | 92.720  | ng    | 0.00      |
| 42) 2,4,6-Tribromophenol    | 15.892 | 330  | 1417278  | 170.401 | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 12.987 | 172  | 3902260  | 87.142  | ng    | -0.01     |
| 79) Terphenyl-d14           | 19.904 | 244  | 6493750  | 93.992  | ng    | -0.01     |
| Target Compounds            |        |      |          |         |       |           |
| 60) Diethylphthalate        | 15.216 | 149  | 110941   | 2.411   | ng    | Qvalue 99 |
| -----                       |        |      |          |         |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP041025\  
Data File : BP024248.D  
Acq On : 10 Apr 2025 11:54  
Operator : RC/JU  
Sample : Q1745-04  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
IB-6A-WC

Quant Time: Apr 10 12:25:59 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP031225.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Apr 10 11:43:37 2025  
Response via : Initial Calibration

