

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041522\
 Data File : BP009855.D
 Acq On : 15 Apr 2022 11:42
 Operator : CG/JU
 Sample : N2373-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BFFD6

Quant Time: Apr 18 01:00:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 00:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	129498	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	584159	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.592	164	429731	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.339	188	970991	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.421	240	1006561	20.000	ng/u1	0.00
88) Perylene-d12	23.886	264	910429	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	14564	5.019	ng/uL	0.00
4) Pyridine-d5	3.805	84	78618	9.827	ng/u1	0.00
7) Phenol-d5	7.110	99	67670	5.952	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	208891	30.863	ng/u1	0.00
11) 2-Chlorophenol-d4	7.487	132	195194	22.761	ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	138265	14.568	ng/u1	0.00
21) Nitrobenzene-d5	9.116	128	137253	32.582	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	141598	30.659	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	246783	25.535	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	361637	26.273	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	1076240	31.883	ng/u1	0.00
49) Acenaphthylene-d8	14.281	160	1156616	28.922	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	34595	5.651	ng/u1	0.00
60) Fluorene-d10	15.581	176	962718	33.902	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	176674	30.714	ng/u1	0.00
73) Anthracene-d10	17.439	188	1712255	36.830	ng/u1	0.00
81) Pyrene-d10	19.669	212	2113612	38.448	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.727	264	1865676	39.350	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041522\
Data File : BP009855.D
Acq On : 15 Apr 2022 11:42
Operator : CG/JU
Sample : N2373-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BFFD6

Quant Time: Apr 18 01:00:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Apr 18 00:59:30 2022
Response via : Initial Calibration

