

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041522\
 Data File : BP009866.D
 Acq On : 15 Apr 2022 17:52
 Operator : CG/JU
 Sample : PB144122BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS122

Quant Time: Apr 18 01:03:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 00:59:30 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/18/2022
 Supervised By :Yogesh Patel 04/18/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	156632	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	725787	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.593	164	521900	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	1178541	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.427	240	1229834	20.000	ng/u1	# 0.00
88) Perylene-d12	23.886	264	1133038	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	17848m	5.085	ng/uL	0.00
4) Pyridine-d5	3.799	84	262716	27.150	ng/u1	0.00
7) Phenol-d5	7.111	99	392541	28.546	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	246122	30.064	ng/u1	0.00
11) 2-Chlorophenol-d4	7.487	132	310100	29.895	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	344554	30.015	ng/u1	0.00
21) Nitrobenzene-d5	9.116	128	163796	31.295	ng/u1	0.00
24) 2-Nitrophenol-d4	9.840	143	182572	31.817	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	348257	29.003	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	456388	26.686	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	1184805	28.901	ng/u1	0.00
49) Acenaphthylene-d8	14.281	160	1337319	27.535	ng/u1	0.00
54) 4-Nitrophenol-d4	14.775	143	218895	29.442	ng/u1	0.00
60) Fluorene-d10	15.581	176	1012592	29.361	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	218311	31.269	ng/u1	0.00
73) Anthracene-d10	17.439	188	1635957	28.992	ng/u1	0.00
81) Pyrene-d10	19.669	212	2036722	30.323	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	1839492	31.175	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	38359	10.722	ng/uL#	32
5) Pyridine	3.817	79	278741	27.609	ng/u1#	48
6) Benzaldehyde	7.093	77	263353	31.227	ng/u1	98
8) Phenol	7.140	94	417570	30.336	ng/u1#	91
10) Bis(2-Chloroethyl)ether	7.375	93	323093	30.134	ng/u1#	78
12) 2-Chlorophenol	7.522	128	319805	30.581	ng/u1	92
13) 2-Methylphenol	8.399	108	319935	29.807	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.493	45	478569	30.196	ng/u1#	85
16) Acetophenone	8.781	105	536482	29.662	ng/u1#	75
17) N-Nitroso-di-n-propyla...	8.775	70	286692	30.575	ng/u1#	74
18) 4-Methylphenol	8.728	108	357369	30.745	ng/u1	93
19) Hexachloroethane	9.046	117	131545	29.793	ng/u1	85
22) Nitrobenzene	9.158	77	415641	30.939	ng/u1#	82
23) Isophorone	9.693	82	803770	29.035	ng/u1	97
25) 2-Nitrophenol	9.875	139	197717	32.306	ng/u1#	87
26) 2,4-Dimethylphenol	9.934	107	413190	28.450	ng/u1#	81
27) Bis(2-Chloroethoxy)met...	10.169	93	470849	28.970	ng/u1#	97
29) 2,4-Dichlorophenol	10.410	162	349894	30.057	ng/u1#	84
30) Naphthalene	10.816	128	1087334	29.130	ng/u1	97
32) 4-Chloroaniline	10.922	127	455142	26.893	ng/u1	97
33) Hexachlorobutadiene	11.110	225	247866	29.564	ng/u1	93
34) Caprolactam	11.693	113	125298m	32.547	ng/u1	
35) 4-Chloro-3-methylphenol	12.040	107	425333	30.154	ng/u1#	70
36) 2-Methylnaphthalene	12.422	142	801786	29.402	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041522\
 Data File : BP009866.D
 Acq On : 15 Apr 2022 17:52
 Operator : CG/JU
 Sample : PB144122BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS122

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/18/2022
 Supervised By :Yogesh Patel 04/18/2022

Quant Time: Apr 18 01:03:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 00:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.640	142	806422	29.287	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.793	216	471552	29.532	ng/ul	95
40) Hexachlorocyclopentadiene	12.775	237	270619	24.423	ng/ul	95
41) 2,4,6-Trichlorophenol	13.022	196	310914	30.490	ng/ul#	84
42) 2,4,5-Trichlorophenol	13.093	196	346419	31.436	ng/ul#	87
43) 1,1'-Biphenyl	13.428	154	1121906	29.462	ng/ul	94
44) 2-Chloronaphthalene	13.469	162	868347	29.666	ng/ul	93
45) 2-Nitroaniline	13.663	65	301148	35.106	ng/ul#	80
47) Dimethylphthalate	14.046	163	1175825	29.800	ng/ul#	92
48) 2,6-Dinitrotoluene	14.163	165	247573	34.112	ng/ul	91
50) Acenaphthylene	14.310	152	1419939	29.625	ng/ul	96
51) 3-Nitroaniline	14.487	138	243140	32.225	ng/ul#	82
52) Acenaphthene	14.651	153	918408	29.265	ng/ul	98
53) 2,4-Dinitrophenol	14.687	184	136244	31.056	ng/ul#	80
55) 4-Nitrophenol	14.793	109	205477	30.023	ng/ul#	51
56) Dibenzofuran	14.987	168	1362230	29.494	ng/ul	98
57) 2,4-Dinitrotoluene	14.945	165	369372	34.659	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.216	232	315456	31.396	ng/ul#	87
59) Diethylphthalate	15.410	149	1231156	30.149	ng/ul	93
61) Fluorene	15.640	166	1133408	29.792	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.634	204	600048	29.664	ng/ul	99
63) 4-Nitroaniline	15.651	138	267313m	35.372	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.710	198	217262	31.307	ng/ul#	89
67) N-Nitrosodiphenylamine	15.845	169	993058	29.436	ng/ul	95
68) 4-Bromophenyl-phenylether	16.528	248	382635	29.994	ng/ul	96
69) Hexachlorobenzene	16.651	284	438942	29.923	ng/ul#	92
70) Atrazine	16.798	200	410935	29.510	ng/ul#	90
71) Pentachlorophenol	16.992	266	287573	28.962	ng/ul#	85
72) Phenanthrene	17.387	178	1897853	30.168	ng/ul	98
74) Anthracene	17.475	178	1910120	29.948	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.393	216	505399	29.661	ng/uL#	86
76) Pentachlorobenzene	14.910	250	489867	29.495	ng/uL	92
77) Carbazole	17.739	167	1726225	29.949	ng/ul#	95
78) Di-n-butylphthalate	18.292	149	2188992	30.884	ng/ul#	93
80) Fluoranthene	19.345	202	2384941	30.848	ng/ul#	95
82) Pyrene	19.698	202	2430352	31.005	ng/ul#	94
83) Butylbenzylphthalate	20.569	149	1032025	32.094	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.333	252	783463	29.895	ng/ul#	97
85) Benzo(a)anthracene	21.410	228	2401468	30.802	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.333	149	1536868	31.774	ng/ul#	82
87) Chrysene	21.463	228	2328616	30.956	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	2604554	32.809	ng/ul	100
90) Benzo(b)fluoranthene	23.139	252	2370594	31.491	ng/ul	98
91) Benzo(k)fluoranthene	23.186	252	2265274m	32.977	ng/ul	
93) Benzo(a)pyrene	23.780	252	2226599	31.909	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.468	276	2283775	31.298	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.486	278	1976554	31.239	ng/ul#	95
96) Benzo(g,h,i)perylene	27.262	276	1852322	30.787	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041522\
 Data File : BP009866.D
 Acq On : 15 Apr 2022 17:52
 Operator : CG/JU
 Sample : PB144122BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS122

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/18/2022
 Supervised By :Yogesh Patel 04/18/2022


Quant Time: Apr 18 01:03:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 18 00:59:30 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 10.722 ng/uL
RT: 3.417 min Scan# 50
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 88 Resp: 38359
Ion Ratio Lower Upper
88 100
43 34.4 6.8 10.2
58 74.4 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#5
Pyridine
Concen: 27.609 ng/uL
RT: 3.817 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52



Tgt Ion: 79 Resp: 278741
Ion Ratio Lower Upper
79 100
52 86.2 35.5 53.3#
51 40.8 20.2 30.4#









#6
Benzaldehyde
Concen: 31.227 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
Client Sample Id :
SLCS122

Tgt Ion: 77 Resp: 26335
Ion Ratio Lower Upper
77 100
105 97.0 77.0 115.4
106 92.7 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#8
Phenol
Concen: 30.336 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion: 94 Resp: 417570
Ion Ratio Lower Upper
94 100
65 32.6 19.0 28.6
66 46.1 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 30.134 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 93 Resp: 323093
Ion Ratio Lower Upper
93 100
63 82.0 45.0 67.4
95 32.6 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#12
2-Chlorophenol
Concen: 30.581 ng/ul
RT: 7.522 min Scan# 754
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:128 Resp: 319805
Ion Ratio Lower Upper
128 100
64 50.7 45.3 67.9
130 31.6 29.0 43.6





#13
2-Methylphenol
Concen: 29.807 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:108 Resp: 31993
Ion Ratio Lower Upper
108 100
107 89.6 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 30.196 ng/ul
RT: 8.493 min Scan# 919
Delta R.T. -0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion: 45 Resp: 478569
Ion Ratio Lower Upper
45 100
77 14.3 17.1 25.7#
79 10.2 13.2 19.8#





#16
Acetophenone
Concen: 29.662 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 105 Resp: 53648
Ion Ratio Lower Upper
105 100
77 84.5 53.4 80.2
51 32.3 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#17
N-Nitroso-di-n-propylamine
Concen: 30.575 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion: 70 Resp: 286692
Ion Ratio Lower Upper
70 100
42 62.1 31.8 47.8#
101 10.6 15.6 23.4#
130 20.3 13.7 20.5





#18
4-Methylphenol
Concen: 30.745 ng/ul
RT: 8.728 min Scan# 958
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:108 Resp: 357369
Ion Ratio Lower Upper
108 100
107 111.7 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#19
Hexachloroethane
Concen: 29.793 ng/ul
RT: 9.046 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:117 Resp: 131545
Ion Ratio Lower Upper
117 100
201 110.0 106.5 159.7
199 67.4 59.0 88.6





#22
Nitrobenzene
Concen: 30.939 ng/u1
RT: 9.158 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 77 Resp: 41564
Ion Ratio Lower Upper
77 100
123 40.7 42.3 63.5
65 15.0 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#23
Isophorone
Concen: 29.035 ng/u1
RT: 9.693 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52



Tgt Ion: 82 Resp: 803770
Ion Ratio Lower Upper
82 100
95 8.2 7.8 11.8
138 16.9 12.9 19.3









#25
2-Nitrophenol
Concen: 32.306 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
ClientSampleId :
SLCS122

Tgt Ion:139 Resp: 19771
Ion Ratio Lower Upper
139 100
65 59.8 49.4 74.2
109 37.3 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#26
2,4-Dimethylphenol
Concen: 28.450 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:107 Resp: 413190
Ion Ratio Lower Upper
107 100
121 46.9 34.1 51.1
122 78.1 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 28.970 ng/ul
RT: 10.169 min Scan# 1204
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 93 Resp: 470849
Ion Ratio Lower Upper
93 100
95 32.0 26.2 39.4
123 11.3 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#29
2,4-Dichlorophenol
Concen: 30.057 ng/ul
RT: 10.410 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:162 Resp: 349894
Ion Ratio Lower Upper
162 100
164 64.3 37.0 55.6
98 37.2 31.0 46.6





#30
Naphthalene
Concen: 29.130 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:128 Resp: 1087334
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.1 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#32
4-Chloroaniline
Concen: 26.893 ng/ul
RT: 10.922 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:127 Resp: 455142
Ion Ratio Lower Upper
127 100
129 32.5 24.6 37.0





#33
Hexachlorobutadiene
Concen: 29.564 ng/ul
RT: 11.110 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:225 Resp: 247860
Ion Ratio Lower Upper
225 100
223 61.2 53.4 80.0
227 61.9 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#34
Caprolactam
Concen: 32.547 ng/ul m
RT: 11.693 min Scan# 1463
Delta R.T. 0.012 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:113 Resp: 125298
Ion Ratio Lower Upper
113 100
55 177.6 58.2 87.2#
56 140.2 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 30.154 ng/ul
RT: 12.040 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:107 Resp: 42533
Ion Ratio Lower Upper
107 100
144 24.3 15.8 23.8
142 74.4 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#36
2-Methylnaphthalene
Concen: 29.402 ng/ul
RT: 12.422 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:142 Resp: 801786
Ion Ratio Lower Upper
142 100
141 88.1 64.3 96.5





#37
1-Methylnaphthalene
Concen: 29.287 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:142 Resp: 80642
Ion Ratio Lower Upper
142 100
141 92.1 69.8 104.8
116 4.3 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.532 ng/ul
RT: 12.793 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:216 Resp: 471552
Ion Ratio Lower Upper
216 100
214 76.8 58.2 87.2
179 20.1 14.2 21.2
108 16.1 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 24.423 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:237 Resp: 270619
Ion Ratio Lower Upper
237 100
235 62.2 53.2 79.8
272 12.1 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#41
2,4,6-Trichlorophenol
Concen: 30.490 ng/ul
RT: 13.022 min Scan# 1689
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:196 Resp: 310914
Ion Ratio Lower Upper
196 100
198 96.1 64.1 96.1
200 30.4 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 31.436 ng/ul
RT: 13.093 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:196 Resp: 346419
Ion Ratio Lower Upper
196 100
198 96.4 68.6 103.0
200 30.8 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#43
1,1'-Biphenyl
Concen: 29.462 ng/ul
RT: 13.428 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:154 Resp: 1121906
Ion Ratio Lower Upper
154 100
153 40.1 28.6 43.0
76 12.0 8.5 12.7





#44
2-Chloronaphthalene
Concen: 29.666 ng/ul
RT: 13.469 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion: 162 Resp: 86834
Ion Ratio Lower Upper
162 100
164 32.4 27.5 41.3
127 40.2 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#45
2-Nitroaniline
Concen: 35.106 ng/ul
RT: 13.663 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion: 65 Resp: 301148
Ion Ratio Lower Upper
65 100
92 64.2 77.9 116.9#
138 94.5 81.5 122.3





#47
Dimethylphthalate
Concen: 29.800 ng/ul
RT: 14.046 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:163 Resp: 117582
Ion Ratio Lower Upper
163 100
194 4.4 4.1 6.1
164 9.7 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#48
2,6-Dinitrotoluene
Concen: 34.112 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:165 Resp: 247573
Ion Ratio Lower Upper
165 100
89 56.4 51.7 77.5
121 20.3 18.3 27.5





#50
Acenaphthylene
Concen: 29.625 ng/ul
RT: 14.310 min Scan# 1908
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:152 Resp: 1419939
Ion Ratio Lower Upper
152 100
151 20.1 18.3 27.5
153 13.1 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#51
3-Nitroaniline
Concen: 32.225 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:138 Resp: 243140
Ion Ratio Lower Upper
138 100
108 12.8 15.8 23.8#
92 123.5 117.1 175.7





#52
Acenaphthene
Concen: 29.265 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:153 Resp: 918408
Ion Ratio Lower Upper
153 100
152 46.3 34.3 51.5
154 88.0 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#53
2,4-Dinitrophenol
Concen: 31.056 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:184 Resp: 136244
Ion Ratio Lower Upper
184 100
63 71.7 34.6 52.0#
154 67.6 57.7 86.5





#55
4-Nitrophenol
Concen: 30.023 ng/u1
RT: 14.793 min Scan# 1988
Delta R.T. 0.012 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:109 Resp: 20547
Ion Ratio Lower Upper
109 100
139 116.3 61.4 92.2
65 128.8 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#56
Dibenzofuran
Concen: 29.494 ng/u1
RT: 14.987 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:168 Resp: 1362230
Ion Ratio Lower Upper
168 100
139 38.9 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 34.659 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:165 Resp: 369372
Ion Ratio Lower Upper
165 100
63 45.8 29.5 44.3#
182 3.0 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 31.396 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52



Tgt Ion:232 Resp: 315456
Ion Ratio Lower Upper
232 100
131 41.8 25.4 38.2#
130 2.1 4.2 6.2#
166 30.0 28.1 42.1





#59
Diethylphthalate
Concen: 30.149 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :

BNA_P

ClientSampleId :





SLCS122

Tgt Ion:149 Resp: 1231150
Ion Ratio Lower Upper
149 100
177 22.2 22.0 33.0
150 12.5 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#61
Fluorene
Concen: 29.792 ng/ul
RT: 15.640 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Tgt Ion:166 Resp: 1133408
Ion Ratio Lower Upper
166 100
165 95.9 70.3 105.5
167 12.7 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 29.664 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
ClientSampleId :
SLCS122

Tgt Ion:204 Resp: 600048
Ion Ratio Lower Upper
204 100
206 33.1 27.6 41.4
141 57.6 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#63
4-Nitroaniline
Concen: 35.372 ng/ul m
RT: 15.651 min Scan# 2136
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:138 Resp: 267313
Ion Ratio Lower Upper
138 100
92 54.5 41.6 62.4
108 92.3 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 31.307 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:198 Resp: 21726
Ion Ratio Lower Upper
198 100
182 3.6 6.0 9.0
77 27.1 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#67
N-Nitrosodiphenylamine
Concen: 29.436 ng/ul
RT: 15.845 min Scan# 2169
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Tgt Ion:169 Resp: 993058
Ion Ratio Lower Upper
169 100
168 65.2 48.3 72.5
167 35.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 29.994 ng/ul
RT: 16.528 min Scan# 2285
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
ClientSampleId :
SLCS122

Tgt Ion:248 Resp: 382635
Ion Ratio Lower Upper
248 100
250 98.6 79.4 119.0
141 77.5 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#69
Hexachlorobenzene
Concen: 29.923 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:284 Resp: 438942
Ion Ratio Lower Upper
284 100
142 30.9 19.2 28.8#
249 32.1 27.4 41.2





#70
Atrazine
Concen: 29.510 ng/ul
RT: 16.798 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:200 Resp: 41093
Ion Ratio Lower Upper
200 100
173 25.6 19.2 28.8
215 48.8 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#71
Pentachlorophenol
Concen: 28.962 ng/ul
RT: 16.992 min Scan# 2364
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:266 Resp: 287573
Ion Ratio Lower Upper
266 100
264 64.1 57.3 85.9
268 63.4 65.7 98.5





#72
Phenanthrene
Concen: 30.168 ng/u1
RT: 17.387 min Scan# 2431
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:178 Resp: 189785
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.7 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#74
Anthracene
Concen: 29.948 ng/u1
RT: 17.475 min Scan# 2446
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:178 Resp: 1910120
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 18.1 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.661 ng/uL
RT: 13.393 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:216 Resp: 505399
Ion Ratio Lower Upper
216 100
214 78.0 50.4 75.6
218 47.6 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#76
Pentachlorobenzene
Concen: 29.495 ng/uL
RT: 14.910 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:250 Resp: 489867
Ion Ratio Lower Upper
250 100
248 62.1 43.5 65.3
252 62.4 46.5 69.7





#77
 Carbazole
 Concen: 29.949 ng/ul
 RT: 17.739 min Scan# 2491
 Delta R.T. 0.000 min
 Lab File: BP009866.D
 Acq: 15 Apr 2022 17:52




Instrument :
 BNA_P
 ClientSampleId :
 SLCS122




Tgt Ion:167 Resp: 1726225
 Ion Ratio Lower Upper
 167 100
 166 21.2 15.8 23.8
 139 13.4 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
 Supervised By :Yogesh Patel 04/18/2022




#78
 Di-n-butylphthalate
 Concen: 30.884 ng/ul
 RT: 18.292 min Scan# 2585
 Delta R.T. -0.006 min
 Lab File: BP009866.D
 Acq: 15 Apr 2022 17:52




Tgt Ion:149 Resp: 2188992
 Ion Ratio Lower Upper
 149 100
 150 9.3 9.6 14.4#
 104 5.6 6.2 9.4#





#80
Fluoranthene
Concen: 30.848 ng/u1
RT: 19.345 min Scan# 2764
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:202 Resp: 238494
Ion Ratio Lower Upper
202 100
101 9.3 9.6 14.4
100 7.1 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#82
Pyrene
Concen: 31.005 ng/u1
RT: 19.698 min Scan# 2824
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:202 Resp: 2430352
Ion Ratio Lower Upper
202 100
101 11.4 10.6 15.8
100 9.4 9.8 14.6





#83
Butylbenzylphthalate
Concen: 32.094 ng/ul
RT: 20.569 min Scan# 2972
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:149 Resp: 103202
Ion Ratio Lower Upper
149 100
91 69.9 71.1 106.7
206 22.6 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#84
3,3'-Dichlorobenzidine
Concen: 29.895 ng/ul
RT: 21.333 min Scan# 3102
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Tgt Ion:252 Resp: 783463
Ion Ratio Lower Upper
252 100
254 63.4 49.8 74.6
126 9.8 4.3 6.5









#85
Benzo(a)anthracene
Concen: 30.802 ng/u1
RT: 21.410 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
ClientSampleId :
SLCS122

Tgt Ion:228 Resp: 2401468
Ion Ratio Lower Upper
228 100
229 19.6 18.2 27.2
226 26.3 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 31.774 ng/u1
RT: 21.333 min Scan# 3102
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:149 Resp: 1536868
Ion Ratio Lower Upper
149 100
167 28.3 32.8 49.2#
279 5.4 5.6 8.4#





#87
Chrysene
Concen: 30.956 ng/u1
RT: 21.463 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122



Tgt Ion:228 Resp: 2328610
Ion Ratio Lower Upper
228 100
226 29.3 22.7 34.1
229 19.9 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#89
Di-n-octyl phthalate
Concen: 32.809 ng/u1
RT: 22.280 min Scan# 3263
Delta R.T. 0.000 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Tgt Ion:149 Resp: 2604554









#90
Benzo(b)fluoranthene
Concen: 31.491 ng/ul
RT: 23.139 min Scan# 3408
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52

Instrument :
BNA_P
ClientSampleId :
SLCS122

Tgt Ion:252 Resp: 2370594
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 7.5 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#91
Benzo(k)fluoranthene
Concen: 32.977 ng/ul m
RT: 23.186 min Scan# 3417
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52


Tgt Ion:252 Resp: 2265274
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.1 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 31.909 ng/u1
RT: 23.780 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:252 Resp: 2226599
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 9.1 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 31.298 ng/u1
RT: 26.468 min Scan# 3975
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Tgt Ion:276 Resp: 2283775
Ion Ratio Lower Upper
276 100
138 19.3 12.3 18.5#
277 25.1 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 31.239 ng/ul
RT: 26.486 min Scan# 3977
Delta R.T. 0.006 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52




Instrument :
BNA_P
ClientSampleId :
SLCS122




Tgt Ion:278 Resp: 1976554
Ion Ratio Lower Upper
278 100
139 12.5 6.6 10.0
279 24.1 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/18/2022
Supervised By :Yogesh Patel 04/18/2022



#96
Benzo(g,h,i)perylene
Concen: 30.787 ng/ul
RT: 27.262 min Scan# 4110
Delta R.T. 0.012 min
Lab File: BP009866.D
Acq: 15 Apr 2022 17:52



Tgt Ion:276 Resp: 1852322
Ion Ratio Lower Upper
276 100
138 17.5 11.6 17.4
277 23.2 15.9 23.9

