

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041823\
 Data File : BP014717.D
 Acq On : 18 Apr 2023 17:57
 Operator : CG/JU
 Sample : SST08051
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080636

Quant Time: Apr 19 01:02:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 00:51:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.158	152	186475	20.000	ng/u1	0.00
20) Naphthalene-d8	10.987	136	759692	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.787	164	460034	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.539	188	983191	20.000	ng/u1	0.00
79) Chrysene-d12	21.616	240	932593	20.000	ng/u1	0.00
88) Perylene-d12	24.198	264	1005699	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.487	96	151770	31.562	ng/uL	0.00
4) Pyridine-d5	3.911	84	1071497	86.952	ng/u1	0.00
7) Phenol-d5	7.293	99	1331445	77.420	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.475	67	795257	72.073	ng/u1	0.00
11) 2-Chlorophenol-d4	7.681	132	1047882	84.303	ng/u1	0.00
15) 4-Methylphenol-d8	8.857	113	1033215	74.481	ng/u1	0.00
21) Nitrobenzene-d5	9.328	128	468332	76.364	ng/u1	0.00
24) 2-Nitrophenol-d4	10.057	143	462433	65.849	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.599	165	1035931	80.999	ng/u1	0.00
31) 4-Chloroaniline-d4	11.116	131	1456205	86.441	ng/u1	0.00
46) Dimethylphthalate-d6	14.193	166	2964809	79.485	ng/u1	0.00
49) Acenaphthylene-d8	14.487	160	3509806	84.801	ng/u1	0.00
54) 4-Nitrophenol-d4	14.963	143	522495	92.722	ng/u1	0.01
60) Fluorene-d10	15.781	176	2510354	80.704	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.892	200	393515	54.395	ng/u1	0.01
73) Anthracene-d10	17.645	188	3874821	81.434	ng/u1	0.00
81) Pyrene-d10	19.857	212	4437410	70.838	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.033	264	4505177	84.316	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.523	88	163507	30.942	ng/uL	96
5) Pyridine	3.934	79	1093796	84.819	ng/u1	93
6) Benzaldehyde	7.281	77	528514	61.828	ng/u1	95
8) Phenol	7.322	94	1378077	77.246	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.569	93	1108090	77.291	ng/u1	99
12) 2-Chlorophenol	7.710	128	1078379	83.354	ng/u1	97
13) 2-Methylphenol	8.593	108	1042362	78.347	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.693	45	1388272	73.753	ng/u1	98
16) Acetophenone	8.987	105	1667901	72.812	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.981	70	801094	64.734	ng/u1#	95
18) 4-Methylphenol	8.928	108	1120828	76.709	ng/u1	97
19) Hexachloroethane	9.257	117	443300	78.422	ng/u1	95
22) Nitrobenzene	9.375	77	1150311	65.290	ng/u1	97
23) Isophorone	9.905	82	2373829	73.093	ng/u1	98
25) 2-Nitrophenol	10.093	139	532272	71.809	ng/u1	95
26) 2,4-Dimethylphenol	10.140	107	1183188	71.317	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.387	93	1461230	75.046	ng/u1	99
29) 2,4-Dichlorophenol	10.622	162	1022651	81.250	ng/u1	97
30) Naphthalene	11.040	128	3459775	81.761	ng/u1	99
32) 4-Chloroaniline	11.140	127	1470574	86.185	ng/u1	99
33) Hexachlorobutadiene	11.322	225	671066	67.298	ng/u1	97
34) Caprolactam	11.916	113	160034	43.039	ng/u1	95
35) 4-Chloro-3-methylphenol	12.246	107	1049465	67.790	ng/u1	96
36) 2-Methylnaphthalene	12.634	142	2299933	80.200	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041823\
 Data File : BP014717.D
 Acq On : 18 Apr 2023 17:57
 Operator : CG/JU
 Sample : SST08051
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080636

Quant Time: Apr 19 01:02:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 00:51:18 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.851	142	2279568	80.150	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.993	216	1264329	77.084	ng/ul	98
40) Hexachlorocyclopentadiene	12.975	237	778663	73.308	ng/ul	99
41) 2,4,6-Trichlorophenol	13.222	196	800147	78.835	ng/ul	98
42) 2,4,5-Trichlorophenol	13.293	196	869454	80.466	ng/ul	99
43) 1,1'-Biphenyl	13.628	154	3047748	82.002	ng/ul	100
44) 2-Chloronaphthalene	13.675	162	2403202	81.778	ng/ul	99
45) 2-Nitroaniline	13.869	65	572723	62.931	ng/ul	95
47) Dimethylphthalate	14.240	163	2965705	79.521	ng/ul	100
48) 2,6-Dinitrotoluene	14.363	165	542140	74.404	ng/ul	86
50) Acenaphthylene	14.516	152	3726734	84.547	ng/ul	99
51) 3-Nitroaniline	14.687	138	490829	75.191	ng/ul#	98
52) Acenaphthene	14.851	153	2532256	81.297	ng/ul	99
53) 2,4-Dinitrophenol	14.887	184	230065	48.758	ng/ul	95
55) 4-Nitrophenol	14.981	109	400690	72.048	ng/ul	93
56) Dibenzofuran	15.187	168	3488438	79.669	ng/ul	97
57) 2,4-Dinitrotoluene	15.140	165	775910	75.071	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.404	232	728562	75.021	ng/ul	95
59) Diethylphthalate	15.604	149	2936995	79.248	ng/ul	98
61) Fluorene	15.839	166	2710832	76.380	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.828	204	1386091	72.692	ng/ul	98
63) 4-Nitroaniline	15.851	138	454681	85.727	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.904	198	417826	59.596	ng/ul#	91
67) N-Nitrosodiphenylamine	16.039	169	2424595	79.260	ng/ul	99
68) 4-Bromophenyl-phenylether	16.722	248	908699	74.274	ng/ul	97
69) Hexachlorobenzene	16.839	284	1031870	73.427	ng/ul	98
70) Atrazine	16.992	200	860177	77.245	ng/ul	100
71) Pentachlorophenol	17.181	266	623481	76.582	ng/ul	100
72) Phenanthrene	17.586	178	4479736	81.898	ng/ul	99
74) Anthracene	17.681	178	4516976	81.934	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.593	216	1290303	74.555	ng/uL	99
76) Pentachlorobenzene	15.104	250	1236044	71.577	ng/uL	99
77) Carbazole	17.933	167	4090094	88.655	ng/ul	99
78) Di-n-butylphthalate	18.475	149	5004006	86.261	ng/ul	99
80) Fluoranthene	19.528	202	5219210	70.467	ng/ul	96
82) Pyrene	19.886	202	5352256	69.216	ng/ul	98
83) Butylbenzylphthalate	20.739	149	1918892	72.042	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.522	252	1755730	85.125	ng/ul	100
85) Benzo(a)anthracene	21.598	228	5243177	79.339	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.510	149	3013118	82.726	ng/ul	99
87) Chrysene	21.657	228	4955919	79.421	ng/ul	99
89) Di-n-octyl phthalate	22.498	149	5230010	80.136	ng/ul	100
90) Benzo(b)fluoranthene	23.404	252	5617541	82.705	ng/ul	98
91) Benzo(k)fluoranthene	23.463	252	5371061	81.418	ng/ul	98
93) Benzo(a)pyrene	24.092	252	4942271	84.534	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.939	276	6553038	81.470	ng/ul	97
95) Dibenzo(a,h)anthracene	26.968	278	5432890	80.674	ng/ul	97
96) Benzo(g,h,i)perylene	27.780	276	5290017	80.684	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041823\
 Data File : BP014717.D
 Acq On : 18 Apr 2023 17:57
 Operator : CG/JU
 Sample : SSTD08051
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080636

Quant Time: Apr 19 01:02:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 00:51:18 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 30.942 ng/uL
 RT: 3.523 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BP014717.D
 Acq: 18 Apr 2023 17:57

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080636

Tgt Ion: 88 Resp: 163507




Ion	Ratio	Lower	Upper
88	100		
43	25.7	25.0	37.6
58	69.9	54.9	82.3




#5
 Pyridine
 Concen: 84.819 ng/uL
 RT: 3.934 min Scan# 127
 Delta R.T. -0.000 min
 Lab File: BP014717.D
 Acq: 18 Apr 2023 17:57

Tgt Ion: 79 Resp: 1093796

Ion	Ratio	Lower	Upper
79	100		
52	65.6	58.3	87.5
51	32.5	27.8	41.6









#6
Benzaldehyde
Concen: 61.828 ng/ul
RT: 7.281 min Scan# 696
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion: 77 Resp: 528514
Ion Ratio Lower Upper
77 100
105 100.8 76.6 114.8
106 96.5 73.3 109.9



#8
Phenol
Concen: 77.246 ng/ul
RT: 7.322 min Scan# 703
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion: 94 Resp: 1378077
Ion Ratio Lower Upper
94 100
65 26.7 21.4 32.0
66 34.0 29.2 43.8









#10
Bis(2-Chloroethyl)ether
Concen: 77.291 ng/ul
RT: 7.569 min Scan# 744
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion: 93 Resp: 1108090
Ion Ratio Lower Upper
93 100
63 73.8 59.1 88.7
95 32.9 25.1 37.7



#12
2-Chlorophenol
Concen: 83.354 ng/ul
RT: 7.710 min Scan# 769
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:128 Resp: 1078379
Ion Ratio Lower Upper
128 100
64 44.2 37.5 56.3
130 31.4 25.8 38.6





#13

2-Methylphenol

Concen: 78.347 ng/ul

RT: 8.593 min Scan# 918

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :





SSTD080636

Tgt Ion:108 Resp: 1042362

Ion Ratio Lower Upper

108 100

107 88.1 70.4 105.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 73.753 ng/ul

RT: 8.693 min Scan# 936

Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion: 45 Resp: 1388272


Ion Ratio Lower Upper

45 100

77 15.2 11.6 17.4

79 12.2 9.2 13.8









#16
Acetophenone
Concen: 72.812 ng/ul
RT: 8.987 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:105 Resp: 1667901
Ion Ratio Lower Upper
105 100
77 73.5 60.6 91.0
51 23.0 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 64.734 ng/ul
RT: 8.981 min Scan# 985
Delta R.T. 0.018 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion: 70 Resp: 801094
Ion Ratio Lower Upper
70 100
42 51.0 42.9 64.3
101 10.5 7.6 11.4
130 22.5 14.2 21.2#









#18
4-Methylphenol
Concen: 76.709 ng/ul
RT: 8.928 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:108 Resp: 1120828
Ion Ratio Lower Upper
108 100
107 110.1 90.8 136.2



#19
Hexachloroethane
Concen: 78.422 ng/ul
RT: 9.257 min Scan# 1032
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:117 Resp: 443300
Ion Ratio Lower Upper
117 100
201 107.8 81.0 121.4
199 66.3 51.5 77.3









#22
Nitrobenzene
Concen: 65.290 ng/ul
RT: 9.375 min Scan# 110
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion: 77 Resp: 1150311
Ion Ratio Lower Upper
77 100
123 46.5 35.4 53.2
65 12.7 10.8 16.2



#23
Isophorone
Concen: 73.093 ng/ul
RT: 9.905 min Scan# 1142
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion: 82 Resp: 2373829
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.8
138 17.9 13.6 20.4









#25
2-Nitrophenol
Concen: 71.809 ng/ul
RT: 10.093 min Scan# 1173
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:139 Resp: 532272
Ion Ratio Lower Upper
139 100
65 45.3 39.4 59.2
109 28.8 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 71.317 ng/ul
RT: 10.140 min Scan# 1182
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:107 Resp: 1183188
Ion Ratio Lower Upper
107 100
121 51.3 40.2 60.2
122 88.7 69.2 103.8





#27

Bis(2-Chloroethoxy)methane

Concen: 75.046 ng/ul

RT: 10.387 min Scan# 1223

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion: 93 Resp: 1461230

Ion Ratio Lower Upper

93 100

95 33.0 25.9 38.9

123 11.4 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 81.250 ng/ul

RT: 10.622 min Scan# 1264

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:162 Resp: 1022651


Ion Ratio Lower Upper

162 100

164 64.9 50.6 76.0

98 33.9 28.8 43.2









#30
Naphthalene
Concen: 81.761 ng/ul
RT: 11.040 min Scan# 1334
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:128 Resp: 3459775
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.5 10.4 15.6



#32
4-Chloroaniline
Concen: 86.185 ng/ul
RT: 11.140 min Scan# 1352
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:127 Resp: 1470574
Ion Ratio Lower Upper
127 100
129 32.0 26.0 39.0









#33
Hexachlorobutadiene
Concen: 67.298 ng/ul
RT: 11.322 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:225 Resp: 671066
Ion Ratio Lower Upper
225 100
223 63.0 49.2 73.8
227 64.3 49.4 74.2



#34
Caprolactam
Concen: 43.039 ng/ul
RT: 11.916 min Scan# 1484
Delta R.T. 0.029 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:113 Resp: 160034
Ion Ratio Lower Upper
113 100
55 161.5 135.4 203.0
56 120.7 101.0 151.4









#35
4-Chloro-3-methylphenol
Concen: 67.790 ng/ul
RT: 12.246 min Scan# 1538
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:107 Resp: 1049465
Ion Ratio Lower Upper
107 100
144 26.8 20.5 30.7
142 81.8 61.9 92.9



#36
2-Methylnaphthalene
Concen: 80.200 ng/ul
RT: 12.634 min Scan# 1606
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:142 Resp: 2299933
Ion Ratio Lower Upper
142 100
141 89.4 70.9 106.3









#37
1-Methylnaphthalene
Concen: 80.150 ng/ul
RT: 12.851 min Scan# 1643
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:142 Resp: 2279568
Ion Ratio Lower Upper
142 100
141 92.3 73.8 110.6
116 3.5 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 77.084 ng/ul
RT: 12.993 min Scan# 1667
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:216 Resp: 1264329
Ion Ratio Lower Upper
216 100
214 78.5 61.7 92.5
179 20.4 15.5 23.3
108 16.1 13.9 20.9









#40
Hexachlorocyclopentadiene
Concen: 73.308 ng/ul
RT: 12.975 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:237 Resp: 778663
Ion Ratio Lower Upper
237 100
235 62.8 50.6 76.0
272 13.3 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 78.835 ng/ul
RT: 13.222 min Scan# 1706
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:196 Resp: 800147
Ion Ratio Lower Upper
196 100
198 95.9 77.8 116.8
200 31.2 25.7 38.5





#42

2,4,5-Trichlorophenol

Concen: 80.466 ng/ul

RT: 13.293 min Scan# 1718

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:196 Resp: 869454

Ion Ratio Lower Upper

196 100

198 96.5 78.0 117.0

200 31.0 24.7 37.1



#43

1,1'-Biphenyl

Concen: 82.002 ng/ul

RT: 13.628 min Scan# 1775

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:154 Resp: 3047748


Ion Ratio Lower Upper

154 100

153 39.9 31.9 47.9

76 13.2 10.8 16.2





#44

2-Chloronaphthalene

Concen: 81.778 ng/ul

RT: 13.675 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:162 Resp: 2403202

Ion Ratio Lower Upper

162 100

164 32.8 26.6 39.8

127 37.3 30.5 45.7



#45

2-Nitroaniline

Concen: 62.931 ng/ul

RT: 13.869 min Scan# 1816

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion: 65 Resp: 572723


Ion Ratio Lower Upper

65 100

92 72.7 57.0 85.4

138 115.4 86.8 130.2









#47
Dimethylphthalate
Concen: 79.521 ng/ul
RT: 14.240 min Scan# 1878
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:163 Resp: 2965705
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.3 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 74.404 ng/ul
RT: 14.363 min Scan# 1900
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:165 Resp: 542140
Ion Ratio Lower Upper
165 100
89 53.9 53.0 79.6
121 20.5 19.8 29.6









#50
Acenaphthylene
Concen: 84.547 ng/ul
RT: 14.516 min Scan# 1925
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:152 Resp: 3726734
Ion Ratio Lower Upper
152 100
151 19.8 15.7 23.5
153 13.3 10.0 15.0



#51
3-Nitroaniline
Concen: 75.191 ng/ul
RT: 14.687 min Scan# 1955
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:138 Resp: 490829
Ion Ratio Lower Upper
138 100
108 12.6 12.9 19.3#
92 113.0 91.5 137.3









#52
Acenaphthene
Concen: 81.297 ng/ul
RT: 14.851 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:153 Resp: 2532256
Ion Ratio Lower Upper
153 100
152 47.9 38.4 57.6
154 88.4 69.8 104.8



#53
2,4-Dinitrophenol
Concen: 48.758 ng/ul
RT: 14.887 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:184 Resp: 230065
Ion Ratio Lower Upper
184 100
63 61.3 51.1 76.7
154 62.6 54.0 81.0









#55
4-Nitrophenol
Concen: 72.048 ng/ul
RT: 14.981 min Scan# 2005
Delta R.T. 0.018 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:109 Resp: 400690
Ion Ratio Lower Upper
109 100
139 138.6 100.5 150.7
65 125.1 102.4 153.6



#56
Dibenzofuran
Concen: 79.669 ng/ul
RT: 15.187 min Scan# 2040
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:168 Resp: 3488438
Ion Ratio Lower Upper
168 100
139 35.6 29.8 44.8









#57
2,4-Dinitrotoluene
Concen: 75.071 ng/ul
RT: 15.140 min Scan# 2031
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:165 Resp: 775910
Ion Ratio Lower Upper
165 100
63 41.3 36.6 54.8
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 75.021 ng/ul
RT: 15.404 min Scan# 2077
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:232 Resp: 728562
Ion Ratio Lower Upper
232 100
131 43.3 32.4 48.6
130 2.3 1.8 2.6
166 30.1 21.9 32.9









#59
Diethylphthalate
Concen: 79.248 ng/ul
RT: 15.604 min Scan# 2111
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:149 Resp: 2936995
Ion Ratio Lower Upper
149 100
177 23.1 17.8 26.8
150 12.5 9.4 14.2



#61
Fluorene
Concen: 76.380 ng/ul
RT: 15.839 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:166 Resp: 2710832
Ion Ratio Lower Upper
166 100
165 98.2 78.6 117.8
167 13.3 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 72.692 ng/ul
RT: 15.828 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:204 Resp: 1386091
Ion Ratio Lower Upper
204 100
206 33.8 26.4 39.6
141 54.9 45.0 67.4



#63
4-Nitroaniline
Concen: 85.727 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:138 Resp: 454681
Ion Ratio Lower Upper
138 100
92 48.5 40.9 61.3
108 74.7 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 59.596 ng/ul
RT: 15.904 min Scan# 2162
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:198 Resp: 417826
Ion Ratio Lower Upper
198 100
182 3.9 3.9 5.9#
77 23.9 23.6 35.4



#67
N-Nitrosodiphenylamine
Concen: 79.260 ng/ul
RT: 16.039 min Scan# 2185
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:169 Resp: 2424595
Ion Ratio Lower Upper
169 100
168 67.3 53.5 80.3
167 37.0 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 74.274 ng/ul
RT: 16.722 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:248 Resp: 908699
Ion Ratio Lower Upper
248 100
250 97.9 79.2 118.8
141 70.5 52.7 79.1



#69
Hexachlorobenzene
Concen: 73.427 ng/ul
RT: 16.839 min Scan# 2321
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:284 Resp: 1031870
Ion Ratio Lower Upper
284 100
142 33.3 25.1 37.7
249 31.3 25.2 37.8





#70

Atrazine

Concen: 77.245 ng/ul

RT: 16.992 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:200 Resp: 860177

Ion Ratio Lower Upper

200 100

173 26.0 21.0 31.6

215 49.2 39.4 59.2



#71

Pentachlorophenol

Concen: 76.582 ng/ul

RT: 17.181 min Scan# 2379

Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:266 Resp: 623481


Ion Ratio Lower Upper

266 100

264 62.6 50.2 75.4

268 62.1 49.8 74.6





#72

Phenanthrene

Concen: 81.898 ng/ul

RT: 17.586 min Scan# 2447

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:178 Resp: 4479736

Ion Ratio Lower Upper

178 100

179 15.3 12.1 18.1

176 19.6 15.3 22.9



#74

Anthracene

Concen: 81.934 ng/ul

RT: 17.681 min Scan# 2464

Delta R.T. 0.012 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:178 Resp: 4516976


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.0 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 74.555 ng/uL

RT: 13.593 min Scan# 1769

Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:216 Resp: 1290303

Ion Ratio Lower Upper

216 100

214 78.6 62.8 94.2

218 48.0 39.6 59.4



#76

Pentachlorobenzene

Concen: 71.577 ng/uL

RT: 15.104 min Scan# 2026

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:250 Resp: 1236044


Ion Ratio Lower Upper

250 100

248 64.1 50.7 76.1

252 63.6 50.1 75.1









#77
 Carbazole
 Concen: 88.655 ng/ul
 RT: 17.933 min Scan# 2506
 Delta R.T. 0.006 min
 Lab File: BP014717.D
 Acq: 18 Apr 2023 17:57




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080636


Tgt Ion:167 Resp: 4090094
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.0 25.4
 139 12.7 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 86.261 ng/ul
 RT: 18.475 min Scan# 2599
 Delta R.T. -0.000 min
 Lab File: BP014717.D
 Acq: 18 Apr 2023 17:57

Tgt Ion:149 Resp: 5004006
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.3 10.9
 104 4.8 3.9 5.9





#80

Fluoranthene

Concen: 70.467 ng/ul

RT: 19.528 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:202 Resp: 5219210

Ion Ratio Lower Upper

202 100

101 12.0 8.4 12.6

100 9.3 6.4 9.6



#82

Pyrene

Concen: 69.216 ng/ul

RT: 19.886 min Scan# 2839

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:202 Resp: 5352256


Ion Ratio Lower Upper

202 100

101 12.9 9.4 14.0

100 10.6 8.0 12.0









#83
Butylbenzylphthalate
Concen: 72.042 ng/ul
RT: 20.739 min Scan# 2984
Delta R.T. -0.000 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:149 Resp: 1918892
Ion Ratio Lower Upper
149 100
91 70.4 58.6 87.8
206 21.0 17.3 25.9



#84
3,3'-Dichlorobenzidine
Concen: 85.125 ng/ul
RT: 21.522 min Scan# 3117
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:252 Resp: 1755730
Ion Ratio Lower Upper
252 100
254 65.0 52.2 78.2
126 10.7 8.0 12.0





#85

Benzo(a)anthracene

Concen: 79.339 ng/ul

RT: 21.598 min Scan# 3115

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:228 Resp: 5243177

Ion Ratio Lower Upper

228 100

229 20.2 15.9 23.9

226 27.8 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 82.726 ng/ul

RT: 21.510 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:149 Resp: 3013118


Ion Ratio Lower Upper

149 100

167 28.0 22.0 33.0

279 4.8 3.6 5.4





#87

Chrysene

Concen: 79.421 ng/ul

RT: 21.657 min Scan# 3139

Delta R.T. 0.006 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:228 Resp: 4955919

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 20.2 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 80.136 ng/ul




RT: 22.498 min Scan# 3283


Delta R.T. -0.000 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Tgt Ion:149 Resp: 5230010









#90
Benzo(b)fluoranthene
Concen: 82.705 ng/ul
RT: 23.404 min Scan# 3436
Delta R.T. 0.006 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57




Instrument :
BNA_P
ClientSampleId :
SSTD080636


Tgt Ion:252 Resp: 5617541
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 10.4 7.0 10.6



#91
Benzo(k)fluoranthene
Concen: 81.418 ng/ul
RT: 23.463 min Scan# 3447
Delta R.T. 0.012 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:252 Resp: 5371061
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.0
125 10.0 7.0 10.4





#93

Benzo(a)pyrene

Concen: 84.534 ng/ul

RT: 24.092 min Scan# 3

Delta R.T. 0.018 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57

Instrument :

BNA_P

ClientSampleId :

SSTD080636





Tgt Ion:252 Resp: 4942271

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 11.0 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 81.470 ng/ul

RT: 26.939 min Scan# 4038

Delta R.T. 0.023 min

Lab File: BP014717.D

Acq: 18 Apr 2023 17:57




Tgt Ion:276 Resp: 6553038


Ion Ratio Lower Upper

276 100

138 23.1 16.3 24.5

277 25.0 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 80.674 ng/ul
RT: 26.968 min Scan# 4038
Delta R.T. 0.029 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Instrument :
BNA_P
ClientSampleId :
SSTD080636

Tgt Ion:278 Resp: 5432890
Ion Ratio Lower Upper
278 100
139 16.2 11.5 17.3
279 24.0 18.6 27.8



#96
Benzo(g,h,i)perylene
Concen: 80.684 ng/ul
RT: 27.780 min Scan# 4181
Delta R.T. 0.023 min
Lab File: BP014717.D
Acq: 18 Apr 2023 17:57

Tgt Ion:276 Resp: 5290017
Ion Ratio Lower Upper
276 100
138 20.3 14.3 21.5
277 24.2 19.5 29.3

