

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041823\
 Data File : BP014719.D
 Acq On : 18 Apr 2023 19:37
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV638

Quant Time: Apr 19 01:47:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 01:37:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.152	152	208214	20.000	ng/ul	0.00
20) Naphthalene-d8	10.987	136	882828	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.786	164	567580	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.539	188	1273681	20.000	ng/ul	0.00
79) Chrysene-d12	21.610	240	1273528	20.000	ng/ul	0.00
88) Perylene-d12	24.192	264	1385367	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.487	96	44133	8.424	ng/uL	0.00
4) Pyridine-d5	3.916	84	286517	20.088	ng/ul	0.00
7) Phenol-d5	7.287	99	373016	20.720	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.469	67	230498	20.754	ng/ul	0.00
11) 2-Chlorophenol-d4	7.675	132	290827	21.753	ng/ul	0.00
15) 4-Methylphenol-d8	8.851	113	295575	21.061	ng/ul	0.00
21) Nitrobenzene-d5	9.322	128	129203	23.356	ng/ul	0.00
24) 2-Nitrophenol-d4	10.051	143	121405	24.298	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.593	165	292799	21.582	ng/ul	0.00
31) 4-Chloroaniline-d4	11.110	131	438886	21.195	ng/ul	0.00
46) Dimethylphthalate-d6	14.187	166	960656	21.376	ng/ul	0.00
49) Acenaphthylene-d8	14.481	160	1109786	21.308	ng/ul	0.00
54) 4-Nitrophenol-d4	14.951	143	158586	22.185	ng/ul	0.00
60) Fluorene-d10	15.775	176	819384	21.248	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	105407	20.450	ng/ul	0.00
73) Anthracene-d10	17.639	188	1301918	21.159	ng/ul	0.00
81) Pyrene-d10	19.851	212	1523250	20.880	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.021	264	1513567	21.080	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.522	88	47816	8.451	ng/uL	94
5) Pyridine	3.934	79	299315	20.916	ng/ul	97
6) Benzaldehyde	7.281	77	224312	25.191	ng/ul	95
8) Phenol	7.316	94	391808	20.851	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.563	93	318776	20.922	ng/ul	98
12) 2-Chlorophenol	7.710	128	298039	21.318	ng/ul	97
13) 2-Methylphenol	8.587	108	291983	20.762	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.687	45	403166	20.725	ng/ul	99
16) Acetophenone	8.981	105	507309	21.731	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.963	70	245047	21.581	ng/ul	98
18) 4-Methylphenol	8.916	108	323076	21.141	ng/ul	99
19) Hexachloroethane	9.251	117	124651	21.166	ng/ul	98
22) Nitrobenzene	9.369	77	335393	22.313	ng/ul	98
23) Isophorone	9.893	82	717184	21.131	ng/ul	99
25) 2-Nitrophenol	10.087	139	143730	24.219	ng/ul	97
26) 2,4-Dimethylphenol	10.134	107	353811	21.382	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.381	93	437869	20.823	ng/ul	99
29) 2,4-Dichlorophenol	10.616	162	295516	21.581	ng/ul	98
30) Naphthalene	11.034	128	1035547	20.930	ng/ul	100
32) 4-Chloroaniline	11.134	127	444907	21.209	ng/ul	99
33) Hexachlorobutadiene	11.322	225	198892	20.535	ng/ul	99
34) Caprolactam	11.887	113	94009	22.673	ng/ul	97
35) 4-Chloro-3-methylphenol	12.234	107	312364	21.827	ng/ul	98
36) 2-Methylnaphthalene	12.628	142	692983	20.980	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041823\
 Data File : BP014719.D
 Acq On : 18 Apr 2023 19:37
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV638


Quant Time: Apr 19 01:47:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 01:37:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.845	142	694470	21.208	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.987	216	389451	20.390	ng/ul	99
40) Hexachlorocyclopentadiene	12.975	237	222942	20.629	ng/ul	98
41) 2,4,6-Trichlorophenol	13.222	196	235289	21.749	ng/ul	99
42) 2,4,5-Trichlorophenol	13.287	196	255835	21.859	ng/ul	98
43) 1,1'-Biphenyl	13.622	154	957902	20.813	ng/ul	100
44) 2-Chloronaphthalene	13.669	162	753831	20.962	ng/ul	99
45) 2-Nitroaniline	13.863	65	166773	24.498	ng/ul	98
47) Dimethylphthalate	14.234	163	961156	21.232	ng/ul	100
48) 2,6-Dinitrotoluene	14.351	165	152653	24.723	ng/ul	94
50) Acenaphthylene	14.510	152	1197599	21.216	ng/ul	99
51) 3-Nitroaniline	14.681	138	164240	23.386	ng/ul	98
52) Acenaphthene	14.851	153	816549	21.216	ng/ul	100
53) 2,4-Dinitrophenol	14.881	184	58531	19.591	ng/ul	99
55) 4-Nitrophenol	14.963	109	129433	22.073	ng/ul	99
56) Dibenzofuran	15.181	168	1139577	21.038	ng/ul	98
57) 2,4-Dinitrotoluene	15.134	165	227879	25.153	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.404	232	224873	22.626	ng/ul	98
59) Diethylphthalate	15.592	149	947260	21.565	ng/ul	99
61) Fluorene	15.833	166	954522	21.588	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.822	204	475237	21.019	ng/ul	99
63) 4-Nitroaniline	15.839	138	174456	24.854	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.898	198	115063	20.969	ng/ul#	90
67) N-Nitrosodiphenylamine	16.033	169	800570	21.093	ng/ul	100
68) 4-Bromophenyl-phenylether	16.722	248	295528	20.703	ng/ul	98
69) Hexachlorobenzene	16.839	284	343163	20.719	ng/ul	98
70) Atrazine	16.980	200	289707	21.491	ng/ul	98
71) Pentachlorophenol	17.175	266	185385	20.562	ng/ul	98
72) Phenanthrene	17.580	178	1501032	20.911	ng/ul	100
74) Anthracene	17.669	178	1535506	21.100	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.592	216	399439	20.083	ng/uL	100
76) Pentachlorobenzene	15.104	250	401285	20.456	ng/uL	98
77) Carbazole	17.927	167	1378169	21.572	ng/ul	100
78) Di-n-butylphthalate	18.475	149	1585356	21.858	ng/ul	100
80) Fluoranthene	19.527	202	1763254	20.774	ng/ul	100
82) Pyrene	19.880	202	1893265	21.079	ng/ul	99
83) Butylbenzylphthalate	20.739	149	550105	22.434	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.516	252	632029	22.167	ng/ul	99
85) Benzo(a)anthracene	21.598	228	1864025	21.057	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.504	149	929195	22.464	ng/ul	99
87) Chrysene	21.651	228	1783115	21.277	ng/ul	99
89) Di-n-octyl phthalate	22.498	149	1475257	19.682	ng/ul	100
90) Benzo(b)fluoranthene	23.398	252	1950002	21.608	ng/ul	99
91) Benzo(k)fluoranthene	23.451	252	1851977	20.519	ng/ul	99
93) Benzo(a)pyrene	24.080	252	1687231	21.051	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.921	276	2186550	20.961	ng/ul	99
95) Dibenzo(a,h)anthracene	26.945	278	1825550	20.897	ng/ul	98
96) Benzo(g,h,i)perylene	27.756	276	1731443	20.753	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 19 01:47:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 19 01:37:13 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.451 ng/uL
RT: 3.522 min Scan# 51
Delta R.T. -0.001 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion: 88 Resp: 47816
Ion Ratio Lower Upper
88 100
43 26.2 25.0 37.6
58 72.1 54.9 82.3



#5
Pyridine
Concen: 20.916 ng/uL
RT: 3.934 min Scan# 127
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion: 79 Resp: 299315
Ion Ratio Lower Upper
79 100
52 69.8 58.3 87.5
51 33.1 27.8 41.6









#6
Benzaldehyde
Concen: 25.191 ng/ul
RT: 7.281 min Scan# 696
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion: 77 Resp: 224312
Ion Ratio Lower Upper
77 100
105 100.6 76.6 114.8
106 96.0 73.3 109.9



#8
Phenol
Concen: 20.851 ng/ul
RT: 7.316 min Scan# 702
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion: 94 Resp: 391808
Ion Ratio Lower Upper
94 100
65 26.6 21.4 32.0
66 34.5 29.2 43.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.922 ng/ul
RT: 7.563 min Scan# 744
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion: 93 Resp: 318776
Ion Ratio Lower Upper
93 100
63 73.3 59.1 88.7
95 33.4 25.1 37.7



#12
2-Chlorophenol
Concen: 21.318 ng/ul
RT: 7.710 min Scan# 769
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:128 Resp: 298039
Ion Ratio Lower Upper
128 100
64 44.9 37.5 56.3
130 33.4 25.8 38.6





#13

2-Methylphenol

Concen: 20.762 ng/ul

RT: 8.587 min Scan# 918

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :





SICV638

Tgt Ion:108 Resp: 291983

Ion Ratio Lower Upper

108 100

107 90.3 70.4 105.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.725 ng/ul

RT: 8.687 min Scan# 935

Delta R.T. -0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion: 45 Resp: 403166


Ion Ratio Lower Upper

45 100

77 14.9 11.6 17.4

79 11.9 9.2 13.8









#16
Acetophenone
Concen: 21.731 ng/ul
RT: 8.981 min Scan# 985
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:105 Resp: 507309
Ion Ratio Lower Upper
105 100
77 74.6 60.6 91.0
51 25.3 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.581 ng/ul
RT: 8.963 min Scan# 982
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion: 70 Resp: 245047
Ion Ratio Lower Upper
70 100
42 54.0 42.9 64.3
101 9.9 7.6 11.4
130 21.1 14.2 21.2









#18
4-Methylphenol
Concen: 21.141 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:108 Resp: 323076
Ion Ratio Lower Upper
108 100
107 112.5 90.8 136.2



#19
Hexachloroethane
Concen: 21.166 ng/ul
RT: 9.251 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:117 Resp: 124651
Ion Ratio Lower Upper
117 100
201 104.2 81.0 121.4
199 63.6 51.5 77.3









#22
Nitrobenzene
Concen: 22.313 ng/ul
RT: 9.369 min Scan# 1051
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion: 77 Resp: 335393
Ion Ratio Lower Upper
77 100
123 46.2 35.4 53.2
65 13.3 10.8 16.2



#23
Isophorone
Concen: 21.131 ng/ul
RT: 9.893 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion: 82 Resp: 717184
Ion Ratio Lower Upper
82 100
95 7.1 5.8 8.8
138 17.6 13.6 20.4









#25
2-Nitrophenol
Concen: 24.219 ng/ul
RT: 10.087 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:139 Resp: 143730
Ion Ratio Lower Upper
139 100
65 52.2 39.4 59.2
109 30.8 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 21.382 ng/ul
RT: 10.134 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:107 Resp: 353811
Ion Ratio Lower Upper
107 100
121 51.6 40.2 60.2
122 88.1 69.2 103.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.823 ng/ul

RT: 10.381 min Scan# 1223

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion: 93 Resp: 437869

Ion Ratio Lower Upper

93 100

95 31.5 25.9 38.9

123 11.1 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 21.581 ng/ul

RT: 10.616 min Scan# 1263

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:162 Resp: 295516


Ion Ratio Lower Upper

162 100

164 63.8 50.6 76.0

98 34.2 28.8 43.2









#30
Naphthalene
Concen: 20.930 ng/ul
RT: 11.034 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:128 Resp: 1035547
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 21.209 ng/ul
RT: 11.134 min Scan# 1351
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:127 Resp: 444907
Ion Ratio Lower Upper
127 100
129 33.2 26.0 39.0









#33
Hexachlorobutadiene
Concen: 20.535 ng/ul
RT: 11.322 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:225 Resp: 198892
Ion Ratio Lower Upper
225 100
223 62.0 49.2 73.8
227 63.6 49.4 74.2



#34
Caprolactam
Concen: 22.673 ng/ul
RT: 11.887 min Scan# 1479
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:113 Resp: 94009
Ion Ratio Lower Upper
113 100
55 172.5 135.4 203.0
56 122.8 101.0 151.4









#35
4-Chloro-3-methylphenol
Concen: 21.827 ng/ul
RT: 12.234 min Scan# 1538
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:107 Resp: 312364
Ion Ratio Lower Upper
107 100
144 25.6 20.5 30.7
142 79.9 61.9 92.9



#36
2-Methylnaphthalene
Concen: 20.980 ng/ul
RT: 12.628 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:142 Resp: 692983
Ion Ratio Lower Upper
142 100
141 88.9 70.9 106.3









#37
1-Methylnaphthalene
Concen: 21.208 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:142 Resp: 694470
Ion Ratio Lower Upper
142 100
141 90.5 73.8 110.6
116 3.7 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.390 ng/ul
RT: 12.987 min Scan# 1666
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:216 Resp: 389451
Ion Ratio Lower Upper
216 100
214 76.1 61.7 92.5
179 19.6 15.5 23.3
108 17.8 13.9 20.9









#40
Hexachlorocyclopentadiene
Concen: 20.629 ng/ul
RT: 12.975 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:237 Resp: 222942
Ion Ratio Lower Upper
237 100
235 62.0 50.6 76.0
272 12.8 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 21.749 ng/ul
RT: 13.222 min Scan# 1706
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:196 Resp: 235289
Ion Ratio Lower Upper
196 100
198 95.9 77.8 116.8
200 31.9 25.7 38.5





#42

2,4,5-Trichlorophenol

Concen: 21.859 ng/ul

RT: 13.287 min Scan# 1717

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:196 Resp: 255835

Ion Ratio Lower Upper

196 100

198 95.3 78.0 117.0

200 30.2 24.7 37.1



#43

1,1'-Biphenyl

Concen: 20.813 ng/ul

RT: 13.622 min Scan# 1774

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:154 Resp: 957902


Ion Ratio Lower Upper

154 100

153 39.8 31.9 47.9

76 13.3 10.8 16.2





#44

2-Chloronaphthalene

Concen: 20.962 ng/ul

RT: 13.669 min Scan# 1782

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:162 Resp: 753831

Ion Ratio Lower Upper

162 100

164 32.7 26.6 39.8

127 37.6 30.5 45.7



#45

2-Nitroaniline

Concen: 24.498 ng/ul

RT: 13.863 min Scan# 1815

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion: 65 Resp: 166773


Ion Ratio Lower Upper

65 100

92 72.5 57.0 85.4

138 110.3 86.8 130.2









#47
Dimethylphthalate
Concen: 21.232 ng/ul
RT: 14.234 min Scan# 1878
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:163 Resp: 961156
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 24.723 ng/ul
RT: 14.351 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:165 Resp: 152653
Ion Ratio Lower Upper
165 100
89 60.6 53.0 79.6
121 23.1 19.8 29.6









#50
Acenaphthylene
Concen: 21.216 ng/ul
RT: 14.510 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:152 Resp: 1197599
Ion Ratio Lower Upper
152 100
151 19.7 15.7 23.5
153 13.2 10.0 15.0



#51
3-Nitroaniline
Concen: 23.386 ng/ul
RT: 14.681 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:138 Resp: 164240
Ion Ratio Lower Upper
138 100
108 13.6 12.9 19.3
92 112.9 91.5 137.3









#52
Acenaphthene
Concen: 21.216 ng/ul
RT: 14.851 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:153 Resp: 816549
Ion Ratio Lower Upper
153 100
152 48.1 38.4 57.6
154 87.7 69.8 104.8



#53
2,4-Dinitrophenol
Concen: 19.591 ng/ul
RT: 14.881 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:184 Resp: 58531
Ion Ratio Lower Upper
184 100
63 62.7 51.1 76.7
154 66.6 54.0 81.0









#55
4-Nitrophenol
Concen: 22.073 ng/ul
RT: 14.963 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:109 Resp: 129433
Ion Ratio Lower Upper
109 100
139 127.9 100.5 150.7
65 127.9 102.4 153.6



#56
Dibenzofuran
Concen: 21.038 ng/ul
RT: 15.181 min Scan# 2039
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:168 Resp: 1139577
Ion Ratio Lower Upper
168 100
139 36.0 29.8 44.8





#57

2,4-Dinitrotoluene

Concen: 25.153 ng/ul

RT: 15.134 min Scan# 2031

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:165 Resp: 227879

Ion Ratio Lower Upper

165 100

63 43.0 36.6 54.8

182 2.9 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.626 ng/ul

RT: 15.404 min Scan# 2077

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Tgt Ion:232 Resp: 224873




Ion Ratio Lower Upper


232 100

131 41.9 32.4 48.6

130 2.0 1.8 2.6

166 28.6 21.9 32.9









#59
Diethylphthalate
Concen: 21.565 ng/ul
RT: 15.592 min Scan# 2110
Delta R.T. -0.006 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:149 Resp: 947260
Ion Ratio Lower Upper
149 100
177 22.5 17.8 26.8
150 12.3 9.4 14.2



#61
Fluorene
Concen: 21.588 ng/ul
RT: 15.833 min Scan# 2150
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:166 Resp: 954522
Ion Ratio Lower Upper
166 100
165 96.6 78.6 117.8
167 13.0 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 21.019 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:204 Resp: 475237
Ion Ratio Lower Upper
204 100
206 32.7 26.4 39.6
141 56.8 45.0 67.4



#63
4-Nitroaniline
Concen: 24.854 ng/ul
RT: 15.839 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:138 Resp: 174456
Ion Ratio Lower Upper
138 100
92 48.6 40.9 61.3
108 77.9 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 20.969 ng/ul
RT: 15.898 min Scan# 2160
Delta R.T. 0.006 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:198 Resp: 115063
Ion Ratio Lower Upper
198 100
182 4.6 3.9 5.9
77 23.2 23.6 35.4#



#67
N-Nitrosodiphenylamine
Concen: 21.093 ng/ul
RT: 16.033 min Scan# 2184
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:169 Resp: 800570
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 36.0 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 20.703 ng/ul
RT: 16.722 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:248 Resp: 295528
Ion Ratio Lower Upper
248 100
250 95.6 79.2 118.8
141 65.3 52.7 79.1



#69
Hexachlorobenzene
Concen: 20.719 ng/ul
RT: 16.839 min Scan# 2321
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:284 Resp: 343163
Ion Ratio Lower Upper
284 100
142 32.2 25.1 37.7
249 30.7 25.2 37.8





#70

Atrazine

Concen: 21.491 ng/ul

RT: 16.980 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:200 Resp: 289707

Ion Ratio Lower Upper

200 100

173 26.0 21.0 31.6

215 50.7 39.4 59.2



#71

Pentachlorophenol

Concen: 20.562 ng/ul

RT: 17.175 min Scan# 2378

Delta R.T. -0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:266 Resp: 185385


Ion Ratio Lower Upper

266 100

264 61.4 50.2 75.4

268 64.1 49.8 74.6





#72

Phenanthrene

Concen: 20.911 ng/ul

RT: 17.580 min Scan# 2447

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:178 Resp: 1501032

Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 18.9 15.3 22.9



#74

Anthracene

Concen: 21.100 ng/ul

RT: 17.669 min Scan# 2462

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:178 Resp: 1535506


Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 19.0 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.083 ng/uL

RT: 13.592 min Scan# 1769

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:216 Resp: 399439

Ion Ratio Lower Upper

216 100

214 78.8 62.8 94.2

218 49.0 39.6 59.4



#76

Pentachlorobenzene

Concen: 20.456 ng/uL

RT: 15.104 min Scan# 2026

Delta R.T. 0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:250 Resp: 401285


Ion Ratio Lower Upper

250 100

248 62.1 50.7 76.1

252 64.1 50.1 75.1









#77
 Carbazole
 Concen: 21.572 ng/ul
 RT: 17.927 min Scan# 2506
 Delta R.T. -0.000 min
 Lab File: BP014719.D
 Acq: 18 Apr 2023 19:37




Instrument :
 BNA_P
 ClientSampleId :
 SICV638


Tgt Ion:167 Resp: 1378169
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.0 25.4
 139 12.4 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 21.858 ng/ul
 RT: 18.475 min Scan# 2599
 Delta R.T. -0.000 min
 Lab File: BP014719.D
 Acq: 18 Apr 2023 19:37

Tgt Ion:149 Resp: 1585356
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 4.9 3.9 5.9









#80
Fluoranthene
Concen: 20.774 ng/ul
RT: 19.527 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:202 Resp: 1763254
Ion Ratio Lower Upper
202 100
101 10.7 8.4 12.6
100 8.0 6.4 9.6



#82
Pyrene
Concen: 21.079 ng/ul
RT: 19.880 min Scan# 2838
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:202 Resp: 1893265
Ion Ratio Lower Upper
202 100
101 12.2 9.4 14.0
100 9.9 8.0 12.0





#83

Butylbenzylphthalate

Concen: 22.434 ng/ul

RT: 20.739 min Scan# 2984

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:149 Resp: 550105

Ion Ratio Lower Upper

149 100

91 71.5 58.6 87.8

206 21.5 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 22.167 ng/ul

RT: 21.516 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:252 Resp: 632029


Ion Ratio Lower Upper

252 100

254 64.6 52.2 78.2

126 10.3 8.0 12.0









#85
Benzo(a)anthracene
Concen: 21.057 ng/ul
RT: 21.598 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:228 Resp: 1864025
Ion Ratio Lower Upper
228 100
229 19.8 15.9 23.9
226 26.8 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.464 ng/ul
RT: 21.504 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:149 Resp: 929195
Ion Ratio Lower Upper
149 100
167 26.9 22.0 33.0
279 4.1 3.6 5.4





#87

Chrysene

Concen: 21.277 ng/ul

RT: 21.651 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:228 Resp: 1783115

Ion Ratio Lower Upper

228 100

226 30.0 24.3 36.5

229 19.6 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 19.682 ng/ul




RT: 22.498 min Scan# 3283


Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Tgt Ion:149 Resp: 1475257









#90
Benzo(b)fluoranthene
Concen: 21.608 ng/ul
RT: 23.398 min Scan# 3436
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37




Instrument :
BNA_P
ClientSampleId :
SICV638


Tgt Ion:252 Resp: 1950002
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.2
125 9.0 7.0 10.6



#91
Benzo(k)fluoranthene
Concen: 20.519 ng/ul
RT: 23.451 min Scan# 3445
Delta R.T. -0.000 min
Lab File: BP014719.D
Acq: 18 Apr 2023 19:37

Tgt Ion:252 Resp: 1851977
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 9.2 7.0 10.4





#93

Benzo(a)pyrene

Concen: 21.051 ng/ul

RT: 24.080 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:252 Resp: 1687231

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 10.3 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.961 ng/ul

RT: 26.921 min Scan# 4035

Delta R.T. 0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37




Tgt Ion:276 Resp: 2186550


Ion Ratio Lower Upper

276 100

138 21.0 16.3 24.5

277 24.5 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 20.897 ng/ul

RT: 26.945 min Scan# 4038

Delta R.T. 0.006 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Instrument :

BNA_P

ClientSampleId :

SICV638





Tgt Ion:278 Resp: 1825550

Ion Ratio Lower Upper

278 100

139 13.8 11.5 17.3

279 24.2 18.6 27.8



#96

Benzo(g,h,i)perylene

Concen: 20.753 ng/ul

RT: 27.756 min Scan# 4177

Delta R.T. -0.000 min

Lab File: BP014719.D

Acq: 18 Apr 2023 19:37

Tgt Ion:276 Resp: 1731443

Ion Ratio Lower Upper

276 100

138 18.1 14.3 21.5

277 23.9 19.5 29.3

