

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041825\
 Data File : BP024342.D
 Acq On : 18 Apr 2025 14:41
 Operator : RC/JU
 Sample : Q1825-04DL 20X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TP-10DL

Quant Time: Apr 18 15:23:57 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 18 12:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	216191	20.000	ng	0.00
21) Naphthalene-d8	10.493	136	850680	20.000	ng	0.00
39) Acenaphthene-d10	14.351	164	516785	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	985118	20.000	ng	0.00
76) Chrysene-d12	21.598	240	1097285	20.000	ng	0.00
86) Perylene-d12	24.939	264	1347132	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	57789	4.420	ng	0.00
7) Phenol-d6	6.910	99	71162	3.975	ng	0.00
23) Nitrobenzene-d5	8.869	82	41511	2.782	ng	0.00
42) 2,4,6-Tribromophenol	15.869	330	32434	4.536	ng	0.00
45) 2-Fluorobiphenyl	12.963	172	102393	3.034	ng	0.00
79) Terphenyl-d14	19.874	244	188836	3.469	ng	0.00
Target Compounds						
						Qvalue
71) Phenanthrene	17.192	178	131632	2.449	ng	96
75) Fluoranthene	19.274	202	188068	2.942	ng	99
78) Pyrene	19.651	202	172761	2.477	ng	98
80) Butylbenzylphthalate	20.610	149	67905	2.273	ng	99
84) Bis(2-ethylhexyl)phtha...	21.515	149	1725279	39.402	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041825\
Data File : BP024342.D
Acq On : 18 Apr 2025 14:41
Operator : RC/JU
Sample : Q1825-04DL 20X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TP-10DL

Quant Time: Apr 18 15:23:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

