

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041825\
 Data File : BP024353.D
 Acq On : 18 Apr 2025 22:11
 Operator : RC/JU
 Sample : Q1826-06
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONCRETE-FLOOR

Quant Time: Apr 19 01:03:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 18 12:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	210625	20.000	ng	0.00
21) Naphthalene-d8	10.493	136	800929	20.000	ng	0.00
39) Acenaphthene-d10	14.351	164	482510	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	968506	20.000	ng	0.00
76) Chrysene-d12	21.586	240	1192239	20.000	ng	0.00
86) Perylene-d12	24.927	264	1219953	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	1766871	138.703	ng	0.00
7) Phenol-d6	6.910	99	2133600	122.322	ng	0.00
23) Nitrobenzene-d5	8.875	82	1333006	94.902	ng	0.00
42) 2,4,6-Tribromophenol	15.863	330	1010436	151.358	ng	0.00
45) 2-Fluorobiphenyl	12.957	172	2810751	89.191	ng	0.00
79) Terphenyl-d14	19.874	244	5328656	90.088	ng	0.00
Target Compounds						
32) Benzoic acid	9.875	122	926466	93.122	ng	99
60) Diethylphthalate	15.186	149	94379	2.610	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041825\
Data File : BP024353.D
Acq On : 18 Apr 2025 22:11
Operator : RC/JU
Sample : Q1826-06
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONCRETE-FLOOR

Quant Time: Apr 19 01:03:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

