




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP009996.D
 Acq On : 21 Apr 2022 02:49
 Operator : CG/JU
 Sample : N2472-04
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0J58

Manual IntegrationsAPPROVED

Quant Time: Apr 21 03:20:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022



TIC: BP009996.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.411min (+ 0.041) 0.10 ng/uL

response 272

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	55.40	64.86
34.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP009996.D
 Acq On : 21 Apr 2022 02:49
 Operator : CG/JU
 Sample : N2472-04
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0J58

Manual IntegrationsAPPROVED

Quant Time: Apr 21 03:20:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022



TIC: BP009996.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.375min (+ 0.006) 2.80 ng/uL m

response 7468

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	55.40	70.32#
34.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
Data File : BP009996.D
Acq On : 21 Apr 2022 02:49
Operator : CG/JU
Sample : N2472-04
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0J58

Manual IntegrationsAPPROVED

Quant Time: Apr 21 03:20:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 20 01:32:14 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



TIC: BP009996.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.763min (-0.000) 5.46 ng/u1

response 30323

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	132.70	77.71#
41.00	23.20	54.07#
42.00	18.00	28.90#




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
Data File : BP009996.D
Acq On : 21 Apr 2022 02:49
Operator : CG/JU
Sample : N2472-04
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0J58

Manual IntegrationsAPPROVED

Quant Time: Apr 21 03:20:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 20 01:32:14 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



TIC: BP009996.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.763min (-0.000) 5.73 ng/u1 m

response 31832

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	132.70	77.71#
41.00	23.20	54.07#
42.00	18.00	28.90#

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP009996.D
 Acq On : 21 Apr 2022 02:49
 Operator : CG/JU
 Sample : N2472-04
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0J58

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022

Quant Time: Apr 21 03:20:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	118948	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	532361	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	390202	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	835735	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.404	240	815138	20.000	ng/ul	# 0.00
88) Perylene-d12	23.851	264	726241	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	7468m	2.802	ng/uL	0.00
4) Pyridine-d5	3.793	84	39919	5.432	ng/ul	0.00
7) Phenol-d5	7.099	99	64690	6.195	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	239476	38.520	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	215563	27.365	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	138880	15.931	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	157449	41.013	ng/ul	0.00
24) 2-Nitrophenol-d4	9.828	143	164512	39.087	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	269526	30.602	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	354072	28.226	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	1203110	39.252	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	1329976	36.626	ng/ul	0.00
54) 4-Nitrophenol-d4	14.763	143	31832m	5.727	ng/ul	0.00
60) Fluorene-d10	15.563	176	1047031	40.607	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.675	200	185523	37.472	ng/ul	-0.01
73) Anthracene-d10	17.422	188	1753943	43.833	ng/ul	0.00
81) Pyrene-d10	19.651	212	2184271	49.064	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.698	264	1724581	45.600	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
Data File : BP009996.D
Acq On : 21 Apr 2022 02:49
Operator : CG/JU
Sample : N2472-04
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0J58

Manual IntegrationsAPPROVED

Quant Time: Apr 21 03:20:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 20 01:32:14 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022

